



STATE GOVERNMENT OF MAHARASHTRA
DIRECTORATE OF MEDICAL EDUCATION &
RESEARCH

Govt.Dental College & Hospital Building, St.George's Hospital Compound, Mumbai - 400 001.

Tel.No. +91-22-22620363-65 Fax:+91-22-22620562/ 22652168

Website : <https://mahatenders.gov.in>

dmer.procurement@maharashtra.gov.in



E- 26 (26-27)

Orthopedics Department Medical Equipment
Turnkey Project

Directorate of Medical Education and Research, Maharashtra, Mumbai Invites Bidders for Under Government Medical College and Hospitals

Tender No.: E – 26 (26-27)

The complete tender document has been published on Maharashtra Tenders for the purpose of downloading. The downloaded tender document shall be considered valid for participation in the electronic bidding process (e- Tendering) subject to the submission of required tender document fee and EMD. willing to participate in this bidding process must register on <https://mahatenders.gov.in> the Agency (hereafter referred as Bidder)

However, Bidder who has been debarred / blacklisted by DMER/Central Govt. /State Govt. / Public Sector Undertaking / any other Local Body on the date of submission of the Bid. The details of the same are available on the above mentioned portal under 'Tenders'. For Registration procedure, please visit <https://mahatenders.gov.in>

The Bidder may get digital signature from any one of the Certifying Authorities (CA's) licensed by the Controller of Certifying Authorities

The technical and commercial bids shall be submitted online up to the end date & time mentioned below.

Tender Schedule

Tender No.	E-26 (26-27)
Name of The Equipment	Orthopedics Department Medical Equipment Turnkey Project
Name of The Medical College & Hospital (Consignee)	Sir J.J. Group of Hospitals, Mumbai
Approximate cost (Rs.)	Rs. 29,60,00,000/-
Date of commencement of sale of Tender document/ download	14.08.2026 -10.00 hrs to 04.09.2026 - 05.00 hrs.
Pre-Bid Meeting	24.08.2026 at 11.30 am
End Date and Time of Bid Submission	04.09.2026 - 05.00 pm...
Date and Time of Opening of Envelope No.1	07.09.2026, 02.00 PM. onwards
Address for Communication	Directorate of Medical Education Research, Mumbai Govt. Dental College Build. 4th floor St. Georges Hospital Compound Mumbai 400 001
Venue for Pre-Bid Meeting & opening of bid	Same as above

The pre-bid meeting will be held on venue of the same is at Conference Hall, DMER, 4th Floor, Govt. Dental College & Hospital Building, St. George's Hospital Compound, Near CSTM, Mumbai, Maharashtra 400001. The prospective bidders should submit their suggestions/observations, if any, in writing before Date 24.08.2026 at. 11.30am. Only suggestions/observations received in writing shall be discussed and clarified in pre-bid meeting and any modification of the tender documents, which may become necessary as a result of pre- bid meeting, within stipulated period shall be made by the Commissioner, Medical Education & Research. Maharashtra, Mumbai exclusively through the issue of addendum /corrigendum, As the Case may be the tender documents uploaded shall be read with any such modification. Authorized representatives of prospective bidders may attend the said meeting and obtain clarification regarding scope of work & tender conditions. **Authorized representatives shall submit authorization letter to attend the pre-bid meeting.**

The tender document is available on <https://mahatenders.gov.in> along with tender notice. However, the tenderer shall have to pay "e-tender price" through online payment gateway before downloading and uploading the tender document in e-tender portal.

Note: -

The Bidders shall have to pay the tender EMD amount through on line payment only. The Commissioner shall not be liable for any omission, mistake or error in respect of any of the above or on account of any matter or thing arising out of or concerning or relating to the tender documents or the bidding process, including any error or mistake therein or in any information or data given by the Directorate of Medical Education and Research, Mumbai

The Commissioner reserves the right to reject all or any of the e-tender(s) without assigning any reason at any stage.

The Bidders shall note that any addendum/corrigendum issued regarding this e-tender notice will be published on the <https://mahatenders.gov.in> only. addendum **No corrigendum will be published in any other manner.**

Commissioner
Medical Education Maharashtra, Mumbai

NOTE : All bid related activities (process) like tender downloading, bid preparation, bid submission and bid opening etc. will be governed by time schedule under the key dates mentioned in above schedule.

**Request for Bids (Two-Envelope Bidding Process) Tender
No. E- 26 (26-27)**

Contract Title: Request for Bids for Orthopedics Department Medical Equipment Turnkey Project under the Medical Education Department Maharashtra.

S.No	Description	Non-refundable Tender Fees Amount to be paid Online	Fee in Rs EMD Amount to be paid online in Rs
1	Orthopedics Department Medical Equipment Turnkey Project	30,000	30,00,000

INSTRUCTIONS TO TENDERER

1. The Tenders shall be submitted online through e-tendering process only. No tender will be accepted manually or through fax etc.
2. Only Manufacturers/ importers/Manufacturer Authorized bidder are eligible to participate in the tender.
3. The Tender form will be available online only.
4. The tender document will be uploaded / released on website www.mahatenders.gov.in The tenderer is required to download the tender document from this website as per the schedule of activities for online tender document download.
5. The tenderer is required to download the tender document within the schedule of activity for online tender document download. After expiry of the date and time for tender document download, authority will not be responsible for any such failure on account of tenderer for not downloading the document within the schedule even though he has paid the cost of the tender to the authority. In such case the cost of the tender paid by the tenderer will not be refunded.
6. During the activity of bid preparation, the tenderer is required to upload all the documents mentioned in the terms & conditions of the technical bid by scanning the documents and uploading it in the PDF. This activity of uploading the documents as well as preparation of commercial bid and other Annexure enclosed with the tender shall be completed within the schedule given for bid preparation. Bidder shall inspect site & submit site survey report signed by appropriate authority along with preliminary drawings in technical bid.
7. After bid preparation, the tenderer is required to submit the bid during the activity of submission of bid.
8. The tenders submitted online will be opened on the schedule date before the participant. At the time of opening of technical bid the tenderer should bring all the original documents which have been uploaded along with the technical bid so that same can be verified at the time of opening of technical bid.
9. As the tenders are invited online, tenderers are required to upload tender as per the schedule given. Authority is not responsible for failure on the part of tenderer to upload tender within the time schedule. In case of any help required for this purpose the tenderer shall contact to help line of e-tendering agency Tenderers are requested to go carefully through the process of e-tendering given on the website of www.mahatenders.gov.in to avoid disqualification of their tender on any ground. In case of any technical problems e.g. payment of EMD, downloading of tender terms, uploading of tender etc. This bidder should contact helpline for resolution of their issues.

10. PRE-BID MEETING:

The pre-bid meeting will be held at the date, time and venue mentioned in the e- Tender Notice. Bidders should send their queries by email on or before 24.08.2026 at 11.30 am. Tenderers shall note that any corrigendum issued regarding this tender notice will be published on the on the website of www.mahatenders.gov.in

The prospective tenderer(s) should submit their suggestions/observations if any, in writing through email before date 24.08.2026 at 11.30 am. Only suggestions / observations received in writing by mail within stipulated time will be discussed and clarified in pre-bid meeting and any modification of the tendering documents, which may become necessary as a result of pre-bid meeting, shall be made by authority through the issue of an addendum/corrigendum. The tender uploaded shall be read along with any modification. Authorized representatives of prospective tenderer(s) can attend the said meeting and obtain clarification regarding specifications, works & tender conditions. Authorized representative should have authorization letter to attend the pre-bid meeting, provided queries are submitted in time.

Any amendment to the tender will be placed on the e-tendering website and the pre bid amendments(addendum/corrigendum) herein called common set of deviations (CSD) should be downloaded by the tenderer and the tenderer should upload these signed & stamped CSD along with the tender terms and conditions as they form a part of tender. PLEASE NOTE THAT THE TENDERER SHOULD NOT BID IN THE TENDER UNLESS AND UNTIL THE CSD (COMMON SET OF DEVIATIONS) ARE UPLOADED OR IT WILL BE LIABLE FOR REJECTION.

- a) The tenderer will not be communicated separately regarding the amendment.
 - b) In case there is single response to this tender the committee will be within its right to take action as per instructions prescribed **G.R.No.SPO-2014/Pra.Kra.82/Part- III/Industry- 4, dated 01.12.2016 issued by Industry, Energy & Labour Department, Mantralaya, Mumbai – 32.**
11. Non-attendance at pre-bid meeting shall not be a cause for disqualification of a tenderer.

12. Manner of Submission of Tender

The tender is divided in two parts and required to be submitted online only.

Technical Bid (Part I):- Contains technical information such as technical specifications, terms & conditions of the tender, qualification criteria, documents required to be submitted to become qualify for opening of commercial bid, general terms & conditions of the tender, instructions to the tenderer for submission of the tender, other Annexures required to be submitted by the tenderer along with the technical bid, **confirming to tender terms & conditions.**

Original Tender Form :-

Tenderers are required to sign and put official seal of the company on every page of the tender form as well as common set of deviations (CSD). They are also required to sign on every page of the specification copy and affix the seal of the company at the right hand corner of the bottom of the page.

IMPORTANT NOTE: - Apart from all the conditions mentioned in this tender document, the tender rules / guidelines published by Industry, Energy & Labour Department, Mantralaya, Mumbai – 32 vide G.R. No. SPO-2014/Pra.Kra.82/Part-III/Industry-4, dated 01.12.2016 are applicable and it is binding for bidders.

TECHNICAL BID (PART - I)

Technical bid shall contain the following documents duly **SELF ATTESTED**, failing which tender will be declared **Invalid and will be Liable for Rejection**. These documents shall be in the form of PDF files, if required in Zip file. These documents needs to be digitally signed by the tenderer and uploaded in the technical bid during online bid preparation stage only.

1. PROFORMA

To be submitted along-with Technical Bid as mentioned in (**Annex- I**).

In every case of offer the name of all manufacturer, address together with brand /make /model should be indicated in the tender. **Any conditional offer of the tenderer will be rejected.**

2. Earnest Money Deposit (EMD) & Tender Fee

The Bidder shall pay EMD & tender fee online only. No EMD will be accepted in the form of Bank Guarantee and Demand Draft.

Bidder exempted from payment of **EMD& Tender Fee** should produce attested documents which make the Bidder eligible for exemption.

Bidders are eligible for EMD & Tender Fee exemption only if the firms are registered with Government as per Micro, Small & Medium Enterprises Development Act 2006 (MSME) in the state of Maharashtra only as the manufacturer for all the Equipment's Mentioned in the concerned Tender. In such event, the Bidder shall submit Self Attested copies of the registration attached with the tender.

The EMD will be forfeited if the tenderer withdraw from tender process at any stage OR the tenderer fails to enter into contract according to the terms & conditions if accepted. EMD of the successful as well as unsuccessful tenderer shall be refunded within due course.

3. Product Catalogue:-

All Catalogues/brochures showing details of Technical Specifications of all Equipment's Mentioned in the concerned Tender shall be enclosed in technical bid, Original catalogues only (photocopies will be rejected). All specification should be substantiated by Technical data sheets. Bidder shall submit Hard copy of technical data at DMER on & before final date of submission.

4. GST /Clearance Certificate:- Copy of G.S.T. Registration Certificate, GST clearance certificate/GST3BR mentioning clearance up to the last month/quarter (as applicable) from the date of publishing of the tender should be attached.

5. Turnover:- - The average turnover of bidder for the last 3 Financial years from the date of publication of tender should be 40% of total tender cost. If past three financial years audited report not available, bidder can attach audited CA certified statement for past three financial years. Turnover exemption is not allowed for MSME, Start up or any other ventures.

6. Past experience : Either OEM (either Directly or through authorized distributor for concerned past project) or bidder shall have completed / delivered at least **Three similar work for Government, Semi Government, PSU's, Local self-government bodies, Autonomous Bodies all over India in last 5 years.** OEM/bidder shall Mandatorily submit Purchase order copy to support past experience. Past experience exemption is not allowed for MSME, Start up or any other ventures.

7. The bidder should submit complete audit report with relevant certificates and schedules for the last 3 financial years from the date of publishing of the tender certified by registered Chartered Accountant. If the audit of last financial year is not complete; bidder can attach CA certified last financial year (Balance Sheet and Profit & Loss Account Statement) signed by registered Chartered Accountant. Products/ equipment's manufactured in countries sharing .

8. Net-worth: The net worth of the Bidder firm (manufacturer or principal of authorized representative) should not be negative on 'The Relevant Date' and also ii) should have not eroded by more than 30% (thirty per cent) in the last three years, ending on 'The Relevant Date'. Note: In case of Indian Bidders/ companies (manufacturer or principal of authorized representative) who have been restructured by Banks in India, under the statutory guidelines, they would be deemed to have qualified the Financial standing criteria

considering the institutional financial backing available to them. Land border with India are not allowed.

9. Authorization Letter:-

A bidder quoting on behalf of manufacturer must attach authorization letter from the manufacturer to quote on behalf of manufacturing & to do after sales services as per **Annexure-II**. (It is mandatory to have a complete address, working e-mail address and contact numbers of the Manufacturer as well as it should mention the Country of manufacture of the equipment. Equipment manufactured in any other place, than the mentioned originally; will be rejected. Incomplete Authorization in any form will not be accepted. It will be liable for rejection. (Original Authorization Letter Inked Copy to be Submitted to DMER). **In case of foreign manufacturer**, the following documents should be **Mandatorily** Submitted for the last 3 Financial years for similar equipment supplied will be required.

- a. Bills of entry
- b. Bills of landing or Airway Bill
- c. Country of Origin Certificate issued by competent authority of that country (Chamber of Commerce) naming the manufacturer, consignee, name of equipment, invoice no, quantity etc.

10. I.S.O., I.S.I., CE from notified body or USFDA Certificates:-

BIS/ISO Valid licences under the Medical Devices Rules, 2017 for the quoted make/model — MD-15 (imported) or MD-5 / MD-9 (indigenous), and in all cases MD-42 (sale, stock, distribution)(MANDATORY).Manufacturer to hold ISO 13485:2016 for the quoted device's manufacturing site. US FDA / CE (EU MDD) desirable, certificates required as per technical specifications Mentioned in the concerned Tender.

11. Users list:-

The tenderer should submit user list of quoted model (**only**) as per **Annexure - IV**. And should also submit Mandatorily Purchase order (quoted model) of **last three Financial years copies** to support the claim.

12. Make / Model:-

Bidder shall submit Mandatorily Make & Model in scope of supply for all Equipment's Mentioned in the concerned Tender. Bidder shall quote latest model as per technical specifications. Capacity or volume of equipment may differ from past experience project but Make & Model shall be same OEM whose past experience is used for qualification .

13. After Sales Service:-

The bidder shall have Mandatorily minimum 2 service centers in Maharashtra with complete address, phone nos., fax nos. and e-mails as per (**Annexure - V**) on or before date of publish of this tender document.

14. Terms & Condition Acceptance Certificate should be submitted as per (Annexure - VI**).**

15. Fall Clause

It is a condition of the contract that all through the currency thereof, the price at which the bidder will the supply stores should not exceed the lowest price charged by the bidder to any customer during the currency of the rate contract and that in the event of the prices going down below the rate contract prices the bidder shall promptly furnish such information to Directorate of Medical Education & Research, Mumbai to enable Directorate of Medical Education & Research, Mumbai to amend the contract rates for subsequent supplies.

16. Commercial bid should not be submitted in Envelope No. 01 (Tech. Bid) else the bid will be rejected.

17. In the case of a Turnkey Tender, A Principal can authorize only one agent, and an agent should not represent or quote on behalf of more than one Principal. However, this shall not debar more than one Authorised distributor (with/ or without the OEM)

from quoting equipment manufactured by an Original Equipment Manufacturer (OEM) in procurements under a Proprietary Article Certificate(PAC).

18. Authentication for Documents:-

The responsibility to produce correct authentication for documents **lies** with the tenderer. If any documents is detected to be forged, bogus etc., the tender shall be rejected and EMD shall be forfeited. Any contract entered under such condition shall also be liable to be cancelled at any time during its currency and further penal action like criminal prosecution, blacklisting etc. against the said contractor and/or the partners shall be instituted.

- 19.** The quantity mentioned in the tender is only approximate estimated quantity, Directorate of Medical Education & Research, Mumbai, if required may Purchase extra 50% of quantity published within 6 months of Purchase Order as repeat order as per GR dated 1/12/2016 with same price agreed in the purchase order.

At the time of submission of tender documents to this office all the photo copies of certificates / documents attached with the tender should be duly Self Attested on and before date of submission of bid.

IF ANY OF THE ABOVE DOCUMENTS / INFORMATION AS ENLISTED FROM NO.1 TO NO.15 IS NOT ATTACHED WITH THE TENDER. THE TENDER IS LIABLE TO BE TREATED AS INVALID.

COMMERCIAL - BID (PART -II)

The Commercial Bid (Part-II) should contain the price. The cost should be quoted in Indian currency only, any fluctuation in the international currency will not be a concern of Directorate of Medical Education Research, Mumbai and the price quoted in the tender will be considered as final. Tenderer should quote the price in proforma mentioned in Annexure-IX mentioning Basic cost of Instrument along with applicable Taxes in percentage.

All financial offers must be prepared and submitted online which should be signed using individual's digital certificate. Date of opening of commercial bid will be informed accordingly. Commercial quote in any other format shall be rejected. Conditional offers shall not be considered and shall be treated as non-responsive.

OTHER TERMS & CONDITIONS CONCERNING THE TENDER

1. No price should be quoted in **part I** (Technical Bid). If any price is quoted in technical bid, the entire tender shall be considered as **INVALID**.
2. Envelope No.2 shall be opened only when the documents as mentioned in Envelope No.1 are in order, as per requirements and as per specifications. Valid tenderer shall be informed accordingly.
3. The Envelop No. 2 (Commercial Bid) shall be opened of only those tenderers who have submitted all the documents as per tender terms and conditions in the envelope no.01 and are technically qualified. If two tenderers have quoted same cost of equipment, then the demonstration of both the tenderer will be taken.
 - a. Unsatisfactory performance at demonstration will disqualify the tenderer.
 - b. The demonstration should be done if possible at the institute specified. In case of heavy / bulky equipment the demonstration may be arranged at a place where the equipment is already functioning and as per convenience of the concern technical committee at the cost of the tenderer. Demonstration of the equipment in the factory premises is not allowed.
4. The tender validity should be for a **minimum period of 120 days** from the date of opening the Envelope No.2 (Commercial Bid). The validity of the tender can be extended up to 180 days at the discretion of Directorate of Medical Education & Research , Mumbai.
5. There will be no change or any type of rectification in quoted cost due to international fluctuation of currency after submission of tender. No communication in this regards will be entertained.
6. **Demonstration:-**
 Directorate may ask for demonstration before the opening of financial bid for technical evaluation or from lowest bidder after financial bid opening and the bidder should arrange for the demonstration in India of the equipment quoted for in the tender within 15 days from the date of intimation of the request for demonstration preferably in Mumbai. However, if complete system of quoted model/complete system is not available in Mumbai, demonstration may be arranged outside Mumbai in any mutually agreed upon hospital at bidders cost. The video recording of the demonstration shall be mandatorily done. Soft copy of the Video Recording shall be handed over to DMER. Arrangement of Video Recording shall be done by the bidder at their own cost. The demonstration report shall be prepared on same day and signed by all Team Members present including representatives of bidder and the report of the demonstration should be scanned and Emailed to DMER Email I.D. on the same day.
 Following Points to be Mandatorily considered while Demonstration :
 - a. The equipment is new and made of virgin material, it is not reconditioned / retrofitted.
 - b. The team shall clearly mention in their report, The name of the equipment manufacturer, model and serial nos. of equipment & country of origin of the manufacturer.
7. **Training:** - The successful tenderer shall have to give sufficient training at his cost to the staffs of the Hospital and Engineers of concern institute to operate the Medical

Equipment. Also it shall be provided as and when required if asked by user department.

8. The successful tenderer will have to pay a security deposit of an amount **equivalent to 3% of the cost** of the equipment offered in the following form within 15 days from the date of issue of supply order from the concerned institute. **Bank Guarantee valid for 2 months after the expiry of warranty period issued by any Nationalized / Scheduled Bank.** Security deposit will be refunded only after completion of warranty period and conformation AMC/CMC.
9. The delivery of goods should be at door and should be made within **36 weeks** for indigenous items and **44 weeks** for imported items, from handover of site. If the date of delivery cannot be strictly adhered to, then extension required by the tenderer should be stated in the bid. No excuses for delay by any statutory authorities like custom etc. will be taken into consideration for extension of the period of delivery. The primary responsibility for supply of goods in time will be rest with the supplier.
10. **Penalty on late delivery:-** In the event of the late delivery of goods, the purchaser will recover from contractor by way of penalty, a sum equal to **half Percent (0.5%)** price of the goods delivered late per week calculated from the next day after the agreed delivery period is over.
11. **Consequences of low quality supply:-** - If the equipment supplied is found of inferior quality or not as per specifications, the contractor shall replace the equipment within one month from the date of intimation at the cost & risk of the contractor and also liable to pay the fine imposed by the consignee, failing which Earnest Money Deposit & Security Deposit of the contractor shall be forfeited and the tenderer shall be liable for penal action including black-listing etc. In addition to the forfeiture of the Earnest Money Deposit & Security Deposit, if any fine is imposed by the consignee same shall be recovered from other dues to the contractor from –his bills payable.
12. **Replacement of Rejected materials:-** - Tenderer / Contractor shall have to replace rejected material with approved one. The supplier shall remove the rejected material within 60 days failing which the same will be disposed off by consignee at the risk and cost of contractor without any further correspondence in this regards.
13. **Risk & Cost Purchase:-** - In case the Contractor/s, shall at any time during the continuance of these presents fails to supply satisfactorily the equipment within the prescribed time as herein provided and or in case shall fail to replace any part/s that may have been rejected with other of approved quality, the consignee shall be at liberty forthwith to procure the same in the open market at the risk and cost of the contractor/s. Similarly if the work underlying the contract is not executed satisfactorily within the stipulated period or after the same having been disapproved wholly or partly is not rectified or re-done to the satisfaction of the Officer in Charge within the said specific period, the consignee shall get the same executed or rectified or re-done through any other agencies, at the entire risk of the supplier and expenses thereby incurred, shall be payable by the supplier and / or may be deducted from any moneys due or become due to the contractor/s and the consignee may, however fix such other subsequent date as he may think fit by which the delivery of the said article and or execution of the said work shall be completed.
14. **Blacklisting:-** - The firm shall be black-listed, if it is found that:-
 - a. Forged documents are submitted OR
 - b. If it becomes responsive on the basis of submission of bogus certificate / information.
 - c. In case of non-supply of equipment / accessories or supply of substandard quality or supply of equipment / accessories found to have been previously used or having reconditioned parts.
15. **Warranty Period :- (including supply of spares).**
The warranty period shall be for **2 years** from the date of commissioning of all equipment supplied as certified by the consignee.

The successful tenderer must ensure 95% uptime during warranty period. In case of downtime, warranty period will be extended for period of downtime. If the equipment is not attended within 24 hours for Mumbai and 48 hours for other places the supplier will be liable to pay a penalty of 0.07% of purchase cost for every day of delay. Such penalty will be recovered from the amount of security deposit. Certificate of such uptime / downtime issued by the end user will be binding for the supplier.

Replacement of parts and service thereof due to manufacturing defects during warranty period will be entirely at the supplier's cost. The expenditure incurred for procurement of space, transport, installation, commissioning and various duties shall be borne by the supplier

16. Annual Maintenance Contract (For rendering services) /Comprehensive Annual Maintenance Contract:-

The tenderer will have to agree to enter into an Annual Maintenance Contract exclusive of all spares & consumables (AMC) 1% per year of the Order value of the machinery / equipment (excluding taxes).

Where required, tenderer will have to agree for Comprehensive Maintenance Contract (CMC) inclusive of all spares @ 5% of the Order value (excluding taxes) of the equipment per year. The period of such AMC / CMC will be of 8 years after completion of warranty period. In case of non-compliance of AMC/CMC the supplier will be liable to pay a penalty. Such penalty shall be recovered from the amount of bank guarantee submitted. Payment for AMC /CMC on yearly basis will be made by the users institution or DMER , at the end of year after satisfactory performance report from the end user.

17. Payment:-

Payment of 50% of the contract value of equipment will be made after supply of material as per scope of supply, 25% after successful installation of all the Equipment's & remaining 25% after successful testing of equipment's on patients for atleast minimum a month.

18. In case of imported machine, the purchase orders will be issued in Indian currency and afterwards the exchange rates are increased; no financial implications will be accepted by the concerned institution or by the Government of Maharashtra.

19. Complaint/s :-

If any bidder has a complaint against any process, decision etc. then the complainant has to give a written complaint Mandatorily within 48 hrs of such a decision or any other grievance to the Directorate of Medical Education & Research, Mumbai to review such a decision or grievance and mention clearly his grievance, after opening of technical bid only those bidders who have participated in tender process can lodge a complaint and after opening of commercial bid only those bidders who have been declared as technically qualified can lodge their grievance.

20. Jurisdiction of the Courts :-

In case of any claim, dispute or differences arising in respect of tender, the cause of action shall be deemed to have arisen in Mumbai and all legal proceedings in respect of any such claim, dispute or difference shall be instituted in a Competent Court in the city of Mumbai.

21. Affidavit on Non-Judicial Stamp Paper of Rs. 500/- stating that the rates quoted in the tender are Prevailing Market Retail Price (MRP) /current prevailing market price and/or rate to be submitted to this office. (To be submitted to this office)

22. "Affidavit on non-judicial stamp paper of Rs.100/-regarding the firm has not been found guilty of malpractices, misconduct or blacklisted/debarred for the quoted product by Public Health Department, Govt. of Maharashtra or by any local authority and other State Government/Central Government's organizations as on the date of submission tender document " (To be submitted to this office)

23. Force Majeure:

- a. For purposes of this Clause, 'Force Majeure' means an event beyond the Control of the Supplier and not involving the Supplier's fault negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
- b. If a Force Majeure situation arises at any time during the subsistence of contract, the Supplier shall promptly but not later than 30 days notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- c. Force Majeure will be accepted on adequate proof thereof.
- d. If contingency continues beyond 30 days, both parties will mutually discuss and decide the course of action to be adopted. Even otherwise contingency continues beyond 60 days then the purchaser may consider for termination of the contract on pro-rata basis.

THE DECISION OF Directorate of Medical Education & Research, Mumbai
WILL BE FINAL AND BINDING. Directorate of Medical Education & Research,
Mumbai RESERVES THE RIGHTS TO REJECT ANY OR ALL TENDERS
WITHOUT ASSIGNING ANY REASON.

I have read all the terms and conditions of the tender carefully and I agree to abide by them.

Signature of Tenderer with Seal

List of Annexure

Annexure No.		Description
Annex No. I	:	Proforma to be submitted along-with Technical Bid
Annex No. II	:	Manufacturer Authorization Letter
Annex No. III	:	Technical Compliance Chart
Annex No. IV	:	User List of Quoted Model
Annex No. V	:	Details of After Sales Service Station
Annex No. VI	:	Terms & Condition Acceptance Certificate
Annex No. VII	:	Consignee List.
Annex No. VIII	:	Proforma of Affidavit.
Annex No. IX	:	Proforma of Affidavit.
Annex No. X	:	A. Annual Turnover proforma & B.Net Worth.
Annex No. XI	:	Format for doubts/queries for Pre-bid meeting
Annex No. XII	:	Power of Attorney
Annex No. XIII	:	Undertaking to be signed by the Bidders
Annex No. XIV	:	Financial proposal Performa (BOQ)

ANNEXURE – I
BIDDER INFORMATION FORM

(To be submitted on the Bidder's Official Letterhead along with the Technical Bid)

Note: Incomplete or unsigned forms are liable to be rejected.

1. Bidder Details

Sr. No.	Particulars	Details to be Furnished
1	Name of the Firm / Company	
2	Registered Office Address	
3	Corporate / Branch Office Address (if applicable)	
4	Telephone Number	
5	Mobile Number	
6	Fax Number (if applicable)	
7	E-mail Address	
8	Website (if applicable)	

2. Constitution of the Firm

Particulars	Details
Constitution of the Firm	Proprietorship / Partnership / LLP / Private Limited / Public Limited / Cooperative / Government Undertaking / Joint Venture / Others (Specify)
Ownership Status	Maharashtra Government / Central Government / Joint Sector / Cooperative / SSI / MSME / Private / Others (Specify)
Year of Establishment	
Registration Number	
PAN	
GSTIN	

3. Details of Proprietor / Partners/ Directors

Name	Designation	Residential Address	Contact Number	Shareholding (%)

4. Bidder Category

The bidder is participating in this tender as:

- ☐ Manufacturer
☐ Authorized Manufacturer
☐ Importer
☐ Authorized Distributor / Dealer
☐ Other (Specify): _____

5. Authorized Contact Person

Particulars	Details
Name	
Designation	
Mobile Number	
E-mail Address	
Name	

6. Manufacturing Unit(s)

Provide details of the original manufacturing unit(s)/factory(ies) from where the goods offered in this tender shall be supplied.

Manufacturing Unit	Full Address	State	Contact Number
Factory 1			
Factory 2 (if applicable)			
Factory 3 (if applicable)			

7. Declaration by the Bidder

I/We hereby declare that the information furnished in this Annexure is true, complete and correct to the best of my/our knowledge and belief.

I/We understand that if any information furnished herein is found to be false, incorrect, misleading, or suppressed at any stage of the tender process or during execution of the contract, the Procuring Entity shall have the right to reject our bid, terminate the contract (if awarded), forfeit the Bid Security/Performance Security (where applicable), and take any other action as deemed appropriate under the terms and conditions of the tender and applicable laws.

I/We further certify that the undersigned is duly authorized to sign and submit this Annexure on behalf of the bidder.

Signature of Bidder: _____

Stamp & Seal: _____

ANNEXURE – II
MANUFACTURER'S AUTHORIZATION FORM
(To be submitted on the Letterhead of Original Equipment Manufacturer/Original Manufacturer)

Tender Reference No.: _____

Tender Title: _____

Date: _____

To,
The Tender Inviting Authority

Subject: Authorization of Bidder by Original Manufacturer for Participation in Tender

Dear Sir/Madam,

We, **M/s.** _____, having our Registered Office / Manufacturing Unit at:

Complete Address:

Country of Origin of Product: _____

are the **Original Manufacturer / Original Equipment Manufacturer (OEM)** of the following items:

Sr. No.	Name of Item / Product	Model No. / Part No
1.		
2.		
3.		

We hereby authorize:

M/s. _____

having its registered office at to participate in the above-mentioned tender and submit the bid for the products manufactured by us.

1. Authorization

We authorize the above-mentioned bidder to:

- a) Submit the tender offer on our behalf.
- b) Quote prices for the products manufactured by us.
- c) Receive purchase orders, if awarded. d) Supply the products and submit invoices as per the terms of the contract. e) Coordinate with the Purchaser for supply, installation, commissioning, inspection, and acceptance of the supplied items, wherever applicable.

2. OEM Declaration and Undertaking

We hereby confirm and undertake that:

1. We are the Original Manufacturer/OEM of the products offered against this tender and possess the necessary

capability and infrastructure for manufacturing and providing technical support.

2.The products offered by the authorized bidder shall be genuine, new, unused, and manufactured by us.

3.We have examined the tender document and agree to comply with all applicable technical specifications, quality requirements, warranty conditions, and other relevant terms and conditions of the tender.

4.We shall provide necessary technical support, warranty services, and after-sales support for the products supplied against this tender either directly or through our authorized service network/importer/authorized representative.

5.We shall ensure availability of genuine spare parts and technical assistance during the warranty period and for the useful life of the equipment, wherever applicable.

6.We accept responsibility for the quality and performance of the products supplied under this tender in accordance with the contractual terms.

7.This authorization shall remain valid for the entire tender process and, if the contract is awarded, throughout the contractual period including the warranty period.

3. Declaration Regarding Multiple Authorization

We confirm that the bidder authorized herein is authorized by us for this tender and that the same authorization shall not be misrepresented or transferred to any other party without our written consent.

4. Undertaking Regarding Compliance

We certify that all information provided in this Manufacturer's Authorization Form is true and correct. In case any information is found to be false or misleading, the Purchaser shall have the right to take appropriate action as per the terms and conditions of the tender.

Note:

- 1.This authorization shall be printed on the official letterhead of the OEM/Manufacturer.
- 2.The signatory signing this document shall be duly authorized by the OEM/Manufacturer.
- 3.Any deviation or incomplete information may lead to rejection of the bid

5.For and on behalf of Original Manufacturer / OEM

Particulars	Details
Name of OEM / Manufacturer	
Registered Office Address	
Manufacturing Unit Address	
Name of Authorized Signatory	
Designation	
Contact No.	
E-mail ID	
Signature of Authorized Signatory	
Company Seal	
Place	

ANNEXURE III
TECHNICAL COMPLIANCE CHART
(To be submitted on Bidder's Letterhead)

Tender Reference No.: _____

Name of Equipment/Instrument: _____

Make & Model Offered: _____

Name of Manufacturer: _____

Name of Bidder/Supplier: _____

Detailed Technical Compliance Statement

Sr. No.	Tender Specification Requirement	Offered Specification (Make/Model/Manufacturer Details)	Compliance Status (YES/NO)	Deviation/Remarks (If Any)	Supporting Document Reference (Catalogue/Page No.)
1	Detailed specification as mentioned in tender	Detailed specification of quoted product including manufacturer name, model number, technical parameters, accessories, features, etc.	YES/NO	Mention exact deviation, if applicable	Catalogue/Data Sheet Page No.

Bidder's Declaration

We hereby certify that:

1. The technical specifications provided above are true and correspond to the product offered in this tender.
2. The quoted equipment/instrument is manufactured by the stated manufacturer and complies with the tender requirements except for deviations explicitly mentioned in the deviation column.
3. All deviations, if any, have been clearly indicated and no deviations are concealed.
4. Supporting technical documents, catalogues, datasheets, and compliance certificates have been submitted along with the bid wherever applicable.
5. In case of any discrepancy between the offered product and submitted compliance statement, the bidder shall be responsible as per tender terms and conditions.

Authorized Signatory of Bidder

Name: _____

Designation: _____

Company Name: _____

Date: _____

Seal & Signature: _____

ANNEXURE IV
USER LIST / PAST PERFORMANCE DETAILS
(Incomplete Annexure is liable for rejection)
(To be submitted on Bidder's Letterhead)

Tender Reference No.: _____**Name of Equipment/Instrument:** _____**Make & Model Quoted:** _____**Name of Manufacturer:** _____**Name of Bidder/Supplier:** _____**Details of Purchase Orders Executed for the Quoted Model During the Last Three Financial Years**

Sr. No.	Model No.	Name & Complete Address of Institute/Organization Where Supplied	Purchase Order No. & Date	Quantity Supplied	Date of Supply/Installation	Authorized Contact Person	Contact Details (Tel./Mobile/Email ID)	Performance/Remarks
1.								

Bidder's Declaration

We hereby certify that:

1. The above-mentioned installations are for the **same/similar model quoted in the present tender**.
2. The information provided is true and can be verified by the purchaser at any time during the tender evaluation process.
3. Copies of relevant purchase orders/work completion certificates, wherever required, shall be submitted for verification.
4. The supplied equipment has been installed and is performing satisfactorily at the mentioned user locations.

Authorized Signatory of Bidder

Name: _____

Designation: _____

Company Name: _____

Date: _____

Seal & Signature: _____

ANNEXURE V
(Incomplete Annexure is liable for Rejection)
DETAILS OF AFTER SALES SERVICE STATION
(On Bidder's Letterhead)

Name of the Equipment/Instrument: _____

Sr. No.	Name of the Agency	Full Address with Pin Code	Contact Person Name	Contact Numbers with STD Code (Tel. Off.)	Email ID	Mobile No.
1.						
2.						

Declaration:

We hereby certify that the above-mentioned details of our after-sales service station(s) are true and correct. We undertake to provide necessary after-sales service, maintenance support, and assistance for the supplied equipment/instrument as per the terms and conditions of the tender.

For (Name of Bidder): _____

Authorized Signatory: _____

Name: _____

Designation: _____

Date: _____

Seal of the Firm: _____

ANNEXURE VI
(Incomplete Annexure is liable for Rejection)
(On Bidder's Letterhead)
TERMS & CONDITION ACCEPTANCE CERTIFICATE

Ref.: _____

Date: _____

To,

Ref.: Tender Document No.: _____

Date: _____

Respected Sir,

I/We, the undersigned, have examined the above-mentioned Tender Enquiry Document, including Amendment/Corrigendum No. _____ dated _____ (if any), the receipt of which is hereby confirmed.

If our tender is accepted, we undertake to supply the goods and perform the services as mentioned in the Tender Enquiry Document in accordance with the delivery schedule specified therein.

We further understand that you are not bound to accept the lowest or any tender received against your Tender Enquiry.

We confirm that we have carefully read and fully agree to all the terms and conditions specified in the above-mentioned Tender Enquiry Document, including any Amendment/Corrigendum issued, if applicable.

For (Name of Tenderer): _____

Signature of Authorized Signatory: _____

Name: _____

Designation: _____

Contact No.: _____

Office Seal of the Tenderer: _____

Date: _____

ANNEXURE – VII
CONSIGNEE LIST

Sr. No.	Name of the Consignee	Number of Device
<u>1.</u>		
<u>2.</u>		
<u>3.</u>		

For (Name of Bidder): _____

Authorized Signatory: _____

Name: _____

Designation: _____

Date: _____

Office Seal: _____

ANNEXURE – VIII
AFFIDAVIT ON NON-JUDICIAL STAMP PAPER OF Rs. 500/-
(To be submitted to this office)

AFFIDAVIT FOR RATE CERTIFICATE

Reference: Tender No.: _____

I/We, _____ (Name of the Authorized Representative),

S/o / D/o _____,

aged _____ years, working as _____ of

M/s _____ (Name of Firm/Company), having its registered office

at _____, do hereby solemnly affirm and declare as under:

1. **That the rates quoted by us against the above-mentioned tender are true and correct and are not higher than the prevailing Market Retail Price (MRP) and the rates offered by us to other Government Departments/Institutions or any other organizations, wherever applicable.**
2. **That we have not quoted higher rates for the same equipment/item to any Government Department, Government Institution, AIIMS, Central/State Government Organization, or any other reputed organization during the applicable period.**
3. **That in case it is found at any stage that the rates quoted by us are higher than the rates supplied by us to other organizations for the same equipment/item, we shall be liable for necessary action as per the terms and conditions of the tender.**
4. **That all information and documents submitted by us in connection with this tender are true and correct to the best of our knowledge and belief.**
5. **That we agree to abide by all terms and conditions mentioned in the tender document, including any amendments/corrigendum issued, if applicable.**

DEPONENT

Signature: _____

Name: _____

Designation: _____

Name of Firm/Company: _____

Office Seal: _____

Date: _____

Place: _____

VERIFICATION

I, _____, the above-named deponent, do hereby verify that the contents of this affidavit are true and correct to the best of my knowledge and belief and nothing material has been concealed therefrom.

Verified at _____ on this _____ day of _____, 20.

DEPONENT

Signature: _____

ANNEXURE – IX
AFFIDAVIT ON NON-JUDICIAL STAMP PAPER OF Rs. 500/-
(To be submitted to this office)
AFFIDAVIT FOR NON-BLACKLISTING / NON-DEBARMENT CERTIFICATE

Reference: Tender No.: _____

I/We, _____ (Name of the Authorized Representative),

S/o / D/o _____,

aged _____ years, working as _____ of

M/s _____ (Name of Firm/Company), having its registered office

at _____, do hereby solemnly affirm and declare as under:

1. **That M/s _____ (Name of the Firm) has not been found guilty of any malpractice, misconduct, or unethical practices.**
2. **That our firm has not been blacklisted, debarred, deregistered, or prohibited from participating in tenders for the quoted product(s) by the Public Health Department, Government of Maharashtra, any Local Authority, State Government Department, Central Government Organization, Public Sector Undertaking, or any other Government institution as on the date of submission of this tender document.**
3. **That there are no pending restrictions or prohibitions imposed on our firm by any Government authority that would prevent us from participating in this tender.**
4. **That the information furnished above is true and correct to the best of my knowledge and belief, and nothing material has been concealed therefrom.**

DEPONENT

Signature: _____

Name: _____

Designation: _____

Name of Firm/Company: _____

Office Seal: _____

Date: _____

Place: _____

VERIFICATION

I, _____, the above-named deponent, do hereby verify that the contents of this affidavit are true and correct to the best of my knowledge and belief and nothing material has been concealed therefrom.

Verified at _____ on this ____ day of _____, **20**.

DEPONENT

Signature: _____

ANNEXURE – X
AVERAGE ANNUAL TURNOVER AND NET WORTH STATEMENT FOR
LAST THREE FINANCIAL YEARS

(As per mentioned in Technical Bid (Part-I) Clause No. 6)

Tender No.: _____

This is to certify that the **Average Annual Turnover of M/s** _____ for the past three financial years is as given below. The statement is true and correct and has been duly submitted in original and online also, as per the tender clause.

A. Average Annual Turnover Statement

Sr. No.	Financial Year	Turnover (Rs. in Lakh)
1	2023–24	
2	2024–25	
3	2025–26	

Turnover (Rs. in Lakh)

Average Annual Turnover for the last three financial years:

Rs. _____ Lakh

B. Net Worth Statement

This is further certified that the **Net Worth of M/s** _____ **has remained positive during each of the last three preceding financial years**, as mentioned below:

Sr. No.	Financial Year	Net Worth (Rs. in Lakh)
1	2023–24	
2	2024–25	
3	2025–26	

We hereby certify that the Net Worth of the firm/company is **positive in all the above-mentioned financial years**.

Date: _____

Place: _____

Seal of Statutory Auditor: _____

Signature of Statutory Auditor: _____

Name of Statutory Auditor (In Capital Letters): _____

Membership No.: _____

UDIN No.: _____

(UDIN is mandatory)

ANNEXURE – XI
FORMAT FOR DOUBTS / QUERIES FOR PRE-BID MEETING
PRE-BID QUERIES FORMAT

Name of the Supplier: _____

Department Name: _____

Tender Reference No.: _____

Tender Name: _____

Due Date: _____

S. No.	Tender Page No.	Tender Clause No.	Clause Title	Queries / Clarification Requirement	Justification
1.					
2.					
3.					
4.					
5.					

Note:

Bidders are requested to submit their queries/clarifications in the above prescribed format only.

Queries received after the stipulated date/time may not be considered.

The bidder should provide proper justification/supporting details wherever clarification or modification is requested.

Seal of the Bidder: _____

Signature of Authorized Signatory: _____

Name: _____

Designation: _____

Date: _____

ANNEXURE – XII
POWER OF ATTORNEY

Bid No.: _____

(To be executed on Non-Judicial Stamp Paper of Rs. 100/- duly attested by Notary Public)

M/s _____

(Name and address of the Registered Office of the Sole Bidder/Lead Member)

do hereby constitute, appoint, and authorize:

Mr./Ms. _____

S/o / D/o _____

Resident of _____ (Address of residence)

who is presently employed with us and holding the position of _____, as our

Authorized Representative, to act for and on our behalf and to do all such acts, deeds, and

things necessary in connection with or incidental to the bid:

Tender for: _____

Tender Reference No.: _____

including submission of bid documents, signing of documents, providing clarifications, attending meetings,

and representing us in all matters connected with our bid for the said project.

The bid is submitted by:

M/s _____

(Name and address of bidder/consortium members, if applicable)

for a period of _____ years.

We hereby agree to ratify and confirm all acts, deeds, and things lawfully done by our said

Attorney pursuant to this Power of Attorney. All acts, deeds, and things done by our aforesaid Attorney shall

always be deemed to have been done by us.

This Power of Attorney shall be effective, binding, and operative until _____, unless
revoked earlier, or until the said Attorney ceases to be in the service of the Company, whichever is earlier.

Accepted by:

Signature: _____

Name: _____

Designation: _____

Address: _____

(Authorized Representative)

For M/s _____

Signature: _____

Name: _____

Designation: _____

Address: _____

Company Seal: _____

Date: _____

Notes:

1. This Power of Attorney shall be executed by the Sole Bidder/Lead Member, as applicable.
2. The mode of execution of the Power of Attorney shall be in accordance with the procedure, if any, laid down under the applicable laws and the charter documents of the executant. Wherever required, the same shall be executed under the common seal of the company as per the prescribed procedure.
3. Wherever required, the executant(s) shall submit supporting documents for verification, including extracts of charter documents, Board Resolution, or any other authorization document empowering the person executing this Power of Attorney on behalf of the company.

Notary Attestation:

Signature & Seal of Notary Public: _____

Date: _____

Place: _____

ANNEXURE – XIII
UNDERTAKING BY THE BIDDER

Bid No.: _____

(To be uploaded in PACKET – A)

AFFIDAVIT

(To be filled in and signed by the Bidder and submitted on Non-Judicial Stamp Paper of Rs. 200/- duly notarized by Notary Public / First Class Magistrate)

I/We, _____

(Full Name in Capital Letters starting with Surname)

the Proprietor / Partner / Managing Director / Holder of Power of Attorney of

M/s _____

(Name of Business Establishment/Firm/Registered Company)

do hereby, in continuation of the terms and conditions contained in the Tender Form and agreed to by me/us, solemnly affirm and undertake as follows:

1.I/We hereby confirm that I/We shall be able to provide the services/supply offered by me/us at the quoted rates and as per the specifications, terms, and conditions mentioned in the tender document after completion of all required formalities within the stipulated time.

2.I/We hereby state and declare that I/We, whose names and addresses are mentioned herein, have not submitted this e-tender under any other name or through any other establishment/firm/company, nor are we directly or indirectly related to or associated with any person, firm, or establishment that has submitted a bid for the same tender.

3.I/We further undertake that if the condition regarding submission of tender under different names of the firm is found to have been violated, the competent authority shall be at liberty to take appropriate action against me/us as per tender conditions.

4. I/We undertake that the prices quoted by us are the best and most competitive prices as per prevailing market conditions and that we have not offered lower prices for the same supply/service to any other Government/Semi-Government organization or any other agency under similar conditions.

5.I/We agree to comply with all applicable labour laws, statutory requirements, and other enactments applicable to the supply/services throughout the contract period.

6.I/We agree to abide by the rules and regulations of the hospital/institution premises currently in force or which may come into force during the currency of the contract.

7.I/We accept the right of the competent authority to restrict entry of any staff/personnel deployed by us into the premises if such person is considered undesirable or likely to create inconvenience. The decision of the authority shall be final and binding on us.

8.I/We understand and accept that our e-tender/contract shall be liable for rejection/termination and the EMD/security deposit may be forfeited shall not sublet or transfer the work to any other agency without prior written approval of the competent authority.

9.I/We understand and accept that our e-tender/contract shall be liable for rejection/termination and the EMD/security deposit may be forfeited in case:

- a) We fail to keep the tender valid/open as required.
- b) We fail to execute the agreement or deposit the required amount when called upon to do so.
- c) We fail to commence supply/services within the specified time mentioned in the work order/indent.
- d) We fail to provide required information, documents, testimonials, or original records when demanded within the stipulated period.

10.I/We hereby declare that we are not debarred or blacklisted by any Central Government Department, State Government Department, Public Sector Undertaking, Local Authority, or any other Government organization as on the date of submission of this bid. We further declare that neither we nor our firm/company have been convicted under the IPC or Prevention of Corruption Act, and no criminal case is pending against us before any court of law.

11.I/We undertake that if any violation of tender terms and conditions, including any act of misrepresentation, cheating, comes to the notice of the competent authority during execution of the contract, the authority shall have the right to take appropriate action as deemed fit.

12.I/We acknowledge that submission of this tender does not create any binding contract between us and the authority, and the authority shall have no obligation towards us until formal acceptance of the tender.

13.I/We confirm that all information/documents submitted regarding registration, tax details, statutory documents, and other requirements are true and correct. In case any information is found false or incorrect after opening of the tender or award of contract, we shall be liable for action including debarment as per tender conditions.

14.I/We confirm that we have submitted this tender with full knowledge of all liabilities and agree not to raise any objection or dispute regarding any action taken by the authority, including forfeiture of deposit or blacklisting, if any information furnished by us is found incorrect.

15.I/We further confirm that all information and documents submitted by us are true and correct to the best of our knowledge and belief. If any information/document submitted by us is found false or incorrect at any stage, we shall compensate the authority for any loss or inconvenience caused and shall not claim any compensation for cancellation/termination of contract.

16.I/We, _____

(Full Name in Capital Letters)

the Proprietor / Managing Partner / Managing Director / Authorized Distributor of:

M/s _____

hereby offer to provide the required supply/services as mentioned in the tender document in accordance with the prescribed specifications and conditions.

17.I/We undertake to maintain complete quality control over the services/supplies provided by us as per tender specifications. In case our explanation is found unsatisfactory for any deficiency, the authority shall be entitled to take appropriate action including forfeiture of deposit and blacklisting as per rules.

I/We solemnly confirm compliance with all requirements and conditions of the tender document.

Details of Bidder

Full Name: _____

Complete Address: _____

Telephone No.: _____

Email ID: _____

Yours Faithfully,
Signature of Bidder: _____
Name: _____
Designation: _____
Office Stamp/Seal: _____
Date: _____

WITNESSES

Witness 1
Full Name: _____
Address: _____
Signature: _____

Witness 2
Full Name: _____
Address: _____
Signature: _____

Annex No. XIV - Financial proposal Proforma (BOQ)
(In complete Annexure is liable for Rejection)
Proforma of Commercial Bid
[COMMERCIAL OFFER BOQ]

Tender Inviting Authority: Director of Medical Education Research, Mumbai						
Name of Work: Orthopedics Department Medical Equipment Turnkey Project						
Contract No: E - 26 (26-27)						
Name of the Bidder/ Bidding Firm / Company						
<u>PRICE SCHEDULE</u> (This BOQ template must not be modified/replaced by the bidder and the same should be uploaded after filling the relevant columns, else the bidder is liable to be rejected for this tender. Bidders are allowed to enter the Bidder Name and Values only)						
NUMBER #	TEXT #	NUMBER #	TEXT #	NUMBER #	NUMBER #	TEXT #
Sl. No.	Item Description	Quantity	Units	Per Unit Price Including GST & other taxes	TOTAL AMOUNT	TOTAL AMOUNT In Words
1	2	4	5	13	53	55
1	Description					
1.01	Orthopedics Department Medical Equipment Turnkey Project		Nos		0.000	INR Zero Only
Total in Figures					0.000	INR Zero Only
Quoted Rate in Words		INR Zero Only				

Note: -

- 1. The rate for items in the tender, need to quote as per Industry Power and Department Rule No.3.1.2.1. Also, L1 will be determined based on the total cost.**
- 2. Total Rate for items should be provided as per Technical Specification**
- 3. Itemwise rate should be Quoted in Pdf BOQ as per scope of work**

CHECK LIST OF THE TENDER DOCUMENTS

List of Documents/ Information should be Uploaded/ submitted with Tender.

The following documents should be uploaded / submitted with the Tender Document with page number on each document as per the order given below.

Sr. No.	PARTICULARS	Page No
Part I (Technical Bid)		
1)	Proforma to be submitted along-with Technical Bid Annex No. I	
2)	EMD: Online/certificate for exemption of EMD	
3)	Copy of GST Registration Certificate	
4)	GST Clearance Certificate	
5)	Copy of memorandum and articles of Association (If Applicable)	
6)	Copy of Certificate of registration under company registration act 1913 & 1956 (If Applicable)	
7)	Manufacturer Authorization Letter as per Annex No. II	
8)	Technical Compliance Chart as per Annex No. III	
9)	User List of quoted model as per Annex No. IV	
10)	Details of After Sales Service Station Annex No. V	
11)	Terms & Condition Acceptance Certificate as per Annex No. VI	
12)	Consignee List Annex No. VII	
13)	Proforma of Affidavit Annex No. VIII	
14)	Proforma of Affidavit Annex No. IX	
15)	A. Annual Turnover proforma Annex No. X (The average turnover for the last 3 years from the date of publication of tender should be 40% of total tender cost.), B. Net Worth Statement.	
16)	Format for doubts/queries for Pre-bid meeting As per Annex No. XI	
17)	Power of Attorney Annex - XII	
18)	AFFIDAVIT (under taking to be signed by the bidder) As per Annex -XIII	
19)	Financial proposal performa (BOQ) as per Annex No. XIV	
20)	Proforma for Price Justification Annex No. XV	
21)	Scope of Supply (Technical Offer)	
22)	Original Product Catalogue / Brochure with technical data sheets	
23)	Original Tender Form duly signed and official seal of the company on every page of the tender.	
24)	The requisite certificate ISO, ISI, UL, BIS, CE, USFDA MD 42 ,MD 5,MD 9,MD 15,etc as per Technical Specification	
25)	Detail of Products supplied to Govt./Pvt. In India in last 03 Financial years.	
	Total Documents in Technical Bid Page No. to	
(Commercial Bid)		
	Financial proposal performa (BOQ) as per Annex No. XIV	

Note :- If, during online bid preparation, any need arises to upload additional documents, apart from the above mentioned documents, an option to upload additional documents has been provided in the e-Tendering software which will be available to bidders during online bid preparation stage.

Technical Specification For Robotic system specification

Overview

Robotic system plays a role in assisting surgeon in surgery with preoperative planning of knee arthroplasty, and the surgical planning software reconstructs the patient's CT image data with 3D images to provide optimal implant size and installation surface information. Before surgery, the surgeon uses the surgical planning software to establish the surgery plan, and the established planning information is stored on removable storage device and transferred to the robot system. The robot system uses the stored plan information according to the procedure of Operation Software to grasp the location and shape of the joint, implant mounting surfaces and cut the bones. When the cutting is completed, the insertion of an implant for the joint is performed by the surgeon.

This robotic system has to be a fully active joint replacement robotic system with 6-axis articulated robotic arm which secures a wider surgical area for superior surgical results. It has to enhance convenience in surgery by providing optimal alignment with an optical tracking sensor.

The real time gap-check function should allow the surgical plan to be modified at any time, even during the surgery.

The Basic features of robotic system should be as follows :

- 1) Robotic system is a surgical robot that automatically cuts the patient's knee bones with a CT image based planning and an optical Tracking camera.
- 2) Planner-Allows you to plan implant sizes in advance with the patient's CT images and helps surgeon find the optimal implant.
- 3) Users can choose between two surgical methods: 'full surface'(Measured Resection) and 'extension surface'(Modified Gap Techniques).
- 4) Gap check and intra-operation plan change are supported to provide more flexible surgical operation.
- 5) User-friendly UI and pendant can operate all functions.
- 6) The Manipulator consists of 6-axis and supports a wide range of surgery operation area. For safe surgery the safety functions are supported such as emergency stop, force sensor-based cutting force detection, device monitoring, and manual guide.

Robotic Vision Cart Specification should have:

- 27" Medical grade touch enabled monitor
- NDI Polaris Optical Tracking System (OTS)
- "ALWAYS ON" Robotics for uninterrupted performance
- Futuristic ergonomic design
- Integrated keyboard & track pad
- Operating Software with dual colour scheme

Robotic Arm specification should have:

- Human Like footprint
- Intelligent sensors for Adaptive performance
- Next generation 6 axis collaborative robot
- Sleek Robotic Arm, Smooth & silent performance
- Adaptive bone tracking
- Multi-directional maneuverability

Planner should have:

- Flawless surgical planning
 - AI based Segmentation to happen within 3-5min
 - Auto implant size detection
 - Auto implant overhang & notching detection
 - CPU i7 or i9 Intel processor -12th generation or above with clock speed 3GHz or above
 - RAM 32 GB or more GPU type Nvidia RTX 3070 or above
 - Planner features like loading of case, DICOM image upload / processing. Software to autobone segmentation, 3D model generation,
- **Product & Class** should be Automatic Robotic Surgical System (Class 3)
 - **Planning** should be 3D CT Base Surgical Planning
 - **Safety** parameters should have Emergency stop & Manual guide
 - **Detection** should be OTS (Optical Tracking System)
 - **Robot Precision** should have Submillimetric positioning accuracy
 - **Cutting** parameters should have the following features :
 - Fully automatic bone resection
 - No cutting guides (jigs) required
 - Tibial cut including keel preparation
 - Fully finished femur with all the cuts including peg holes
 - Robot Cutting: Max. 50mm/sec
 - Should have milling tool for all the cuts including box, peg holes and tibial keel
 - Should have speed control during cutting
 - **The dimension and weight should be as follows**
 - Vision Cart - 580x725x2100mm (Width x Depth X Height)
 - Arm Cart - 450x530x1650mm (Width x Depth X Height)
 - Vision Cart Weight: 210 KG
 - Arm Cart Weight: 160 KG
 - Compact hardware, minimize product space
 - **The system should be intraoperatively able to perform the following:**
 - Gap check/ Gap Balancing
 - Modification in plan during surgery

- Various cutting options (Full or partial)
- Fine-tune implant placement with accuracy during the procedure

➤ **In addition the system should be able to offer:**

- Quick calibration, verification & registration
- Better stability while taking cuts
- Bone movement monitoring - if bone is moved more than 4 mm robot should stop, and also should have auto recovery option (robot will start from where it has stopped)
- It should also have backup facility for uninterrupted surgical experience

TECHNICAL SPECIFICATION FOR MODULER ICU

1	Ahu (Air Handling Unit) – According to Room CFM Capacity with EC (Electronically Commutated) Fan The AHU shall be of standard floor-mounted design, suitable for indoor installation, delivering a total air volume of 3500 CFM at a static pressure of 125 mm WG. The unit shall be equipped with high-efficiency EC (Electronically Commutated) fans, designed for energy efficient and reliable air handling performance in clean environments. #Double-skin construction with 45 mm thick PUF insulation. #Outer skin: 0.55 mm thick pre-coated G.I. sheet. #Inner skin: 0.55 mm thick plain G.I. sheet. #Frame: Aluminium extruded profiles, modular design without thermal bridge. #Base frame: 100 mm deep, constructed from 16-gauge thick G.I. sheet, with aluminium die cast corner joints for structural rigidity. #High-efficiency backward-curved EC fans, dynamically balanced for low vibration and noise levels. #Fan speed electronically controlled, #Fan design shall ensure uniform airflow and reduced power consumption. #6-row Deep Dx Cooling Coil, copper tubes with aluminium fins, designed for optimal heat transfer. #Pre-filter: 10-micron rating, flange-type, non-woven synthetic media, installed before the cooling coil. #Fine filter: 3-micron rating, flange-type, non-woven synthetic media, placed after the blower section. #Provision for return air and fresh air mixing, using aerofoil design aluminium dampers for efficient airflow control.	NOS
2	Condensing Unit (Air-Cooled): Factory-assembled condensing unit complete with Danfoss / Copeland or equivalent compressor, air-cooled condenser coil, and weatherproof housing. Unit to include copper interconnecting piping, expansion valve, liquid line drier, moisture indicator, and hand shut off valves. Suitable for R-22 / R-407C / R-410A refrigerant. Complete with all necessary accessories and electricals for proper functioning .Make Emerson/Voltas/Carrier	NOS

3	Copper piping: Copper tube piping of size 7/8" x 1/2", 20-gauge wall thickness, suitable for refrigerant application. Piping to be insulated with closed-cell nitrile rubber insulation for thermal efficiency and condensation prevention. All joints to be properly brazed using Harris (USA) make silver brazing rods or equivalent, ensuring leak-proof connections. Installation shall include necessary supports, clamps, and sleeves for a secure and vibration-free layout.	
----------	--	--

4	AHU Electrical Control panel : AHU's PCGI 1.2mm thick sheet constructed Electrical panel with Schneider make contactor, MCB, Relay & other controller, GIC/L&T Make Phase preventer, Minilec/Solectron make Timer & TC (UL app.) make failure & other indicator. CG Make	NOS
----------	--	-----

5	GI Sheet Metal ducting works: Ducts shall be made of GI with curves & bends where indicated for easy flow of air an ensured to be air tight by applying silicon sealant after fabrication hangers shall be provided to GI ducts & shall be suspended by means of MS black painted rods. Hangers shall be not more than 2.5 m apart. 24 Gauge (0.63 mm) Thermal Insulation with 13 mm thick For Supply Air Duct Cross link polyethylene for supply air duct. Joints will be lapped with Nitrile tape for better thermal insulation. Thermal Insulation with 9 mm thick For Return Air Duct Cross link polyethylene for Return air duct. Joints will be lapped with Nitrile tape for better thermal insulation.	SQFT
6	Hepa Housing 2X2: Powder Coated 1.5mm thick HEPA Filter mounting, SS304 perforated sheet and Top body construction with Flange . # Global / FIne airsys eqv makes HEPA Filters (EU-14 Standard) , 99.99 Efficiency @ 15mm W/C IPD and 35mm W/C FPD. Knife edge mouting arrangement. # Face velocity 90-110FPM @ Grill level and design to meet class 100 (ISO-5) @ grill level . # SS304 Filter mounting studs with J hook # SS304 CSK Bolts, Zinc coted MS Ceiling Mounting Anchor fastener & 12mm threaded rod and L angle.	NOS
7	Riser Module Filter and Grill: Supply, Installation, Testing and commissioning of Return Air Riser with SS Perforated Grill along with f final Return air riser.	NOS

8	PPGL WALL PANEL :Supply, Installation, Testing & Commissioning of Pre-fabricated Operating Room of freestanding structure, constructed from composite, free standing insulated steel wall puff panels. The pre-fabricated Operating Room will be constructed by Wall panels. The wall panels height will be of 2400mm & thickness of 50mm & the width will be maximum of 1180mm and minimum of 100mm, All PUF panels PPGL sheet of 0.5 mm thick at both side of the panel, in which they are firmly bonded together. The double skinned panel shall be sandwiched with core consisting of rigid polyurethane foam, which has been injected under high pressure, with a minimum density of 40 kg/m3. The individual wall panels shall use with Silicon sealant to provide seamless operating room for joining two panels, no welding should be allowed. All wall-mounted equipment shall be flush mounted and sealed into theatre. The wall panel design and construction shall allow for installation and support of all equipment and provision of openings required for the installations, without affecting rigidity and strength. Access boxes shall be fitted to the rear of all the wall -mounted equipment to enable maintenance to be carried out from outside the operating room.	SQFT
9	PPGL Ceiling Panel : Supply, Installation, Testing & Commissioning of Pre-fabricated modular construction shall be designed and constructed for exact size. The prefabricated Operating Room will be free standing structure from composite free standing insulated steel ceiling panels. Ceiling panel is placed at an angle of 90° with the wall panel. The ceiling will be constructed using 0.5mm Antibacterial PPGL sheet panel at both sides. The thickness of the ceiling panel should be minimum 50mm. The core between two panels will consist of rigid polyurethane foam, which has been injected under high pressure, with a minimum, density of 40 kg/m3. The individual wall panels shall use with Silicon sealant to provide seamless operating room for joining two panels, no welding should be allowed.	SQFT

10	Antibacterial paint: Filling of all joints and cavities with Screeding& putty for metal surface and sanded Flush to provide a joint less Finish primer with metal primer to give a bonding of paint and then sprayed with a water based liquid plastic aseptic and self- sterilizing wall coating system which protects against growth of micro-organism such as mouldid fungi, Yeast and bacteria on its surface Equivalent to a minimum dry film thickness of 300 microns.	SQFT
11	Vnyle Anti Static flooring : Supply & Installation of Anti Static seamless flooring system,2mm thick , with perfectly curved flush coving all joints welded properly Roll.	SQFT

12	Self Levelling for above Flooring: Epoxy based self-levelling - 2 to 3mm for laying the Vinyl .	SQFT
13	PCGI Flushed Double Door: Door complete with air tight door frame made from1 mm thick PCGI sheet and shutters made from 0.8 mm thick PCGI in double skin construction duly insulated/infill with puff having density 40 Kg/m3 Hardware including Door closure, Hinges, Handle, Vision panels, Lock and door Bottom gasket seal etc.	NOS
14	Led Light Fixture : Supply & Installation of Recess mounted IP55 protocol non- hygroscopic based peripheral LED Lights to be mounted on false ceiling . SIZE 600X600	NOS
15	INTERNAL ELECTRICAL WIRING, CONDUITING WITH FIXTURES INSIDE THE OPERATION THEATRE, Wiring with Low leakage current wires of FRLS wires as per requirements including providing and fixing of conduits and boxes, etc. to complete the work in all respect.	LOT
16	Sky light : 600mm*600mm with led lights and design of skies or nature as required at back side of the bed and at false ceiling for nurse station.absorption .	NOS
17	Medical Gas Line Installation Only Existing Line Connection on Bed head Panel . They should ensure that all works carried out are to here commendation made in the Department of Health and Social Securities should provide Oxygen, Vacuum, supply from existing lines .Bidder shall be responsible for supply, installation, testing and commissioning of complete MGPS system inside Connection Bed Head Panel Only including Essential accessories.Terminal units should be gas specific and only accept the correct Medical Gas probe.	LOT
18	Control Panel: Providing Supply, Installation of membrane type control panel flush in the theater wall with distribution Board complete with all accessories etc. The Control Panel Comprising of the Following :- A)HVAC ON/OFF B)Temperature & Humidity displayed	NOS
19	Monitor Stand	LOT
20	Gas Outlet with Bed Head Panel on Wall Gas of Oxygen x1 nos , Vac x1 No, and Switch Socket.	LOT
21	ICU Curtain With Track	LOT
22	Demolition of Brick wall & Plaster work	LOT

Table Specifications for Orthopedic OT Table with attachments

Operating Table TENDER SPECIFICATIONS
Mobile Operating table should be an Electromechanical/Electrohydraulic Operating Table with C-Arm Image Intensifier compatible with Electrically positioned tabletop suitable for all surgical specialties
The table should be both US FDA/510K exempted and European CE Approved.
The table base and column should be made of high-quality Stainless steel and should have easy to clean surface
The table should have 5 segments, head section, Upper Back Section, Back section, Seat section and non-split leg section. The back and leg section should be motorized, and leg section should be individually Motorized.
The table should have 4 easy running rotatable Double joints, electro conductive Wheels for easy maneuverability of the table. The diameter of the wheel should be 96mm or more
The table base should be equipped with a manual 5th wheel with the gas spring support, is always on and supports directional driving and table maneuvering
Table should also have self-compensatory floor locks for uneven surfaces in the OR to provide better stability to the table during the surgery. The Floor Lock should compensate for at least 10mm of uneven surfaces. In case of emergency, a quick emergency button should be located at the base of the table
The total weight capacity of the table should not be less than 454kgs Centered Column and 270 kgs in dynamic weight
The table base is equipped with a battery charge indicator and service display informing technical staff about the need to perform a table maintenance and possible types of faults (service codes for quick fault diagnosis).
TABLETOP
The dimensions of the table column of 323 mm x 323 mm.
Sockets for connecting the remote control on both sides of the column from the legs and from the head side.
The operating tabletop should have 4 motorized joint, the back section and leg section of the table-top should be motorized. The leg section should be individually motorized
The table-top is equipped with seamless, removable mattress, with anti-decubitus properties, with 90mm thickness . Table-top translucent to X-ray for its entire length, no metal cross rails which could cause obstruction for imaging area.
The table-top, column and table base are easy to keep clean. Side rails of the table- top without visible screws and a one-piece base cover without any joints, gluing or welding.
The table is equipped with a collision Prevention, which prevents collision of the table-top components and prevents, the table-top components against hitting to the floor and itself.
The table should have synchronized Zero Leveling position, all the segments should move simultaneously to the Zero Position
The table is equipped with a hook coupling points system protecting against incorrect assembly of table-top components, with sensors recognizing the mounted component and its length

Should have motorized Longitudinal sliding of 400mm, 200mm towards head section & 200mm towards leg section to facilitate access for C-arm image intensifier.
The table should have reverse orientation for better C arm imaging capabilities
The head and leg section of the table should be detachable and interchangeable
The remote control should have following functions Up/Down, Longitudinal Shift, Trend/Rev Trend, Rt/Lt Tilts, Flex/Reflex, Back section up/Down, Leg Section Up/Down (Simultaneous and Individual), Beach Chair Position, Zero Positions and Floor locks
The table should have back-up system present on the column in the form a display. The following movements should be controlled from the column Patient orientation, Longitudinal Shift, Up/Down, Trend/Rev Trend, Rt/Lt Tilts, Back section up/Down, Leg Section Up/Down and Zero Positions and Floor locks. The Column keypad should also have battery status indicators.
The base of the table should have LED Strip as a safety feature which gives visual and audible feedback on the status of floor locking and unlocking and Battery status If the table is in unlock position the LED strip should lit up Yellow IF the table is in locked position the LED strip should lit up green If the battery is low or there is an error the LED strip should lit up in Red
ELECTRICAL SPECIFICATIONS
The table should be equipped with maintenance free rechargeable batteries with a capacity for about a week's use in the operating room i.e., 5 days with 8 Hrs. working
Recharging of the batteries and supply of the operating table by means of a mains cord
Nominal mains voltage (Selectable) 220/230-240V AC Via mains cord.
The maintenance-free batteries contained in the column are charged by an internal power pack, allowing use with a direct power supply. Housing made from disinfectant-compatible, easy-to-clean rust-free stainless steel.
TECHNICAL SPECIFICATIONS:
Length: 2270mm without upper back section and 2580mm or more with upper back section
Width of the top with side Rails – 600mm
Minimum Height (without mattresses) 598mm or less
Maximum Height (Without Mattresses) 1148mm or more
Maximum Lateral Tilt: 25/25 Degree
Trendelenburg: 35 Degree
Anti-Trendelenburg: 35 Degree
Longitudinal Shift - 400mmcms, 200mm towards head section & 200mm towards leg section to
Leg Section - 90 Degree/105 Degree
Back Section - 90/90 Degree
Facility for single button levelling, levelling of all the segments should be simultaneously
CERTIFICATIONS
The Table must meet the most international regulatory requirement EN 60601-1, IEC 60601-2-46, EN ISO 13845, EN 60601-1-6, EN ISO 19054,
The table should have IP X4
The table should be European CE Certified and US FDA Approved

Accessories to be provided with OT table – configuration with each table.

Accessories to be provided with OT table – configuration with each table	Qty
Remote control	1
Upper Back Section with Pad	1
Head section double joint with pad	1
Leg section two-part with spread with pad	1
Pad for head positioning	1
Anesthesia screen with radial Clamp	1
Body strap rotatable	1
Arm support with pads	2
Orthopedic Accessories	Qty
Extension adapter	1
Universal support	1
Rectangle Shape Supporting pad universal	1
Counter traction post 150	1
Double joint strut, pair	1
Extension strut medium	2
Adapter spindle traction mechanism	2
Spindle traction mechanism	2
Traction boot	2
Docking trolley for extension unit	1
Rail clamp for extension strut	1
Clamp basic	1
Side rail extension	1
Clamp radial setting	1
Base Support	1
Tibia Accessories	Qty
Counter traction post for tibia	1
Adapter counter traction post pivoted	1
Counter traction post 60	1
Condylar fixation device	1
Leather extension sandal right	1
Leather extension sandal left	1
Hip Endo Prosthesis	Qty
Universal lateral support height adjustable with Back, Buttocks & Pubic support (1 each- total 3 supports)	3
Tunnel pillow II	1
Femur in Lateral Position	Qty
Counter traction post for femur	1

Arm holder for lateral position with radial clamp	1
Hand Operating Table	Qty
Hand operating table	1
Traction device for Humerus	1
Counter Traction for Humerus	1
WEINBERGER hand traction device	1
Arthroscopy Knee Accessories	Qty
Arthroscopy positioning device basic	1
Clamp radial setting	1
Meniscus positioning device	1

TECHNICAL SPECIFICATION FOR TOURNIQUET

The Dual Channel Electronic Tourniquet System should have:

- Ability to operate two single bladder cuffs or one dual bladder cuff during a procedure
- Large Display- Easy to see pressure and time on tourniquet.
- Displays and stores reliable, real-time data
- A 5-hour internal battery backup
- Visual indicators when cuff inflating and reached necessary pressure.
- Default Readings
- Time range should be 1-240 minutes
- Pressure range should be 100-475mmHg
- IVRA Lock feature - Used in Bier Block procedures. Ensures tourniquet cuff not deflated until unlocked.
- Avoids user accidentally deflating one cuff and causing local anesthetist to rush into patient's system.
- Printer capability – gives inflation and deflation time of the tourniquet
- Re-Usable and Disposable Cuff Option
- Single port and Dual port option in cuff
- USFDA approval.

Technical Specification For Anaesthesia Workstation

General

- Should have provision for delivery of Oxygen, Nitrous oxide and medical Air
- The machine should be capable of delivering Low flow anesthesia.
- The anesthesia machine with circle absorber, Ventilator and Vaporiser should be from same manufacturer and should be CE approved.
- Should have independent attachments for connecting central gas supply and pin indexed cylinders.
- Should have Non Interchangeable pipeline hose inlet connection to pipelines for medical Oxygen, Nitrous Oxide, and medical AIR
- Should have large size pressure gauges, for easy visibility, colour coded, two each for Oxygen and Nitrous oxide; One for AIR
- Anesthesia machine frame shall be manufactured in strong but lightweight material.
- The machine shall have a maximum of four castors/wheels for manoeuvrability with front two castors having brakes
- Should have minimum two non-lockable and spacious drawers for storing accessories
- The frame shall have GCX compatible channels incorporated for mounting other accessories.
- The option for an integrated independent Oxygen flow meter for Oxygen delivery is mandatory.
- Frame shall accommodate up to two backup cylinders one each for Oxygen and Nitrous Oxide
- On activation of the system on/off switch gas flow and vaporization shall immediately be available
- Should have selector switch for selecting Open and Close circuit operation
- The common gas outlet shall be mounted at the front of the machine (not rear or side) in order to be easily accessible in the event of an emergency and for use of alternate breathing circuits
- The option for illumination of the writing table/work surface is mandatory
- The frame should have integrated power outlets to supply a minimum of three external devices
- Should have Top shelf ,Manoeuvring handle and foot rest
- Machine should have table top work space to keep drugs.
- The unit should have a battery back-up facility for the ventilator in the event of power loss. Minimum 90 minutes battery backup required.
- Input Power : 100 – 240 VAC

Gas Flow

- Antistatic and Cascaded dual flow tubes should be available for all gases (O₂, N₂O & AIR) to allow suitable resolution and accurate control at low total fresh gas flows.
- The flow range shall be 150ml-10 lpm
- Should have N₂O cut off facility if O₂ supply fails.
- Should have Oxygen failure alarm both Visual and Audible.
- Should have Oxygen Flush facility bypassing Vaporiser. O₂ flush switch should be conveniently placed for easy accessibility. O₂ flush switch should be non lockable.
- The unit shall have a mechanical Anti-Hypoxic Device (AHD) system to control the ratio of Oxygen and Nitrous oxide. A completely mechanical AHD system, that requires no electricity, is mandatory
- The mechanical AHD system must limit minimum Oxygen levels to approximately 25%
- Should have minimum mandatory Oxygen flow of 100 ml when switched on
- It shall be possible to deliver Air with only basal flow oxygen independent of the abovementioned hypoxic control.
- Gas flow shall be controlled mechanically only
- Visual display of individual gas flows is mandatory, this shall be by physical means such as glass flowmeters independent of electrical power
- Flow meters should have the option of backlight illumination
- In the event of complete power loss and battery failure it shall still be possible to set the fresh gas flow accurately for each gas, manually ventilate (with bag) and continue to keep the patient anaesthetized.

Vaporizers

- The unit should accommodate at least two vaporizers for Anesthetic agent delivery.
- The manifold should only accept Vaporizers with approved Back bar connections and prevent usage of more than one vaporizer simultaneously. Preferably selectatec compatible back bar.
- Vaporizers supplied with the unit shall be routine maintenance free for the life of the product
- Vaporizers supplied with the unit shall be manufactured from lightweight materials to aid in fitting & removal
- Should have an operating flow range of 0.2 – 15 L/min and nominal capacity 250 ml

Ventilator

- Ventilator shall cater for a diverse range of patient groups from neonates to adult patients with restrictive airways
- Ventilator should be Pneumatically driven and controlled electronically

- Ventilator shall have a large 8.4" colour TFT display, for exclusive use of ventilator control and monitoring.
- Control of the ventilator user interface shall be by touch screen and rotary dial
- Ventilator shall have the following ventilation abilities, volume control (VCV), Pressure Control ventilation (PCV), SIMV-V and SIMV-P , Spont/Pressure Support Ventilation with apnoea backup(PSV) ,Manual
- Assisted modes of breathing shall be flow triggered.
- On power up, in the case of an emergency mechanical ventilation shall be available without the need to carry out user or machine self-checks
- Ventilator shall compensate for fresh gas flow and compliance of the entire circuit.
- Should have user adjustable alarms for major parameters such as Tidal Volume, Minute Volume, FiO₂, Airway Pressure, Frequency,
- Apnea alarms must be user adjustable to allow for all operating conditions and phases during Anesthesia
- Ventilator should have the ability to display Patient Spirometry loop.
- Ventilator should also display waveforms for Flow, Volume and Airway Pressure
- Ventilator shall display Compliance and Resistance measurement
- Volume measurement shall be by separate flow sensors for inspiratory & expiratory breathing paths
- The volume measurement flow sensors/transducers shall be housed completely within the breathing system absorber & not remoted via tubes or channels
- Volume measurement sensors should not be disposable type.

Ventilator Parameters

Tidal Volume	-	10ml-1600 ml
Frequency	-	4-100 bpm
I:E Ratio	-	4:1 to
1:10 Inspiratory Pause of Ti	-	0-60%
PEEP	-	OFF, 3-30 cmH ₂ O
Pressure Limit	-	5-70
cmH ₂ O Minute Volume lpm	-	0.5 to 60

Breathing System

- The breathing system designed so that it can be removed & replaced as a complete unit without the use of tools
- All parts of the breathing system that are in contact with patient gas shall be latex free and Autoclavable except for non autoclavable removable part like O₂ sensor and Pressure manometer.
- Bag/Vent switch shall be integrated on the absorber and should activate ventilator in vent mode and vice versa (One step operation).
- Breathing system should have heater system to avoid water condensation.
- Should have quick release canister for soda lime, capacity minimum 1.5 litre
- The breathing system absorber canisters shall have a bypass system to allow for canister change mid-case without loss of ventilation pressure. The requirement for an automatic bypass without extra input from the user is mandatory
- The ventilator bellows shall be clearly visible and be of upright design. The bellows should ascend on expiration to provide a quick visual indicator for assisted breathing, spontaneous breathing and system leaks.
- Same bellows shall be used for Adult and Pediatric application
- Should have provision for FiO₂ monitoring cell and FiO₂ value should be monitored on the main screen.
- A bag arm with positional adjustment shall be available as standard

Optional

- An option for integrated Sidestream CO₂ monitoring is Mandatory
- An option for integrated SpO₂ monitoring is Mandatory
- Central Brake
- Vaporiser of customer choice
- Suction

TECHNICAL SPECIFICATION FOR HIGH SPEED ORTHO DRILL SYSTEM	
Motor Specs- Qty-1 Nos	
1.	The drill should have an electric motor driven through a console
2.	Maximum motor speed should be 75,000 Rpm
3.	It should reduce thermal necrosis by providing Integrated irrigation.
4.	It should have individual Standalone motor inside.
Console specs- Qty 1 Nos	
1.	It should be a multispecialty console with provision to attach Neuro Drill, ENT Shaver(microdebrider), Small bone saws, Bone Morsilser (Used in cutting of bones in granular level)
2.	It should have touchscreen capability
3.	should have the provision to attach 3 hand pieces and should be able to select any one handpiece by a click of the footswitch
4.	The console should have the provision to use 2 handpieces simultaneously with 2-foot switches.
5.	Power Supply should be 220- 240V only 50-60Hz.
6.	It should identify different hand pieces with display on console
7.	It should be programmable as per surgeon preference.
8.	It should be compatible to more than 7 handpieces.
9.	It should have function of controlling brightness, contrast and alarms on the console
10.	It should have the option of controlling the speeds and features of the different types of hand pieces from the same unit.
FOOTSWITCH: 1 Nos	
Fully programmable footswitch as user needs Able to control following function via footswitch	
Footswitch motion – Forward, Reverse. Oscillation. Switch over to high/low speed. Increase or decrease speed. Can operate 2-3 hand pieces sequentially attached to the same motor control unit.	
Attachments	
Metal cutting attachment straight - 9 CM & 14CM - 1 Nos Each	
There should be Straight & angled attachment- 12 cm, 14cm-1 Nos. each	
Consumables	
1.	Diamond Bur- Size between 1mm to 5mm- Total 10 Nos
2.	Cutting Burr - Size between 1mm to 6 mm- Total 10 Nos
3.	Metal Cutting Burr - 5 Nos
5.	Sterilization Tray- 1 Nos
All products quoted should be from the same manufacturer to offer compatibility of system & should be USFDA & CE approved.	
Dura-Repair- Qty 4 Nos	
1)	It should be Onlay & Suturable
2)	Thickness should be 0.3mm
3)	Prehydrate – 5 Min
4)	Material should be purified Type 1 collagen derived from bovine Achilles Tendon

TECHNICAL SPECIFICATION FOR SUCTION MACHINE

SUCTION MACHINE

1. Oil free noiseless aspiration unit with rapid vacuum build up mounted on a well-balanced trolley with at least four good quality antistatic castors with brakes
2. The machine should be based on aluminium and glass free heat resistant piston cylinder technology made up of long lasting high-tech material
3. System should be oil free and should not contain any glass or stainless-steel components
4. System should have membrane vacuum regulator & easily readable vacuum gauge in kPa and mmHg
5. System should be capable of operating at vacuum levels between 0 to -712mmHg.
6. System should have Clean Touch buttons providing One-touch flow.
7. System should provide multiple preset flow rates of 40 ltr/min, 50 ltr/min and 60 ltr/min with one touch button.
8. The noise level should be below 43 dB at the highest flow rate.
9. Should have two levels of mechanical overflow protection along with an autoclavable Safety Jar made up of PSU (Polysulfone) with auto shut off feature to prevent entry of any type of fluid into the machine.
10. System should have two autoclavable PSU (Polysulfone) collection jars of 3 litres capacity with changeover valve for easier switching from one jar to another jar without turning off the machine
11. The lid of the jars should be autoclavable in order to ensure complete sterility.
12. The safety jar and collection jars along with the lids should be capable of withstanding 134 degree Celsius temperature while autoclaving and without getting deformed or losing clarity.
13. System should have clampholders for directly clamping the jars on the machine rail without the use of screws.
14. System should have high quality housings materials to withstand potent disinfectants and a smooth and a single piece hood without gaps and grooves for maintaining hygiene, easier cleaning and preventing growth of pathogens
15. The machine should have continuous duty cycle
16. System should be free of any disposable filter/part/accessory

17. System should not have any recurring cost
18. System should have Type CF certification for providing protection against electric shock.
19. System should be compliant with IP 21- Ingress Protection to prevent ingress of foreign bodies and any type of fluid
20. System should have color coded LED indicators for status of the machine- Standby, Running and Error
21. System should have valid European CE Certificate as per 93/42/EEC directive on MDD and US FDA Certificate
22. The Equipment shall have valid ISO 13485 Certificate

TECHNICAL SPECIFICATIONS**ELECTRO CAUTERY MACHINE**

SN	SPECIFICATION DESCRIPTION
1	Unit should be Advanced Microcontroller based Technology.
2	Unit should have isolated Monopolar and Bipolar Outputs
3	Unit should have HF leakage monitoring system.
4	Unit should have Equipotential socket for Earth Potential Equalization
5	Unit should have HF Time-Out Facility to prevent accidental activation
6	Unit should be 400W In Monopolar Cut and max. 120W in Bipolar
7	System should have LCD Touchscreen Display.
8	System should have atleast 2 Monopolar sockets, 1 Bipolar socket.
9	Unit should have facility to store 99 programs with surgeons name and procedure name displayed
10	Unit should have Split Type Patient Plate Contact-Quality Monitoring System for maximum Patient Safety. Unit Should Have Audio-Visual Error indication.
11	System should have atleast 4 Monopolar Cut modes including the following modes:
a.	Pure Cut
b.	Lap Cut
c.	Blend (10 Level)
d.	Endo Cut (10 Level)
12	System should have atleast 4 Monopolar Coagulation modes including the following modes:
a.	Dessicate
b.	Fulgurate
c.	Spray
d.	Soft
13	System should have atleast 4 Bipolar modes including following modes:
a.	Standard
b.	Micro
c.	Macro
d.	Auto Start/ Stop
14	System should be safety certifications of IEC 60601-1, IEC 60601-1-2, IEC 60601-2-2.
15	System should be European CE certified with 4 digit Notified Body.
16	Company should be ISO 9001 & ISO 13485 certified company for manufacturing and supply of Electro-Medical Equipment
17	System should be provided with following accessories:
a.	Footswitch - 1 no.
b.	Reusable Monopolar Handswitch Pencil - 1 no.
c.	Reusable Foot-Operated Monopolar Pencil with 5 Electrodes - 1 no.
d.	Reusable Bipolar Cable - 1 no.
e.	Reusable Dual/Split Pad Silicone Patient Plate with Cable - 1 no.
f.	Reusable Bayonet Bipolar forceps, - 1 no.

TECHNICAL SPECIFICATION FOR SHOULDER TRACTION DEVICE

- Should have full range of positioning systems to fit all beach chair shoulder, lateral decubitus shoulder, elbow, wrist, knee, ankle positioning.
- Should be battery operated limb positioner.
- Should offer instantaneous limb repositioning within sterile zone itself
- Should deliver exceptional control in beach chair position for shoulder. Should keep patient stable throughout the surgery. Should deliver reproducible position for patient's arm. Should provide easy access to surgical site.
- Should deliver exceptional control lateral decubitus position for shoulder. Should deliver internal and external rotation while maintaining traction adduction and abduction as easy as stepping on a pedal linear and lateral distraction.
- Should deliver exceptional control in Elbow positioning. Should provide positioning option for lateral prone and supine positions. Provide superior access for both open and arthroscopic procedures. Should provide control of supination and pronation.
- Should deliver exceptional control in Wrist positioning. Should control and measure traction. Should provide control of supination and pronation. Should provide positioning of hand, wrist & distal forearm procedures should provide 360 degree access to wrist and hand.
- Should deliver exceptional control in Knee positioning. Should provide stable limb positioning. Should be able to deliver superb range of motion and access to knee joints. Should be simple in handling to achieve multiple orientations for repositioning and access of knee joints.
- Should deliver exceptional control in Ankle positioning. Set up for ankle should be quick and easy. Should provide measurable traction throughout procedure.
- Should come with a distal activation switch which can be attached to the sterile drape of the patient and can be used to reposition the shoulder by applying small pressure
- Should come with footswitch.
- Should come with universal square rail clamps to be connected in available universal OT table.
- Should have the provision for additional battery backup.
- Should come with the battery charger
- Should come with a cart to make it comfortable to move from one place to another.
- Should be provided with 1 Box of Disposables for Shoulder, Knee, Ankle and Elbow Joints.

TECHNICAL SPECIFICATION FOR ORTHOPAEDIC INSTRUMENT SET

S. No.	Nomenclature	Size and specification of item	Qty needed (in Nos)
1	Artery Forceps Straight	18 cm	16
2	Artery Forceps Curved	18 cm	16
3	Mayo Haeger Needle Holder with Tungsten Carbide Jaws	25cm/10 Inch	4
4	Dietrich/Ryder Needle Holder with Tungsten Carbide Jaws	23cm/9 Inch	4
5	Kocher Artery Forceps Straight	30 cm/12 Inch	24
6	Mcindoe Forceps (Plain)	15cm/ 6 Inch	4
7	Mcindoe Forceps (Tooth)	15 cm/ 6 Inch	4
8	Tooth Forceps with Tungsten Carbide Jaws	25cm/10 Inch	4
9	Mosquito Forceps Curved	12 cm/ 5 Inch	16
10	Allies Tissue Forceps with Tungsten Carbide Jaws	20cm/ 8 Inch	18
11	Mayo Scissors Curved with Tungsten Carbide Jaws	20cm/08 Inch	4
12	Mixer Right Angled Artery Forceps	30cm/ 12 Inch	4
13	BP Handle size No 4	Universal	4
14	BP Handle size No 7	Universal	4
15	Mosquito Artery Forceps	Universal	16
16	Kilner Scissor	Universal	4
17	Pituitary Forceps Straight	12 inches	1
18	Pituitary Forceps. Curved anterior	12 inches	1
19	Verbrugge Self centering bone holding forceps for Forearm Bone	Universal	4
20	Verbrugge Self centering bone holding forceps for Femur	Universal	4
21	Lane's Bone Holding forceps	Universal	4
22	Bone nibbler/rounguer (Jenson's) double action	Universal	4
23	Double action bone cutter (large)	Universal	2
24	Farabeuf periosteal elevator	Universal	4
25	Cannulated Screw Driver 4.5mm Long	18 inches	4
26	Cannulated Screw Driver 3.5mm	12 inches	4
27	Teflon Punch for HRA	Universal	1
28	Screw driver 4.5mm (Hexagonal)	9 inches	4
29	Depth Gauge for Small fragement SS	Universal	2
30	Depth Gauge for Large fragement SS	Universal	2
31	Wire Passer	Universal	2
32	Wire Cutter	Universal	4

33	Wire Twister	Universal	2
34	Reduction Forceps	140mm	2
35	Reduction Forceps	170mm	2
36	Reduction Forceps	220mm	2
37	Reduction Forceps	240mm	2
38	Done Reduction Forceps with Point	130mm	4
39	Bone Reduction Forceps with Point	180mm	4
40	Bone Reduction Forceps with Point	200mm	4
41	Retractor Short Narrow Tip Small	With 16mm	8
42	Retractor Long Narrow Tip (For Hip Surgery)	With 18mm	8
43	Osteotome with Fiber Handle, Straight	5mm	1
44	Osteotome with Fiber Handle, Curved	5mm	1
45	Osteotome with Fiber Handle, Straight	10mm	1
46	Osteotome with Fiber Handle, Curved	10mm	1
47	Osteotome with Fiber Handle, Straight	15mm	1
48	Osteotome with Fiber Handle, Curved	15mm	1
49	Osteotome with Fiber Handle, Straight	20mm	1
50	Osteotome with Fiber Handle, Cury	20mm	1
51	Osteotome with Fiber Handle, Straight	25mm	1
52	Osteotome with Fiber Handle, Curved	25mm	1
53	Osteotome with Fiber Handle, Straight	30mm	1
54	Osteotome with Fiber Handle, Curved	30mm	1
55	Osteotome with Fiber Handle, Straight	35mm	1
56	Osteotome with Fiber Handle, Curved	35mm	1
57	Small Chiser 5 1/2"	Long 4mm	1
58	Small Chiser 5 1/2"	Long 6mm	1
59	Small Chiser 5 1/2"	Long 8mm	1
60	Small Chiser 5 1/2"	Long 10mm	1
61	Small Chiser 5 1/2"	Long 12mm	1
62	Chisel With Fibre Handle, Straight	5mm	1
63	Chisel With Fiber Handle, Curved	5mm	1
64	Chisel With Fibre Handle, Straight	10mm	1
65	Chisel With Fiber Handle, Curved	10mm	1
66	Chisel With Fibre Handle, Straight	15mm	1
67	Chisel With Fiber Handle, Curved	15mm	1
68	Chisel With Fibre Handle, Straight	20mm	1
69	Chisel With Fiber Handle, Curved	20mm	1
70	Chisel With Fibre Handle, Straight	25mm	1
71	Chisel With Fiber Handle, Curved	25mm	1
72	Chisel With Fibre Handle, Straight	30mm	1
73	Chisel With Fiber Handle, Curved	30mm	1
74	Chisel With Fibre Handle, Straight	35mm	1
75	Chisel With Fiber Handle, Curved	35mm	1
76	Hammer With Fiber Handle	500g	2

77	T-Handle With Chuck	Universal	4
78	Bone Hook	Universal	4
79	Periosteal Elevator With Fiber Handle Curved St. Edge, Width	3mm	4
80	Periosteal Elevator With Fiber Handle Curved Round Edge	20mm	4
81	Gouge With Fiber Handle, Curved	13mm	4
82	Gouge With Fiber Handle, Curved	14mm	4
83	Gouge With Fiber Handle, Curved	5mm	1
84	Gouge With Fiber Handle, Curved	10mm	1
85	Gouge With Fiber Handle, Curved	15mm	1
86	Gouge With Fiber Handle, Curved	20mm	1
87	Gouge With Fiber Handle, Curved	25mm	1
88	Gouge With Fiber Handle, Curved	30mm	1
89	Gouge With Fiber Handle, Curved	35mm	1
90	Plaster Cutting Saw- Electric	Universal	1
91	Lorenz Plaster Shears	Universal	1
92	Galpie retractors	Small	4
93	Galpie retractors	Medium	4
94	Galpie retractors	Large	4
95	Mastoid self retaining retractors	Small/ 5 in	4
96	Mastoid self retaining retractors	Medium/ 7 in	4
97	Mastoid self retaining retractors	Large/ 9 in	4
98	SS Scale	12 Inch	2
99	SS Scale	6 Inch	4
100	Screw Holding Forceps	SS Scale	2
101	Screw Holding Forceps	8mm	2
102	Harrington Rod Pin Cutter	Universal	2
103	Pin Cutter 18.5" Stainless Steel	Universal	2
104	K wire cutter	Universal	4
105	Cannulated tap	6.5mm	2
106	Cannulated Screw Driver	6.5mm	2
107	Screw driver 3.5mm (Hexagonal)	3.5mm	2
108	Murphy Skid	Universal	2
109	Putty Rasp	Universal	2
110	Nylon Face Impactor	Universal	2
111	K Nail Extractor with 4 Hooks	Universal	2
112	T-Handle with Quick coupling	Universal	4
113	Flat Nosed Pliers Parallel Tip	Universal	2
114	Plier cum wire Bender cum wire cutter	Medium	2
115	Plier cum wire Bender cum wire cutter	Large	2
116	Bone AWL	Straight	2
117	K wire Bender	Universal	2
118	Bandage cutting Scissor	Universal	10
119	Cobra retractor for hip surgery	Large	4

120	Bone Plate Bender	Universal	2
121	Self centering Bone Holding Forceps	Large	2
122	Reduction Forceps-serrated-speed Lock	200mm	4
123	Bone File	Universal	1
124	Bone Curette (Scoop)	4x6	1
125	Bone Curette (Scoop)	6x9	1
126	Volkman Bone Curette, double ended	Universal	2
127	Mallet	Small	1
128	Mallet	Medium	1
129	Mallet	Large	1
130	Hohman Retractor	Small	4
131	Hohman Retractor	Medium	4
132	Hohman Retractor	Large	4
133	Dressing Scissor	6 inch	10
134	Dressing Scissor	8 inch	10
135	Mayo Scissor	5 inch	2
136	Mayo Scissor	7 inch	2
137	Mayo Scissor	9 inch	2
138	Stitch removal scissor	5 inch	5
139	Perforated Sterilization box, stainless Steel	9 inch x 9 inch	4
140	Sterilization Drum, stainless Steel	11 inch x 9 inch	4
141	Surgical instrument tray, stainless steel for ortho instruments	Large size (approx 16 x 8 in)	6
142	Plain Forceps	7 inch	4
143	Plain Forceps	9 inch	4
144	Sponge Holding Forceps	Double	10
145	Towel Clip	8", 10"	16
146	Towel Forceps	4,5,6	16
147	Skin Hook	Serrated	4
148	Ilizarov Wire tensioner	Universal	1
149	Periarticular reduction forceps, ball tipped- large fragments	Universal	2
150	Bristow Periosteal Elevator	Universal	4
151	Verbrugge-Muller bone lever 65 mm wide	65 mm wide	2
152	Humeral head retractor	Universal	2
153	Midi LAMBOTTE Osteotome	12 mm	4
154	Liston Bone Cutter- 7.5in Straight	7.5 mm	2
155	Broken Screw retriever	3.5 mm	1
156	Broken Screw retriever	4.5 mm	1
157	Broken Screw retriever	6.5 mm	1
158	wire guide medial malleolar clamp with	Universal	1

159	TENS extractor	Universal	1
160	Double drill guide	4.5mm/3.2mm drill bits	2
161	Drill sleeve	2.8mm drill bits	2
162	Universal drill guide for neutral and load position	4.5mm/3.2mm drill bits	1
163	Universal dill guide for neutral and load position	2.8mm drill bits	1
164	Holding sleeve for screw	3.5 mm screws	2
165	DHS/DCS Triple reamer set	Complete set, long 38mm barrel length	1
166	DHS/DCS Triple reamer set	Complete short	1
167	Langenbeck right angled retractor	$\frac{1}{4}" \times 1"$	2
168	Langenbeck right angled retractor	$\frac{1}{2}" \times 1\frac{1}{2}"$	2
169	Langenbeck right angled retractor	$1" \times 2"$	2
170	Langenbeck right angled retractor	$\frac{3}{4}" \times 1\frac{3}{4}"$	2
171	Langenbeck right angled retractor	$\frac{1}{4}" \times 1"$	2

Digitally signed by
ANIL BHANDARI
Date: 13-08-2026
15:20:16

Commissioner
Medical Education, Research & Ayush