

ID	Assessment	Evidence retained for the defense
REQ-MOTION	DEMONSTRATED	Closed-loop DF/PF tracking with sub-degree RMSE in the reported tests.
REQ-USE	PLATFORM	Integrated operation on a seated / non-weight-bearing mannequin setup.
REQ-FIT	DESIGN	Adjustable tibial fixation, sliding foot length, and mechanism positioning implemented.
REQ-ALIGN	DESIGN	Tibial rail and adjustable tracking mechanism allow pre-operation repositioning.
REQ-COMP	DESIGN	Spring-linked tracking mechanism provides limited passive out-of-plane accommodation.
REQ-ACT	FUNCTIONAL	Opposing flexor/extensor SMA groups generated bidirectional sagittal motion.
REQ-TRANS	FUNCTIONAL	Bowden/PTFE transmission operated throughout tracking, repeatability, and endurance protocols.
REQ-SENSE	FUNCTIONAL	Hall-effect encoder supplied the relative-angle feedback used in closed loop.
REQ-CTRL	PARTIAL	BPID tracking, directional allocation, and mutually exclusive final commands were evidenced.
REQ-HEAT	PARTIAL	PWM modulation, zero-command intervals, mutual exclusion, and scheduled recovery reduce unnecessary command-level heating.
REQ-EVAL	ENGINEERING	Campaign covered transmission condition, tracking, repeatability, and prolonged active/recovery operation.

Main verification still pending
Human fit/comfort and pressure; independent angle metrology; force/torque and thermal/electrical measurement; fault-injection and robustness tests; multi-prototype and ethics-approved human evaluation.