CERTIFICATE OF QUALITY ASSURANCE

Davis

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Sensor: 6327 Wireless Vantage Pro2 Plus, ISS
Unit Serial #: 250815N07
Date of Calibration: 2025-08-15
Date Calibration Expires: 2026-08-15
Specified Humidity Accuracy: ±2% RH: We tested and certified accuracy at 33, 80, and 90%. The design of the product and physics of the sensor assures that the 2% accuracy will be maintained down to the minimum operating range limit.
Specified Temperature Accuracy: ±0.5°F:



Type of Calibration: New Instrument
Received Condition: N/A
As left condition: In Tolerence

Humidity Reference	Calibration Date	Expiration Date	Serial #:
Vaisala HMT-333	13-MAR-2025	13-MAR-2026	P4050812

Temperature Reference	Calibration Date	Expiration Date	Serial #:
Vaisala HMT-333	24-OCT-2024	24-OCT-2025	R14540911

Procedures Used :

Procedure Name	Description
NIST- HUM Rev. E	NIST – Humidity Calibration
NIST-TEMP Rev. H	NIST – Temperature Calibration

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained. I certify that this sensor has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____ DATE ___ 08/15/2025____

File last revised 11/13/15

Powered by CamScanner

CERTIFICATE OF QUALITY ASSURANCE

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Product: 6327 Wireless Vantage Pro Plus ISS
Sensor Serial #: 250815N08
Date of Calibration: 2025-08-15
Date Calibration Expires: 2026-08-15

Specified Accuracy: ±3% + 1 rainfall count at 2.21"/hr. with 1.00" delivered.

Test Condition (Temp/RH) : 75 Deg. F / 55 %

Type of Calibration: New Instrument
Received Condition: N/A
As left condition: In Tolerence

Standards used to Calibrate Equipment

Volumetric Reference	Calibration Date	Expiration Date	Serial #:
CAVRO XLP6000 Pump	04-DEC-2024	04-DEC-2025	0602252467

Procedures Used :

Procedure Name	Description
NIST- RAIN Rev. D	NIST – Rain Calibration

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained.

I certify that this sensor has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____Toha_____ DATE ___ 08 /15/2025_____

File last revised 3/21/24

Powered by CamScanner

CERTIFICATE OF QUALITY ASSURANCE

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Sensor: 6410 Anemometer
Sensor Serial #: 250815N09
Date of Calibration: 2025-08-15
Date Calibration Expires: 2026-08-15

Specified Accuracy: Greater of ±2 mph or ±5%, whichever is greater (as tested facing into the wind)

Test Condition (Temp/RH) : 75 Deg. F / 55 %

Type of Calibration: **New Instrument**
Received Condition: **N/A**
As left condition: **In Tolerence**

Standards used to Calibrate Equipment

Wind Speed Reference	Calibration Date	Expiration Date	Serial #:
Alnor AXD610	09-OCT-2024	09-OCT-2025	AXD611336002

Procedures Used :

Procedure Name	Description
NIST- WIND SPEED Rev. D	NIST – WIND Calibration

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained.

I certify that this sensor has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____Toha_____ DATE _____ 08/15/2025_____

File last revised 3/21/2024

Powered by CamScanner