CERTIFICATE OF QUALITY ASSURANCE

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Product: 6262 Wireless Vantage Pro Plus
Product Serial #: 6313U22500200873
Date of Calibration: 2023-08-07
Date Calibration Expires: 2024-08-07

Accuracy: ±0.03" Hg between 20.00" Hg and 30.00" Hg at room temperature

Type of Calibration: New Instrument
Final Calibration Status: Meets Specifications

Barometer Reference	Calibration Date	Expiration Date	Serial #:
Vaisala PTB220 Class A	28-FEB-2023	28-FEB-2024	V5120004

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained.

I certify that this barometer has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____ DATE ____8/7/2023___

CERTIFICATE OF QUALITY ASSURANCE

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Sensor: 6262 Wireless Vantage Pro2 Plus
Unit Serial #: 230807N04
Date of Calibration: 2023-08-07
Date Calibration Expires: 2024-08-07

Specified Humidity Accuracy: ±2% RH: We tested and certified accuracy at 33, 80, and 90%. The design of the product and physics of the sensor assures that the 2% accuracy will be maintained down to the minimum operating range limit.
Specified Temperature Accuracy: ±0.5°F:



Type of Calibration: New Instrument
Final Calibration Status: Meets Specifications

Humidity Reference	Calibration Date	Expiration Date	Serial #:
Vaisala HMT-333	02-DEC-2022	02-DEC-2023	P4050812

Temperature Reference	Calibration Date	Expiration Date	Serial #:
Vaisala HMT-333	02-DEC-2022	02-DEC-2023	P4050812

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained. I certify that this sensor has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____ DATE ___8/7/2023_____

CERTIFICATE OF QUALITY ASSURANCE

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Product: 6262 Wireless Vantage Pro Plus
Sensor Serial #: 230807N05
Date of Calibration: 2023-08-07
Date Calibration Expires: 2024-08-07

Specified Accuracy: ±3% + 1 rainfall count at 2.21"/hr. with 1.00" delivered.

Type of Calibration: New Instrument
Final Calibration Status: Meets Specifications

Volumetric Reference	Calibration Date	Expiration Date	Serial #:
CAVRO XLP6000 Pump	17-FEB-2023	17-FEB-2024	802008695

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained.

I certify that this sensor has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____ DATE ___8/7/2023_____

CERTIFICATE OF QUALITY ASSURANCE

Davis Instruments · 3465 Diablo Ave., Hayward, CA 94545, U.S.A.
Tel. (510) 732-9229 · Fax (510) 732-9188

Sensor: 6410 Anemometer
Sensor Serial #: 230807N06
Date of Calibration: 2023-08-07
Date Calibration Expires: 2024-08-07

Specified Accuracy: Greater of ±2 mph or ±5%, whichever is greater (as tested facing into the wind)

Type of Calibration: New Instrument
Final Calibration Status: Meets Specifications

Wind Speed Reference	Calibration Date	Expiration Date	Serial #:
MKS Baratron	12-JUL-2023	12-JUL-2024	000335058

This instrument has been calibrated to specifications using a reference traceable to the National Institute of Standards and Technology (NIST). Davis Instruments Corp. certifies that this instrument meets or exceeds published measurement specifications (unless otherwise noted). Calibration records for inspection, measuring, and test equipment are maintained.

I certify that this sensor has been inspected and tested in accordance with the Davis Instruments Corp. quality assurance program.

SIGNED _____ DATE _____8/7/2023_____