
Final test checklist - AD2CP/UNS





Certificate of Calibrations and Tests

Page 1 of 2

Instrument Information

Customer Reference No.	R50849-10
RMA No.	8574
Instrument Type	Echosounder 70-120kHz Glider
Instrument Frequency	100 kHz
Instrument S/N	105712
Head S/N	D-5712
Interface Board S/N	105712
Digital Board Mfr. S/N	D D0 13887
Analog Board Mfr. S/N	HzA B0 03709
Interface Board Mfr. S/N	EI CO 7774
Digital Board Rev.	Admin to the substantial and a
Analog Board Rev.	
Interface Board Rev.	

Calibrations and tests performed

	and the second		
1	Maria a access	Daggod	
	Pressure	Passeu	
		The state of the s	

All the tested values are within Nortek AS specifications

April 24, 2025

Date

(im Junioudhagin Reviewed and approved (sign.)



Details

Instrument Type	Echosounder 70-120kHz Glider	
Instrument S/N	105712	
Pressure Range	1500 dBar	
Date	April 8, 2025	
Operator	Nicolas Etchart	
Location	Nortek Factory Norway	
Result	Passed	

Description

Verification is performed in an automated pressure chamber. Fixed-point measurements are collected to verify the sensor. Criteria of acceptance is \pm 0.1% of full scale.

Reference: Paroscientific, Inc. - Digiquartz 9000-10K-101. Accuracy 0.01% of 6895dBar.

Verification Results

Reference (dBar)	Pressure Diff. (dBar)	Pressure Diff. (% of FS)
155:45	-2,80	-0.19
345.96	-1.75	-0.12
472,95	-1.37	-0.09
648.93	-1.27	-0.08
781.97	-1.10	-0.07
909.18	-0.65	-0.04
1076.15	-0.08	-0,01
1235.36	0.67	0.04
1369.98	1.57	0.10
1496.68	2.80	0.19