PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

ECO Chlorophyll Fluorometer Characterization Sheet

Date: 8/2/2023 S/N: FLBBCDSLC-8411

Chlorophyll concentration expressed in µg/l can be derived using the equation:

CHL (µg/I) = Scale Factor * (Output - Dark counts)

Dark counts Scale Factor (SF) Maximum Output Resolution

Ambient temperature during characterization

Digital

45 counts

0.0072 µg/l/count

4130 counts

1.1 counts

23.0 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: SF = x + (output - dark counts), where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyli-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (Thalassiosira welssflogil). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

Scattering Meter Calibration Sheet

1/0/1900

Wavelength: 700

S/N FLBBCDSLC-8411

Use the following equation to obtain either digital or analog "scaled" output values:

 $\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor } \times \text{ (Output - Dark Counts)}$

Scale Factor for 700 nm

1.825E-06 (m⁻¹sr⁻¹)/counts

Output

= meter output counts

Dark Counts

48 counts

Instrument Resolution

1.0 counts

Definitions:

- Scale Factor: Calibration scale factor, $\beta(\theta_o)$ /counts. Refer to User's Guide for derivation.
- · Output: Measured signal output of the scattering meter.
- Dark Counts: Signal obtained by covering detector with black tape and submersing sensor in water. Instrument Resolution: Standard deviation of 1 minute of collected data.

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

ECO CDOM Fluorometer Characterization Sheet

Date: 8/2/2023 S/N: FLBBCDSLC-8411

CDOM concentration expressed in ppb can be derived using the equation:

CDOM (ppb) = Scale Factor * (Output - Dark Counts)

Dark Counts Scale Factor (SF) Maximum Output Resolution

Ambient temperature during characterization

Digital

48 counts 0.0909 ppb/count 4130 counts

1,3 counts

23.0 °C

Dark Counts: Signal output of the meter in clean water with black tape over detector.

SF: Determined using the following equation: SF = x + (output - dark counts), where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

Maximum Output: Maximum signal output the fluorometer is capable of.

Resolution: Standard deviation of 1 minute of collected data.