
Final test checklist - AD2CP/UNS





Certificate of Calibrations and Tests

Page 1 of 2

Instrument Information

Customer Reference No.	50334-1585 8575 Echosounder 70-120kHz Glider	
RMA No.		
Instrument Type		
Instrument Frequency	100 kHz	
Instrument S/N	105815	
Head S/N	D-5815	
Interface Board S/N	105815	
Digital Board Mfr. S/N	D D0 13901	
Analog Board Mfr. S/N	HzA B0 03702	
Interface Board Mfr. S/N	EI CO 7767	
Digital Board Rev.	A Section of the Control of the Cont	
Analog Board Rev.		
Interface Board Rev.		

Calibrations and tests performed

and the second contract of the second contrac	
Bussesses	Pascod
Pressure	rasseu

All the tested values are within Nortek AS specifications

April 24, 2025

Date

um <u>| Lumerudhagen</u> Reviewed and approvep/(sign.)



Details

Instrument Type	Echosounder 70-120kHz Glider
Instrument S/N	105815
Pressure Range	1500 dBar
Date	April 22, 2025
Operator	Nicolas Etchart
Location	Nortek Factory Norway
Result	Passed

Description

Verification is performed in an automated pressure chamber. Fixed-point measurements are collected to verify the sensor. Criteria of acceptance is \pm 0.1% of full scale.

Reference: Paroscientific, Inc. - Digiquartz 9000-10K-101. Accuracy 0.01% of 6895dBar.

Verification Results

Reference (dBar)	Pressure Diff. (dBar)	Pressure Diff. (% of FS)
166,62	-1.70	-0.11
329.15	-1.26	-0.08
483.43	-0.84	-0.06
637.54	-0.39	-0.03
759.07	-0.06	0.00
939.60	0.51	0.03
1078.34	0,53	0.04
1221.91	0.91	0.06
1371.65	1,45	0.10
1522.40	1.70	0.11