AANDERAA
a xylem brand

CALIBRATION CERTIFICATE

Form No 854, May 5th 2020

Certificate no: Oxygen Optode 4831_670_249724
FoilID: 1517M

Product: **Oxygen Optode 4831**
Serial no: 670
Calibration date: 23.01.2026

2-POINT RECALIBRATION, MULTIPOINT CALIBRATED OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	338

Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 µM	0 - 120%	-5 to +40°C
Accuracy	< ±3µM or ±2% 1)	±2% 2)	±0.03°C 3)
Resolution	< 1 µM	< 0.4%	0.01°C
Settling Time (63%)	< 25 seconds	< 25 seconds	< 2 seconds

1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter
2) Within calibrated range 0 - 120% / 0 - 30°C
3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	35.59	61.44
Temperature reading (°C)	10.04	22.61
Air Pressure (hPa)	986.85	

Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-2.789648E+00	1.173401E+00

With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.80963E-03	1.18012E-04	2.47494E-06	2.30237E+02	-3.78396E-01	-5.77698E+01	4.54717E+00
PhaseCoef	-7.560000E-01	1.000000E+00	0.000000E+00	0.000000E+00			
TempCoef	2.884118E+01	-3.211506E-02	3.617435E-06	-5.017796E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						
Firmware Version	4.9.1						

Date:23.01.2026

Sr. Repair Technician