

ECG Anomaly Detection — Session Report


Prepared for: ff • Session ID: 1b33e835 • Generated: 2026-07-11 11:04:39

Session Information

Patient / User	Sriyantha Sadaruwan
Source	arduino (COM6)
Started	2026-07-11 11:04:39
Ended	2026-07-11 11:07:49
Duration	190.0 s
Status	completed
Total beats analyzed	23,552
Anomalous beats	0 (0.00%)
Model	ecg-ae-v2 (v1783189189) • Dilated-U-Net-Autoencoder
Detection threshold (MAE)	0.1341

ECG Recording & Detection Results

The following plot shows the recorded ECG signal (z-score normalized). Red-shaded regions indicate samples flagged as anomalous by the autoencoder (reconstruction error exceeding the trained threshold).



Main Arrhythmia Finding

No alerts were triggered during this session. The ECG signal was classified as within normal limits by the autoencoder model.

Clinical Summary

The recorded ECG segment was predominantly within normal limits, with fewer than 1% of analyzed samples exceeding the anomaly threshold. No clinically significant arrhythmias were detected by the model during this session.

Physician Notes

Physician signature: _____

Date: _____

This report is generated automatically by an unsupervised deep-learning anomaly detection system and is intended for decision support only. It does not constitute a medical diagnosis and should not replace clinical judgment by a qualified physician.