

One temporal test set; three evaluation questions

20,969 later beats from 45 patient clusters

12,085 normal + 7,958 supported abnormal + 926 other abnormal

A Binary detection: 20,969 beats

→ All normal and abnormal beats, including 926 other abnormal labels.
Question: does the gate distinguish normal from abnormal?

B Subtype classification: 7,958 beats

→ True supported abnormal beats; score independently of the binary gate.
Question: does the subtype model choose the correct arrhythmia group?

C Complete system: 20,043 beats

→ Normal and supported abnormal beats; apply the gate, then subtype.
Question: does the full hierarchy return the correct final class?

Supported arrhythmias: PVC, PAC, LBBB, RBBB and AFib.

The three groups overlap; each answers a separate evaluation question.