Field Width Calculation

HI-ART

- RP. Plan 1 - 50 mm - Fixed_ rev5. dcm (DYNAMIC)
  EXPECTED FIELD WIDTH = 50.48 mm (50 mm)
  CALCULATED FIELD WIDTH[1] = 50.48 mm (max(ΔASYMY) = 50.48 mm)

- RP. Plan 2 - 50mm_ rev5. dcm (DYNAMIC)
  EXPECTED FIELD WIDTH = 50.48 mm (50 mm)
  CALCULATED FIELD WIDTH = 50.48 mm (max(ΔASYMY) = 50.48 mm)

- RP. Plan 3 _ 25mm _ Fixed_ rev5. dcm (DYNAMIC)
  EXPECTED FIELD WIDTH = 25.12 mm (25 mm)
  CALCULATED FIELD WIDTH = 25.12 mm (max(ΔASYMY) = 25.12 mm)

- RP. Plan 4 _ 25 mm _ rev5. dcm (DYNAMIC)
  EXPECTED FIELD WIDTH = 25.12 mm (25 mm)
  CALCULATED FIELD WIDTH = 25.12 mm (max(ΔASYMY) = 25.12 mm

- RP. Plan 5 _ 10 mm _ Fixed _ rev5. dcm (DYNAMIC)
  EXPECTED FIELD WIDTH = 10.48 mm (10 mm)
  CALCULATED FIELD WIDTH = 10.48 mm (max(ΔASYMY) = 10.48 mm)

RAYSTATION

- RP. Plan 1 _ 10 mm _ rev5. dcm (STATIC)
  EXPECTED FIELD WIDTH = 10 mm
  CALCULATED FIELD WIDTH[2] = 7 mm (max(ΔASYMY) = 7 mm)

- RP. Plan 2 _ 25mm _ rev 5. dcm (STATIC)
  EXPECTED FIELD WIDTH = 25mm
  CALCULATED FIELD WIDTH = 20 mm (max(ΔASYMY) = 20 mm)

- RP. Plan 3 _ 50 mm _ rev5. dcm (STATIC)
  EXPECTED FIELD WIDTH = 50 mm
  CALCULATED FIELD WIDTH = 42 mm (max(ΔASYMY) = 42 mm)

1. Using Eq. (2) 2. Using Info. Beam Sequence. Item_1. Private_4001_1028/1017

- RP. Plan 4 _ 25 mm _ rev 5. dcm (DYNAMIC)

  EXPECTED FIELD WIDTH = 25 mm

  CALCULATED FIELD WIDTH = 20 mm (max (ΔASYMY) = 20 mm)

- RP. Plan 5 _ 50 mm _ rev 5 . dcm (DYNAMIC)

  EXPECTED FIELD WIDTH = 50 mm

  CALCULATED FIELD WIDTH = 42 mm (max(ΔASYMY) = 42 mm)
