



AI-Based Smart Cardiopulmonary Blood Flow Control System

Fuzzy Logic Controller Report

Student Name: Sawab Hussein

Instructor: Dr. Omar Abdulghafoor

Course: Artificial Intelligence, Spring 2026

Contents

1	Problem Statement	2
2	Membership Functions and Rules	2
	Explicit IF-THEN Rules	3
3	Inference at the Required Operating Point	3
4	Centroid Defuzzification	4
5	Conclusion	5

1 Problem Statement

A cardiopulmonary bypass rotary pump must maintain safe blood flow by adjusting the pump speed N . The required operating point is

$$\Delta P = 84 \text{ mmHg}, \quad Q = 6 \text{ L/min.}$$

The work completed for this project was to convert the assignment rule table into a complete fuzzy controller, evaluate the controller at the required operating point, draw the intermediate and final fuzzy sets, and compute the crisp rotary-pump speed. The implementation is reproducible from `fuzzy_cpb_controller.py`, which defines the membership functions, stores the nine IF–THEN rules, performs Mamdani inference, applies centroid defuzzification by numerical integration, and saves the report diagrams in the `figures` folder.

The controller uses the standard Mamdani sequence: fuzzify the two inputs, compute each rule firing strength with the logical AND operator as a minimum, clip each output membership function by its rule strength, combine all output clips with a maximum operator, and finally defuzzify the aggregated output set.

2 Membership Functions and Rules

The triangular membership functions were read from the assignment figures. The universes are $Q \in [0, 10]$ L/min, $\Delta P \in [50, 150]$ mmHg, and $N \in [2800, 3200]$ rpm; the vertices used in the implementation are summarized below.

Table 1: Triangular membership vertices used by the controller.

Variable	Low / Slow	Medium	High / Fast
Blood flow (L/min)	$Q (0, 3.5, 6)$	$(3.5, 6, 8.5)$	$(6, 8.5, 10)$
Delta pressure (mmHg)	$\Delta P (50, 72, 100)$	$(72, 100, 120)$	$(100, 120, 142)$
Pump speed (rpm)	$N (2800, 2900, 3000)$	$(2900, 3000, 3100)$	$(3000, 3100, 3200)$




Figure 1: Membership functions for flow, pressure difference, and pump speed.

Figure 1 is the controller definition in graphical form. The first two panels show how measured flow and pressure are translated into linguistic values, while the third panel shows the possible linguistic pump-speed outputs. These curves are the numerical basis used by the Python file when it calculates memberships for $Q = 6$ and $\Delta P = 84$.

Table 2: Fuzzy rule base used for the rotary pump controller.

$\Delta P \backslash Q$	Low Flow	Medium Flow	High Flow
Low ΔP	Fast	Medium	Slow
Medium ΔP	Medium	Medium	Slow
High ΔP	Medium	Medium	Slow

Explicit IF–THEN Rules

1. IF ΔP is low AND Q is low THEN Speed = fast
2. IF ΔP is low AND Q is medium THEN Speed = medium
3. IF ΔP is low AND Q is high THEN Speed = slow
4. IF ΔP is medium AND Q is low THEN Speed = medium
5. IF ΔP is medium AND Q is medium THEN Speed = medium
6. IF ΔP is medium AND Q is high THEN Speed = slow
7. IF ΔP is high AND Q is low THEN Speed = medium
8. IF ΔP is high AND Q is medium THEN Speed = medium
9. IF ΔP is high AND Q is high THEN Speed = slow

3 Inference at the Required Operating Point

The input $Q = 6$ lies exactly at the peak of the medium-flow triangle. Therefore, the flow input belongs completely to the medium set and has zero membership in the low and high sets. The pressure value $\Delta P = 84$ lies in the overlap between low and medium pressure, so both pressure labels are partially active.

For $Q = 6$, the medium-flow set is fully active:

$$\mu_{Q,\text{low}}(6) = 0, \quad \mu_{Q,\text{medium}}(6) = 1, \quad \mu_{Q,\text{high}}(6) = 0.$$

For $\Delta P = 84$, the two active pressure memberships are

$$\mu_{\Delta P,\text{low}}(84) = \frac{100 - 84}{100 - 72} = 0.571, \quad \mu_{\Delta P,\text{medium}}(84) = \frac{84 - 72}{100 - 72} = 0.429.$$

Only two rules fire with nonzero strength:

Table 3: Nonzero rule activations.

Rule	Consequent	Firing strength
R2	Low ΔP , Medium Q : Speed is Medium	0.571
R5	Medium ΔP , Medium Q : Speed is Medium	0.429

Both active rules point to the same consequent, so the aggregated output is simply the medium-speed membership function clipped at

$$\alpha_{\text{medium}} = \max(0.571, 0.429) = 0.571.$$

The remaining rules have zero firing strength because one of their required input labels has zero membership at this operating point.

Figure 2 visualizes the full rule matrix and confirms that only two cells are active: low pressure with medium flow, and medium pressure with medium flow. Figure 3 then shows the consequence

of these active rules: the medium-speed output set is clipped at 0.571, while the slow and fast output sets do not contribute to the aggregate. The vertical dashed line marks the centroid of the aggregated output.




Figure 2: Rule firing strengths.




Figure 3: Aggregated output fuzzy set.

4 Centroid Defuzzification

The centroid method computes

$$N^* = \frac{\int_{2800}^{3200} n \mu_{\text{agg}}(n) dn}{\int_{2800}^{3200} \mu_{\text{agg}}(n) dn}.$$

For this operating point, the active output is the medium-speed triangle clipped at $\alpha = 0.571 = 4/7$. The unclipped medium-speed set is centered at 3000 rpm and has the vertices (2900, 0), (3000, 1), and (3100, 0). The clipping level intersects the triangle at

$$n_L = 2900 + 100 \left(\frac{4}{7} \right) = 2957.143, \quad n_R = 3100 - 100 \left(\frac{4}{7} \right) = 3042.857.$$

Therefore, the aggregated membership function is

$$\mu_{\text{agg}}(n) = \begin{cases} \frac{n - 2900}{100}, & 2900 \leq n \leq 2957.143, \\ \frac{4}{7}, & 2957.143 \leq n \leq 3042.857, \\ \frac{3100 - n}{100}, & 3042.857 \leq n \leq 3100, \\ 0, & \text{otherwise.} \end{cases}$$

First, evaluate the denominator, which is the area under the clipped output set:

$$\begin{aligned} A &= \int_{2900}^{2957.143} \frac{n - 2900}{100} dn + \int_{2957.143}^{3042.857} \frac{4}{7} dn + \int_{3042.857}^{3100} \frac{3100 - n}{100} dn \\ &= 16.327 + 48.980 + 16.327 \\ &= 81.633. \end{aligned}$$

Next, evaluate the numerator, which is the first moment of the clipped output set:

$$\begin{aligned} M &= \int_{2900}^{2957.143} n \left(\frac{n - 2900}{100} \right) dn + \int_{2957.143}^{3042.857} n \left(\frac{4}{7} \right) dn + \int_{3042.857}^{3100} n \left(\frac{3100 - n}{100} \right) dn \\ &= \left[\frac{n^3}{300} - \frac{29n^2}{2} \right]_{2900}^{2957.143} + \left[\frac{2n^2}{7} \right]_{2957.143}^{3042.857} + \left[\frac{31n^2}{2} - \frac{n^3}{300} \right]_{3042.857}^{3100} \\ &= 47,968.902 + 146,938.776 + 49,990.282 \\ &= 244,897.959. \end{aligned}$$

Finally, divide the first moment by the area:

$$N^* = \frac{M}{A} = \frac{244,897.959}{81.633} = \boxed{3000.0 \text{ rpm}}.$$

The exact arithmetic gives the same value because $M = 12,000,000/49$ and $A = 4000/49$, so the ratio is exactly 3000.




Figure 4: Defuzzified control surface with the required operating point highlighted.

Figure 4 is a broader check of the controller behavior over the input space. Each point on the surface is obtained by running the same fuzzy inference and centroid calculation for a different pair of $(\Delta P, Q)$. The highlighted point corresponds to the assignment case, showing where the computed 3000 rpm result sits relative to nearby operating conditions.

5 Conclusion

At $\Delta P = 84$ mmHg and $Q = 6$ L/min, the fuzzy controller activates the medium-speed consequent only. After aggregation and centroid defuzzification, the recommended rotary-pump speed is

$$\boxed{3000 \text{ rpm}}.$$

This value is consistent with the controller goal of maintaining the reference flow near 6 L/min under moderate pressure loading. The submitted Python file documents the computational workflow and regenerates every diagram used in this report, so the written solution and generated figures are tied to the same implementation.