

Devicestatus	Link	devicepicture	device apps	device
Active	https://www.jeol.co.jp/en/products/detail/JSM-7610F.html		The possibility of examining the surface of biological and non-biological samples with the availability of elemental analysis technology Magnification power: 1000000x Contrast strength: 1nm	Scanning Electron Microscope JEOL (JSM-7610F)
non Active	https://www.biotech.iastate.edu/wp_single/wp-content/uploads/2012/01/quanta.pdf		The possibility of examining the surfaces of biological samples Magnification power: 500 000 X Contrast strength: 1nm Voltage: 1-30KV	Scanning Electron Microscope FEI (QUANTA-250)

Active	http://faculty.olin.edu/~jstolk/matsci/Operating%20Instructions/SEM%20operating%20instructions.pdf		The possibility of examining the surfaces of biological and non-biological samples Magnification power 300,000 X	Scanning Electron Microscope JEOL (JSM-6060LV)
Active	https://www.datum.jeol.co.jp/en/related_items/jfc-1600.html		biological and non-biological sample with gold or platinum	Auto Fine Coater JEOL (JFC-1600)
Active	http://www.avba.co.il/image/users/390742/ftp/my_files/JEOL-%20JEM-1400Plus%20-%20Brochure.pdf?id=28918316		• Possibility to study and photograph super clips Raqqa for non-biological biological samples, as well as the study and imaging of living organisms minute. • It is used to study nanoparticles and drugs. • 0.2 . nm high contrast strength • X1200 . high magnification power	Transmission Electron Microscope JEOL (JEM 1400 Plus)

Active	https://www.leica-microsystems.com/products/sample-preparation-for-electron-microscopy/p/leica-em-uc7/		Preparing ultra-thin slices	Ultra Microtome
Active	https://www.leica-microsystems.com/products/sample-preparation-for-electron-microscopy/p/leica-em-txp/showcase/		Resin block trimming and polishing	Trimmer
Active	https://downloads.leica-microsystems.com/Leica%20EM%20KMR3/Brochures/Leica_EMKMR3_Brochure_06_16_SCR_EN.pdf		Making glass knives	Glass knife maker

Active	https://downloads.leica-microsystems.com/Leica%20OEM%20KMR2/Brochures/Leica_EMKMR2-Brochure_EN.pdf		Making glass knives	Glass knife maker
Active	https://www.leica-microsystems.com/products/sample-preparation-for-electron-microscopy/p/leica-em-tp/		automated tissue processor	Tissue processor
non Active	https://www.moleculardevices.com/sites/default/files/en/assets/user-guide/dd/img/imagexpress-ultra-confocal-high-content-screening-system.pdf		The robotic cone microscope is capable of fluorescent imaging in the case of analysis of fixed cells and tissues Measuring important event	Confocal High Content Screening System ImageXpress Ultra

Widefield High Content Screening System

ImageXpress Micro XLS



Automated wide range microscope is capable of imaging fluorescent, fluorescent light and phase-contrast in the case of analysis of live or fixed cells and measurement of important events

non Active

<https://www.moleculardevices.com/sites/default/files/en/assets/brochures/dd/img/imageXpress-micro-xls-widefield-high-content-screening-system.pdf>