

# Panda™ iRes Warmer

## Operation and Maintenance Manual



**© 2007-2017 General Electric Company**

All rights reserved. General Electric Company reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your GE Representative for the most current information. Panda is a trademark owned by Datex-Ohmeda, Inc. GE and GE Monogram are trademarks of General Electric Company. All other company and product names mentioned may be trademarks of the companies with which they are associated.

# Table of Contents

---

|                                                                |      |
|----------------------------------------------------------------|------|
| <b>Important Safety Information .....</b>                      | v    |
| Warnings, Cautions and Notes .....                             | v    |
| Symbols.....                                                   | vii  |
| <b>About this Manual .....</b>                                 | ix   |
| Scope and intended users .....                                 | ix   |
| Organization of this manual.....                               | ix   |
| References.....                                                | x    |
| <b>Chapter 1 - Introduction .....</b>                          | 1-1  |
| System Description and Indications for Use .....               | 1-1  |
| Mechanical controls and cable connectors.....                  | 1-2  |
| Controls and Displays.....                                     | 1-5  |
| Temperature Regulation .....                                   | 1-6  |
| Control Modes.....                                             | 1-7  |
| On Screen Help.....                                            | 1-8  |
| Home Screen.....                                               | 1-9  |
| Scale (optional) .....                                         | 1-11 |
| Trends.....                                                    | 1-12 |
| Setup .....                                                    | 1-13 |
| Alarms.....                                                    | 1-15 |
| Temperature regulation alarms .....                            | 1-17 |
| Alarm Table .....                                              | 1-21 |
| <b>Chapter 2 - System Setup and Checkout Procedures.....</b>   | 2-1  |
| Mechanical checks .....                                        | 2-1  |
| Controller checks.....                                         | 2-4  |
| Optional equipment checks .....                                | 2-7  |
| <b>Chapter 3 - Operating the Warmer .....</b>                  | 3-1  |
| Basic Operating Procedure .....                                | 3-1  |
| Start up.....                                                  | 3-3  |
| Panda iRes Warmer Flow Chart .....                             | 3-7  |
| Attaching the Skin Temperature Probe .....                     | 3-8  |
| Setting the manual temperature alarm.....                      | 3-11 |
| Raising or lowering the bedside panels.....                    | 3-13 |
| Tilting the bed .....                                          | 3-14 |
| Raising and lowering the bed .....                             | 3-15 |
| Using the X-ray tray .....                                     | 3-16 |
| Using the observation and procedure lights .....               | 3-16 |
| Uninterruptible Power Supply (UPS).....                        | 3-17 |
| Using the In-Bed Scale (optional) .....                        | 3-17 |
| Weighing procedure .....                                       | 3-18 |
| Using the Integrated SpO <sub>2</sub> Monitor (optional) ..... | 3-22 |
| Using the Integrated Resuscitation System (optional).....      | 3-22 |
| Mounting Optional equipment.....                               | 3-22 |
| ThermaLink option.....                                         | 3-22 |
| Using the Serial Data interface.....                           | 3-22 |
| Using the Nurse Call System interface.....                     | 3-22 |

|                                                                   |                   |
|-------------------------------------------------------------------|-------------------|
| <b>Chapter 4 - Maintenance and Cleaning .....</b>                 | 4-1               |
| Repair Policy .....                                               | 4-1               |
| Maintenance schedule .....                                        | 4-2               |
| Cleaning Instructions.....                                        | 4-3               |
| Cleaning the Warmer.....                                          | 4-4               |
| Cleaning Solutions .....                                          | 4-5               |
| Cleaning and disinfecting individual components .....             | 4-6               |
| Patient Probe (reusable) .....                                    | 4-6               |
| Cleaning other components.....                                    | 4-7               |
| <b>Appendix A - Options .....</b>                                 | A-1               |
| <b>Appendix B- Specifications .....</b>                           | B-1               |
| Power Requirements and Accessory Outlets.....                     | B-1               |
| Standards.....                                                    | B-1               |
| Operating Environment.....                                        | B-1               |
| User Control Settings.....                                        | B-1               |
| Performance.....                                                  | B-2               |
| Mechanical Specifications .....                                   | B-2               |
| Uninterruptible Power Supply (UPS) Specifications .....           | B-2               |
| <b>Appendix C - Electromagnetic Compatibility .....</b>           | C-1               |
| <b>Appendix D - Conforming with Standards and Directives.....</b> | D-1               |
| <b>Appendix E - Additional Safety Information.....</b>            | E-1               |
| <b>GE Service Centers.....</b>                                    | inside back cover |

# Important Safety Information

---

Before using the Panda iRes Warmer, read through this entire manual. As with all medical equipment, attempting to use this device without a thorough understanding of its operation may result in patient or user injury. This device should only be operated by personnel trained in its operation under the direction of qualified medical personnel familiar with the risks and benefits of this type of device. Additional precautions specific to certain procedures are found in the text of this manual.

Complete the checkout procedures in this manual before putting the unit into operation. If the unit fails any portion of the checkout procedure it must be removed from use and repaired.

Do not use the warmer in the presence of flammable anesthetics; an explosion hazard exists under these conditions.

Device alarms are readable only at a viewing angle of 129° from the front of the device. Position the device for appropriate view of the display.

In the event of loss of Mains power the unit will enter a power fail alarm condition. While in power fail alarm condition a single red LED will be lit at the left side of the alarm light display. The audible alarm is a repeating series of three rapid beeps separated by a short pause. The unit will remain in power fail alarm condition for up to ten minutes, or until mains power is restored. If mains power is restored in less than 10 minutes the unit will return to operation with the same settings as when power was lost. If mains power has not been restored within ten minutes the unit will shut down and the LED indicator and alarm tones will cease. Upon restart of the system after a loss of power of greater than 10 minutes the user may need to reset SpO<sub>2</sub> parameters.

Always disconnect the power before performing service or maintenance procedures detailed in this manual. Apply power only if you are specifically instructed to do so as part of the procedure. Thoroughly air dry the warmer after cleaning it with flammable agents. Small amounts of flammable agents, such as ether, alcohol or similar cleaning solvents left on the warmer can cause a fire.

Only competent individuals trained in the repair of this equipment should attempt to service it as detailed in the Service Manual.

Detailed information for more extensive repairs is included in the service manual solely for the convenience of users having proper knowledge, tools and test equipment, and for service representatives trained by GE Healthcare.

**WARNINGS, CAUTIONS, NOTES, and SYMBOLS:** This section introduces the various types of warnings, cautions information notes and symbols used in this guide to alert you to possible safety hazards and to provide you with additional information.



**WARNING:** A Warning statement is used when the possibility of injury to the patient or the operator exists.

---



**CAUTION:** A CAUTION statement is used when the possibility of damage to the equipment exists.

---



**NOTE:** A Note provides additional information to clarify a point in the text.

---

## Warnings, Cautions and Notes

This section identifies general warnings, cautions and Notes associated with the use of the Panda iRes Warmer.



**WARNING:**

As with all medical equipment, carefully route patient cabling to reduce the possibility of patient entanglement or strangulation.

---



**WARNING:**

Do not lean against side of warmer. Leaning against the side of the warmer may cause it to tip over.

---



**CAUTION:**

U.S. Federal law restricts this device to sale by, or on order of, a licensed medical practitioner.

---



**CAUTION:**

This device is for professional use only, by trained clinicians.

**For professional use only**

---



**NOTE:**

Ranges listed in this manual represent the operational ranges of the equipment. The gauge ranges may exceed operational ranges.

---



**NOTE:**

Air always means medical grade air.

---



**NOTE:**

Additional copies of this manual are available on request from the GE Healthcare office listed on the inside back cover of this manual.

---

**SYMBOLS:** This section identifies the symbols that are displayed on the device. Equipment configurations can differ. Labels or symbols may not appear on every device.

| Symbol | Description                                                                                              |
|--------|----------------------------------------------------------------------------------------------------------|
|        | Type BF Equipment                                                                                        |
|        | Functional Earth Terminal                                                                                |
|        | Protection Earth Terminal                                                                                |
|        | Alternating Current                                                                                      |
|        | Alarm Silence                                                                                            |
|        | European Union Representative                                                                            |
|        | Consult instructions for use                                                                             |
|        | Catalog Number                                                                                           |
|        | Serial Number                                                                                            |
|        | Manufacturer                                                                                             |
|        | Do not reuse                                                                                             |
|        | Help menu                                                                                                |
|        | Patient temperature                                                                                      |
|        | Observation light                                                                                        |
|        | Scale                                                                                                    |
|        | Date of manufacture                                                                                      |
|        | Do not lean on warmer                                                                                    |
|        | Dispose of electrical and electronic equipment as unsorted municipal waste; must be collected separately |
|        | Increase/decrease, up/down                                                                               |
|        | Equipment shall be used only by qualified, trained medical personnel.                                    |

|     |                                                                                                 |
|-----|-------------------------------------------------------------------------------------------------|
|     | Do not place items in path of heat.                                                             |
|     | Pinch hazard                                                                                    |
|     | Ground Equalization Potential Post                                                              |
| UDI | Unique Device Identifier. The UDI is a unique marking for identification of the medical device. |

# About this Manual

---

## Scope and intended users

This manual describes the features and operation of the Panda iRes Warmer. The Panda iRes Warmer is used in hospital delivery rooms, hospital newborn nurseries and neonatal intensive care units (NICU).

The intended users for this manual consists of end users of the equipment, primarily care providers in infant delivery rooms and NICUs, and hospital biomedical engineering services.

## Organization of this manual

The organization of this manual is:

**Important Safety Information** provides important safety information for the safety of both the patient and the care providers.

**Chapter 1 - Introduction** describes the Panda iRes Warmer system and the system's indications for use.

**Chapter 2 - System Setup and Checkout Procedures** describes the procedures to set up and check out the Panda iRes Warmer prior to use.

**Chapter 3 - Operating the Warmer** describes the procedures to operate the Panda iRes Warmer.

**Chapter 4 - Maintenance and Cleaning** provides information on performing maintenance and cleaning of the Panda iRes Warmer.

**Appendix A - General Use Items** is a list of general use and ancillary items that may be used with the Panda iRes Warmer.

**Appendix B - Specifications** provides information on power requirements and accessory outlets, standards, operating environment, storage conditions, user control settings, performance and mechanical specifications.

**Appendix C - Electromagnetic Compatibility (EMC)** provides information on the Panda iRes Warmer's electromagnetic emissions and the electronic environment that is compatible with the use and operation of the Panda iRes Warmer.

**Appendix D - Conforming with Standards and Directives** provides information on the European Council Directive and the disposal of electrical/electronic equipment.

**Appendix E - Additional Safety Information** provides information on IEC 60601-1 3rd edition requirements

**Inside back cover** provides a geographically organized list of Datex-Ohmeda service centers.

## References

References to other manuals pertaining to the Panda iRes Warmer are:

- Giraffe Warmer Operation and Maintenance Manual, M1110734
- SpO<sub>2</sub> - Masimo SET Option, Operation and Maintenance Supplement, M1110918
- Resuscitation System - Bag and Mask Option, Operation and Maintenance Supplement M1110023
- Resuscitation System - T-piece Option Operation and Maintenance Supplement, M1109204
- Service Manual - Resuscitation Giraffe and Panda Warmers, M1128929
- Service Manual - Giraffe and Panda Warmers, M1128921

# Chapter 1

## Introduction

---

### The Panda iRes Warmer includes:

- Recessed heater dish
- Dimmable observation lights
- Hands Free Alarm Silence
- Graphical trending features
- APGAR Timer
- Full color control panel
- Resuscitation mattress
- Progressive, adjustable alarms
- Bed tilt
- Dovetail rail system

### Optional features include:

- Integrated SpO<sub>2</sub> monitor
- Aimable procedure light
- Integrated Resuscitation
- Tubing management wall
- In-bed weighing scale
- Elevating base
- Multiple drawer packages
- Wood accents
- Options



Figure 1: Panda iRes Warmer

## System Description and Indications for Use

Infant radiant warmers provide infrared heat in a controlled manner to neonates who are unable to thermoregulate based on their own physiology. Infant radiant warmers may be used to facilitate the neonate's transition to the external environment or to provide a controlled open environment. An optional integrated SpO<sub>2</sub> monitoring feature may be used for continuous noninvasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO<sub>2</sub>) and pulse rate (measured by an SpO<sub>2</sub> sensor). An optional integrated resuscitation system may be used to provide the basic equipment required for pulmonary resuscitation of infants. Pulmonary resuscitation includes practices necessary to establish a clear airway and provide oxygen or air/oxygen mixtures and/or manual ventilation to the infant.

## Mechanical Controls and Cable Connections




Figure 2: Panda iRes Warmer, front oblique view

| Feature number | Description                     |
|----------------|---------------------------------|
| 1              | Two dimmable observation lights |
| 2              | Aimable procedure light         |
| 3              | Procedure light "On/Off" switch |
| 4              | Color display screen            |
| 5              | Resuscitation system (optional) |
| 6              | Bed, with integrated scale      |
| 7              | Front bedside panel             |
| 8              | Bed tilt control lever          |




Figure 3: Panda iRes Warmer, side view

| Feature number | Description                                   |
|----------------|-----------------------------------------------|
| 9              | Recessed radiant heater                       |
| 10             | Side bedside panel                            |
| 11             | Front handle                                  |
| 12             | Pass trough drawer                            |
| 13             | Two bed height adjustment pedals, up and down |
| 14             | Four caster wheels                            |
| 15             | Four brakes                                   |
| 16             | Elevating column                              |
| 17             | Tank guard (optional)                         |
| 18             | Scale cable connector                         |
| 19             | Temperature probe Jack                        |
| 20             | High pressure air/oxygen yoke (optional)      |
| 21             | Maneuvering handle and cord wrap              |
| 22             | Dovetail rail                                 |

## Mechanical Controls and Cable Connections (continued)




Figure 4: Panda iRes Warmer, rear view

| Feature number | Description                  |
|----------------|------------------------------|
| 23             | RS 232 connector             |
| 24             | Two accessory power outlets  |
| 25             | SpO <sub>2</sub> Jack        |
| 26             | Removable rear bedside panel |
| 27             | Power cord inlet             |
| 28             | Mains Power Switch           |



**NOTE:** For equipment installation instructions, refer to the Service Manual.

## Controls and Displays




Figure 5: Control panel and color screen display

| Feature number | Description                                                                                                                                                                              |
|----------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1              | <b>Power stand-by switch</b> - On the left below the display turns the power to the warmer "On" and "Off".                                                                               |
| 2              | <b>Oximetry key</b> (optional) - This key retains its same function at all times.                                                                                                        |
| 3              | <b>Mode key</b> - To select manual or baby mode. This key retains its same function at all times                                                                                         |
| 4              | <b>Temperature/power increase/decrease key</b> - This key retains its same function at all times                                                                                         |
| 5              | <b>Baby temperature</b> - Can be displayed in degrees Celsius or degrees Fahrenheit. The default setting is degrees Celsius.                                                             |
| 6              | <b>Alarm silence key</b> - Alarms can be silenced by pushing the key above the display or by a wave of your hand directly in front of the alarm light.                                   |
| 7              | <b>Help key</b> - The key with the "?" in the upper right corner brings up the help screen that explains alarms and functions.                                                           |
| 8              | <b>Task keys</b> - The five keys on the right are <b>"soft" keys</b> that change their function depending on what task you wish to perform. They control equipment settings and options. |
| 9              | <b>Dimmer knob</b> - On the right controls the brightness of the observation lights.                                                                                                     |
| 10-11          | <b>Date and time</b> - Displayed at the bottom of the screen.                                                                                                                            |


# Temperature Regulation



## **WARNING:**

Do not place the baby in the bed while in Warmup Mode. Warmup Mode is used to quickly warm and maintain heat to an empty bed. Warmup Mode is not designed for clinical use with a baby since the Check Baby Alarm is disabled.


## Warmup Mode

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                         |
|-------------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   | 1    | <p>The Panda Warmer automatically starts in Warmup Mode. During Warmup Mode <b>"Check Baby"</b> alarms are deactivated.</p> <p>If you do not press the <b>"Admit Baby"</b> key or the <b>"APGAR On/Off"</b> key during the first minutes after start up, the display will show <b>"Warming..."</b> and the radiant heater will automatically run at 100% power.</p> |
|  | 2    | <p>After 7 minutes, the display indicates <b>"Warmup Complete..."</b> and the default heater power setting decreases to 50%. The display continues to flash "Admit Baby" until you press the <b>"Admit Baby"</b> key or the <b>"APGAR On/Off"</b> key and start normal operation.</p>                                                                               |

# Control Modes

The Panda Warmer has two control modes, **Manual Mode** and **Baby Mode**. In Manual Mode, the warmer controls radiant heater output from a heater power percentage setting that you enter using the control panel. In Baby Mode, the warmer controls radiant heater output based on temperature readings from a probe attached to the baby's skin (skin probe) and a set temperature (set temp) you enter using the control panel.

## Manual Mode

| Screen                                                                            | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
|-----------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <b>Manual Mode - Set heater power.</b><br>After completing warm up and pushing the "Admit Baby" key, or the "APGAR On/Off" key the warmer starts in the Manual Mode with a default 100% heater power setting.<br><br>The percentage of full power can be adjusted in 5% increments using the temperature/power keys on the left of the display.<br><br>You can monitor the baby's temperature with a skin probe in the Manual Mode, shown in the upper left corner of the display, but it does not control heater output. If you intend to use Manual Mode for a prolonged period of time a skin probe should be attached to the baby. |






### WARNING:

Do not leave baby unattended while the warmer is in Manual Mode. Whenever possible, use the warmer in the servo-control mode (Baby Mode). The Manual Mode requires constant monitoring of the baby's condition by the caregiver.

## Baby Mode



| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
|-------------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <b>Baby Mode - Adjust set temperature.</b><br>In the Baby Mode the warmer controls heater output to a desired baby set temp you enter. Heater output is based on measured differences between the set temperature and the skin probe temperature reading. You can enter desired set temperatures from 34 to 37.5 °C using the increase/decrease keys on the left of the display, and this value is shown above the power indicator bar graph in <b>white</b> numerals slightly smaller than the baby's actual temperature. Baby temperature is shown in <b>green</b> in the upper left corner of the display. |

## On Screen Help

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                              |
|-------------------------------------------------------------------------------------|------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 1    | <p>Pushing the “Help” (?) key brings up Help screen that provides information about alarms and functions.</p> <p>When an alarm sounds, pushing the “Help” key automatically displays information about that alarm.</p>                                                                   |
|   | 2    | <p>The help screen main menu contains four sections:</p> <ol style="list-style-type: none"> <li>1. “Operating Modes” (see Step 2, below),</li> <li>2. “Functions” (see Step 3, below),</li> <li>3. “Alarms” (see Step 4, below), and</li> <li>4. “About” (see step 5, below).</li> </ol> |
|  | 2    | <p><b>“Operating Modes”</b></p> <p>Pushing “Operating Modes” brings up information on warmer modes such as Baby Mode, etc.</p>                                                                                                                                                           |

| Screen | Step | Description                                                                                                                                   |
|--------|------|-----------------------------------------------------------------------------------------------------------------------------------------------|
|        | 3    | <p><b>"Functions"</b></p> <p>Pushing <b>"Functions"</b> brings up information on warmer features such as APGAR, etc.</p>                      |
|        | 4    | <p><b>"Alarms"</b></p> <p>Pushing <b>"Alarms"</b> brings up information on alarms.</p>                                                        |
|        | 5    | <p>Pushing <b>"About"</b> brings up information on the current software revision.</p> <p>Pushing <b>"Exit"</b> brings up the Home Screen.</p> |

## Home Screen

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                                            |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 1    | <b>Mode</b><br>Current mode setting, "Manual" or "Baby" mode, is highlighted in blue on the left side of screen (arrow).                                                                                                                                                                                                                                                               |
|   | 2    | <b>APGAR On/Off</b><br>Press the "APGAR On/Off" key to start the APGAR timer and continuously display time elapsed. A tone will sound at 1, 5, 10 minutes, and every five minutes thereafter to remind you to evaluate the baby's APGAR score. A visual message will appear for the 1, 5, and 10 minute APGAR times only. Press the "APGAR On/Off" key a second time to end the timer. |
|  | 3    | <b>Menu</b><br>Push the "Menu" key to access the "APGAR On/Off", "Scale", "Trends", "SpO2", and "Setup" keys.<br><b>Note:</b> If the Scale or SpO2 options are not installed on your unit, these keys will not appear on the display.                                                                                                                                                  |


## Scale (optional)

Push the “**Menu**” key to access the “**Scale**” key. If the scale jack is not plugged in, the scale key will not appear. Push the “**Scale**” soft key to start the weighing procedure and access the next menu. For more detail on performing the weighing procedure and using the scale (refer to “Using the In-Bed Scale (optional)”) on page 3-17.


## Trends

The warmer trends temperature, and SpO<sub>2</sub> (optional). Push “**Menu**” key then the “**Trends**” key to display the trending screens.

### Temperature trending

| Screen                                                                            | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                       |
|-----------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <p>Push the “<b>Temp</b>” key to view a graph of set temperatures, baby temperatures and heater output. Manual Mode graphs show percentage of power and baby temperatures. If a probe is attached to the baby, Baby Mode graphs display set temperatures, baby temperatures and heater output.</p> <p>Push the “<b>Time Scale</b>” key to change the time period shown on the graph from the default 15 minutes to 30 minutes, or 60 minutes.</p> |

### SpO<sub>2</sub> trending (with SpO<sub>2</sub> option installed)

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |
|-------------------------------------------------------------------------------------|------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <p>Push the “<b>SpO<sub>2</sub></b>” key to view a graph that plots both the baby's blood oxygen saturation percentage and the baby's pulse rate.</p> <p><b>Note:</b> the “<b>SpO<sub>2</sub></b>” key only appears on the display if the SpO<sub>2</sub> option is installed.</p> <p>Push the “<b>Time Scale</b>” key to change the time period shown on the graph from the default 15 minutes to 30 minutes, or to 60 minutes.</p> <p>Push the “<b>Back</b>” key to return to the trending screen and the “<b>Exit</b>” key to return to the Home Screen.</p> |

# Setup

| Screen                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  <p>The Home Screen displays the current Baby Temp (C) and Set Temp (C), both set to 36.8. It includes a color bar for SpO2 levels (25%, 50%, 75%, 100%) and a Mode selection between Manual and Baby. Other buttons include Pulse Rate and SpO2. The screen also shows the date (Monday, June 11, 2007) and time (03:27). Control buttons on the left include up/down arrows and a central button, while the right side has a question mark icon, a power button, a light button, and a gear icon.</p> | 1    | <p>Push “<b>Menu</b>” on the Home Screen to access the warmer “<b>Setup</b>” key.</p> <p>Push the “<b>Temp Units</b>” key to change from the default Celsius (C) to Fahrenheit (F) degrees.</p> <p>Push the “<b>Alarm Volume</b>” key to choose from 4 volume levels.</p> <p>Push the SpO<sub>2</sub> “<b>Pulse Tone Volume</b>” key to choose from 4 pulse tone volume levels and a tone off setting. The default is “On” at low volume.</p> <p>Push the “<b>Set Clock</b>” key to set the Panda iRes Warmer clock. You can adjust the clock for time zone and seasonal variations.</p> <p>Push the “<b>Next</b>” key to access a second setup menu.</p> |
|  <p>The second setup menu screen shows various configuration options: Elevate (Enabled/Disabled), Hands Free Silence (Enabled/Disabled), Pt. Temp Alarm (0.5 to 10), and Man. Temp Alarm (Off to 0.5 to 10). It also includes an “Exit” button. The layout is similar to the Home Screen with control buttons on the left and right.</p>                                                                                                                                                              | 2    | <p>Push the “<b>Elevate</b>” key to enable and disable the bed elevating feature.</p> <p>Push “<b>Hands Free Silence</b>” key to enable and disable the hands free alarm silence feature.</p>                                                                                                                                                                                                                                                                                                                                                                                                                                                             |

| Screen                                                                             | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|------------------------------------------------------------------------------------|------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   | 3    | <b>Patient Temperature Alarm</b><br>Push the “Pt. Temp Alarm” key to choose between two settings in the warmer’s Baby Mode. The 0.5 setting activates an alarm when the difference between the set Temperature and the baby’s skin temperature is greater than 0.5°C. The default 1.0 setting activates an alarm when the difference between the set temperature and the baby’s skin temperature is greater than 1.0°C.                                                                                                                                                                                                                                                                                                                                                                                                                                    |
|  | 4    | <b>Manual Temperature Alarm</b><br>You can use the manual temperature alarm when you wish to manage the baby’s temperature in the Manual Mode, but want to be notified when the baby’s temperature drops below or rises above a threshold setting.<br><br>Place the temperature probe on the baby’s skin. Push the “Man Temp Alarm” key to choose from 3 settings in the warmer’s Manual Mode. The default Off setting deactivates the alarm, the 0.5 setting activates an alarm when the difference between the set temperature and the baby’s skin temperature is greater than 0.5°C. The 1.0 setting activates an alarm when the difference between the set temperature and the baby’s skin temperature is greater than 1.0°C. For more details on operation of the Manual Temperature Alarm, (see “Setting the manual temperature alarm” in chapter3). |



**NOTE:** After power down, only the time setting will remain.

## Alarms

| Illustration                                                                                                                                                                                                                                                                                                                                                                                                                        | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |                |             |   |             |   |                |   |                   |   |               |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------|-------------|---|-------------|---|----------------|---|-------------------|---|---------------|
|  <p>The illustration shows the control panel of the Panda iRes Warmer. Feature 1 is a red blinking alarm light at the top. Feature 2 consists of two motion sensor icons on the left. Feature 3 is a yellow 'Silence' key icon on the right. Feature 4 is a red alarm message 'Confirm Probe Jack Connection' displayed on the central screen.</p> | <p>There are two types alarms:</p> <p><b>Low Priority Alarms</b> are indicated by a blinking red alarm light, a red alarm message on the central alarm area of the display, and an intermittent audio tone.</p> <p><b>High Priority Alarms</b> have the same blinking red alarm light and red alarm message. The difference is that the audio tone for high-priority alarms is continuous.</p> <p><b>Alarm features</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr style="background-color: #d3d3d3;"> <th style="text-align: center; padding: 2px;">Feature number</th><th style="text-align: center; padding: 2px;">Description</th></tr> </thead> <tbody> <tr> <td style="text-align: center; padding: 2px;">1</td><td style="text-align: center; padding: 2px;">Alarm light</td></tr> <tr> <td style="text-align: center; padding: 2px;">2</td><td style="text-align: center; padding: 2px;">Motion sensors</td></tr> <tr> <td style="text-align: center; padding: 2px;">3</td><td style="text-align: center; padding: 2px;">Alarm silence key</td></tr> <tr> <td style="text-align: center; padding: 2px;">4</td><td style="text-align: center; padding: 2px;">Alarm message</td></tr> </tbody> </table> | Feature number | Description | 1 | Alarm light | 2 | Motion sensors | 3 | Alarm silence key | 4 | Alarm message |
| Feature number                                                                                                                                                                                                                                                                                                                                                                                                                      | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |                |             |   |             |   |                |   |                   |   |               |
| 1                                                                                                                                                                                                                                                                                                                                                                                                                                   | Alarm light                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |                |             |   |             |   |                |   |                   |   |               |
| 2                                                                                                                                                                                                                                                                                                                                                                                                                                   | Motion sensors                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |                |             |   |             |   |                |   |                   |   |               |
| 3                                                                                                                                                                                                                                                                                                                                                                                                                                   | Alarm silence key                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                     |                |             |   |             |   |                |   |                   |   |               |
| 4                                                                                                                                                                                                                                                                                                                                                                                                                                   | Alarm message                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                         |                |             |   |             |   |                |   |                   |   |               |
|  <p>The illustration shows the control panel of the Panda iRes Warmer. A hand is pressing the yellow 'Silence' button located on the right side of the screen. The screen displays temperature readings: Baby Temp (C) at 37.7 and Control Temp at 36.6.</p>                                                                                      | <p><b>To silence the alarm.</b> There are two methods of silencing the alarm:</p> <ol style="list-style-type: none"> <li>1. Press the yellow button at the top of the control panel.</li> <li>2. Use the Hands Free Alarm Silence (see below).</li> </ol> <p>When you silence an alarm, the text message gets smaller and moves to the top of the screen and the red alarm light goes from blinking to solid. When the alarms have been silenced, the audio tone is temporarily turned off, but the warmer remains in alarm status until the alarm condition clears. When the silence period expires, the audio tone will resume, but the silence period varies with the alarm (for greater detail about alarm priorities and silence periods, see the Alarm Table at the end of this chapter). If there are multiple alarms, the text messages alternate every 2 seconds.</p>                                                                                                                                                                                                                                                                                                                                                                        |                |             |   |             |   |                |   |                   |   |               |

| Illustration                                                                      | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
|-----------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | <p><b>Hands Free Alarm Silence</b></p> <p>The motion sensor for the Hands Free Alarm Silence is located in the alarm light panel. Silence the alarm by gently waving your hand approximately 2-6 inches in front of the sensor.</p> <p>Silencing an alarm with the Hands Free Alarm Silence works exactly the same as pressing the alarm silence button for all alarms. Using the Hands Free Alarm Silence will allow you to silence alarms without touching a panel surface.</p> |




**WARNING:**


Do not block the alarm speaker located on the back of the control panel near the power outlets. Doing so may interfere with audio alarms.

# Temperature regulation alarms

## Baby Hot - Check Temp Probe

| Screen                                                                            | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |
|-----------------------------------------------------------------------------------|------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <p><b>Baby Mode</b></p> <p>The “Baby Hot - Check Temp Probe” alarm activates when the baby’s skin temperature is greater than 1.0°C above the Set Temperature. Silencing the alarm suspends the low priority audio alarm for twelve (12) minutes. When the baby’s temperature is greater than 2.0°C above the Set Temperature a high priority alarm activates. Silencing the alarm suspends the high priority audio alarm for five (5) minutes.</p> <p>The low priority alarm setting can be narrowed to 0.5°C using the “Setup” screen and “Pt Temp Alarm” keys.</p> <p><b>Manual Mode</b></p> <p>The alarm can be activated in the Manual Mode using the “Setup” and “Man Temp Alarm” keys. (See Chapter 3 for detailed description.)</p> <p><b>Action:</b> monitor the baby carefully and check the attachment of the probe on the baby’s skin.</p> |

## Baby Cold-Check Temp Probe

| Screen                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  <p>The screen displays the current baby temperature (35.3°C), the set temperature (36.5°C), and a red rectangular message box containing the text "Baby Cold - Check Temp Probe". The screen also shows a yellow bell icon at the top right, a blue "APGAR On/Off" button, and a menu bar with "Mode", "Manual", "Baby", "Pulse Rate", and "SpO2". At the bottom, there is a power switch, a GE logo, a sun icon, and a gear icon.</p> | 1    | <p><b>Baby Mode</b></p> <p>The "Baby Cold- Check Temp Probe" alarm activates when the baby's skin temperature is greater than 1.0°C below the Set Temperature. Silencing the alarm suspends the low priority audio alarm for twelve (12) minutes. When the baby's temperature is greater than 2.0°C below the Set Temperature a high priority alarm activates. Silencing the alarm suspends the high priority audio alarm for five (5) minutes.</p> <p>The low priority alarm setting can be narrowed to 0.5°C using the "Setup" screen and "Pt Temp Alarm" keys.</p> <p><b>Manual Mode</b></p> <p>The alarm can be activated in the Manual Mode using the "Setup" and "Man Temp Alarm" key. (See Chapter 3 for detailed description).</p> <p><b>Action:</b> monitor the baby carefully and check the attachment of the probe on the baby's skin.</p> |

## Check Baby

| Screen                                                                                                                                                                                                                                                                                                                                                                                                                                                             | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  <p>The screen displays the heater power level (100%), and a red rectangular message box containing the text "Check Baby". The screen also shows a yellow bell icon at the top right, a blue "APGAR On/Off" button, and a menu bar with "Mode", "Manual", "Baby", "Pulse Rate", and "SpO2". At the bottom, there is a power switch, a GE logo, a sun icon, and a gear icon.</p> | 1    | <p><b>Baby Mode</b></p> <p>The low priority "Check Baby" alarm activates, and an intermittent tone sounds, after the radiant heater has been running at 100% power for more than twelve minutes.</p> <p><b>Manual Mode</b></p> <p>The low priority "Check Baby" alarm activates, and an intermittent tone sounds after the warmer has been operating at greater than 25% heater power* has been exceeded for more than twelve minutes.</p> <p><b>Action:</b> check the attachment of the skin probe to the baby and monitor the baby's temperature closely.</p> <p>*Default power setting is 25%, but this can be adjusted on the Service Screen. Refer to the Service Manual.</p> |

## Check Baby-Heat Off

| Screen                                                                                                                                                                                                                                                                                                                                   | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  <p>The control panel displays the 'Check Baby - Heat Off' alarm in red on the main screen. The screen shows 'Baby Temp (C)' at 36.5, 'Heater Power' at 0%, and the mode is set to 'Baby'. The 'Check Baby - Heat Off' alarm is highlighted in red.</p> | 1    | <p><b>Baby Mode</b><br/>A continuous tone sounds after the radiant heater has been running at 100% power for more than fifteen minutes, and the "Check Baby" alarm" has gone unanswered for three minutes. The radiant heater then shuts off automatically and the high priority "Check Baby-Heat Off" alarm activates.</p> <p><b>Manual Mode</b><br/>A continuous tone sounds after the warmer has been operating at greater than 25% heater power* for more than fifteen minutes, and the "Check Baby" alarm has gone unanswered for three minutes. The radiant heater then shuts off automatically and the high priority "Check Baby-Heat Off" alarm activates.</p> <p><b>Action:</b> silence the alarm and check the attachment of the skin probe to the baby and monitor the baby's temperature closely.</p> <p>*Default power setting is 25%, but this can be adjusted on the Service Screen. Refer to the Service Manual.</p> |

## Confirm Probe Jack Connection

| Screen                                                                                                                                                                                                                                                                                                                                                   | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  <p>The control panel displays the "Confirm Probe Jack Connection" alarm in red on the main screen. The screen shows 'Baby Temp (C)' at 36.5, 'Set Temp' at 36.5, and the mode is set to 'Baby'. The "Confirm Probe Jack Connection" alarm is highlighted in red.</p> | 1    | <p><b>Baby Mode</b><br/>The "Confirm Probe Jack Connection" alarm activates when the skin probe jack is unplugged.</p> <p><b>Manual Mode</b><br/>It will also activate when there is no probe in the jack in the Manual Mode, if the Man Temp Alarm has been set.</p> <p><b>Action:</b> check that the skin probe is fully inserted into the temperature probe jack. If this does not cancel the alarm, check if the Man Temp Alarm has been set. If the alarm is still not cancelled, then replace the temp probe.</p> |

## Temperature Probe Failure

| Screen                                                                            | Step | Description                                                                                                                                                                                              |
|-----------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <p>This “Temp Probe Failure” alarm activates when the difference in the two thermistors in the skin temperature probe is greater than 0.5°C for over 6 minutes.</p> <p><b>Action:</b> replace probe.</p> |

## System Failure




### WARNING:

Do not use the warmer if the system failure alarm is activated. Remove the unit from service and refer to qualified personnel for repair.

If an electrical failure is detected, the system failure automatically shuts off the heater and triggers a two tone audio alarm that can not be silenced.

**Action:** make note of the error message and remove unit from service.

## Scale Alarms

| Screen                                                                              | Step | Description                                                                                                                                                                           |
|-------------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <p><b>Weight on scale above maximum</b><br/>This alarm means that the weight on the scale is above 8 kilograms (17.6 lb).</p> <p><b>Action:</b> check for other objects on scale.</p> |

## Alarm Table

| Alarm                         | Activation Criteria                                                                                                                      | Alarm Silence | Audio Signal*           | Mode                                                        |
|-------------------------------|------------------------------------------------------------------------------------------------------------------------------------------|---------------|-------------------------|-------------------------------------------------------------|
| Check Baby                    | Radiant % power @ 100% for > 12 min.                                                                                                     | 12 min.       | 2                       | Baby Mode                                                   |
|                               | Radiant heater % above "check baby alarm disabled" limit for > 12 min.                                                                   | 12 min.       | 2                       | Manual Mode                                                 |
| Check Baby - Heat off         | Radiant % power @ 100% for >15 min. and the "Check Baby" alarm has gone unanswered for three minutes.                                    | 15 min.       | 1                       | Baby Mode                                                   |
|                               | Radiant heater % above "check baby alarm disabled" limit for > 15 min. and the "Check Baby" alarm has gone unanswered for three minutes. | 15 min.       | 1                       | Manual Mode                                                 |
| Baby Cold - Check Temp. Probe | <2.0° C from Set Temp.                                                                                                                   | 5 min.        | 1                       | Baby Mode                                                   |
|                               | <1.0° C (<0.5° C) from Set Temp.                                                                                                         | 12 min.       | 2                       | Baby Mode                                                   |
| Baby Hot - Check Temp. Probe  | >2.0° C from Set Temp.                                                                                                                   | 5 min.        | 1                       | Baby Mode                                                   |
|                               | > 1.0° C (>0.5° C) from Set Temp.                                                                                                        | 12 min.       | 2                       | Baby Mode                                                   |
| Confirm Probe Jack Connection | No longer getting a reading from the temperature probe.                                                                                  | 2 min.        | 2 to 1 (after 30 secs.) | Baby Mode or Manual Mode when Manual Temperature Alarm set. |
| Temp. Probe Failure           | Two thermistors in a probe differ by 0.5° C or more.                                                                                     | 2 min.        | 2 to 1 (after 30 secs.) | Baby Mode or Manual Mode when Manual Temperature Alarm set. |

\* 1= high priority audio signal  
 2= low priority audio signal

This page intentionally blank.

# Chapter 2

## System Setup and Checkout Procedures

### Mechanical checks



#### WARNING:

Before using the Panda Warmer, read this entire manual. Attempting to use device without a thorough understanding of its operation may result in patient or user injury.



#### WARNING:

Do not perform the preoperational checkout procedure while the patient occupies the unit.



#### WARNING:

Complete the preoperational checkout procedure section of this manual before putting the unit into operation. If the equipment fails any portion of the checkout procedure it must be removed from use and repaired.



#### WARNING:

The use of accessories, replacement parts, or power cords other than those specified by the manufacturer may affect the performance of the device and could result in damage to the device or unsafe operating conditions.

| Illustration                                                                        | Step     | Description                                                                                    |
|-------------------------------------------------------------------------------------|----------|------------------------------------------------------------------------------------------------|
|  | <b>1</b> | Disconnect the power cord for the mechanical portion of the preoperational checkout procedure. |
|  | <b>2</b> | Examine the power cord for any signs of damage. Replace the cord if damage is evident.         |

| Illustration                                                                        | Step | Description                                                                                                                                                                                                              |
|-------------------------------------------------------------------------------------|------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 3    | Check that the plug retaining brackets for the power cord and the accessory outlets on the back of the control panel are in place.                                                                                       |
|   | 4    | Examine the unit overall for any damaged or missing parts.                                                                                                                                                               |
|  | 5    | Check that all the casters are in firm contact with the floor and that the unit is stable. Lock the caster brakes and check that they hold the unit in place. Release the brakes and check that the unit moves smoothly. |

| Illustration                                                                       | Step | Description                                                                                                                                                                                                                                                                                                                                          |
|------------------------------------------------------------------------------------|------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   | 6    | Check the operation of all four bedside panels. The bedside panels should lock securely in the upright position.                                                                                                                                                                                                                                     |
|  | 7    | Check the operation of the bed tilt mechanism. When you squeeze the tilt control and push down on the foot of the bed the head of the bed should raise easily, and should stay in position at any angle along its tilt path when you let go of the tilt control. The bubble levels on the side bedside panels should indicate the mattress is level. |

## Controller checks




### **WARNING:**

Do not use the warmer in the presence of flammable anesthetics: an explosion hazard exists under these conditions.



### **WARNING:**

Always connect the warmer directly to a hospital grade wall outlet. Connecting to a power strip or another piece of equipment may result in power failure errors.

| Illustration                                                                        | Step | Description                                                                                                                                                                                                                                                                                                   |
|-------------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   | 1    | Make sure the power cord is connected to the outlet on the unit and to the rated power supply.                                                                                                                                                                                                                |
|  | 2    | Switch on the power at the mains switch on the back of the unit, and at the standby switch on the front control panel. Verify the following: <ul style="list-style-type: none"><li>• All the displays and indicators light</li><li>• The software revision appears</li><li>• The prompt tone begins</li></ul> |

| Illustration                                                                        | Step | Description                                                                                                                                                                                        |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 3    | <p>Press the "Admit Baby" key to enter normal operation in Manual Mode. Select Heater Power percentage to silence the prompt tone.</p>                                                             |
|   | 4    | <p>Connect the baby temperature probe to the jack on the right side of the bed. If using a Panda Warmer with an in-bed scale, the temperature probe jack is above the In-bed scale connection.</p> |
|  | 5    | <p>Check the skin probe. Warm it by placing it between your fingers, and verify that the baby temperature reading increases.</p>                                                                   |

| Illustration                                                                        | Step | Description                                                                                                                                                                                                                                                                                                         |
|-------------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 6    | Unplug the skin probe and change the warmer to Baby Mode. Check that both visual and audio alarms trigger in the Baby Mode. Silence the alarm by using either the alarm silence key or the Hands Free Alarm silence.                                                                                                |
|   | 7    | Return the warmer to Manual Mode by pressing the "Mode" key. Select a Heater Power setting to silence the prompt tone.                                                                                                                                                                                              |
|  | 8    | If the unit is equipped with an elevating bed, check the operation of the bed elevation mechanism. Raise and lower the bed along its entire travel range, checking that the mechanism operates smoothly. Check that the two bed height adjustment pedals on either side of the unit raise and lower the bed height. |



**NOTE:** If readying the bed to admit a new patient, power down the warmer for 10 seconds, then turn power back on. Allow warmer to preheat the bed in Warmup Mode.

## Optional equipment checks

| Illustration                                                                      | Step     | Description                                                                                                                |
|-----------------------------------------------------------------------------------|----------|----------------------------------------------------------------------------------------------------------------------------|
|  | <b>1</b> | Check that all optional equipment is securely mounted.                                                                     |
|                                                                                   | <b>2</b> | Check the operation of any optional equipment with reference to their appropriate operation manuals.                       |
|                                                                                   | <b>3</b> | Setup any required suction or gas supply systems. Check them for leaks as described in their respective operation manuals. |



**WARNING:**

Do not place the warmer in a high ambient air flow environment.



**WARNING:**

Do not place the baby in the bed while in Warmup Mode, Service Mode, or while servicing or calibrating the unit.

This page intentionally blank.

# Chapter 3

## Operating the Warmer

---

### Basic Operating Procedure

---

**WARNING:**

Always set the brakes before placing a patient in the unit.

---

**WARNING:**

Do not leave the patient unattended when using the Panda iRes Warmer.

---

**WARNING:**

Check the patient's temperature periodically with an independent monitor to ensure the comfort and the safety of the patient. If the warmer is used for an extended time, it is recommended that the baby control mode be used. When an alarm is silenced, close monitoring of the patient's condition is required.

---

**WARNING:**

Use of electrosurgical units or other electrical field radiating equipment can affect the operation of the unit. Keep the patient probe lead as far away as possible from electrosurgical cables. Do not allow excess electrical cables to be laid on the bed platform. Use of electrosurgical units or other instruments that radiate electrical fields can cause indirect heating, by several tenths of a degree of the skin temperature probe due to absorbed electrical energy. When using these devices near the radiant warmer, operate the warmer in Manual Mode for maximum safety and use a skin temperature probe to monitor patient temperature. For added safety, you should consider use of the manual temperature alarm found in the Setup menu.

---

**WARNING:**

The use of phototherapy equipment or heated mattresses may raise the patient's temperature. Monitor patient temperature using a skin probe.

---

**WARNING:**

Radiant warmers may increase an infant's insensible water loss. Take appropriate measures to maintain the patient's fluid balance while caring for them in a radiant warmer.

---

**WARNING:**

Radiant energy can adversely affect blood components. When using intravenous tubing systems for delivery of blood components to patients occupying a warmer, shield any tubing with aluminum foil.

---

**WARNING:**

Radiant energy may cause more rapid urine evaporation, and may lead to inaccurate urine diagnostic test analysis and inaccurate weight measurements. Frequent diaper changes are recommended.

---





**WARNING:**

Use Baby Mode unless Manual Mode is specifically prescribed. While both modes require patient monitoring, Manual Mode requires constant attention. In Manual Mode, you must take the responsibility for detecting changes in the environment (high air flow, direct sunlight, phototherapy lamp usage, etc.) or the patient's condition and make heater power adjustments in response to these changes. In Baby Mode, the warmer automatically adjusts heater output to maintain the desired skin temperature, reducing (but not eliminating) the need to monitor the patient and make adjustments to the Baby Mode control temperature.



**WARNING:**

Do not leave the baby unattended while any bedside panels are lowered or removed.



**WARNING:**

When bedside panels are in an upright position, ensure that the bedside panels are properly locked in place.



**WARNING:**

The patient probe is not isolated from earth ground. Any additional equipment used with the Panda Warmer must comply with IEC 601.



**WARNING:**

Bed-to-heater spacing less than approximately 85 cm will result in incorrect operation and may adversely affect the patient's condition.



**WARNING:**

Do not place objects in the radiant heat path. Objects will be heated and could block heat to the baby.



**WARNING:**

Do not hang items from the heater head.



**WARNING:**

Do not connect unapproved equipment into accessory outlets.



**WARNING:**

Radiant energy can adversely affect blood components. When using intravenous tubing systems for delivery of blood components to patients occupying a warmer, limit the volume of a single transfusion to a quantity that can be infused within four hours or per hospital protocol. If necessary, tubing may be protected by infrared radiance reflective material.



**WARNING:**

Do not use other thermoregulation methods in conjunction with this device.



**WARNING:**

Use of this device is restricted to one patient at a time.




**WARNING:**

Cover patient's eyes if an observation light is used for an extended time. Prolonged exposure to an observation light can damage patient's eyes.

# Operating the Warmer (continued)

## Start up

| Illustration                                                                        | Step | Description                                                                                                                                                                   |
|-------------------------------------------------------------------------------------|------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 1    | Plug the unit into a properly rated AC power outlet and set the caster brakes.                                                                                                |
|   | 2    | Place the mains power switch, located by the outlets on the back of the unit, in the "On" (I) position.                                                                       |
|  | 3    | Place the power Stand By switch, located on the left side of the controller, in the <b>On (I)</b> position. Check the control panel to ensure the screen is working properly. |

| Illustration                                                                        | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                             |
|-------------------------------------------------------------------------------------|------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 4    | <p>Always preheat the bed before placing the baby in the bed. The Panda Warmer automatically starts in Warmup Mode. During Warmup Mode check baby alarms are deactivated. If you do not press the <b>"Admit Baby"</b> key or the <b>"Apgar On/Off"</b> key during the first minutes after start up, the display will show "Warming..." and the radiant heater will automatically run at 100% power. After 7 minutes, the display indicates <b>"Warmup Complete..."</b> and the default heater power setting drops to 50%. The display continues to flash <b>"Admit Baby"</b> until you press the <b>"Admit Baby"</b> key and start normal operation</p> |
|   | 5    | <p>Pressing the <b>"Admit Baby"</b> key or the <b>"Apgar On/Off"</b> key brings the Panda Warmer to Manual Mode at 100% power. An audible tone alerts the user to select a heater power setting.</p> <p>Press the <b>"Admit Baby"</b> key prior to placing the baby on the bed. You may press the <b>"Admit Baby"</b> key to start normal operation of the warmer at any time</p>                                                                                                                                                                                                                                                                       |
|  | 6    | <p>Select heater power setting using the up or down temperature keys.</p>                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |

| Illustration                                                                        | Step | Description                                                                                                                                                                                                                                              |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 7    | <p>Place the baby in the bed</p> <hr/> <p><b>WARNING:</b><br/>Do not place baby in bed during Warmup Mode.</p>                                                                                                                                           |
|   | 8    | <p>For Baby Mode operation, connect the skin temperature probe to the temperature probe jack. In the Manual Mode, use a probe if you wish to display the patient skin temperature. See "Attaching the Skin Temperature Probe" later in this section.</p> |
|  | 9    | <p>If Baby Mode is desired, press the "Mode" key to toggle to Baby Mode. Current mode setting, "Manual" or "Baby" mode, is highlighted in blue on the left side of screen.</p>                                                                           |


| Illustration                                                                      | Step | Description                                                                       |
|-----------------------------------------------------------------------------------|------|-----------------------------------------------------------------------------------|
|  | 10   | Select a temperature setting using the <b>up</b> or <b>down</b> temperature keys. |




**WARNING:**

Do not connect unapproved equipment into accessory outlets.

# Panda iRes Warmer Flowchart



## Attaching the Skin temperature Probe

| Illustration                                                                       | Step     | Description                                                                                                                                                                                                                                                                                                                                                                                                                                |
|------------------------------------------------------------------------------------|----------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   | <b>1</b> | <p>Place the skin probe on the baby's skin. If the baby is prone, place the probe on the baby's back. If the baby is supine, you should place the probe on the baby's abdomen (or at a site consistent with your unit's policy and procedures).</p> <p>Ensure that the baby's skin is clean and dry before applying the probe.</p> <p>Before using the probe, inspect it to make sure it is clean and free from any foreign materials.</p> |
|  | <b>2</b> | <p>Select a location on the baby's abdomen (or back) so that the probe is directly in the path of the radiant heat.</p> <p><b>1</b> shows the heat reflective patch.</p> <p><b>2</b> shows a reusable patient temperature probe.</p> <p><b>3</b> shows a disposable patient temperature probe.</p>                                                                                                                                         |



### **WARNING:**

Do **not** place the probe between the baby and the mattress; inaccurate readings will result.



### **WARNING:**

Do **not** remove the reflective foil from the disposable probe or reflective patch. Do **not** use a probe without a reflective patch. Replace the patch when repositioning the probe or when adhesive strength degrades.



### **WARNING:**

Do **not** pull on the probe wire. Remove the probe from the skin by gently lifting the adhesive patch. Remove the probe from the jack panel by grasping the plug at the panel



### **WARNING:**

Regularly check that the probe is attached- if the probe is **not** in contact with the baby's skin inaccurate readings will result



**WARNING:**

Only use Ohmeda probes; other manufacturers probes are not calibrated to GE Healthcare equipment. Using probes from other manufacturers may cause inaccurate temperature readings, may not comply with safety standards, and will void your GE Healthcare equipment warranty.



**WARNING:**

When using phototherapy lamps, the probe must be directly in the path of the radiant heat of the lamp; do **not** place the probe in an area shielded from the lamp's light. The phototherapy lamp may raise the baby's skin temperature



**WARNING:**

Do **not** use rectal temperatures to control the baby's temperature



**WARNING:**

Do **not** remove the probe from its storage bag until required for use. Replace the probe if the cable or tip becomes damaged.



**WARNING:**

Do **not** re-use a disposable probe. Cleaning and reusing a single use probe may damage the probe and result in inaccurate readings.

Attaching the Skin Temperature (continued next page)

## Attaching the Skin Temperature Probe (continued)

| Illustration                                                                       | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                             |
|------------------------------------------------------------------------------------|------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   | 2    | To attach the disposable probe (single use only), peel the paper backing from the adhesive side and apply to the baby's skin, with the reflective foil side up. To attach the reusable probe, use the heat reflecting patch. Place the metal side of the probe against the baby's skin, peel the paper backing from the adhesive side of the patch and place the patch over the probe with the reflective foil side up. |
|                                                                                    | 3    | Route the probe wire through the gaps at the corner of the bedside panels or through the one of the slots in the rear bedside panel.                                                                                                                                                                                                                                                                                    |
|  | 4    | Plug the temperature probe into the jack.                                                                                                                                                                                                                                                                                                                                                                               |



### NOTE:

The Panda iRes Warmer cannot differentiate between an increase in core temperature with cold skin (fever), and low core and skin temperatures (hypothermia). Patient temperature should be verified with an axillary thermometer.

## Setting the manual temperature alarm

The “Man.Temp Alarm” is a safety mechanism that you can activate whenever you manage the baby’s temperature in Manual Mode. It provides a selectable alarm set point that will notify you whenever the baby is getting too hot or too cold, just as the warmer does in Baby Mode.

To activate the alarm in Manual Mode, you must take the following steps:

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
|-------------------------------------------------------------------------------------|------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 1    | Insert the temperature probe into the probe jack.                                                                                                                                                                                                                                                                                                                                                                                                                                         |
|   | 2    | Attach the skin temperature probe onto the baby's skin.                                                                                                                                                                                                                                                                                                                                                                                                                                   |
|  | 3    | <p>Press the <b>“Menu”</b> soft key on the right side of the control panel:</p> <ol style="list-style-type: none"><li>Press the <b>“Setup”</b> soft key on the submenu</li><li>Press the <b>“Next”</b> soft key on the submenu</li><li>Press the <b>“Man. Temp Alarm”</b> soft key</li><li>In most cases, the <b>“Man.Temp Alarm”</b> soft key will be in the off position. Select whether you want a safety notification of 0.5 or 1.0 °C around the desired baby temperature.</li></ol> |

| Illustration                                                                        | Step | Description                                                                                                                                                                                                                                                                                                                                                |
|-------------------------------------------------------------------------------------|------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 4    | <p>If in Manual Mode, briefly go to the Baby Mode on the left side of the control panel. Press the Mode key to change the warmer from Manual Mode to Baby Mode. Note: If in Baby Mode, skip to Step #6.</p> <hr/> <p> <b>NOTE:</b> If in Baby Mode, skip to Step #6.</p> |
|   | 5    | <p>Enter the set temp that will serve as the "Man. Temp Alarm" set point.</p>                                                                                                                                                                                                                                                                              |
|  | 6    | <p>Return to Manual Mode and adjust the heater power up or down to achieve the desired baby's skin temperature.</p>                                                                                                                                                                                                                                        |

## Raising or lowering the bedside panels

| Illustration                                                                        | Step | Description                                                                                                                                                                            |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 1    | To lower a bedside panel, pull it up and then pull the top edge away from the bed.                                                                                                     |
|   |      | To raise a bedside panel, swing it to the upright position, then allow it to engage in the latched position.                                                                           |
|  | 2    | To remove a bedside panel, lower the panel and hold it parallel with the floor. Press the side panel release button (arrow) with one hand and remove the panel.                        |
|                                                                                     |      | To replace a bedside panel, insert the panel into the fixed end pin first. Then press the side panel release button in with one hand, place the panel in position and release the pin. |


| Screen                                                                            | Step | Description                                                                                      |
|-----------------------------------------------------------------------------------|------|--------------------------------------------------------------------------------------------------|
|  | 3    | <p>Remove the rear bedside panel by pressing the release button and lifting up on the panel.</p> |



#### **WARNING:**

The Panda iRes Warmer and Giraffe Warmer bedside panels are different heights. Use only the Panda iRes Warmer bedside panels on the Panda iRes Warmer.

## Tilting the bed

| Illustration                                                                       | Step | Description                                                                                                                                                                                                                                                                                                                                                                        |
|------------------------------------------------------------------------------------|------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | 1    | <p>The bed tilts twelve degrees to allow feet up or head up positioning of the baby. The tilt release is located underneath the foot of the bed. Grasp both the handle and metal release in one hand and squeeze. Push down or lift up on the foot of the bed to the desired position. Letting go of the release at any point on the tilt path locks the bed in that position.</p> |



#### **WARNING:**

Tilting the mattress from its horizontal position can result in uneven heat distribution.

## Raising and lowering the bed



### **WARNING:**

Before raising or lowering the bed, check that there is adequate slack in tubing and leads and that no obstructions limit the range of motion.



### **CAUTION**

Objects placed on the warmer legs may be damaged by raising or lowering the bed and could create a tripping hazard.

| Illustration | Step |                                                                                                                                                                                                                                                                                                                                                                                                                                        |
|--------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|              | 1    | <p>If the unit is equipped with an elevating base, the bed can be raised or lowered using the bed height pedals located on the legs on either side of the warmer. The bed height can be adjusted low enough for use by a seated caregiver, or can be raised high enough for procedures performed by standing caregivers.</p> <p>Fixed height bed models can be adjusted by following the instructions found in the service manual.</p> |



### **WARNING:**

Do not attempt to adjust the fixed base while a patient is in the bed. After adjustments to the fixed base height have been made, ensure the fixed base is locked in place before placing a patient in the bed.

## Using the X-ray tray



### **WARNING:**

Never place the baby on the X-ray tray.

| Illustration | Step     |                                                                                                                                                                                                                                                                                                                                               |
|--------------|----------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|              | <b>1</b> | <p>Lower the side bedside panel and place a film cassette on the optional tray that slides out from under the mattress for X-ray procedures. The tray slides out on either side of the bed. The cassette can be slid into the cavity under the mattress without moving the baby.</p> <p>Note: Mattress removed for illustrative purposes.</p> |

## Using the observation and procedure lights

| Illustration | Step     | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|--------------|----------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|              | <b>1</b> | <p>Two observation lights are located in the heater head on either side of the heater dish (1). The observation lights serve as general illumination of the bed surface and can be set to the desired light intensity using the observation light dimmer knob located on the lower right corner of the control panel.</p> <p><b>Using the procedure light (optional)</b><br/>An optional aimable procedure light (2) is located just below the heater dish. The procedure light provides a more intense spot light to illuminate a precise area during procedures. The procedure light can be directed at any point on the mattress and is activated using the switch on the light handle (3). The procedure light is not intended for use during surgical procedures.</p> |

## Uninterruptible Power Supply (UPS)

**NOTE:** The UPS is available only in select markets. To understand availability within your region, contact your sales representative.

The Giraffe UPS is a medical grade power supply offering uninterrupted power supply to the Panda iRes Warmer. The UPS is designed to securely mount to a single Panda iRes Warmer via specific mounting hardware provided in the Giraffe UPS kit, or separately in the Shelf Hardware Mounting Kit. The UPS may serve to benefit customers experiencing power line voltage disturbances, or Brown Outs, or temporary power outages, or Black Outs. Per the UPS Specifications, the Giraffe UPS provides power to the Panda iRes Warmer in situations where the product is disconnected from a wall receptacle.

For specifications and a detailed description of the installation procedure, refer to the Giraffe UPS Installation Instructions (provided with the UPS).

## Using the In-Bed Scale (optional)



The Panda Warmer can be equipped with its own in-bed scale that is operated from the control panel screen.



### Scale installation



| Illustration                                                                       | Step     |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|------------------------------------------------------------------------------------|----------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|  | <b>1</b> | <ol style="list-style-type: none"><li>1. Remove the clear Plexiglas plate and mattress from the bed.</li><li>2. Fit the scale weighing platform into the recesses in the bed.</li><li>3. Route the scale cable though the gap at the corner of the bedside panels and plug the connector into the jack.</li><li>4. Place the clear plate and mattress on the weighing platform. Make sure they are properly seated on the platform.</li></ol> <hr/> <p> <b>NOTE:</b> The scale may be connected or disconnected at any time.</p> <hr/> |


## Weighing procedure

The baby should be in approximately the center of the bed. Stuffed animals and other objects should not lean against bedsides. All leads, I.V. tubes and ventilator tubes should be secured. Blankets may be tucked under the mattress, but must not be tucked under the weighing platform.

| Screen                                                                             | Step | Description                                                                                                                       |
|------------------------------------------------------------------------------------|------|-----------------------------------------------------------------------------------------------------------------------------------|
|   | 1    | Press the "Menu" key, then select the "Scale" key to bring up the scale menu. If the scale is not connected, no icon will appear. |
|  | 2    | Select "Weigh" from the menu to initiate weighing.                                                                                |

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                            |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 3    | <p>You will be prompted to check “<b>Level bed...</b>” to ensure that the bed is level during the weighing procedure.</p>                                                                                                                                                              |
|   |      | <p>If the bed is not level, then level the bed by sighting the bubble level on the side bedside panel and using the bed tilt control lever.</p> <p>If the bed is level, proceed to the next step.</p>                                                                                  |
|  | 4    | <p>In order to “zero” the scale, the display will prompt you to lift the baby by showing the “<b>Lift Baby</b>” icon while a tone sounds. Lift the baby and any tubing or leads attached to the baby. Make sure that arms, legs, blankets, and clothing are clear of the mattress.</p> |

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                                            |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 5    | Hold the baby while the "Hold Baby" icon appears.                                                                                                                                                                                                                                                                                                                                      |
|   | 6    | A second tone sounds and the "Replace Baby" icon appears on the screen. Place the baby back down on the mattress while holding up any leads or tubes attached to the baby.                                                                                                                                                                                                             |
|  | 7    | <p>The scale will now calculate the baby's weight which will appear in the display.</p> <p> <b>NOTE:</b> The scale weighs any object on the platform, so if you replace the baby without holding up lead and tubes, the weight of the leads and tubes will be included with the baby's weight.</p> |

| Screen                                                                              | Step | Description                                                                                                                                                                                                                                                                                                                                                                                  |
|-------------------------------------------------------------------------------------|------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | 8    | <p>Push the “Reweigh” key within five minutes of the last weighing sequence if the weight is in doubt</p> <hr/> <p> <b>NOTE:</b> The most likely cause for using <b>Reweigh</b> is failure to keep lines, tubes and medical equipment clear of the mattress when placing the baby back on the mattress.</p> |
|   | 9    | Selecting “Last weight” will display the most recent weight value.                                                                                                                                                                                                                                                                                                                           |
|  | 10   | Push the “g/oz” key to toggle the weight display between “grams” (default) and “ounces”.                                                                                                                                                                                                                                                                                                     |

## Using the Integrated SpO<sub>2</sub> Monitor (optional)

The Panda iRes Warmer can be equipped with its own integrated SpO<sub>2</sub> Monitor that is operated from the control panel screen.

Refer to the SpO<sub>2</sub> Monitor setup, check out and use instructions in the SpO<sub>2</sub> - Masimo SET Option, Operation and Maintenance Supplement.

## Using the Integrated Resuscitation System (optional)

Refer to the integrated Resuscitation System for setup, check out and use instructions in the Bag and Mask Option, Operation and Maintenance Supplement or the Resuscitation System - T-piece Option Operation and Maintenance Supplement, depending on your Panda iRes Warmer's resuscitation equipment setup.

## Mounting optional equipment



### **WARNING:**

Overloading the shelves and rails can affect the stability of the unit. Always try to evenly distribute the weight of options on both sides of the unit for a more balanced load. Limit the total load on the rails to 22 kg. Do not mount more than 19 kg to one rail. Do not use more than one GCX Mount on the rails at one time.




---



### **WARNING:**

Do not transport with pass through drawer(s) open.

---

| Illustration                                                                        | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
|-------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|    | <p>Rail system components mount to the uprights and provide access to commonly used equipment such as suction regulators, flowmeters, collection bottles, instrument shelves, etc. This patented design consists of a dovetail shaped aluminum extrusion and a positive locking mounting block. Mounting blocks attach various options to the rail system.</p>                                                                                                                                           |
|   | <p>Mounting rail system components:</p> <ol style="list-style-type: none"> <li>1. Loosen the mounting screw on the mounting block.</li> <li>2. Place the mounting block in position on the rail</li> <li>3. Tighten the mounting screw</li> </ol> <p>Release the rail system component by loosening the mounting screw.</p>                                                                                                                                                                              |
|  | <p><b>Mounting universal adapter plate.</b><br/>The Universal adapter plate allows bottle slides and additional bracket to mount to the rail system.</p> <ol style="list-style-type: none"> <li>1. Loosen the two mounting screws (1) on the side of the adapter plate with the hex key provided with the plate.</li> <li>2. Place the adapter plate in position on the rail.</li> <li>3. Tighten the two mounting screws.</li> </ol> <p>Release the adapter plate by loosening the mounting screws.</p> |

# ThermaLink

The ThermaLink Serial data interface and Nurse Call connections are options offered with the Panda Warmer. Your unit has these features if there is a 9 pin connector on the back of the heater housing, near the power receptacles.

The cable length shall not exceed 15 feet (4.5 meters).

## Using the Serial Data interface



### **WARNING:**

Remote monitoring does not replace the need for direct patient observation by qualified medical personnel.

The ThermaLink serial data output can be used with a computer or a commercial RS-232 monitor. Because of the wide variety of applications and systems, detailed information on decoding the data stream appears in the Service manual. For details of the RS-232 protocol and the connector pinout, refer to the Service manual.



### **WARNING:**

The computer or RS-232 monitor's user program must continuously check the data link. The program should constantly verify connection to the warmer controller and check for updated data.

## Using the Nurse Call System interface



### **WARNING:**

Remote monitoring does not replace the need for direct patient observation by qualified medical personnel.



### **WARNING:**

If you connect the Nurse Call output to system which uses the normally open connection, a disconnected Nurse Call cable will not trigger an alarm.

The Nurse Call connector lets you use the warmer with your current remote alarm system. Nurse Call alarms trigger for but are not limited to:

- Patient Temperature Alarms
- Control Temperature Alarms
- Probe Failure Alarms
- System Failure Alarms

The Nurse Call alarms works with the warmer's audible alarm. Silencing the audible alarm on the warmer stops the Nurse Call alarm even if the alarm condition still exists. At the end of the silence period, the Nurse Call alarm and the audible alarm reactivate unless the condition has been resolved. The alarm silence period ends prematurely if another alarm triggers. Refer to the Service manual for additional information on Nurse Call connections.

## Nurse Call checkout

| Steps | Description                                                                                                                                  |
|-------|----------------------------------------------------------------------------------------------------------------------------------------------|
| 1     | Complete the basic operating procedure in chapter 3.                                                                                         |
| 2     | Verify proper operation of the Nurse Call station.                                                                                           |
| 3     | Connect the Nurse Call connector to the warmer.                                                                                              |
| 4     | Place the unit into Baby Mode and unplug the patient probe to trigger an alarm. Verify that you also get an alarm at the Nurse Call station. |

This page intentionally blank.

# Chapter 4

## Maintenance and Cleaning

---



### **WARNING:**

The heater, lamps and surrounding areas are hot enough to cause skin burns. Unplug the unit and allow the heater to cool before disassembly or cleaning.



### **WARNING:**

Always disconnect the power before performing service or maintenance procedures detailed in this manual. Apply power only if you are specifically instructed to do so as part of the procedure.



### **WARNING:**

Thoroughly air dry the unit after cleaning with flammable agents. Small amounts of flammable agents, such as ether, alcohol or similar cleaning solvents left in the unit can cause a fire.



### **WARNING:**

Never oil or grease oxygen equipment unless a lubricant that is made and approved for this type of service is used. Oils and grease oxidize readily, and in the presence of oxygen, will burn violently.

## Repair Policy

Warranty repair and service should be performed by a GE Healthcare Service Representative or at the GE Healthcare Service and Distribution Center. To contact a GE Healthcare Service Representative, call the GE Healthcare Service Office listed on the back cover.

Do not use malfunctioning equipment, including equipment that does not pass the checkout procedure. Non-warranty repairs may be performed by a GE Healthcare Representative or by a competent, trained person who has experience in repairing devices of this nature. Refer to the Service Manual for a list of service parts and for instructions on how to service and calibrate the unit. After service, follow the checkouts procedures prior to returning the unit to service.



### **CAUTION:**

Detailed information for more extensive repairs is included in the service manual solely for the convenience of users having proper knowledge, tools and test equipment, and for service representatives trained by General Electric.

# Maintenance schedule

The unit should be maintained in accordance with the preventative maintenance procedures detailed in the Service Manual. Service maintenance must be performed by a technically competent individual.

## Operator maintenance

This schedule lists the minimum frequencies. Always follow hospital and local regulations for required frequencies.

### Every week

Clean the warmer and disinfect as required, or between each patient.

## Service maintenance

This schedule lists the minimum frequencies. Always follow hospital and local regulations for required frequencies.

### Annually

Perform the electrical safety checks as described in the service manual.

Calibrate the scale as described in the service manual.

### Every two years

Replace the battery as described in the service manual.



**NOTE:** The battery is used to sound the power failure alarm and to power memory circuits during a power failure.

---

## Cleaning Instructions

After each patient use, follow your hospital's infection control procedures for surface disinfection.

Wipe down the surfaces of the warmer with a soft cloth dampened with a disinfectant-detergent solution. Always follow the cleaning solution manufacturer's direction for use. Dry all surfaces with a soft cloth to remove any cleaner residue.



**WARNING:**

Do not clean the radiant heating element inside the protective grid.



**WARNING:**

Disconnect the power cord before cleaning the warmer.



**CAUTION:**

Electronic devices in the microprocessor controller are susceptible to damage from discharges of static electricity. These devices are adequately protected, but can be damaged if the unit is disassembled beyond that recommended for cleaning and maintenance.



**WARNING:**

The heater, lamps and surrounding areas are hot.



**WARNING:**

Do not spray cleaning solution into the vents on the back of the heater housing; this can damage electronics inside the unit.

## Cleaning the Warmer

| Steps | Description                                                                                                                                |
|-------|--------------------------------------------------------------------------------------------------------------------------------------------|
| 1     | Lower elevating base (or raise it depending upon bed height) to ergonomically comfortable cleaning position.                               |
| 2     | Turn bed off at standby power switch.                                                                                                      |
| 3     | Unplug from power outlet.                                                                                                                  |
| 4     | Move bed to cleaning room/area.                                                                                                            |
| 5     | If bed was previously on, allow it to cool for at least 30 minutes.                                                                        |
| 6     | Obtain usual cleaning/disinfectant solution.                                                                                               |
| 7     | Remove all ancillary equipment (e.g. patient temperature probe, shelves, etc.)                                                             |
| 8     | Empty drawer module.                                                                                                                       |
| 9     | Remove all 4 bedside panels.                                                                                                               |
| 10    | Unplug scale if installed.                                                                                                                 |
| 11    | Remove the clear Plexiglas plate and mattress from the bed.                                                                                |
| 12    | Remove scale.                                                                                                                              |
| 13    | Remove x-ray tray.                                                                                                                         |
| 14    | Wipe down component parts of bed and chassis base with recommended cleaning solution per hospital infection control policy. Rinse and dry. |
| 15    | Reinstall chemically disinfected component parts in reverse order from breakdown steps.                                                    |
| 16    | Chemically disinfect and dry control panel.                                                                                                |
| 17    | Chemically disinfect and dry dove tail rails.                                                                                              |
| 18    | Chemically disinfect and dry drawer module.                                                                                                |
| 19    | Chemically disinfect elevating base and legs.                                                                                              |
| 20    | Re-plug scale into probe jack panel, if not already done.                                                                                  |
| 21    | Reattach all clean ancillary equipment (e.g. patient temperature probe, shelves, etc.).                                                    |
| 22    | Plug bed into power outlet.                                                                                                                |
| 23    | Turn bed on at standby power switch.                                                                                                       |
| 24    | Prepare bed for next admission.                                                                                                            |

## Cleaning Solutions

Cleaning solutions that may be used safely:

| Generic Formulation          | Maximum Concentration |
|------------------------------|-----------------------|
| Sodium Hypochlorite (bleach) | 0.5% Aqueous Solution |
| Glutaraldehyde               | 2%                    |
| Hydrogen Peroxide            | 6%                    |
| Iodophor Solution            | 0.27%                 |
| Cavicide®                    | 100% spray            |

Do not use the following cleaners; they will damage the parts you are cleaning and are not recommended:

- Isopropyl Alcohol (in concentrations greater than 15%)
- Quaternary Ammonium (such as Virex)
- Solvents (such as acetone)



**CAUTION:**

Use of cleaning/disinfecting solutions containing chemicals not listed above, i.e. alcohol, acetone, etc., or chemicals in greater concentrations than those listed above, may damage the probe.

# Cleaning and disinfecting individual components

## Patient Probe (reusable)

**CAUTION:**

Avoid placing excessive strain on the probe lead. When cleaning, be careful not to pull on or bend the lead at the probe tip. Always remove the probe from the incubator by grasping the plug at the panel. Do not pull on the probe lead.

**CAUTION:**

Do not apply cold sterilization or cleaning solutions to the probe connector.

**CAUTION:**

Do not autoclave or gas sterilize the skin temperature probe. Do not immerse the probe in liquid cleaner. Avoid placing excessive strain on the probe lead. Always remove the probe by grasping the plug at the panel. Do not pull on the probe lead. These precautions will prevent damage to the probe.

**CAUTION:**

Do not allow cleaning fluid to leak into probe and electrical connectors. Equipment damage may occur.

| Steps | Description                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
|-------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1     | <p>Determine if the patient probe is disposable or reusable:</p> <ul style="list-style-type: none"><li>Reusable probes use a separate, heat reflecting patch, are gray and have a round metal disk at the patient end.</li><li>Disposable probes come with a smaller heat reflecting patch already attached, are white and have no metal disk at the patient end.</li></ul> <p> <b>NOTE:</b> Disposable skin temperature probes cannot be cleaned and are intended for single patient use</p> |
| 2     | Clean the reusable patient temperature probe by gently wiping with a soft damp cloth containing a disinfecting agent safe for use on the probe materials. Always be sure to wipe dry all cleaning agents after cleaning.                                                                                                                                                                                                                                                                                                                                                         |

## Cleaning other components

**CAUTION:**

Do not clean the unit with organic solvents, scouring compounds, strong acids, or strong bases. These compounds may damage components.

---

**CAUTION:**

To minimize the generation of static electricity, do not polish the side panels with a dry cloth.

---

**CAUTION:**

Do not autoclave or gas sterilize any of the plastic parts.

---

Divide the components according to cleaning methods. Methods other than those detailed in this section may damage the unit. Always be sure to wipe dry all cleaning agents after cleaning.

Apply the cleaning solutions with a clean cloth or sponge. Dry the parts with a clean damp soft cloth to avoid scratches.



**NOTE:** Do not soak parts in cleaning solutions. Always wipe parts dry of all cleaning solutions. Following these two recommendations will greatly extend the life of the parts.



**NOTE:** Any parts you clean with iodophor solution will stain yellow.



**NOTE:** Do not allow excess cleaning solution to seep in between plastic parts (for example: between the side panels and lock or hinges) where it can not be easily wiped dry with a cloth.

This page intentionally blank.

# Appendix A

## Options

The following accessory options are available for use with the device. Other options available for use with this device may be obtained through your sales representative



### **WARNING:**

The use of accessories, sensors, and cables other than those specified may result in increased emission and/or decreased immunity.

| Accessory                                        | Part Number                       |
|--------------------------------------------------|-----------------------------------|
| Disposable patient probe (box of 10)             | 2074816-001                       |
| Disposable patient probe (box of 50)             | 2074817-001                       |
| Reusable patient probe                           | 2075796-001                       |
| Heat reflecting patch (box of 50)                | 0203-1980-300                     |
| Giraffe In-Bed Scales                            | Gravity Zone Specific Part Number |
| Corner trays SE & NW                             | 6600-1793-500                     |
| Corner trays SW & NE                             | 6600-1794-500                     |
| Swivel instrument shelf (12 in x 12 in) 90°      | 6600-0865-700                     |
| Swivel instrument shelf (12 in x 12 in) 360°     | 6600-0513-801                     |
| Giraffe tubing management arm                    | 6600-0837-800                     |
| Giraffe silo support                             | 6600-0853-800                     |
| Patient restraint (pack of 5)                    | M1169540                          |
| Giraffe pressure diffusing mattress              | 2076066-001                       |
| Pressure Diffusing mattress cover sheet          | 6600-0688-800                     |
| Dovetail basket assembly                         | M1187887                          |
| Dovetail rail extension                          | 6600-0852-800                     |
| Ventilator mounting pole                         | 0217-5357-800                     |
| IV pump mounting post (20 in x 1 in)             | 0217-5376-800                     |
| Rotating and adjustable IV pole                  | 6600-0851-800                     |
| IV pole dual hook 24 inch                        | 6600-0491-801                     |
| IV pole dual hook 12 inch                        | 0217-5378-800                     |
| Utility post (3.5 in x 1 in) dovetail            | 0217-5374-800                     |
| North end dovetail handle                        | M1159084                          |
| Cord wrap holders (2)                            | M1151830                          |
| Retaining clips (6)                              | 6600-0055-851                     |
| Resuscitation bag and mask holder                | 6600-2150-500                     |
| Giraffe Easy-load E-cylinder holder              | 6600-0836-800                     |
| E-cylinder holder 2 sliding parts                | M1151570                          |
| Cylinder holder XL International 2 sliding parts | M1151571                          |

| <b>Accessory</b>                                    | <b>Part Number</b>            |
|-----------------------------------------------------|-------------------------------|
| Giraffe Stand-Alone Resuscitation System T-Piece    | Language Specific Part Number |
| Giraffe Stand-Alone Resuscitation System Bag & Mask | Language Specific Part Number |
| Giraffe Blue Spot PT Lite                           | Language Specific Part Number |
| BiliSoft LED Phototherapy System                    | Language Specific Part Number |
| Giraffe Shuttle                                     | Language Specific Part Number |
| Tubing Mangement South Wall Panda                   | M1092506                      |
| Tubing Mangement North Wall Panda                   | 6600-2145-500                 |
| HFOV side walls(2) east/west Giraffe                | M1092332                      |
| Higher Side Wall Kit                                | M1179223                      |

# Appendix B

## Specifications

### Power Requirements and Accessory Outlets

| Power Requirements        | Accessory Outlets      |
|---------------------------|------------------------|
| 5.25 A @ 100v ~, 50/60 Hz | 2 A @ 100v ~, 50/60 Hz |
| 4.57 A @ 115v ~, 50/60 Hz | 2 A @ 115v ~, 50/60 Hz |
| 2.39 A @ 220v ~, 50/60 Hz | 1A @ 220v ~, 50/60 Hz  |
| 2.28 A @ 230v ~, 50/60 Hz | 1 A @ 230v ~, 50/60 Hz |
| 2.19 A @ 240v ~, 50/60 Hz | 1A @ 240v ~, 50/60 Hz  |

### Standards

| Designed to meet requirements of: |                                 |
|-----------------------------------|---------------------------------|
| IEC 60601-2-21 with amendment     | 21 CFR CH-1 Section 1020.30 (n) |
| IEC 60601-1 with amendment        | UL 60601-1                      |
| IEC 60601-1-2 with amendment      | BSEN - 45501 with amendment     |
| CSA/CAN C22.2 # 601.1             |                                 |

### Operating Environment

| Temperature        | 18 to 30°C                                |
|--------------------|-------------------------------------------|
| Humidity           | 5 to 75% Non-condensing relative humidity |
| Pressure           | 70 - 106 kPa                              |
| Air Velocity       | up to 0.3 m/sec.                          |
| Storage conditions |                                           |
| Temperature        | -25 to 60°C                               |
| Humidity           | 0 to 85% Non-condensing relative humidity |
| Pressure           | 50 to 106 kPa                             |

### User Control Settings

|                                                          |
|----------------------------------------------------------|
| Patient control temperature 34-37.5°C in 0.1° increments |
| Radiant heat power 0-100% in 5% increments               |

## Performance

| System                                   |                                                                                              |
|------------------------------------------|----------------------------------------------------------------------------------------------|
| Warmer expected service life             | Approx. 8 years (see Note, below).                                                           |
| Heater Element                           | 360 Watts                                                                                    |
| Heater Output                            | 27 mW/cm <sup>2</sup>                                                                        |
| VA rating                                | 805 VA                                                                                       |
| Patient temperature measurement accuracy | ± 0.3°C @ 30°C to 42°C                                                                       |
| Temperature probe accuracy               | ± 0.1°C @ 30°C to 42°C                                                                       |
| Observation Light                        | 2 dimmable 35W halogen bulbs; estimated life 3000 hrs based on manufacturer's specifications |
| Procedure Light                          | 2000 lux* average; estimated life 3000 hrs.                                                  |

\*At nominal voltage.

| Weight scale     |               |
|------------------|---------------|
| Functional range | 300 g to 8 kg |
| Accuracy         | ± 10 g        |



**NOTE:** The warmer is designed to last at least 8 years in normal use when operated, maintained and serviced in accordance with the instructions provided in both the Operations and Maintenance, and Service Manuals.

## Mechanical Specifications

|                |                                |
|----------------|--------------------------------|
| Height:        | 193 - 218 cm                   |
| Width:         | 64 cm.                         |
| Depth:         | 119 cm                         |
| Weight:        | 100 kg                         |
| Mattress Size: | 66 x 48 x 2 cm                 |
| Bed Capacity:  | 14 kg                          |
| Bed Tilt:      | +/- 12 degrees continuous tilt |

| Options                |                     |
|------------------------|---------------------|
| Storage drawer package | 6.8 kg maximum load |
| Instrument shelf       | 3.6 kg maximum load |

## Uninterruptible Power Supply (UPS) Specifications

For UPS specifications, refer to the Giraffe UPS Installation Instructions (provided with the UPS).

# Appendix C

## Electromagnetic Compatibility

---

### Electromagnetic compatibility (EMC) guidance

**Safety Standards:** IEC 60601-1, IEC 60601-2-19 and IEC 60601-2-21,

**EMC Standards:** IEC 60601-1-2 2nd ed.



#### **WARNING:**

Medical Electrical Equipment needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided below.



#### **WARNING:**

Portable and mobile RF communication equipment can affect Medical Electrical Equipment. Caution should be used when operating such device around Medical Electrical Equipment



#### **WARNING:**

This equipment/system is intended for use by healthcare professionals only. This equipment/system may cause radio interference or may disrupt the operation of nearby equipment. It may be necessary to take mitigation measures, such as re-orienting or relocating the equipment/system or shielding the location.

### Electromagnetic Compatibility Essential Performance

The following device performance features were considered to be essential performance.

- Temperature sensing
- Elevating base movement
- Display operation
- In-bed scale
- Pulse Oximetry
- Pulse Rate
- Oxygen Saturation reading

## **Guidance and Manufacturer's declaration - Electromagnetic emissions**

The Giraffe warmer is intended for use in the electronic environment specified below. The user of the Giraffe warmer should ensure that it is used in such an environment.

| <b>Electromagnetic Emissions</b>                            |                   |                                                                                                                                                                                                                   |
|-------------------------------------------------------------|-------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Emissions test</b>                                       | <b>Compliance</b> | <b>Electromagnetic environment guidance</b>                                                                                                                                                                       |
| RF emissions<br>CISPR 11                                    | Group 1           | The Giraffe Warmer uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.                           |
| RF emissions<br>CISPR 11                                    | Class A           | The Giraffe Warmer is suitable for use in all establishments, other than domestic and those directly connected to the public low voltage power supply network that supplies buildings used for domestic purposes. |
| Harmonic emissions<br>IEC 61000-3-2                         | Class A           |                                                                                                                                                                                                                   |
| Voltage fluctuations/<br>flicker emissions<br>IEC 61000-3-3 | Complies          |                                                                                                                                                                                                                   |

## Manufacturer's guidance and declaration regarding electromagnetic immunity

The Giraffe warmer is intended for use in the electronic environment specified below. The user of the Giraffe warmer should ensure that it is used in such an environment.

| Electromagnetic immunity                                                                            |                                                                                                                                                               |                                                                                                                                                                |                                                                                                                                                                                                                                                                                              |
|-----------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Immunity Test                                                                                       | IEC 60601 Test Level                                                                                                                                          | Compliance level                                                                                                                                               | Electromagnetic environment guidance                                                                                                                                                                                                                                                         |
| Electrostatic discharge (ESD)<br>IEC 61000-4-2                                                      | +/- 6kV contact<br>+/- 8kV air                                                                                                                                | +/- 6kV contact<br>+/- 8kV air                                                                                                                                 | Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material the relative humidity should be at least 30%.                                                                                                                                                 |
| Electrical fast transient/burst<br>IEC 61000-4-4                                                    | +/- 2kV for power supply line.<br>+/- 1kV for input/output line.                                                                                              | +/- 2kV for power supply line.<br>+/- 1kV for input/output line.                                                                                               | Mains power quality should be that of a typical commercial or hospital environment.                                                                                                                                                                                                          |
| Surge<br>IEC 61000-4-11                                                                             | +/- 1kV differential Mode.<br>+/- 2kV common mode line.                                                                                                       | +/- 1kV differential Mode.<br>+/- 2kV common mode line.                                                                                                        | Mains power quality should be that of a typical commercial or hospital environment.                                                                                                                                                                                                          |
| Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11 | <5 % Ut(>95 % dip in Ut) for 0.5 cycle<br>40 % Ut (60 % dip in Ut) for 5 cycles<br>70 % Ut(30 % dip in Ut) for 25 cycles<br><5% Ut(>95% dip in Ut) for 5 sec. | <5 % Ut(>95 % dip in Ut) for 0.5 cycle<br>40 % Ut(60 % dip in Ut) for 5 cycles<br>70 % Ut (30 % dip in Ut) for 25 cycles<br><5% Ut (>95% dip in Ut) for 5 sec. | Mains power quality should be that of a typical commercial or hospital environment. If the user of the Giraffe warmer requires continued operation during power mains interruptions, it is recommended that the Giraffe warmer be powered from an uninterruptible power supply or a battery. |
| Power frequency (50/60 Hz) magnetic field environment<br>IEC 61000-4-8                              | 3 A/m                                                                                                                                                         | 3 A/m                                                                                                                                                          | Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital.                                                                                                                                                                |



**NOTE:** Ut is the main voltage before application of the test level.

## International Electronic Commission (IEC) guidance and manufacturer's declaration regarding electronic immunity

The Giraffe warmer is intended for use in the electronic environment specified below. The user of the Giraffe warmer should ensure that it is used in such an environment.

| Electromagnetic immunity           |                                                           |                                                      |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|------------------------------------|-----------------------------------------------------------|------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Immunity Test                      | IEC 60601 Test Level                                      | Compliance level                                     | Electromagnetic environment guidance                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| Conducted RF<br>IEC 61000-4-6      | 3 Vrms<br>150 KHz to<br>80 MHz                            | 3V                                                   | Portable and mobile RF communications equipment should be used no closer to any part of the Giraffe warmer, including cables, than the recommended separation distance calculated from the equation applicable for the frequency of the transmitter.                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| Radiated RF<br>IEC 61000-4-3       | 3 V/m<br>80 MHz to<br>2.5 GHz                             | 3 V/m                                                | <b>Recommended separation distance</b><br><br>$d = 1.2\sqrt{P}$<br><br>Radiated RF can affect the accuracy of in-bed-scale readings. However, the in-bed-scale is not critical to the performance of the Giraffe warmer unit (see Note 1 on page B-4).                                                                                                                                                                                                                                                                                                                                                                                                                                         |
| IEC 60601-2-19<br>& IEC 60601-2-21 | 3 V/m<br>26 MHz to 1 GHz<br><br>10 V/m<br>26 MHz to 1 GHz | 3 V/m<br>normal operation<br><br>10 V/m<br>no hazard | $d = 1.2\sqrt{P} = 26 \text{ MHz to } 800 \text{ MHz}$<br><br>$d = 2.3\sqrt{P} = 800 \text{ MHz to } 2.5 \text{ GHz}$<br><br>where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m).<br><br>Field strengths for fixed RF transmitters as determined by an electromagnetic site survey (see Note 3a) should be less than the compliance level in each frequency range (see Note 3b).<br><br>Interference may occur in the vicinity of equipment marked with the following symbol  |



---

**NOTE 1:** Portable and mobile equipment can affect medical electronic equipment.

---



---

**NOTE 2:** At 80 MHz and 800 MHz, the higher frequencies applies.

---



---

**NOTE 3:** These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.

- a -** Field strengths from fixed transmitters such as base stations for radio, cellular/cordless telephones and land mobile radios, amateur radios, AM and FM radio broadcasts and TV broadcasts cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Giraffe warmer unit is used exceeds the applicable RF compliance level above, the unit should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the Giraffe warmer.

- b -** Over the frequency range 150 KHz to 80 MHz field strengths should be less than 3 V/m.
-

## **Recommended separation distance between portable and mobile RF communications equipment and the Giraffe warmer**

The Giraffe warmer is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Giraffe warmer can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Giraffe warmer as recommended below, according to the maximum output power of the communications equipment.

| Rated maximum output power of transmitter (W) | Separation distance in meters (m) according to frequency of transmitter |                                        |                                         |
|-----------------------------------------------|-------------------------------------------------------------------------|----------------------------------------|-----------------------------------------|
|                                               | 150 kHz to 80 MHz<br>$d = 1.2\sqrt{P}$                                  | 80 MHz to 800 MHz<br>$d = 1.2\sqrt{P}$ | 800 MHz to 2.5 GHz<br>$d = 2.3\sqrt{P}$ |
| 0.01                                          | 0.12                                                                    | 0.12                                   | 0.23                                    |
| 0.1                                           | 0.38                                                                    | 0.38                                   | 0.73                                    |
| 1.0                                           | 1.2                                                                     | 1.2                                    | 2.3                                     |
| 10.0                                          | 3.8                                                                     | 3.8                                    | 7.3                                     |
| 100.0                                         | 12.0                                                                    | 12.0                                   | 23.0                                    |

For transmitters rated at a maximum output power not listed above, the recommended separation distance ( $d$ ) in meters (m) can be determined using the equation applicable to the frequency of the transmitter, where  $P$  is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.



**NOTE:** At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.



**NOTE:** These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

## Appendix D

# Conforming with Standards and Directives

---



Systems with this mark agree with the European Council Directive (93/42/EEC) for Medical Devices when they are used as specified in their Operation and Maintenance Manuals. The XXXX is the certification number of the Notified Body.



This symbol indicates that the waste of electrical and electronic equipment must not be disposed as an unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of your equipment.



# Appendix E

## Additional Safety Information

---

### Statements



#### **WARNING:**

To avoid the risk of electrical shock, this equipment must only be connected to a supply mains with protective earth.

The warmer does not adjust for patient temperature in warmup mode. When placing patient in the unit, immediately change modes to manual mode or baby mode.

No additional tasks are required to power down the device after clinical use, other than turning off the unit at the standby power switch on the front of the unit.

Isolation of the unit from the supply mains can only be achieved by turning off the device using the mains power switch on the rear of the unit. The standby switch on the probe panel will not isolate the device from the supply mains.



#### **WARNING:**

Electrical shock hazard: Before servicing, always unplug the unit from wall power.

### User Responsibility

This Product will perform in conformity with the description thereof contained in this manual and accompanying labels and/or inserts, when assembled, operated, maintained and repaired in accordance with the instructions provided. This Product must be checked periodically. A defective Product should not be used. Parts that are broken, missing, plainly worn, distorted or contaminated should be replaced immediately. Should such repair or replacement become necessary, GE Healthcare recommends that a telephone or written request for service advice be made to the nearest GE Healthcare Regional Service Center. This Product or any of its parts should not be repaired other than in accordance with written instructions provided by GE Healthcare and by GE Healthcare trained personnel. The Product must not be altered without GE Healthcare's prior written approval. The user of this Product shall have the sole responsibility for any malfunction that results from improper use, faulty maintenance, improper repair, damage or alteration by anyone other than GE Healthcare.

## Patient Temperature Control Mode

In baby mode, the system performs the following adjustment every 30 seconds. The Baby Condition column indicates the difference between the baby skin temperature and the baby mode set temperature.

| Baby Condition           | Heater Output Adjustment |
|--------------------------|--------------------------|
| Warm by 0.15°C or more   | Decrease by 5%           |
| Warm by 0.06°C to 0.14°C | Decrease by 1%           |
| Warm by 0.05°C or less   | No adjustment            |
| Cold by 0.05°C or less   | No adjustment            |
| Cold by 0.06°C to 0.14°C | Increase by 1%           |
| Cold by 0.15°C to 0.24°C | Increase by 5%           |
| Cold by 0.25°C to 0.34°C | Increase by 10%          |
| Cold by 0.35°C or more   | Increase by 15%          |

## Specifications

The product is designed to meet a life span of 8 years. However, with proper maintenance and repairs, the service life can be extended as long as service parts are available.

The maximum patient weight is 40 kg (88 lbs).

During the seven minute warmup period, the radiant heater is at 100% power (~27 mW/cm<sup>2</sup>). After the radiant heater has been running in warmup mode for seven minutes, the heater power setting drops to 50% (10 to 15 mW/cm<sup>2</sup>).





EC REP

#### **EC Representative**

GE Medical Systems SCS  
283 Rue de la Miniere  
78530 BUC - FRANCE

#### **World Headquarters**

GE Healthcare  
9900 Innovation Drive  
Wauwatosa, WI 53226 USA  
Tel 1 800 345 2700

#### **Europe, Middle East, Africa**

GE Healthcare  
P.O. Box 900  
FIN-00031 GE  
Finland  
Tel +358 10 39411  
Fax +358 9 146 3310

#### **Latin America Representatives**

GE Healthcare  
3350 SW 148 Avenue  
Suite 301  
Miramar, Florida, 33027 USA  
Tel + 1 954 744 5600

#### **Germany**

GE Medical Systems Information Technologies GmbH  
Munzinger Str. 3-5  
79111 Freiburg  
Tel. 49 761 4543 570  
Fax 49 761 4543 571  
Service 0800 4343258

#### **Asia Representative**

GE Healthcare  
Shanghai GE (China) Hi-tech Park  
No1 Huatuo Road, Zhangjiang Hi-tech Park Pudong, Shanghai,  
P.R.China 201203  
上海GE中国科技园  
地址：中国上海市浦东张江高科技园华佗路1号, 201203  
Tel + (8621) 38777888  
Fax + (8621) 38777402

#### **Brazil Only**

GE Healthcare do Brasil Comércio e Serviços para  
Equipamentos Medicos-Hospitalares Ltda.  
Av. Magalhães de Castro, 4800, Andar 12 cj 121 e Andar 13 cj  
131 e 132, Torre 3 - Jardim Panorama  
CEP: 05676-120 - São Paulo/SP - Brasil  
Tel +55 (11) 3004 2525 (Capitais e Regiões Metropolitanas)  
Tel +55 (11) 0800 165 799 (Demais Localidades)

|                    |                     |
|--------------------|---------------------|
| <b>China</b>       | <b>800 810 8188</b> |
| Australia          | 1300 722 229        |
| India              | 1 800 425 7255      |
| Korea              | (02) 1544 4564      |
| South Eastern Asia | (65) 6277 3444      |

 Datex-Ohmeda, Inc.  
9900 Innovation Drive  
Wauwatosa, WI 53226 United States

Panda iRes Warmer  
Operation and Maintenance Manual  
Printed in USA  
© 2012-2017 General Electric Company  
All rights reserved.

M1110737  
Revision D  
Language: English