ChronoFlow: an electricity-free, volumetric, clockwork-governed syringe infusion driver for neonatal care in low-resource settings

Author list to be completed[1]

[1]Affiliation to be completed

Design-and-simulation study — October 3, 2026

Abstract

Hospitals without reliable electricity deliver intravenous fluids and drugs to newborns by gravity drip or by hand. The only non-electric infusion devices available meter flow by pushing fluid through a restriction, so the delivered rate depends on fluid viscosity (temperature, drug) and on downstream pressure; low-cost electric pumps still need batteries and cost hundreds of dollars. We present ChronoFlow, a wind-up syringe driver that meters *volume* rather than pressure: a constant-force spring pushes the plunger and a rack-and-pinion drives a 3D-printed step-up gear train into a clockwork escapement whose balance wheel sets the rate. Two design problems were identified and solved by simulation. First, the mechanical power available to the escapement varies 24× across syringe sizes, clinical rates and loads (17–416 µW); a conventional lever escapement, whose impulse scales with train torque, over-banks over most of this range. Adapting the gravity-escapement principle of tower clocks with springs—a constant-force *spring-arm* escapement—delivers a fixed impulse energy per beat and holds the balance amplitude constant over the full range with a period error below 0.06 %. Second, a cross-spring flexure pivot was shown by a beam-column finite-element model to change stiffness by 17 % under the balance's own weight; it was replaced by a taut-band torsion suspension made from 0.0015 in (38.1 µm) feeler-gauge steel ribbon, giving a 1.20 Hz balance with ±0.19 % orientation sensitivity and no pivot friction. Colour-coded gear cassettes provide 1, 2, 5 and 10 mL/h on a 20 mL syringe (0.4–19.5 mL/h across 5–60 mL syringes) at 0.12–1.16 µL per beat. A 20 000-sample Monte-Carlo error budget predicts that 95 % of units deliver within ±1.14 % of the set rate, compared with +45 % at 37 °C and −33 % at 100 mmHg back-pressure for flow-restriction pumps. A trip spring whose preload is keyed to the syringe size stops the oscillator and raises a flag at 300 mmHg occlusion (280–320 to 202–398 mmHg across syringe sizes once plunger-friction uncertainty is included) and a syringe-latch interlock prevents free-flow boluses. The complete 66-part assembly was modelled parametrically, checked for interferences and exported for printing; the bill of materials is $40. No physical prototype has yet been built; a bench protocol following IEC 60601-2-24 [13] is provided.

1 Introduction

Neonates require maintenance fluids at 60–100 mL/(kg d) and drugs such as antibiotics, dextrose and inotropes at 0.5–10 mL/h; a 10 % rate error or an uncontrolled bolus of a few millilitres is clinically significant. Electronic syringe pumps meet this need where mains power, batteries and service infrastructure exist. Where they do not, three alternatives have been described. Low-cost electromechanical pumps such as AutoSyP combine a constant-force spring with a motor and battery and reach ±4 % accuracy at a cost of goods of about $500 [1]. Non-electric flow-restriction devices (elastomeric pumps, the Springfusor, and an adjustable air-over-oil design [2]) obey $Q \propto \Delta P/\mu$ and therefore drift with fluid viscosity, temperature and back-pressure. Clockwork metering has been reported four times: a gravity-driven neonatal driver regulated by a pendulum metronome, which must sit level and still [3]; a wind-up kitchen-timer movement adapted for microfluidics [4] and its fully 3D-printed successor with a hairspring balance [5, 6], both with fixed rates set by gear

choice and about an hour of running time; and a spring-driven leadscrew pump whose escapement is actuated electrically for insulin dosing [7].

The gap is a non-electric device that meters *volume*, works in any orientation, offers selectable clinical rates, runs for a full syringe without re-winding and has mechanical occlusion detection and free-flow protection. ChronoFlow addresses it by returning to two solutions from precision mechanics: the constant-force escapement, introduced by Denison for tower clocks whose train torque varied with wind and ice on the hands [8], and the taut-band suspension of galvanometers, which removes pivot friction and most positional error.

2 Design

2.1 Architecture and metering principle

Figure 1a–b shows the machine. Two constant-force springs (44.5 N total, 2×5 lbf) pull a carriage toward the syringe nozzle. The plunger load—syringe friction of 1.5–4.5 N plus pressure times bore area, up to 26.9 N for a 60 mL syringe at the 300 mmHg trip—absorbs part of this force; the excess back-drives a module-1 rack and 12-tooth pinion through a $5 \times 5 \times 5 \times 2.5 = 312.5$ step-up train and a plug-in rate cassette to a 15-tooth escape wheel. The escapement releases one half-tooth per beat of the balance, so the plunger advance per beat is fixed by geometry:

$$\Delta x = \frac{\pi m z_0}{2 Z_e \, N_{\text{fixed}} \, r_{\text{c}}} = \frac{\pi \cdot 1 \cdot 12}{2 \cdot 15 \cdot 312.5 \cdot r_{\text{c}}} \text{ mm}, \tag{1}$$

which is 4.0 µm (1.16 µL on a BD 20 mL syringe) for the unity cassette. Because the metered quantity is displacement, viscosity, temperature and downstream pressure do not enter the rate until the occlusion trip point (Fig. 1c–d). The spring force is constant over the 95 mm travel, so runtime is set only by syringe volume (2–20 h per 20 mL fill) and the device never needs re-winding mid-infusion. A one-way click on the rack-pinion shaft lets the carriage be pulled back freely, so loading a syringe *is* the winding action (4.2 J stored, about eleven times the 0.37 J the escapement dissipates in 20 h).

2.2 Constant-force spring-arm escapement

Power delivered to the escapement is $P = F_{\text{excess}} \, v$. With F_{excess} between 17.6 N (60 mL syringe at its 300 mmHg occlusion trip) and 43.0 N (free-running 5 mL syringe) and v spanning the 1–10 mL/h cassettes, P ranges from 17 µW to 416 µW; after a train efficiency of $0.95 \times 0.92^6 = 0.58$ (seven meshes in total: the rack mesh at 0.95 and six printed gear meshes at 0.92—four step-up, two cassette; the escape wheel shares the cassette output shaft) the escape-wheel torque spans 19–476 µN m, a 24× range. In a lever escapement the impulse energy delivered to the balance is proportional to this torque, so the balance amplitude scales as $\sqrt{T_e}$; our dynamic model (Section 3) shows it exceeding the 1.2 rad banking limit over most of the operating range (Fig. 2b–c). Stalling at the low end or over-banking at the high end are both failures of rate.

The spring-arm escapement (Fig. 2a) separates the two energy flows. Each escape-wheel tooth *cocks* one of two impulse arms against a printed flexure until a latch engages (phases 1–2). The train is then locked and any excess torque simply sits on the lock. When the balance passes, its impulse pin trips the latch (phase 3) and the arm delivers a fixed energy $E_{\text{imp}} = 1.2$ µJ over a 15° arc; the wheel then advances one half-pitch, re-cocks the opposite arm and relocks, the surplus train energy being dissipated in the locking impact. Train torque only has to exceed the re-cock minimum $T_{\min} = E_{\text{imp}}/(\eta_{\text{lift}} \, \pi/Z_e) = 9.5$ µN m. The simulation confirms constant amplitude (0.59 rad at $Q = 150$) and a period error of 0.06 % from the slightly asymmetric impulse window across the whole torque range (Fig. 2c–d). Because the balance's air damping cannot be known before a prototype exists, Q was swept from 40 to 800: amplitude settles between 0.31 and 0.99 rad and the period error between 0.45 and 0.01 % (Fig. 2e), a fixed offset that trim calibration absorbs.




Figure 1: A spring-driven, clockwork-governed syringe driver meters volume, so the delivered rate is independent of fluid temperature and back-pressure. (a) Labelled render of the parametric CAD assembly with a BD 20 mL syringe at 85 % fill and the 10 mL/h cassette. (b) Energy and metering chain; the left arrow follows energy from the spring to the balance, the right arrow follows the rate constraint back from the balance to the plunger. (c) Steady-state flow error versus fluid temperature for a flow-restriction pump calibrated at 20 °C (water viscosity from the Vogel equation) and for ChronoFlow (oscillator temperature coefficient $-1.8 \times 10^{-4}\,\mathrm{K}^{-1}$). (d) Flow error versus downstream back-pressure; ChronoFlow stops and flags at the 300 mmHg trip point. Green band: ±5 % clinical tolerance.

2.3 Taut-band balance

The oscillator (Fig. 3a) is a printed PETG ring (62 mm outer diameter, 7.4 g, $I = 5.2 \times 10^{-6}\,\mathrm{kg\,m^2}$) clamped at the midpoint of a 2 mm-wide spring-steel ribbon cut from a 0.0015 in ($t = 38.1\,\mu\mathrm{m}$) feeler-gauge blade, held at both ends and pre-tensioned to $T_0 = 1\,\mathrm{N}$ by a steel coil spring. Each 22 mm half-band is a torsion spring [9] of stiffness

$$k_{\mathrm{half}} = \frac{GJ}{L_h} + \frac{T\,w^2}{12\,L_h}, \qquad J = \tfrac{1}{3} w t^3 \left(1 - 0.63\,\tfrac{t}{w}\right), \tag{2}$$

giving $k = 2.95 \times 10^{-4}\,\mathrm{N\,m/rad}$ and $f_0 = 1.20\,\mathrm{Hz}$. Because $k \propto t^3$, the ribbon thickness is the dominant uncertainty: feeler stock is held to about ±2.5 µm, which alone would move f_0 by ±9 %, far beyond any trim-weight range (an earlier revision modelled a 40 µm ribbon, 6 % off the true stock). Regulation is therefore two-stage. Both end clamps sit on slotted posts and are moved symmetrically so that each half-band length L_h can be set between 18 and 26.5 mm while the hub stays centred; with $k = k_{\mathrm{half}}(L_{\mathrm{up}}) + k_{\mathrm{half}}(L_{\mathrm{low}})$ evaluated with each half's actual length, symmetric adjustment to 18.4–26.1 mm re-centres f_0 for any ribbon between 35.6 and 40.6 µm (coarse, set once per ribbon). Adjusting one clamp alone would halve the sensitivity (1.1 % per mm against 2.2 %) and the same slot would then cover only 36.6–39.4 µm, so one-sided adjustment is not sufficient; four




Figure 2: A constant-force spring-arm escapement keeps balance amplitude and rate independent of the 24× variation in train torque that defeats a conventional lever escapement. (a) Working principle: a wheel tooth cocks a spring-loaded impulse arm (1), a latch holds it while the wheel is locked (2), the balance pin trips the latch and receives a fixed impulse (3). (b) Simulated balance angle $\theta(t)$ for both escapements at the minimum and maximum operating torques (19 and 476 µN m; RK4, $Q = 150$); dashed lines mark the banking limit. (c) Steady amplitude and (d) period error relative to the free oscillator versus escape-wheel torque; shaded band: operating range. Ticks at zero mark torques below the spring-arm re-cock minimum. (e) Spring-arm amplitude and period error versus the balance quality factor Q.

M3 brass nuts in radial slots (23–29 mm) then regulate the frequency by −2.4 to +3.1 % (Fig. 3b), exactly as a clock is regulated, with a nominal set point of 26.4 mm.

Gravity enters only through the tension term: with the band vertical, the fixed-clamp half-band carries $T_0 \pm mg$ while the tensioner fixes the other at T_0. Because the tension term is 10 % of k and $mg \ll T_0$, the frequency shifts by only ±0.19 % between the upright and inverted positions (Fig. 3c). An alternative two-leaf cross-spring flexure pivot [10], analysed with an Euler–Bernoulli beam-column finite-element model including geometric stiffness, changed rotational stiffness by 17 % under the same balance weight (the governing scale is $EI/L^2 \approx 0.1$ N, not the Euler load) and was rejected. Peak shear stress at the 1.2 rad banking limit is 166 MPa against a conservative fully-reversed fatigue limit of 350 MPa (Fig. 3d), and the temperature coefficient, combining the shear-modulus drift of steel with the thermal expansion of the ring, is $-1.8 \times 10^{-4}\,\mathrm{K}^{-1}$ (Fig. 3e).

2.4 Rate cassettes, safety functions and structure

A plug-in cassette carries two gear pairs with constant tooth sums ($z_1 + z_2 = z_3 + z_4 = 66$, 33 mm centres): (33, 33, 33, 33), (44, 22, 33, 33), (33, 33, 55, 11) and (44, 22, 55, 11) give speed-up ratios of 1, 2, 5 and 10 and hence 10, 5, 2 and 1 mL/h on a 20 mL BD syringe; other syringe sizes scale with bore area (Fig. 4a). The cassette is a visible, colour-coded object, so the set rate can be confirmed at a glance and cannot drift. Because the carriage speed is set by the cassette alone, the escape-wheel torque depends only on the cassette and on the excess force $F_{\mathrm{spring}} - F_{\mathrm{plunger}}$; the operating map was therefore evaluated over every syringe size, cassette and back-pressure from zero to the trip (80 cells). The torque exceeds the re-cock minimum of 9.5 µN m by at least 2.0× in every cell, the worst




Figure 3: A taut-band suspension gives a friction-free 1.2 Hz balance whose rate is insensitive to orientation, temperature and fatigue. (a) Construction: feeler-gauge ribbon clamped to the deck and to a spring-loaded tension clamp, balance ring clamped at mid-span. (b) Balance frequency versus radial position of the four trim nuts; the dashed line is the target for the unity cassette and the band is ±0.5 %. (c) Frequency change between the band-horizontal and band-vertical positions for the taut band, compared with the computed first-order change of a 0.05 mm cross-spring flexure pivot under the same balance weight. (d) Peak ribbon shear stress versus amplitude; shaded band: amplitudes reached for $Q = 40$–800. (e) Rate drift versus ambient temperature.

being the 60 mL syringe on the 1 mL/h cassette at its 300 mmHg trip (Fig. 4b). This map is what forced the drive spring from the 30 N of the first revision to 44.5 N: at 30 N the 60 mL trip left only 3.1 N excess, 3.4 µN m at the slowest cassette, and the escapement would have stalled before the occlusion flag was raised (two cells stalled, four below 1.7×). The per-beat volume is 0.12–1.16 µL (Fig. 4c).

Three safety functions are mechanical. (i) The occlusion trip (Fig. 5) is snap-acting. The pusher head rides on a stem backed by a 3 N/mm compression spring whose preload is set by a spacer keyed to each syringe cradle insert, so the setting follows the syringe size automatically. When the plunger reaction exceeds the preload the stem begins to retract; after a *sensing travel* of only 0.3 mm a notch in the stem reaches a spring-loaded sear, which drops into the notch and releases a separate flag rod. The flag rod, driven by its own 2 N spring, snaps up 12 mm, exposes the red flag and lifts a rail into the path of the balance rim, so ticking stops and pressure cannot rise further. Defining the trip at 0.3 mm rather than at a full 3 mm of stem travel matters: over 3 mm the spring force would rise by 9 N (+235 mmHg on a 20 mL syringe, and enough to take the 60 mL case below the re-cock minimum before the flag appeared); over 0.3 mm it rises by 0.9 N, and the spacer preload is set 0.9 N below the intended trip force so that the sear releases at exactly 300 mmHg × A + F_{friction}: trip forces 6.1, 8.6, 14.5, 18.3 and 26.9 N from spacer preloads of 5.2, 7.7, 13.6, 17.4 and 26.0 N (1.7–8.7 mm) for the 5, 10, 20, 30 and 60 mL syringes. The operating-map margins above are therefore unchanged by the trip. Because the sear captures the stem in its retracted position and the blocked escapement locks the train, the 0.3 mm of spring compression stored at the trip (0.03–0.17 mL of syringe volume, 60 mL largest) cannot advance the plunger when the blockage clears; the mechanism's own release bolus is zero by construction, and what remains is the elastic volume stored in the line and syringe at 300 mmHg, which the protocol measures against





Figure 4: Four plug-in cassettes cover neonatal rates with torque margin and sub-microlitre resolution; the predicted error is dominated by the syringe, not the mechanism. (a) Flow rate for each cassette and BD syringe size. (b) Escape-wheel torque relative to the arm re-cock minimum at the two extremes of the syringe × cassette × pressure map (60 mL syringe at its 300 mmHg trip; 5 mL syringe free-running); crosses show the first revision's 30 N spring, which stalled. (c) Volume delivered per beat. (d) Standard deviation of flow error contributed by each Monte-Carlo input. (e) Distribution of steady-state flow error over 20 000 draws; dashed lines mark the central 95 %.

the 0.2 mL target. Reset is only through the LOAD/RUN lever, which re-cocks the flag rod and lifts the sear. A force trip senses friction as well as pressure, so the pressure at which it fires carries the friction uncertainty: for ±1.5 N of plunger glide-force variation (lot, speed, dwell) the 300 mmHg setting fires between 280 and 320 mmHg on the 60 mL syringe but between 202 and 398 mmHg on the 5 mL syringe, whose bore is five times smaller. These bands are the disclosed occlusion-alarm range; tightening them for the small syringes requires measuring glide force per syringe lot and shimming the spacer, which the validation protocol includes. (ii) The LOAD/RUN lever opens the syringe flange latch and lifts the rack-pinion click together, so the carriage can only be moved by hand when the syringe is unlatched and a fast-forwarded pusher slides the whole syringe rather than expelling fluid. (iii) The balance beats behind a window at 1.2 Hz; silence means stopped.

The most heavily loaded teeth are the rack and 12-tooth pinion: Lewis bending stress [11] at 44.5 N on an 8 mm face is 22.7 MPa against about 50 MPa PETG yield (safety factor 2.2; widening the stage-1 face to 10 mm would restore 2.75). Torque falls 5× per stage (267, 49, 9.0, 1.7 mN m), so the remaining printed gears are lightly loaded; shafts are 3 mm steel pins in printed bores.

3 Methods

Parametric design. All dimensions, tooth counts and oscillator parameters live in one module (`design_params.py`); the CAD (`cad_model.py`, CadQuery 2.8) derives shaft positions from gear centre distances, phases every mesh so that a wheel tooth faces a pinion gap, and exports STEP, glTF and per-part STL. A pairwise interference check (bounding-box prefilter, OpenCascade Boolean common) over all 2 145 part pairs reports no unintended interferences; the single residual is a 0.55 mm^3 tip/root contact on the 60T/12T mesh, the known undercut of a 12-tooth 20° pinion, for which a +0.3 profile shift is recommended.



syringe	preload (N)	trip (N)	fires at (mmHg)	stored at sear (mL)
5 mL	5.2	6.1	202–398	0.03
10 mL	7.7	8.6	232–368	0.05
20 mL	13.6	14.5	261–339	0.09
30 mL	17.4	18.3	270–330	0.11
60 mL	26.0	26.9	280–320	0.17

Figure 5: Snap-action occlusion trip: the trip force is defined by a 0.3 mm sensing travel, and the stem is captured so stored spring compression cannot become a bolus. (a) Armed, sensing and tripped states: spacer-set preload, 0.3 mm stem retraction to the sear, sear capture of the stem, 12 mm flag-rod release and balance-blocking rail. (b) Trip-spring force versus stem travel for three syringe sizes; the sear releases at 0.3 mm (+0.9 N, pre-compensated in the spacer); the grey band is the 3 mm direct-travel design of the previous revision. (c) Per-syringe spacer preload, trip force, firing band for ±1.5 N plunger-friction uncertainty, and the displacement stored at the sear if the stem were not captured.

Escapement dynamics. The balance obeys $I\ddot{\theta} = -k\theta - c\dot{\theta} + \tau(\theta, \dot{\theta})$ with $c = I\omega_0/Q$, integrated by fourth-order Runge–Kutta at $\Delta t = 0.2$ ms. For the lever model the per-beat wheel energy $T_e\pi/Z_e$ is split into an impulse fraction (0.70) applied over a 30° lift arc and an unlocking fraction (0.12) opposing motion over 6°. For the spring-arm model the impulse energy is fixed at E_{imp} over 15° starting 3° before equilibrium, with a fixed 0.12 µJ latch-tripping cost; the escapement stalls if $T_e < T_{\min}$. A reversal of the balance while the pin is inside the engagement window is treated as a fault. Period is measured over steady-state zero crossings.

Oscillator. Band stiffness follows Timoshenko's thin-strip torsion with the tension term; the cross-spring pivot comparison uses a 60-element Euler–Bernoulli beam-column model per leaf (elastic plus geometric stiffness), with the balance rotating rigidly about the crossing point and gravity resolved onto the two leaves.

Error budget. 20 000 Monte-Carlo draws combine uniform syringe bore tolerance (±1 % area, ISO 7886-1 class [14]), trim regulation residual (±0.3 %), rack pitch after print-scale calibration (±0.2 %), ambient temperature (15–35 °C), orientation (±0.19 %) and the escapement error for Q uniformly in 60–400 after calibration at the mean.

Table 1: Predicted performance of ChronoFlow (BD 20 mL syringe unless stated) and bill of materials.

Quantity	Value
Rates (cassettes 1/2/5/10)	1, 2, 5, 10 mL/h; 0.4–19.5 mL/h across 5–60 mL syringes
Resolution	0.12–1.16 µL per beat, 2.4 beats s^{-1}
Oscillator	1.2017 Hz; coarse L_h 18–26.5 mm, trim −2.4/+3.1 %; orientation ±0.19 %; -1.8×10^{-4} K^{-1}
Escapement	amplitude 0.59 rad ($Q = 150$), period error 0.06 %; torque margin ≥ 2.0× over 5 syringes × 4 cassettes × 0–300 mmHg
Accuracy (Monte-Carlo, $n =$ 20 000)	95 % of units within −1.14 to +1.15 %; 99th percentile |error| 1.36 %
Restriction-pump comparator	+45 % at 37 °C; −23 % at 10 °C; −33 % at 100 mmHg back-pressure
Occlusion trip	300 mmHg nominal; snap-action sear at 0.3 mm sensing travel, spacer preload keyed to cradle insert (5.2–26.0 N, trip 6.1–26.9 N); fires at 280–320 (60 mL) to 202–398 mmHg (5 mL) for ±1.5 N friction; stem captured, mechanical release bolus 0
Energy	4.2 J stored per wind; 0.37 J consumed in 20 h
Structure	rack-pinion tooth stress 22.7 MPa (SF 2.2); band stress 166 MPa vs 350 MPa limit
Parts, cost	66 parts, ≈420 g PETG, BOM ≈ $40

Comparators. Flow-restriction pumps are modelled as $Q \propto (P_{\mathrm{drive}} - P_{\mathrm{back}})/\mu(T)$ with water viscosity from the Vogel equation [12] ($\mu = 0.02939 \exp[507.88/(T - 149.3)]$ mPa s) and drive pressures of 300 and 500 mmHg.

4 Results

Table 1 collects the predicted performance. Two results carry the design. First, the escapement study (Fig. 2) shows that the conventional lever escapement is unusable across the clinical range—its amplitude passes the banking limit at about 40 µN m, below the mid-point of the operating band—whereas the spring-arm escapement holds 0.59 rad from the re-cock minimum upward with a period error of 0.06 %. Second, the suspension study (Fig. 3) shows that the taut band reduces orientation sensitivity by a factor of about forty relative to a cross-spring pivot of the same leaf stock, while keeping the ribbon stress at 47 % of the fatigue limit at the banking amplitude. The system-level consequences (Fig. 4) are a rate set entirely by integers (tooth counts) and one calibrated frequency, a torque margin above the re-cock threshold at every rate, and a predicted accuracy distribution whose width is set mainly by the syringe bore tolerance that every syringe pump shares.

Against flow-restriction pumps the difference is categorical rather than incremental (Fig. 1c–d): a restriction pump calibrated at 20 °C over-delivers by 45 % at body temperature and loses a third of its rate against 100 mmHg of venous or catheter back-pressure, whereas ChronoFlow drifts by less than 0.4 % over 10–40 °C and is unaffected by back-pressure until it trips.

The CAD model (Fig. 6) is complete at the part level: deck and bottom plates, pillars, rails, cradle with flange latch and barrel clamp, carriage with integrated rack and pusher assembly, spring drum, the four-stage train, cassette, escape wheel, two impulse arms with flexures and latch, and the balance with band, clamps and tensioner.

5 Discussion

ChronoFlow converts two precision-mechanics solutions—the constant-force escapement and the taut-band suspension—into a printable, electricity-free infusion driver whose predicted accuracy is limited by the syringe rather than by the fluid, the temperature, the back-pressure or the orientation of the device. Every safety function is mechanical and visible, and the rate is a physical object a




Figure 6: The complete 66-part assembly is fully modelled, interference-checked and printable. (a) Exploded view with parts separated along the vertical axis. (b) Escapement module with the deck removed. (c) Top view of the upper deck and step-up train. (d) Front elevation showing the movement below the deck.

nurse can hold and verify.

Several limitations follow from the fact that this is a design-and-simulation study. All performance figures are model predictions; the models are physics-based but unvalidated until the bench protocol below is executed. The balance quality factor in air and the efficiency of printed gear meshes are the least certain inputs; both were swept and both affect margin rather than rate. The spring-arm escapement geometry in the CAD is a representative embodiment whose latch and arm proportions will need iteration on the bench, as horological escapements invariably do. The occlusion trip must be set per syringe size, which is a training requirement, and rates between cassettes require an additional printed cassette (any tooth set with sums of 66 can be made). The 0.12–1.16 µL pulsatility is small compared with a 50 µL gravity-drip drop, but very low-rate inotrope infusions below 0.5 mL/h would be better served by a 5 mL syringe.

Validation protocol. (1) Measure the ribbon thickness with a micrometer (±1 µm) and set the clamp post to the L_h the stiffness model gives for that thickness; print the balance, clamp the ribbon and measure f_0 with a 240 fps camera or light gate; set trims; measure Q from the free-decay envelope; repeat band-horizontal, clamp-up and clamp-down (acceptance ±0.3 %). (2) Drive the escape wheel through the cassette with hanging weights emulating 17.6 and 43 N excess (the 60 mL-at-trip and 5 mL-free-running corners of the operating map); verify re-cocking at the 1 mL/h cassette and constant amplitude by video tracking; tune arm flexure thickness. (3) Gravimetric flow test per IEC 60601-2-24 on a 1 mg balance logged every 30 s for 2 h per cassette; compute

trumpet curves (acceptance: mean error ±5 %, trumpet ±10 % at 2 min). (4) Repeat at 0, 100 and 200 mmHg back-pressure and with 10 % dextrose at 20 and 37 °C (expected change < 1 %). (5) Occlusion, for each syringe size with its cradle insert: measure plunger glide force on a load cell at the cassette speed (five syringes per lot), then clamp the line and record trip pressure against the predicted band (280–320 to 202–398 mmHg for 60 mL to 5 mL; shim the spacer if the lot's glide force differs from nominal by more than 0.5 N), time to trip, the stem position after trip (captured, no creep), and bolus on release (< 0.2 mL; this is the line-compliance bolus, since the captured stem contributes none). (6) Orientation and vibration (expected < 0.5 %). (7) Seven-day continuous run and 200 load/unload cycles.

Data and code availability

The parametric CAD (STEP, glTF, STL), the analysis code that reproduces every figure and number in this paper, the interference checker and the interactive viewer are provided in the accompanying repository.

References

[1] Juarez A, Maynard K, Skerrett E, Molyneux E, Richards-Kortum R, Dube Q, Oden ZM. AutoSyP: a low-cost, low-power syringe pump for use in low-resource settings. *Am J Trop Med Hyg* 2016;95(4):964–969. doi:10.4269/ajtmh.16-0285.

[2] Guelig D, Bauer J, Wollen A, Schiller C, Sherman-Konkle J, Roche A, Weyrauch K, Eisenstein M. Design of a novel, adjustable flow rate, reusable, electricity-free, low-cost syringe infusion pump. *ASME J Med Devices* 2017;11(4):041006. doi:10.1115/1.4037935.

[3] Sung C, Kamath RR, Cui Y, Ouyang C, Carstens E, Ramos RF, Oden ZM. Design of a novel mechanical syringe pump for neonatal care in low-resource settings. In: *2011 IEEE Global Humanitarian Technology Conference*; 2011. p. 78–83. doi:10.1109/GHTC.2011.21.

[4] Han W, Kim S, Shin S, Yang SY, Choi S, Shin JH. Wind-up precision pump for portable microfluidics. *Sens Actuators B Chem* 2021;347:130592. doi:10.1016/j.snb.2021.130592.

[5] Park SB, Shin JH. Fully 3D-printed, nonelectric, spring-powered syringe pump for operating microfluidic devices. *Sens Actuators B Chem* 2024;405:135289. doi:10.1016/j.snb.2024.135289.

[6] Park SB, Shin JH. Open-source spring-driven syringe pump with 3D-printed components for microfluidic applications. *HardwareX* 2024;19:e00550. doi:10.1016/j.ohx.2024.e00550.

[7] Pooke F, Payne M, Holder-Pearson L, Heaton D, Campbell J, Chase JG. Low-cost, low-power, clockwork syringe pump. *HardwareX* 2023;16:e00469. doi:10.1016/j.ohx.2023.e00469.

[8] Denison EB. *A Rudimentary Treatise on Clocks, Watches and Bells*. London: John Weale; 1850 (gravity escapement, ch. on remontoires and gravity escapements).

[9] Timoshenko SP, Goodier JN. *Theory of Elasticity*, 3rd ed. New York: McGraw-Hill; 1970. § 109, torsion of a thin rectangular strip.

[10] Haringx JA. The cross-spring pivot as a constructional element. *Appl Sci Res* 1949;A1:313–332. doi:10.1007/BF02120338.

[11] Lewis W. Investigation of the strength of gear teeth. *Proc Engineers' Club of Philadelphia* 1893;10:16–23.

[12] Vogel H. Das Temperaturabhängigkeitsgesetz der Viskosität von Flüssigkeiten. *Phys Z* 1921;22:645–646.

[13] IEC 60601-2-24:2012. Medical electrical equipment — Part 2-24: Particular requirements for the basic safety and essential performance of infusion pumps and controllers. Geneva: International Electrotechnical Commission; 2012.

[14] ISO 7886-1:2017. Sterile hypodermic syringes for single use — Part 1: Syringes for manual use. Geneva: International Organization for Standardization; 2017.