SH 21 BANGLADESH



Additional Information

Download the LIBERO software from www.elpro.com/downloads

- Use liberoCONFIG to configure LIBERO with your own settings
- Use elproVIEWER to access all recorded data and create own reports

LIBERO Configuration

Type: LIBERO CB V4.32 Current State: Stopped

 Device ID:
 7208 0189 832
 Start at:
 29.Jul.2021 14:07:03

 Log Interval:
 5 m
 First Measurement:
 29.Jul.2021 14:07:03

 Report Time Base:
 UTC
 Checksum:
 C001 / 757.579.706

Configured by: C20122, GVA11D0090057/aitikhlefk, 29.Jul.2021 08:54:28

Alarm Zone	Allowed Time	Total Time	No. of Excursions	Longest Exc.	Status
H1: over 40.0 °C	5 m (sin)	0 m	0 / unlimited	0 m	OK
G: 2.0 to 40.0 °C	unlimited	48 d 13 h 55 m			
L1: below 2.0 °C	5 m (sin)	0 m	0 / unlimited	0 m	OK

MKT Limits and Maximum Duration

Alarm Conditions

Duration: 48 d 13 h 55 m none

MKT: 25.8 °C

Logging Results

Highest Temperature: 39.4 °C; 03.Sep.2021 10:42:03 Stop at: 16.Sep.2021 03:59:21

Lowest Temperature: 17.9 °C; 02.Sep.2021 08:47:03 Alarm at: none

Average Temperature: 24.6 °C File created: 16.Sep.2021 03:59:22

