

PRECISION NDE SERVICES INC.

Unit 12, 485 Speers Road, Oakville, ON L6K 3S4
Tel: (905) 555-0200 | calibration@precisionnde.com

CALIBRATION CERTIFICATE

Certificate Number: PNDE-CAL-2024-0102
Date of Calibration: January 10, 2024
Calibration Due Date: January 10, 2025
Calibration Procedure: PNDE-CAL-UT-001 Rev. 3

EQUIPMENT INFORMATION

Description:	4-Step Calibration Block
Manufacturer:	Precision NDE Services
Model:	CS-4STEP
Serial Number:	PNDE-CS-001
Range:	0.100" to 0.400"
Resolution:	N/A
Owner:	Precision NDE Services Inc.

REFERENCE STANDARDS

Standard	Description	Certificate	Due Date
PNDE-CS-001	4-Step Carbon Steel Block	NIST-2024-4521	Jan 2025
PNDE-SS-003	2-Step 316SS Block	NIST-2024-4522	Jan 2025

CALIBRATION RESULTS

Nominal (in)	Measured (in)	Error (in)	Tolerance	Status
0.100	0.1001	+0.0001	±0.002	PASS
0.200	0.2002	+0.0002	±0.002	PASS
0.300	0.2999	-0.0001	±0.002	PASS
0.400	0.4001	+0.0001	±0.002	PASS

Environmental Conditions: Temperature 68°F ± 5°F, Humidity 45% RH

CERTIFICATION STATEMENT: This equipment has been calibrated using standards traceable to the National Institute of Standards and Technology (NIST). The calibration was performed in accordance with Precision NDE Services Inc. procedures and meets the requirements of ASME Section V, Article 5.

Calibrated By:

Approved By:

External Lab (NIST Traceable)

David Park, UT Level III

Calibration Technician

Quality Manager