Larkhall NDE Services Inc.

Calibration Certificate

Site: Hamilton Plant

No.	Status	Date
LNDE-CAL-2016-0042	Issued	2016-01-25

Calibration Certificate - Olympus 37DL PLUS 0611452

Instrument Identification

Document Number: LNDE-CAL-2016-0042

Facility: Hamilton Plant

Instrument: Olympus 37DL PLUS

Serial Number: 37DL-0611452

Calibration Standards

Reference Block 1: LNDE-CS-001 4-step carbon steel calibration block, 0.100-0.400 in

Reference Block 2: LNDE-SS-003 calibration block

Traceability: Larkhall NDE Services Inc. Master Standards

Calibration Schedule

Calibration Date: 2016-01-25

Calibration Interval: 12 months

Valid Until: 2017-01-25

Sign-off	Name	Function	Date	Signature
Prepared	David Park	Level III / Quality Manager, SNT-TC-1A UT Level III	2016-01-25	_____