Larkhall NDE Services Inc.

Calibration Certificate

Site: Hamilton Plant

No.	Status	Date
LNDE-CAL-2017-0702	Issued	2017-07-10

Calibration Certificate - Olympus 38DL PLUS 2847193

Instrument Identification

Instrument: Olympus 38DL PLUS

Serial Number: 38DL-2847193

Status: New gauge placed in service

Calibration Standards

Reference Block: LNDE-CS-001

Block Description: 4-step carbon steel, 0.100-0.400 in

Traceability: Larkhall NDE Services Inc. Master Standards

Certification Period

Calibration Date: 2017-07-10

Valid Until: 2018-07-10

Interval: 12 months

Sign-off	Name	Function	Date	Signature
Prepared	Robert Kim	calibration technician	2017-07-10	_____