Larkhall NDE Services Inc.

Calibration Certificate

Site: Hamilton Plant

No.	Status	Date
LNDE-CAL-2024-0847	Issued	2024-10-15

Calibration Certificate - Olympus 38DL PLUS 2847193

Instrument Identification

Instrument: Olympus 38DL PLUS

Serial Number: 38DL-2847193

Calibration Date: 2024-10-15

Valid Until: 2025-10-15

Reference Standards

Primary Reference: LNDE-CS-001 (4-step carbon steel calibration block, 0.100-0.400 in)

Secondary Reference: LNDE-SS-003

Traceability: NIST traceable via Larkhall NDE Services Inc. master standards

Calibration Details

Procedure: LNDE-CAL-UT-001 Rev 3

Scope: Gauge calibrated ahead of the 5-year survey

Location: Hamilton Plant

Sign-off	Name	Function	Date	Signature
Prepared	Robert Kim	calibration technician	2024-10-15	_____