# Calibration-Aware Anisotropic 3D Hysteresis and Leakage Auditing for Volumetric Lower-Grade Glioma Segmentation and Radiogenomic Profiling in Brain MRI

Shayan Ganji Nataj[a], Keivan Borna[a,*]

[a]*Department of Computer Science, Faculty of Mathematics and Computer Science, Kharazmi University, Tehran, Iran*

**Abstract**

Automated lower-grade glioma (LGG) segmentation in brain MRI is vital for neurosurgical planning and radiotherapy. However, clinical translation of 2D deep learning models is hindered by slice-level data leakage, severe probability miscalibration under overlap losses, and axial boundary discontinuity. We introduce CAL-3DH (Calibration-Aware, Leakage-Audited Anisotropic 3D Hysteresis), uniting quality-aware preprocessing, nested probability calibration, and anisotropic 3D post-processing. Evaluated on the complete TCGA-LGG benchmark (110 patients, 3,929 axial slices) under patient-grouped 5-fold cross-validation with nested inner-fold tuning (75/13/22 patient split), our controlled audit empirically demonstrates that conventional random slice splitting artificially inflates positive-slice Dice by +8.79 percentage points (pp) (+11.88% relative inflation) via patient identity memorization. Nested temperature scaling ($T^* = 1.844$) reduces pooled Expected Calibration Error by 46.9% without shifting decision boundaries. By decoupling in-plane planar segmentation from axial rescue along the $z$-axis, CAL-3DH achieves positive-slice Dice of 74.92% [95% CI: 70.82%–78.39%], overall Dice of 86.12% [95% CI: 84.32%–87.75%], and a 42.4% relative reduction in tumor-free slice false alarms. Patient-level 3D volumetric analysis yields 77.73% Dice, 10.97-voxel 95% Hausdorff distance, and good-to-excellent clinical measurement reliability ($\mathrm{ICC}(2,1) = 0.8894$, 95% CI: [0.7782, 0.9521]). Furthermore, radiogenomic profiling confirms that CAL-3DH-derived 3D tumor volumes preserve statistically significant separation across molecular consensus clusters (Kruskal-Wallis $p = 0.0268$) with consistent biological rank ordering. CAL-3DH establishes a standardized, leakage-free benchmark for translational neuro-oncology.

*Keywords:* Lower-grade glioma, Brain MRI segmentation, Probability calibration, Anisotropic 3D hysteresis, Data leakage auditing, Radiogenomic profiling

*Corresponding author.

*Email addresses:* shayanganji@khu.ac.ir (Shayan Ganji Nataj), borna@khu.ac.ir (Keivan Borna)

## 1. Introduction

Lower-grade gliomas (LGG), encompassing World Health Organization (WHO) grade II and grade III astrocytomas and oligodendrogliomas, represent diffusely infiltrative primary brain neoplasms characterized by substantial histological and genetic heterogeneity (Louis et al., 2021; Ceccarelli et al., 2016). Accurate volumetric delineation on magnetic resonance imaging (MRI) is essential for image-guided maximal surgical resection, conformal radiation planning, and monitoring therapeutic response (Buda et al., 2019). Over the past decade, deep learning architectures—most notably 2D U-Net variants (Ronneberger et al., 2015; Zhou et al., 2018; Oktay et al., 2018)—have emerged as the prevailing standard for automated lesion delineation.

Despite reported segmentation accuracies exceeding an 80% Dice similarity coefficient across benchmark datasets, two methodological weaknesses undermine empirical validity in the medical computing literature. First and most severely, slice-level data leakage occurs when volumetric MRI scans are decomposed into independent 2D axial slices and randomly partitioned into training and test sets. Contiguous slices from the same subject inevitably appear across both partitions. Because anatomical morphology, contrast dynamics, and spatial coordinates change gradually across contiguous $z$-slices, deep networks easily memorize subject-specific features. This slice leakage constitutes the primary engine of artificial performance inflation in the literature, falsely boosting reported Dice scores by up to +8.79 pp (+11.88% relative inflation). Second, post-hoc threshold optimization leakage arises when the binary decision threshold $\tau$ is chosen by sweeping values directly over pooled out-of-fold test predictions. While the wide empirical Dice plateau across reasonable threshold ranges bounds the resulting optimism bias to a modest $\Delta\text{Optimism} = +0.063$ pp, establishing nested inner-fold thresholding is an essential methodological standard that allows researchers to legitimately transition from the inefficient default $\tau = 0.50$ to the optimal $\tau^*_{xy} = 0.74$ without test snooping.

Beyond evaluation methodology, 2D networks face fundamental domain-specific limitations when applied to volumetric scans. Overlap-based loss formulations like Focal Tversky prioritize rare foreground voxels by heavily down-weighting easy background terms, causing output probabilities to become overconfident and miscalibrated. Furthermore, because 65.05% of all axial slices in the TCGA-LGG cohort are healthy brain tissue containing zero neoplastic voxels, standard 2D networks lacking cross-slice volumetric context frequently trigger false-positive alarms on non-tumor slices while failing to detect small infiltrative margins at apical and basal tumor boundaries.

To resolve these challenges, this study presents **CAL-3DH** (*Calibration-Aware, Leakage-Audited Anisotropic 3D Hysteresis*), an end-to-end volumetric framework evaluated on the complete TCGA-LGG cohort (110 patients, 3,929 slices) (Buda et al., 2019). The contributions of this work are fourfold:

1. **Controlled Leakage Audit:** We implement a matched 5-fold cross-validation experiment directly comparing patient-grouped partitioning against random-slice partitioning under identical training hyperparameters, empirically quantifying the performance

inflation caused by slice memorization (+8.79 pp on positive-slice Dice and +6.20 pp on overall Dice).

2. **Nested Inner-Fold Calibration:** We integrate scalar temperature scaling fitted exclusively on internal validation logits (`inner_val`), reducing pooled Expected Calibration Error (ECE) by 46.9% on unseen test data without shifting classification boundaries.
3. **Anisotropic 3D Z-Continuity Hysteresis in CAL-3DH:** We decouple horizontal in-plane segmentation ($\tau^*_{xy}$) from axial margin rescue ($\tau^*_{z,\text{low}}$) and suppress isolated 3D false positives ($V^*_{\text{min}}, \tau^*_{\text{seed}}$), improving overlap accuracy ($\text{Dice}_{\text{pos}} = 74.92\%$, $\text{Dice}_{\text{all}} = 86.12\%$) and positive precision (76.86%, a +4.91 pp gain) while reducing empty-slice false alarms by 42.4% relative to baseline ($FPR_{\text{empty}} = 7.86\%$).
4. **Volumetric Radiogenomic Validation:** We conduct 3D patient-level surface and volumetric assessments ($n = 110$), demonstrating good-to-excellent clinical measurement reliability ($\text{ICC}(2,1) = 0.8894$, 95% CI: $[0.7782, 0.9521]$) and confirming that model-derived tumor volumes preserve statistically significant separation across molecular consensus clusters ($p = 0.0268$).

The synergistic combination of automated image quality screening ($Q(I)$), multi-scale wavelet denoising, post-hoc probabilistic calibration, and radiogenomic digital biomarker extraction aligns directly with the scope of *Computerized Medical Imaging and Graphics*, advancing computational neuro-oncology toward translational clinical readiness.

## 2. Dataset Audit and Quality-Aware Preprocessing

### *2.1. Cohort Characteristics and Clinical Variables*

Experiments were conducted using the standardized TCGA-LGG MRI collection introduced by Buda et al. (2019), comprising 110 patients diagnosed with WHO grade II or III lower-grade gliomas. Each patient study consists of axial multi-sequence MRI slices (fluid-attenuated inversion recovery [FLAIR], T1-pre, and T1-post contrast) mapped into $256 \times 256$ matrices with matched expert manual segmentations verified by board-certified radiologists. The full imaging volume encompasses 3,929 paired axial slices. Per SAGER reporting guidelines, the cohort includes 56 male patients (50.9%) and 53 female patients (48.2%), with sex recorded as unknown for one patient. Clinical records extracted from `data.csv` document patient demographics, histological subtype, tumor grade, and genomic clusters (Table 1).

### *2.2. Quality-Aware Screening and Wavelet Denoising*

Prior to model training, raw axial slices were analyzed for pronounced imaging artifacts and skull-stripping failures using three complementary morphological descriptors. First, parenchymal texture insufficiency was quantified via Edge Density ($ED$), defined as the proportion of Canny edge pixels (thresholds 50 and 150) following morphological dilation with a $5 \times 5$ square structuring element, normalized by image area:

$$ED = \frac{|E_{\text{dilated}}|}{H \times W} \tag{1}$$

**Table 1**: Clinical, histological, and genomic characteristics of the 110 patients in the TCGA-LGG cohort.

| **Clinical Variable** | **Category / Unit** | **Value / Count** |
|---|---|---|
| Total Patients ($N$) | Subjects | 110 |
| Total Axial Slices | Slices | 3,929 |
| Tumor-Bearing Slices | Slices ($y > 0$) | 1,373 (34.95%) |
| Tumor-Free Slices | Slices ($y = 0$) | 2,556 (65.05%) |
| Patient Sex | Male ($n = 56$) | 50.9% |
| (SAGER Reporting) | Female ($n = 53$) | 48.2% |
| | Unknown ($n = 1$) | 0.9% |
| Patient Age | Mean ± SD (years) | 47.1 ± 13.4 |
| | Range (Min – Max) | 20.0 – 75.0 |
| WHO Histologic Grade | Grade II ($n = 51$) | 46.4% |
| | Grade III ($n = 58$) | 52.7% |
| | Unspecified ($n = 1$) | 0.9% |
| Molecular Subtypes | COC 1 (IDH-wt like) | 56 (50.9%) |
| (*COCCluster*) | COC 2 (IDHmut-noncodel) | 24 (21.8%) |
| | COC 3 (IDHmut-codel) | 30 (27.3%) |

where slices satisfying $ED < 0.003$ exhibit negligible brain tissue. Second, anatomical field-of-view truncation was evaluated through the Contour Ratio ($CR$), defined as the area of the largest closed external contour extracted from the dilated edge map divided by image area:

$$CR = \frac{\text{Area}(\mathcal{C}_{\text{largest}})}{H \times W} \tag{2}$$

where $CR < 0.01$ flags severe field truncation. Third, decentering was assessed by the Centroid Distance ($d_c$), defined as the Euclidean distance from the external contour centroid $(c_x, c_y)$ to the geometric slice center normalized by the half-diagonal:

$$d_c = \frac{\sqrt{(c_x - W/2)^2 + (c_y - H/2)^2}}{\sqrt{(W/2)^2 + (H/2)^2}} \tag{3}$$

where $d_c > 0.30$ captures severe peripheral displacement.

Slices satisfying the conjunction $(ED < 0.003) \wedge (CR < 0.01) \wedge (d_c > 0.30)$ were flagged as morphological artifacts. A rigorous cohort audit across all 3,929 slices identified 200 artifact slices (5.09%, Table 2). Exactly 199 of these 200 slices (99.50%) contained zero neoplastic tissue ($y = 0$). Only a single slice (0.50%, Slice 23 of TCGA_DU_5855) contained a marginal peripheral tumor component truncated by skull-stripping boundary erosion. Crucially, these 200 artifact slices (5.09%) were excluded strictly from the 75-patient training set during model weight optimization; all internal validation (13 patients) and untouched outer test evaluations (22 patients) are reported across the complete, unfiltered 3,929-slice cohort (with Table 2 and Table 3 additionally displaying the 3,729-slice screened test comparison). Retaining these 200 noisy slices in the untouched test evaluation confirmed an

**Table 2**: Screening audit of the 200 artifact slices excluded during training.

| **Screening Metric** | **Flagged Slices** | **Ratio (%)** |
|---|---|---|
| Total Slices Screened | 3,929 | 100.0% |
| Total Artifact Slices Excluded | 200 | 5.09% |
| Empty Slices (Zero Tumor, $y = 0$) | 199 / 200 | **99.50%** |
| Marginal Slices (Atypical Clipping) | 1 / 200 | 0.50% |
| Edge Density Violation ($ED < 0.003$) | 200 / 200 | 100.0% |
| Contour Ratio Violation ($CR < 0.01$) | 200 / 200 | 100.0% |
| Centroid Distance Violation ($d_c > 0.30$) | 200 / 200 | 100.0% |
| Triple Conjunction Violation | 200 / 200 | **100.0%** |
| Positive-Slice Dice on Full Cohort (3,929) | 73.96% [70.10 – 77.15] | |
| Positive-Slice Dice on Screened (3,729) | 73.96% [70.10 – 77.13] | |
| **Absolute Impact on Positive Dice** | **0.00% (Identical)** | |

absolute 0.00 pp change in positive-slice Dice (73.96% in both full and screened cohorts), verifying zero empirical bias. All 2D and 3D volumetric and surface distance metrics ($HD_{95}$, ASD, volumes) are evaluated on the standardized $128 \times 128$ in-plane grid in voxel units because the public Kaggle/TCIA TIFF distribution does not retain raw DICOM physical millimeter spacing metadata.

To suppress high-frequency acquisition noise while retaining crisp anatomical borders, non-flagged slices were processed using BayesShrink soft-thresholded discrete wavelet denoising (Chang et al., 2000) conditioned on a localized slice quality metric $Q(I)$:

$$Q(I) = \frac{1}{2}\Big[\sigma\big(I - \mathcal{M}_{3\times3}(I)\big) + \sigma\big(I - \mathcal{G}_{3\times3}(I)\big)\Big] \tag{4}$$

representing the average standard deviation of high-frequency residuals under $3 \times 3$ median ($\mathcal{M}_{3\times3}$) and Gaussian ($\mathcal{G}_{3\times3}$) spatial filters. Wavelet denoising (Symlet-4 wavelet, 2-level decomposition) was applied when $Q(I) < 4.0$ (3,738 slices, 95.14%). Slices with high intrinsic contrast ($Q(I) \geq 4.0$, 191 slices, 4.86%) bypassed filtering to avoid unnecessary smoothing of delicate parenchymal boundaries (Figure 1). Slices were normalized to zero mean and unit variance and resized to $128 \times 128 \times 3$.

## 3. Network Architecture and Nested Evaluation Protocol

### *3.1. Attention Residual U-Net Architecture*

The segmentation backbone of CAL-3DH employs an Attention Residual U-Net (Attn Res-UNet) with deep supervision (Figure 2). Residual units replace standard two-layer convolutional blocks, utilizing identity shortcut mappings that mitigate vanishing gradients across deep stages. Channel capacities progress as $[32, 64, 128, 256]$ in the encoder, transitioning to a 512-channel bottleneck.





**Figure 1**: Quality-aware preprocessing and screening audit across the TCGA-LGG cohort (3,929 axial slices). (a) Parenchymal noise quality index $Q(I)$ distribution and adaptive quality gate ($Q = 4.0$), routing 3,738 slices (95.14%) to BayesShrink wavelet denoising while bypassing 191 high-contrast slices (4.86%). (b) Representative real MRI slice examples: top row displays three of the 200 artifact slices excluded during training (illustrating field-of-view truncation, scanner noise fringe, and non-parenchymal tissue with exact $ED, CR, d_c$ metrics); bottom row compares raw, wavelet-denoised, and high-contrast bypass slices.

To ensure numerical stability when training with small mini-batch sizes ($B = 16$), all batch normalization layers are replaced with Group Normalization (GN) (Wu and He, 2018) using $G = 8$ groups:

$$x_{\text{norm}} = \gamma \cdot \frac{x - \mu_G}{\sqrt{\sigma_G^2 + \epsilon}} + \beta \tag{5}$$

where $\mu_G$ and $\sigma_G$ are computed independently across spatial coordinates and group channels. Feature non-linearities utilize Gaussian Error Linear Units (GELU) (Hendrycks and Gimpel, 2016):

$$\text{GELU}(x) = x \cdot \Phi(x) = x \cdot P(X \leq x), \quad X \sim \mathcal{N}(0, 1) \tag{6}$$

Regularization is reinforced via Stochastic Depth (DropPath) (Huang et al., 2016) across residual branches:

$$\begin{aligned} H &= \text{GELU}\left(\text{shortcut}(x) + \frac{b}{1-p}\mathcal{F}(x)\right), \\ b &\sim \text{Bernoulli}(1-p) \end{aligned} \tag{7}$$

with linear decay rates $p \in [0.00, 0.05, 0.10, 0.15]$ in the encoder stages and $p = 0.20$ at the bottleneck.

In the decoder, additive attention gates (Oktay et al., 2018) dynamically recalibrate intermediate feature representations before concatenation. Given spatial features $x^l$ and





**Figure 2**: End-to-end CAL-3DH processing pipeline showing Quality-Aware Preprocessing (Stage 1), the Attention Res-UNet backbone with Group Normalization ($G = 8$), GELU, Stochastic Depth, and Deep Supervision (Stage 2), followed by Nested Temperature Scaling ($T^* = 1.844$) and Anisotropic 3D Z-Continuity Hysteresis (Stage 3).

gating signal $g$, attention coefficients $\alpha^l \in [0, 1]$ are obtained via:

$$\alpha^l = \sigma\left(\psi^T\left(\text{GELU}\left(W_x^T x^l + W_g^T g + b_g\right)\right) + b_\psi\right) \tag{8}$$

where linear projections $W_x$ and $W_g$ align channel dimensions, and $\sigma$ denotes the sigmoid activation. This gating suppresses background noise in peripheral brain regions while amplifying infiltrative tumor signals. Deep supervision auxiliary heads at decoders 2 and 3 provide intermediate gradients during backpropagation. The final network comprises exactly 8.20 million trainable parameters (8,202,005 parameters).

### *3.2. Loss Function and Training Dynamics*

To address acute class imbalance (positive lesion pixels comprise less than 1.5% of total voxels), networks were trained using a compound production loss combining Focal Tversky Loss (FTL) (Salehi et al., 2017; Abraham and Khan, 2019) and Binary Focal Loss (BFL) (Lin et al., 2017):

$$\mathcal{L}_{\text{total}} = 0.70\,\mathcal{L}_{\text{FTL}} + 0.30\,\mathcal{L}_{\text{BF}} \tag{9}$$

The Focal Tversky formulation is defined as:

$$\text{TI}(y, \hat{p}) = \frac{\sum_i y_i \hat{p}_i + \epsilon}{\sum_i y_i \hat{p}_i + \alpha \text{FN}(y, \hat{p}) + \beta \text{FP}(y, \hat{p}) + \epsilon} \tag{10}$$

$$\mathcal{L}_{\text{FTL}} = (1 - \text{TI})^\gamma \tag{11}$$

where $\gamma = 0.60$, $\epsilon = 1.0$, and the voxel-level error components are defined as $\text{FN}(y, \hat{p}) = \sum_i y_i(1 - \hat{p}_i)$ and $\text{FP}(y, \hat{p}) = \sum_i (1 - y_i)\hat{p}_i$.

In this formulation, $\alpha = 0.80$ is the penalty weight assigned to **False Negatives (FN)**, while $\beta = 0.20$ weights **False Positives (FP)**. Heavily penalizing false negatives ($\alpha = 0.80 \gg \beta = 0.20$) prioritizes the detection of diffuse, low-contrast tumor boundaries, achieving high baseline Recall (82.98%). However, this asymmetric penalty leads raw network probabilities to exhibit an intermediate-confidence false-positive halo surrounding lesions (raw Precision = 71.95%) and substantial probability miscalibration ($T^* = 1.844$). As detailed in Section 4, this peripheral false-positive halo is systematically eliminated in the CAL-3DH framework through nested calibration and anisotropic 3D Z-hysteresis filtering, elevating final Precision to 76.86%.

The Binary Focal Loss component is configured with focusing parameter $\gamma_{\text{BF}} = 2.0$ and label smoothing $\epsilon_{\text{LS}} = 0.02$. Auxiliary heads participate via multi-scale deep supervision:

$$\mathcal{L}_{\text{multi}} = 1.0\,\mathcal{L}_{\text{total}}^{(\text{main})} + 0.30\,\mathcal{L}_{\text{total}}^{(d2)} + 0.30\,\mathcal{L}_{\text{total}}^{(d3)} \tag{12}$$

Optimization utilized the AdamW optimizer (Loshchilov and Hutter, 2019) with weight decay ($10^{-4}$) and a Cosine Annealing learning rate schedule featuring an 8-epoch linear warmup up to $\eta_{\max} = 3 \times 10^{-4}$, decaying to $\eta_{\min} = 10^{-6}$ over 100 epochs with early stopping patience of 20 epochs. Mini-batch size was fixed to 16. Stochastic data augmentation (random rotation $\pm 15°$, scaling 0.90–1.10, and elastic deformation with $\alpha = 15, \sigma = 3$) was

active strictly during training. Horizontal flipping was intentionally excluded to preserve biological hemispheric asymmetry of gliomas. Validation and testing were performed in pure deterministic inference mode. All experiments were implemented in Python 3.11 with PyTorch 2.4 and CUDA 12.2 on an NVIDIA RTX 4090 GPU (24GB VRAM). A fixed global random seed (42) was enforced across all patient-grouped folds, weight initialization, and stochastic data augmentations to guarantee exact computational reproducibility.

### *3.3. Patient-Grouped 5-Fold Cross-Validation*

To prevent slice-level data leakage, 110 patients were partitioned into 5 balanced folds using stratified group splitting. All slices from any given patient were assigned strictly to a single fold.

In each fold $k \in \{1, \ldots, 5\}$, exactly 22 patients (20% of the 110-subject cohort, $\approx 786$ slices per fold) were set aside as an untouched outer test partition (`outer_test`), strictly blinded and evaluated only once per fold. The remaining 88 non-test patients were systematically partitioned in an 85%/15% ratio into 75 training patients ($\approx 2,535$ clean slices per fold following artifact exclusion) utilized to optimize network parameters $\theta_k$, and an internal validation partition (`inner_val`) of 13 patients (467 slices per fold, 2,338 slices pooled across the 5 folds). While reserving 13 patients specifically for inner calibration reduces the training pool from 88 to 75 subjects, this design rigorously prevents any hyperparameter, temperature ($T^*$), or threshold ($\tau^*$) leakage into the 22 held-out test patients.

Every test slice was evaluated strictly by the single dedicated model $\mathcal{M}_k(\cdot; \theta_k^*)$ trained for that fold:

$$\hat{Y}_i = \mathcal{M}_k(X_i; \theta_k^*), \quad \forall i \in \mathcal{D}_{\text{test}}^{(k)} \tag{13}$$

This 1-to-1 patient-to-fold mapping eliminates ensemble averaging ambiguity: no soft blending or multi-model averaging was utilized.

To quantify the precise degree of optimistic bias induced by slice-level data leakage, a matched parallel 5-fold Random-Slice cross-validation control experiment was conducted concurrently. Under this control protocol, the 3,929 axial slices were randomly partitioned into five equal folds without patient identity grouping, allowing slices from all 110 patients (100% patient overlap) to co-exist simultaneously across training, validation, and testing sets. Identical network architectures, learning rate schedules, loss formulations ($\alpha = 0.8, \beta = 0.2, \gamma = 0.6$), and stopping criteria were strictly enforced to isolate the effect of patient-level data leakage from model capacity.

### *3.4. Slice-Level Metric Definitions*

In MRI volumes containing large proportions of healthy slices, evaluation metrics must handle non-tumor slices without distortion. We adhere to the standard clinical convention:

$$\text{Dice}(y, \hat{y}) = \begin{cases} 1.0, & \text{if } |y| = 0 \ \wedge \ |\hat{y}| = 0 \text{ (TN)} \\ 0.0, & \text{if } |y| = 0 \ \wedge \ |\hat{y}| > 0 \text{ (FP)} \\ \frac{2|y \cap \hat{y}|}{|y| + |\hat{y}|}, & \text{if } |y| > 0 \end{cases} \tag{14}$$

Three primary slice-level metrics are reported to comprehensively capture clinical utility: $\text{Dice}_{\text{pos}}$, calculated strictly over the 1,373 tumor-bearing slices ($|y| > 0$); $\text{Dice}_{\text{all}}$, computed over all 3,929 slices including healthy non-tumor tissue; and the empty-slice false-positive rate ($\text{FPR}_{\text{empty}}$), quantifying the percentage of healthy slices erroneously predicted to contain lesion voxels ($|\hat{y}| > 0$ when $|y| = 0$).

## 4. Calibration and Anisotropic 3D Z-Hysteresis (CAL-3DH)

### *4.1. Nested Temperature Scaling Calibration*

Because overlap losses yield uncalibrated probability estimates, raw network logits $z(x)$ were calibrated via post-processing temperature scaling (Guo et al., 2017). Temperature parameter $T_k > 0$ was optimized strictly on inner validation logits by minimizing negative log-likelihood (NLL):

$$\hat{T}_k^* = \arg\min_{T>0} \sum_{j \in \mathcal{D}_{\text{val}}^{(k)}} \Big[ - y_j \log \sigma(z_j/T) - (1 - y_j) \log \big(1 - \sigma(z_j/T)\big) \Big] \tag{15}$$

Because temperature scaling applies a monotonic scalar transformation ($T_k > 0$), it preserves the relative ranking of logits while softening extreme confidence values. Expected Calibration Error (ECE) was calculated across $M = 15$ equally spaced probability bins:

$$\text{ECE} = \sum_{m=1}^{M} \frac{|B_m|}{N} \left|\text{acc}(B_m) - \text{conf}(B_m)\right| \tag{16}$$

To quantify miscalibration within the clinically critical infiltrative margin region while preventing background dilution, we define the active-margin Overconfidence Gap across probability bins $\mathcal{M}_{\text{active}} = \{m \mid \text{bin}_m \subseteq [0.2, 0.9]\}$:

$$\text{Gap}_{\text{active}} = \frac{1}{|\mathcal{M}_{\text{active}}|} \sum_{m \in \mathcal{M}_{\text{active}}} \left|\text{conf}(B_m) - \text{acc}(B_m)\right| \tag{17}$$

The functional role of calibration in the CAL-3DH pipeline is twofold: because raw, uncalibrated confidence fluctuates substantially across training folds ($T_k^* \in [1.520, 2.387]$), scalar temperature scaling maps raw logits onto a stable, probabilistically consistent likelihood manifold before applying downstream spatial hysteresis rules ($\tau_{z,\text{low}}^*, \tau_{\text{seed}}^*$), while simultaneously providing clinicians with reliable voxel-level uncertainty maps along ambiguous infiltrative tumor margins.

### *4.2. Anisotropic 3D Z-Continuity Hysteresis Algorithm*

In 2D axial segmentation, high decision thresholds suppress false alarms but truncate apical and basal tumor margins, whereas low thresholds preserve boundaries at the

expense of false alarms on healthy slices. CAL-3DH resolves this trade-off through an anisotropic 3D Z-continuity hysteresis algorithm operating across calibrated probabilities $\hat{P} = \sigma(z/\hat{T}_k^*)$. All post-processing hyperparameters were optimized strictly on internal validation splits (`inner_val`, 13 patients): (1) scalar temperature $T_k \in [0.5, 4.0]$ via bounded scalar minimization of validation NLL; (2) planar threshold $\tau_{xy,k} \in [0.20, 0.90]$ with step 0.02 maximizing validation $\text{Dice}_{\text{pos}}$; and (3) 3D spatial parameters via grid search over $\tau_{z,\text{low}} \in \{0.35, 0.40, 0.45, 0.50\}$, $V_{\min} \in \{20, 50, 80\}$, and $\tau_{\text{seed}} \in \{0.70, 0.75\}$ maximizing validation patient-volume $\text{Dice}_{\text{pos}}$:

1. **In-Plane Planar Binarization:** Each slice is initially thresholded with fold-specific inner-validation threshold $\tau_{xy,k}^*$ ($0.62 \leq \tau_{xy}^* \leq 0.84$, mean: 0.74):
$$M^{(0)}(x, y, z) = \mathbb{I}\left(\hat{P}(x, y, z) \geq \tau_{xy,k}^*\right) \tag{18}$$
preserving horizontal tumor boundary fidelity and shaving the FP halo.
2. **Z-Continuity Boundary Rescue:** For slices where $M^{(0)}$ detects zero (or near-zero, $< 10$ pixels) tumor tissue, adjacent slices $z-1$ and $z+1$ are inspected. If an adjacent slice contains a verified tumor ($\geq 20$ pixels), a dilated spatial footprint $\mathcal{S}_{\text{prior}}$ is projected onto slice $z$:
$$\mathcal{S}_{\text{prior}} = \text{Dilate}_{2D}\left(M_{z-1}^{(0)} \cup M_{z+1}^{(0)}, \mathcal{K}_{3\times 3}\right) \tag{19}$$
Voxels within $\mathcal{S}_{\text{prior}}$ satisfying a relaxed axial threshold $\tau_{z,\text{low},k}^*$ (mean $\tau_{z,\text{low}}^* = 0.35$, where $\tau_{z,\text{low}}^* < \tau_{xy}^*$) are rescued:
$$M_z^{(1)} = M_z^{(0)} \cup \left(\mathcal{S}_{\text{prior}} \cap \mathbb{I}(\hat{P}_z \geq \tau_{z,\text{low},k}^*)\right) \tag{20}$$
3. **3D Volumetric Suppression:** On the resulting 3D volume, 26-connected components $\{C_l\}$ are labeled. Components are discarded if their 3D volume is smaller than $V_{\text{min},k}^*$ (mean $V_{\min}^* = 38$ voxels) or their peak probability is below $\tau_{\text{seed},k}^*$ (mean $\tau_{\text{seed}}^* = 0.70$):
$$\mathcal{C}_{\text{valid}} = \{C_l \,||C_l| \geq V_{min,k}^* \wedge \max_{v \in C_l} \hat{P}(v) \geq \tau_{seed,k}^*\} \tag{21}$$
4. **Fallback Safeguard:** To prevent small apical tumors from collapsing to empty volumes (which imposes maximal boundary distance penalties), if $\mathcal{C}_{\text{valid}} = \emptyset$, components with $\max \hat{P} \geq \tau_{seed}^*$ are retained.

## 5. Experimental Results and Leakage Auditing

### *5.1. Baseline Segmentation Performance*

Table 3 summarizes baseline Attn Res-UNet performance across all 3,929 out-of-fold test slices at fixed threshold $\tau = 0.50$. Ninety-five percent confidence intervals were computed via 10,000 patient-clustered bootstrap iterations (Efron and Tibshirani, 1994). The model achieved a positive-slice Dice of 73.96% [95% CI: 70.10%–77.15%], an overall Dice of 82.02% [95% CI: 79.75%–84.10%], an intersection-over-union ($\text{IoU}_{\text{pos}}$) of 63.97%, and an empty-slice false-positive rate of 13.65%.





**Figure 3**: Probability calibration and voxel confidence dynamics across 15 probability bins. (a) Reliability diagram comparing the uncalibrated patient-grouped model (ECE = 0.369%), the temperature-calibrated model ($T^* = 1.844$, reducing pooled ECE by 46.9% to 0.196%), and the data-leaked random-slice model against ideal diagonal calibration. (b) Logarithmic voxel frequency distribution across confidence bins for patient-grouped and random-slice regimes.

**Table 3**: Baseline out-of-fold performance at fixed $\tau = 0.50$ across full (3,929) and screened (3,729) cohorts (10,000 patient-clustered bootstrap 95% CIs).

| **Evaluation Metric** | **Full Cohort (3,929)** | **Screened Cohort (3,729)** |
|---|---|---|
| $\text{Dice}_{\text{pos}}$ (%) | 73.96 [70.10 – 77.15] | 73.96 [70.10 – 77.13] |
| $\text{Dice}_{\text{all}}$ (%) | 82.02 [79.75 – 84.10] | 82.21 [80.01 – 84.27] |
| $\text{IoU}_{\text{pos}}$ (%) | 63.97 [59.78 – 67.89] | 63.97 [59.78 – 67.88] |
| $\text{Precision}_{\text{pos}}$ (%) | 71.95 [67.80 – 75.80] | 71.95 [67.80 – 75.80] |
| $\text{Recall}_{\text{pos}}$ (%) | 82.98 [78.60 – 86.70] | 82.98 [78.60 – 86.70] |
| $\text{FPR}_{\text{empty}}$ (%) | 13.65 (349 / 2556) | 13.41 (316 / 2357) |

### *5.2. Nested Temperature Scaling and Thresholding*

Table 4 presents fold-by-fold calibration and threshold optimization results. Fitted optimal temperatures ranged from $T^* = 1.520$ to 2.387 (mean $T^* = 1.844$). On untouched test sets, temperature scaling consistently reduced NLL from 0.0230 to 0.0154 ($-32.9\%$) and Brier score (Brier, 1950) from 0.00375 to 0.00338 ($-9.9\%$). The mean per-fold ECE decreased from 0.372% to 0.254% (a 31.7% relative reduction), while pooled cohort ECE dropped from 0.369% to 0.196% (a 46.9% relative reduction), visually confirmed in Figure 3.

Crucially, the small absolute magnitude of cohort ECE (0.369%) stems from voxel dilution: over 98.5% of MRI voxels belong to background air and healthy, non-enhancing brain tissue ($y = 0, \hat{p} \approx 0$, clearly reflected in the first bin of the logarithmic histogram in Figure 3(b) encompassing $> 6 \times 10^7$ voxels). Within the clinically vital active probability bins corresponding to the ambiguous infiltrative tumor margins ($\hat{p} \in [0.2, 0.9]$ in Figure 3(a)), the raw uncalibrated network exhibits severe overconfidence, manifested by an Overconfidence Gap of 0.253 (25.3%, Section 5.5). Nested temperature scaling ($T^* = 1.844$) contracts this margin overconfidence gap down to 0.200 (a 20.9% relative reduction), substantially shifting empirical confidence toward the ideal diagonal while active-margin weighted ECE decreases from 36.14% to 29.57%. The corresponding training and validation convergence dynamics across folds are illustrated in Figure 4. In Figure 4(a), validation loss is slightly lower than training loss during early epochs because heavy stochastic data augmentation, DropPath stochastic depth ($p \leq 0.20$), and label smoothing ($\epsilon_{\text{LS}} = 0.02$) act as strong regularizers during training and are deactivated during deterministic validation; in contrast, in the random-slice regime (Figure 4(c)–(d)), validation Dice unnaturally surpasses training Dice by nearly 15 pp due to subject memorization across contiguous slices.

### *5.3. Quantification of Optimism Bias ($\Delta$Optimism)*

To rigorously disentangle sources of performance overestimation, we quantified the empirical difference between nested inner-fold thresholding and pooled test threshold sweeping (Figure 5). When decision thresholds were selected strictly on independent inner validation partitions (mean calibrated $\tau^*_{xy} = 0.74$, corresponding to raw uncalibrated threshold 0.86), the model achieved an out-of-fold test $\text{Dice}_{\text{pos}}$ of 74.60% [95% CI: 70.45%–78.08%]. Conversely, sweeping thresholds directly across the pooled outer test set peaked at $\tau^*_{\text{leak}} = 0.76$

**Table 4**: Fold-by-fold nested calibration, threshold optimization, and out-of-fold performance.

| Fold | Optimal $T^*$ | Inner $\tau^*_{xy}$ | Inner $\tau^*_{z,low}$ | Inner $V^*_{min}$ | Inner $\tau^*_{seed}$ | Outer ECE (%) | | Outer NLL | | Outer Test $Dice_{pos}$ (%) | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | Raw | Calibrated | Raw | Calibrated | 2D Base ($\tau = 0.50$) | 2D Nested ($\tau^*_{xy}$) | CAL-3DH (Ours) |
| 1 | 2.019 | 0.76 | 0.35 | 50 | 0.70 | 0.354 | 0.153 | 0.0267 | 0.0159 | 77.02 | 77.00 | **77.26** |
| 2 | 2.387 | 0.62 | 0.35 | 20 | 0.70 | 0.330 | 0.227 | 0.0191 | 0.0112 | 74.96 | 75.01 | **75.04** |
| 3 | 1.649 | 0.82 | 0.35 | 80 | 0.70 | 0.391 | 0.367 | 0.0186 | 0.0141 | 79.39 | 81.91 | **82.39** |
| 4 | 1.520 | 0.84 | 0.35 | 20 | 0.70 | 0.476 | 0.335 | 0.0325 | 0.0232 | 70.44 | 70.50 | **70.96** |
| 5 | 1.644 | 0.68 | 0.35 | 20 | 0.70 | 0.308 | 0.186 | 0.0181 | 0.0127 | 68.15 | 68.61 | **69.01** |
| **Mean** | **1.844** | **0.74** | **0.35** | **38** | **0.70** | **0.372** | **0.254** | **0.0230** | **0.0154** | **73.96** | **74.60** | **74.92** |





**Figure 4**: Training and validation convergence dynamics across 5 folds under identical training hyperparameters. (a) Patient-Grouped: Loss convergence showing steady generalization to unseen subjects without overfitting. (b) Patient-Grouped: Dice progression reaching stable plateau. (c) Random-Slice: Loss collapse where validation loss falls sharply below training loss due to slice memorization. (d) Random-Slice: Curve inversion where validation Dice exceeds training Dice across all epochs.





**Figure 5**: Decision threshold sensitivity analysis on out-of-fold predictions plotted across calibrated probabilities $\hat{P} = \sigma(z/\hat{T}^*_k)$. The curve illustrates the operational trade-off between tumor-slice Dice ($\text{Dice}_{\text{pos}}$), overall Dice ($\text{Dice}_{\text{all}}$), and false alarm rate on empty slices ($\text{FPR}_{\text{empty}}$). Vertical markers denote the standard baseline ($\tau = 0.50$), the nested inner-fold calibrated threshold ($\tau^*_{xy} = 0.74$, corresponding to raw uncalibrated threshold 0.86), and the post-hoc leaked threshold ($\tau^*_{\text{leak}} = 0.76$, raw equivalent 0.88).

(raw threshold 0.88), yielding an apparent $\text{Dice}_{\text{pos}}$ of 74.66%. The resulting empirical optimism bias is:

$$\Delta\text{Optimism} = 74.66\% - 74.60\% = +0.063 \text{ pp} \tag{22}$$

This modest optimism bias (+0.063 pp) contrasts sharply with the massive +8.79 pp inflation generated by slice-level data leakage (Section 5.5). The primary engine of artificial performance inflation in prior literature is subject identity memorization across contiguous slices, whereas nested inner-fold thresholding represents an essential methodological best practice. Because the Dice plateau is remarkably stable across calibrated probabilities $\hat{P} \in [0.68, 0.85]$ (Figure 5), the empirical risk of post-hoc threshold tuning is negligible. This plateau stability guarantees that transitioning from the default, suboptimal threshold ($\tau = 0.50$) to the nested optimal threshold ($\tau^*_{xy} = 0.74$) legitimately captures +0.64 pp in $\text{Dice}_{\text{pos}}$ (73.96% to 74.60%), a +2.43 pp jump in overall $\text{Dice}_{\text{all}}$ (82.02% to 84.45%), and a 24.9% relative reduction in empty-slice false alarms (349 to 262 slices) with zero test partition snooping.

*5.4. CAL-3DH Performance and Step-by-Step Anisotropic 3D Z-Hysteresis Ablation*

In clinical volumetric FLAIR protocols where 65.05% of axial slices are entirely tumor-free ($2,556/3,929$ slices), the primary clinical advantage of CAL-3DH lies in eliminating **148 false alarm slices relative to the baseline ($-42.4\%$, dropping from 349 to 201 slices)** and **61 fewer false alarm slices compared to 2D Nested ($-23.3\%$ additional relative reduction, from 262 to 201 slices)**. This selective pruning propels overall volumetric accuracy ($\text{Dice}_{\text{all}}$) by +4.10 pp (+5.00% relative gain) over baseline and +1.67 pp over 2D Nested ($p < 0.0001$, patient-level paired Wilcoxon signed-rank test).

Crucially, unlike conventional isotropic 3D volumetric filters that cause axial-transverse blurring and degrade horizontal in-plane sensitivity on anisotropic thick-slice scans ($z$-spacing typically 4–6 mm vs. 1 mm in-plane), CAL-3DH preserves high-fidelity planar boundaries while recovering discontiguous apical and basal slices via confident 2D spatial seeds. Consequently, positive-slice Dice reaches 74.92% [95% CI: 70.82%–78.39%], uniformly improving across all five cross-validation folds simultaneously (Table 4).

To disentangle the specific biophysical contributions of each algorithmic component in Stage 3, Table 5 details the step-by-step ablation across all 3,929 out-of-fold test slices:

1. **Step 1 Only (2D Nested Thresholding, $\tau_{xy}^*$):** Raising the in-plane decision threshold from 0.50 to the calibrated inner-fold optimum $\tau_{xy}^* = 0.74$ elevates $\text{Dice}_{\text{pos}}$ from 73.96% to 74.60% and improves Precision from 71.95% to 76.51%, reducing empty-slice false alarms by 24.9% (349 to 262 slices). However, it leaves missed apical/basal boundary slices unaddressed (0/220 rescued).
2. **Step 1 + Step 2 (+ Axial Z-Continuity Rescue, $\tau_{z,\text{low}}^*$):** Activating axial continuity rescue ($\tau_{z,\text{low}}^* = 0.35$ on slices adjacent to confident 2D seeds) boosts Recall from 80.14% to 80.62%, elevates boundary slice Dice from 29.85% to 30.39%, and successfully rescues 4 previously completely missed apical boundary slices (4/220, 1.82%). Because low-threshold propagation can expand into adjacent slices, empty-slice false alarms slightly increase to 291 slices (11.38%) in the absence of volume filtering.
3. **Step 1 + Step 3 (+ 3D Component Filtering, $V_{\text{min}}^*, \tau_{\text{seed}}^*$):** Enforcing 3D connected-component volume filtering ($V_{\text{min}}^* = 38$ voxels, $\tau_{\text{seed}}^* = 0.70$) without Z-rescue aggressively suppresses spurious noise, dropping empty-slice false alarms to 176 slices (6.89%) and increasing $\text{Dice}_{\text{all}}$ to 86.66% and Precision to 77.34%. However, without Z-rescue, missed boundary slices remain unrecovered (0/220).
4. **Full CAL-3DH (Step 1 + Step 2 + Step 3 + Fallback):** Combining both steps harmonizes sensitivity and specificity: all 4 rescued apical slices are preserved (boundary slice Dice reaches 30.48%), $\text{Dice}_{\text{pos}}$ attains its peak at 74.92% [95% CI: 70.82%–78.39%], and empty-slice false alarms are maintained at a low 201 slices (7.86%, a 42.4% relative reduction over baseline).

To account for inter-slice clustering within patients ($n = 110$), we evaluated valid patient-level paired Wilcoxon signed-rank tests with paired bootstrap 95% confidence intervals:

- **Patient-Averaged $\text{Dice}_{\text{all}}$:** CAL-3DH achieves $86.16 \pm 8.78\%$ [95% CI: 84.53%–87.75%], significantly outperforming both 2D Baseline ($82.15 \pm 10.58\%$ [95% CI: 80.20%–

**Table 5**: Step-by-step ablation of Stage 3 (Anisotropic 3D Z-Hysteresis) across all 3,929 out-of-fold test slices ($n = 110$ patients). Boundary slices ($n = 220$) represent the first and last tumor-bearing axial slices adjacent to healthy tissue across all 110 volumes ($2 \times 110$).

| **Configuration / Ablation Stage** | $\text{Dice}_{\text{pos}}$ **(%)** | $\text{Dice}_{\text{all}}$ **(%)** | **Precision (%)** | **Recall (%)** | $\text{FPR}_{\text{empty}}$ **% (count)** | **Boundary Dice (%)** | **Rescued (/220)** |
|---|---|---|---|---|---|---|---|
| (1) 2D Baseline ($\tau = 0.50$) | 73.96 | 82.02 | 71.95 | **82.98** | 13.65 (349 / 2556) | 29.68 | 0 / 220 |
| (2) Step 1: 2D Nested ($\tau^*_{xy} = 0.74$) | 74.60 | 84.45 | 76.51 | 80.14 | 10.25 (262 / 2556) | 29.85 | 0 / 220 |
| (3) Step 1 + Step 2 (+ Axial Z-Rescue, $\tau^*_{z,\text{low}}$) | 74.88 | 83.81 | 75.82 | 80.62 | 11.38 (291 / 2556) | 30.39 | **4 / 220** |
| (4) Step 1 + Step 3 (+ 3D Filtering, $V^*_{\text{min}}, \tau^*_{\text{seed}}$) | 74.64 | **86.66** | **77.34** | 79.45 | **6.89 (176 / 2556)** | 29.85 | 0 / 220 |
| **(5) CAL-3DH (Full)** (Step 1 + 2 + 3 + Fallback) | **74.92** | 86.12 | 76.86 | 79.74 | 7.86 (201 / 2556) | **30.48** | **4 / 220** |

*Patient-Level Paired Testing ($n = 110$ subjects):* CAL-3DH vs. Baseline: $\text{Dice}_{\text{all}}$ $W = 668, p < 0.0001$; $\text{FPR}_{\text{empty}}$ $W = 75.5, p < 0.0001$; 3D Volumetric Dice $W = 927, p < 0.0001$. CAL-3DH vs. 2D Nested: $\text{Dice}_{\text{all}}$ $W = 687, p < 0.0001$; $\text{FPR}_{\text{empty}}$ $W = 240, p = 0.0002$; 3D Volumetric Dice $W = 645, p < 0.0001$ (paired Wilcoxon signed-rank tests).

84.10%], $W = 668, p < 0.0001$) and 2D Nested ($84.48 \pm 9.33\%$ [95% CI: 82.74%–86.20%], $W = 687, p < 0.0001$).

- **Patient Empty-Slice False Alarms ($\text{FPR}_{\text{empty}}$):** CAL-3DH suppresses false alarms to $7.68 \pm 10.37\%$ [95% CI: 5.81%–9.67%], demonstrating highly significant reductions compared to 2D Baseline ($13.25 \pm 14.05\%$ [95% CI: 10.66%–15.93%], $W = 75.5, p < 0.0001$) and 2D Nested ($10.09 \pm 11.45\%$ [95% CI: 8.01%–12.31%], $W = 240, p = 0.0002$).

- **Patient 3D Volumetric Dice:** As shown in Table 7, CAL-3DH yields $77.73 \pm 18.65\%$ [95% CI: 74.20%–81.18%], achieving statistically significant superiority over 2D Baseline ($76.00 \pm 17.94\%$, $W = 927, p < 0.0001$) and 2D Nested ($77.57 \pm 18.66\%$, $W = 645, p < 0.0001$).

### *5.5. Data Leakage Audit: Patient-Grouped vs. Random-Slice*

Table 6 contrasts patient-grouped partitioning against random-slice partitioning under identical network architectures and loss formulations. As demonstrated by the empirical convergence dynamics (Figure 4(c)–(d)), random slice splitting leads to an unnatural loss collapse and curve inversion, where validation performance surpasses training performance because adjacent slices are shared across partitions.

Random slice splitting artificially inflated positive-slice Dice by **+8.79 pp** (+11.88% relative inflation, from 73.96% to 82.74%) and overall Dice by **+6.20 pp** (+7.56% relative inflation, from 82.02% to 88.22%). Because contiguous slices share identical scanner noise and lesion geometry, the network reproduces adjacent slice masks from memory rather than learning generalizable tumor semantics. Multi-scenario qualitative segmentation results across representative test slices are shown in Figure 6.

**Table 6**: Systematic data leakage audit comparing patient-grouped against random-slice 5-fold cross-validation on the TCGA-LGG cohort.

| Protocol | Leakage Status | $\text{Dice}_{\text{pos}}$ (%) | $\text{Dice}_{\text{all}}$ (%) | Pooled ECE (%) | Brier Score | Overconf. Gap |
|---|---|---|---|---|---|---|
| Patient-Grouped (Uncalibrated) | Zero Leakage | 73.96 | 82.02 | 0.369 | 0.00375 | 0.253 |
| Patient-Grouped (Calibrated $T^*$) | Zero Leakage | 73.96 | 82.02 | **0.196** | **0.00338** | **0.200** |
| Random-Slice (Conventional) | **Slice Memorization** | **82.74** | **88.22** | 0.230 | 0.00228 | 0.278 |
| **Leakage Impact** | Memorization Distortion | **+8.79 pp** | **+6.20 pp** | – | – | – |
| **Calibration Impact** | Uncertainty Regularization | 0.00 pp | 0.00 pp | **-46.9%** | **-9.9%** | **-20.9%** |





**Figure 6**: Multi-scenario qualitative segmentation comparison across representative test slices. Row A (Axial Rescue): Boundary slice rescue along the Z-axis, where 2D Nested misses apical tumor (Dice = 0.0%) while CAL-3DH restores detection (Dice = 67.2%). Row B (Empty-Slice TN): Empty-slice false positive suppression, where 2D models trigger false alarms on skull artifacts while 3D spatial filtering preserves clean non-tumor classifications. Row C (Core Contour): High geometric fidelity on large, complex infiltrative core lesions (Dice = 96.3%).

**Table 7**: Patient-level 3D volumetric and surface boundary metrics across all 110 TCGA-LGG patients. Pairwise comparisons of 3D Volumetric Dice confirm that CAL-3DH statistically significantly outperforms both 2D Baseline ($W = 927, p < 0.0001$) and 2D Nested ($W = 645, p < 0.0001$, paired Wilcoxon signed-rank test).

| Method | 3D Volumetric Dice (%) | $HD_{95}$ (voxels) | ASD (voxels) | 3D Prec. (%) | 3D Rec. (%) |
|---|---|---|---|---|---|
| 2D Baseline ($\tau = 0.50$) | 76.00 ± 17.94 (Med: 82.96) | 12.89 ± 19.21 (Med: 3.57) | 3.59 ± 12.47 (Med: 0.88) | 71.89 | **85.96** |
| 2D Nested ($\tau^*_{xy}$) | 77.57 ± 18.66 (Med: 84.42) | 11.26 ± 18.32 (Med: 2.34) | 3.18 ± 12.48 (Med: 0.76) | 77.40 | 82.98 |
| **CAL-3DH (Ours)** | **77.73 ± 18.65 (Med: 84.42)** | **10.97 ± 18.28 (Med: 2.24)** | **3.10 ± 12.48 (Med: 0.63)** | **77.70** | 82.99 |





**Figure 7**: Clinical 3D volumetric agreement and calibration analysis across 110 patients. (a) Volumetric correlation and linear regression fit between automated 3D predictions and radiologist ground truth (ICC$(2,1) = 0.8894$, Pearson $r = 0.8886, p = 2.44 \times 10^{-38}$, Spearman $\rho = 0.8819$). (b) Bland-Altman agreement plot with minimal mean bias (+143.7 voxels, 2.38%, under 2.5% of mean lesion volume) and 95% limits of agreement $[-4{,}526.6, +4{,}814.1]$ voxels.

## 6. 3D Volumetric Agreement and Radiogenomic Fidelity

### *6.1. Patient-Level 3D Surface and Volumetric Assessment*

Individual 2D predictions were stacked into full 3D volumes across all 110 subjects. Table 7 details 3D Volumetric Dice, 95% Hausdorff Distance ($HD_{95}$), and Average Surface Distance (ASD).

The proposed CAL-3DH framework consistently outperforms both 2D baselines across all patient-level volumetric metrics. 3D Volumetric Dice reaches 77.73%±18.65% with a median of 84.42% (interquartile range [IQR]: 70.85%–91.24%). Furthermore, surface boundary errors are substantially constrained: $HD_{95}$ drops to 10.97±18.28 voxels (median: 2.24, IQR: 1.00–12.03), while ASD improves to 3.10±12.48 voxels (median: 0.63, IQR: 0.29–2.21), confirming high morphological fidelity along complex infiltrative tumor margins.





**Figure 8**: Downstream radiogenomic fidelity and histological stratification with clinical records. (a) 3D tumor sphericity across WHO histologic grades (Grade II vs. Grade III). (b) 3D tumor volume stratification across TCGA molecular consensus clusters (COC 1, COC 2, COC 3), demonstrating statistically significant differentiation ($p = 0.0268$) with exact biological rank ordering preservation.

### *6.2. Clinical Agreement and Bland-Altman Analysis*

Agreement between automated 3D volumes and expert manual ground truth was evaluated using standardized clinical biostatistical tests (Figure 7). Two-way random absolute agreement yielded an Intraclass Correlation Coefficient of ICC$(2,1) = 0.8894$ (95% CI: $[0.7782, 0.9521]$), indicating good-to-excellent clinical measurement reliability according to standard guidelines (Shrout and Fleiss, 1979; Koo and Li, 2016). Linear and monotonic agreements were corroborated by Pearson correlation ($r = 0.8886, p = 2.44 \times 10^{-38}$) and Spearman rank correlation ($\rho = 0.8819, p = 4.72 \times 10^{-37}$). Bland-Altman agreement analysis (Bland and Altman, 1986) revealed a minimal mean volumetric bias of +143.7 voxels (accounting for 2.38%, under 2.5% of mean lesion volume $\approx$ 6,027 voxels), with 95% limits of agreement spanning $[-4{,}526.6, +4{,}814.1]$ voxels. As illustrated in Figure 7(b), these limits of agreement reflect heavy-tailed variance driven by a small subset of large, diffusely infiltrative grade III tumors ($>$ 15,000 voxels), whereas bias remains tightly centered near zero for small and moderate lesion burdens.

### *6.3. Radiogenomic Stratification with Clinical Data*

To verify whether automated segmentations support downstream clinical biomarker extraction, we evaluated whether predicted 3D geometries reproduce known biological associations documented in the TCGA pan-glioma atlas (Figure 8). Genomic and histological variables were utilized strictly for post-hoc validation and were never seen during model training, feature extraction, or hyperparameter selection:

*6.3.1. WHO Histologic Grade Differentiation (Grade II vs. III)*

In clinical neuro-oncology, diffuse LGG grade is governed by microscopic cellular atypia and mitotic index rather than macroscopic tumor burden. A non-parametric Mann-Whitney U test confirmed that tumor volume does not differentiate WHO Grade II ($n = 51$) from Grade III ($n = 58$) in either manual ground truth ($p = 0.8104$; Grade II mean: 5,928.3 voxels; Grade III mean: 6,120.2 voxels) or CAL-3DH segmentations ($p = 0.8151$; Grade II mean: 6,241.8 voxels; Grade III mean: 6,125.9 voxels). In contrast, 3D sphericity $\Psi$ revealed a statistically significant distinction in radiologist ground truth (Grade II: $0.7145 \pm 0.1404$, median 0.7404 [IQR: 0.5850–0.8104]; Grade III: $0.6392 \pm 0.1190$, median 0.6436 [IQR: 0.5508–0.7189], Mann-Whitney $U = 1954.0, p = 0.0040$), reflecting greater boundary irregularity in higher-grade neoplasms. For CAL-3DH segmentations, sphericity exhibited an aligned directional shift (Grade II: $0.6730 \pm 0.1421$, median 0.6502 [IQR: 0.5770–0.7671]; Grade III: $0.6222 \pm 0.1339$, median 0.6359 [IQR: 0.5469–0.7165], Mann-Whitney $U = 1740.0, p = 0.1137$).

*6.3.2. Molecular Subtype Stratification (COCCluster)*

The TCGA Pan-Glioma consortium stratifies LGG into three distinct Cluster-of-Clusters (COC) molecular platforms: COC 1 ($n = 56$, IDH-wildtype-like tumors exhibiting clinically aggressive growth), COC 2 ($n = 24$, IDH-mutant non-codel astrocytomas), and COC 3 ($n = 30$, IDH-mutant 1p/19q co-deleted oligodendrogliomas).

A Kruskal-Wallis non-parametric test on 3D tumor volumes established statistically significant stratification in both manual ground truth ($H = 8.16, p = \mathbf{0.0169}$, effect size $\epsilon^2 = 0.0576$) and automated CAL-3DH segmentations ($H = 7.24, p = \mathbf{0.0268}$, effect size $\epsilon^2 = 0.0489$). Group comparisons confirm that the reported volume values represent cohort group means preserving consistent biological rank ordering:

$$\text{COC 1 Mean (7,098.0)} > \text{COC 3 Mean (5,509.6)} > \text{COC 2 Mean (4,834.5)} \tag{23}$$

Specifically, ground truth volumes yielded: COC 1 mean $6{,}780.5 \pm 4{,}816.2$ voxels (median 6,278.0, IQR: 2,546.5–8,825.5); COC 2 mean $4{,}390.2 \pm 5{,}226.1$ voxels (median 2,680.0, IQR: 1,243.5–4,709.2); and COC 3 mean $5{,}930.7 \pm 4{,}656.1$ voxels (median 3,857.5, IQR: 1,983.0–8,776.0). For CAL-3DH, the segmented volumes yielded: COC 1 mean $7{,}098.0 \pm 5{,}082.1$ voxels (median 6,535.5, IQR: 2,748.5–9,582.5); COC 2 mean $4{,}834.5 \pm 5{,}690.1$ voxels (median 2,566.5, IQR: 1,492.8–5,467.0); and COC 3 mean $5{,}509.6 \pm 4{,}126.6$ voxels (median 4,537.0, IQR: 2,507.2–7,989.0). Pairwise post-hoc Mann-Whitney U testing confirmed significant volumetric separation between COC 1 and COC 2 ($U = 926.0, p = 0.0078$), with expected moderate overlap between intermediate clusters (COC 1 vs. COC 3, $p = 0.2128$; COC 3 vs. COC 2, $p = 0.2007$).

## 7. Discussion and Benchmark Comparison

*7.1. Comparison with State-of-the-Art Literature*

Table 8 benchmarks our results against published investigations evaluated strictly on the TCGA-LGG benchmark dataset (Buda et al., 2019).

**Table 8**: Systematic benchmark comparison against published deep learning studies on the TCGA-LGG dataset, stratified by data partitioning protocol and evaluation metric scope.

| Study | Year | Architecture | Post-Proc. / Calibration | Params | Overall $\text{Dice}_{\text{all}}$ (%) | Tumor $\text{Dice}_{\text{pos}}$ (%) | 3D Vol. Dice / ICC |
|---|---|---|---|---|---|---|---|
| ***Part A: Random-Slice Partitioning (Data-Leaked Regime)*** | | | | | | | |
| LiteMRINet (Yuan and Kang, 2025) | 2025 | LiteMRINet | None / None | 1.61M | 85.50 (Soft)† | NR | NR |
| PoSAM-ULTRA (Alhatamleh et al., 2026) | 2026 | PoSAM-ULTRA | Prompt Tuning | NR | 91.40 (Soft)† | NR | NR |
| This Work (Fixed $\tau = 0.50$) | 2026 | Attn Res-UNet | None / None | 8.20M | 88.22 [Soft: 88.18]† | 82.74† | Invalid (Leaked)‡ |
| This Work (Tuned $\tau^*$) | 2026 | Attn Res-UNet | None / Inner $\tau^*$ | 8.20M | 89.83† | 83.75† | Invalid (Leaked)‡ |
| ***Part B: Patient-Grouped Partitioning (Zero-Leakage Clinical Standard)*** | | | | | | | |
| Buda et al. (Buda et al., 2019) | 2019 | Standard 2D U-Net | None / Fixed $\tau = 0.50$ | 7.80M | NR | 72.10 | NR |
| This Work (2D Baseline) | 2026 | Attn Res-UNet | None / Fixed $\tau = 0.50$ | 8.20M | 82.02 | 73.96 | 76.00% / 0.8652 |
| This Work (2D Nested) | 2026 | Attn Res-UNet | Temp ($T^*$) / Nested $\tau_k^*$ | 8.20M | 84.45 | 74.60 | 77.57% / 0.8860 |
| **CAL-3DH (Ours)** | **2026** | **Attn Res-UNet** | **Temp ($T^*$) / 3D Hyst.** | **8.20M** | **86.12** | **74.92** | **77.73% / 0.8894** |

†Performance inflated by slice-level data leakage (inter-slice correlation and subject feature memorization across training and testing partitions).

‡3D reconstruction is anatomically invalid in random-slice partitioning due to scattered, non-contiguous slice assignment.

**NR**: Not Reported. **Note**: CAL-3DH evaluated with zero leakage achieves $\text{Dice}_{\text{all}} = 86.12\%$, exceeding the 85.50% of LiteMRINet (which benefited from slice leakage).

As detailed in Table 8, recent investigations on the TCGA-LGG benchmark such as LiteMRINet (Yuan and Kang, 2025) (reporting 85.50% Dice) and PoSAM-ULTRA (Alhatamleh et al., 2026) (reporting 91.40% Dice) rely on random slice-level partitioning evaluated using unthresholded soft Dice formulations. As demonstrated by our controlled audit (Section 5.5), random slice partitioning induces substantial performance inflation (+8.79% on positive slices and +6.20% on all slices) because contiguous slices from the same brain volume leak into both training and evaluation subsets, allowing networks to memorize subject anatomy. When audited under identical random-slice conditions, our baseline architecture achieves an overall Dice of 88.22% (global soft Dice of 88.18%) and reaches 89.83% under threshold calibration, directly confirming that elevated scores reflect partition design rather than algorithmic superiority.

In Table 8, we benchmark CAL-3DH alongside published studies on the TCGA-LGG cohort. Notably, Buda et al. (2019) reported a patient-level volumetric Dice of approximately 82% using a standard 2D U-Net under patient-level isolation, whereas our slice-level evaluation under identical patient isolation yields a positive-slice Dice of 74.92% and patient-level 3D Volumetric Dice of 77.73% ($HD_{95} = 10.97$ voxels, $\text{ICC}(2,1) = 0.8894$). Furthermore, while studies like LiteMRINet (Yuan and Kang, 2025) (85.50% soft Dice) and PoSAM-ULTRA (Alhatamleh et al., 2026) (91.40% soft Dice) report high overlap scores under random-slice partitioning, direct numerical equivalence is precluded by differences in data partitioning protocols and empty-slice Dice inclusion conventions. Rather than claiming unqualified algorithmic superiority over leaked models, Table 8 primarily serves to

illuminate the substantial methodological divergence between random-slice and zero-leakage patient-grouped regimes.

*7.2. Limitations and Future Scope*

Four primary limitations warrant consideration. First, while the TCGA-LGG cohort represents a multi-institutional benchmark, prospective validation across external centers with heterogeneous scanner field strengths (1.5 T vs. 3.0 T) and varying acquisition protocols will further substantiate generalizability. Future studies should also implement center-stratified cross-validation using institution codes where DICOM headers permit. Second, the public TCGA-LGG TIFF benchmark distribution lacks raw DICOM/NIfTI physical millimeter spacing headers, necessitating evaluation in normalized voxel space; acquiring full volumetric DICOM metadata in future cohorts will enable isotropic millimeter re-sampling. Third, our cross-validation utilized a fixed single seed (42); while extensive 10,000-sample bootstrap confidence intervals verify metric stability, multi-seed repetitions across massive multi-center registries will provide deeper insight into initialization variance. Fourth, in accordance with Sex and Gender Equity in Research (SAGER) and SGBA reporting guidelines, while demographic sex distributions were documented across the cohort (50.9% male, 48.2% female), sex- and gender-stratified subgroup calibration was not performed due to sample size constraints ($N = 110$ patients); evaluating potential sex-dependent morphological calibration shifts represents an important direction for larger cohorts.

## 8. Conclusion

The clinical translation of automated deep learning systems for lower-grade glioma management requires rigorous adherence to methodological integrity, reliable uncertainty quantification, and anatomical continuity across volumetric scans. In this study, we addressed the pervasive vulnerabilities of 2D segmentation pipelines through CAL-3DH, an end-to-end framework uniting leakage auditing, nested calibration, and anisotropic volumetric post-processing.

Our empirical audit conclusively demonstrated that conventional slice-level random splitting induces a substantial +8.79% artificial Dice inflation by allowing deep networks to memorize patient-specific morphological features across contiguous slices rather than extracting generalizable neoplastic patterns. Furthermore, we showed that uncalibrated overlap losses generate overconfident probabilistic halos that can be effectively regularized via nested temperature scaling without shifting decision boundaries. By explicitly decoupling in-plane planar binarization from axial boundary rescue along the $z$-axis, the CAL-3DH algorithm eliminated 42.4% of false alarms on tumor-free tissue while recovering infiltrative apical margins. Beyond geometric metrics, the extracted 3D volumes demonstrated excellent reproducibility against radiologist assessments (ICC = 0.8894) and preserved statistically significant stratification across WHO-defined molecular consensus clusters ($p = 0.0268$).

By establishing a transparent, leakage-audited, patient-grouped benchmark on the complete TCGA-LGG cohort, CAL-3DH provides a standardized methodological template and a robust volumetric framework to support surgical planning, radiation oncology, and quantitative neuro-oncology diagnostics.

### CRediT authorship contribution statement

**Shayan Ganji Nataj:** Conceptualization, Methodology, Software, Formal analysis, Investigation, Data curation, Visualization, Writing – original draft. **Keivan Borna:** Supervision, Project administration, Writing – review and editing, Validation.

### Declaration of competing interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.


### Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.


### Data statement

The TCGA-LGG MRI collection is openly accessible via The Cancer Imaging Archive (TCIA) (Clark et al., 2013; Pedano et al., 2016) and mirrored on Kaggle at https://www.kaggle.com/datasets/mateuszbuda/lgg-mri-segmentation. The complete CAL-3DH open-source codebase, patient-grouped 5-fold cross-validation splits, precomputed caching modules, model weights, and leakage auditing scripts are openly hosted on GitHub at https://github.com/Shayan-Ganji/lgg_segmentation.


### Acknowledgements

The authors thank the National Cancer Institute and the TCGA and TCIA consortia for curating and providing open access to the lower-grade glioma imaging and genomic repository.


### Declaration of generative AI and AI-assisted technologies in the manuscript preparation process

During the preparation of this work, the authors used Claude (Anthropic) solely for language editing, grammatical polishing, and code debugging assistance. All experimental designs, mathematical formulations, statistical audits, and scientific interpretations were conducted by the authors. After using these tools, the authors reviewed and edited the content as needed and take full responsibility for the content of the published article.

## References


Abraham, N., Khan, N.M., 2019. A novel focal Tversky loss function with improved attention U-Net for lesion segmentation, in: IEEE International Symposium on Biomedical Imaging (ISBI), IEEE. pp. 683–687. doi:10.1109/ISBI.2019.8759329.

Alhatamleh, S., Abu Mhanna, H.Y., et al., 2026. Attention-enhanced SAM with PBFO tuning: advancing glioma MRI segmentation. Front. Med. 13, 1730353. doi:10.3389/fmed.2026.1730353.

Bland, J.M., Altman, D.G., 1986. Statistical methods for assessing agreement between two methods of clinical measurement. Lancet 327, 307–310. doi:10.1016/S0140-6736(86)90837-8.

Brier, G.W., 1950. Verification of forecasts expressed in terms of probability. Mon. Weather Rev. 78, 1–3. doi:10.1175/1520-0493(1950)078<0001:VOFEIT>2.0.CO;2.

Buda, M., Saha, A., Mazurowski, M.A., 2019. Association of genomic subtypes of lower-grade gliomas with shape features automatically extracted by a deep learning algorithm. Comput. Biol. Med. 109, 218–225. doi:10.1016/j.compbiomed.2019.05.002.

Ceccarelli, M., Barthel, F.P., Malta, T.M., Sabedot, T.S., Salama, S.R., Murray, B.A., Morozova, O., Newton, Y., Radenbaugh, A., Pagnotta, S.M., et al., 2016. Molecular profiling reveals biologically discrete subsets and pathways of progression in diffuse glioma. Cell 164, 550–563. doi:10.1016/j.cell.2015.12.028.

Chang, S.G., Yu, B., Vetterli, M., 2000. Adaptive wavelet thresholding for image denoising and compression. IEEE Trans. Image Process. 9, 1532–1546. doi:10.1109/83.862633.

Clark, K., Vendt, B., Smith, K., Freymann, J., Kirby, J., Koppel, P., Moore, S., Phillips, S., Maffitt, D., Paden, M., Kaping, T., Tarbox, L., Bennett, F., Carey, B., Jarosz, F., Gierada, D., Henderson, S., Valk, J., Sharma, A., Marks, G., Foltz, C., Miller, D., O'Donnell, B., Farhat, M.A., Bernstein, J., Smith, M., Vance, G., Clunie, D., Lu, C., Rubin, D., Freymann, J., Erickson, B., 2013. The Cancer Imaging Archive (TCIA): Maintaining and operating a public information repository. J. Digit. Imaging 26, 1045–1057. doi:10.1007/s10278-013-9622-7.

Efron, B., Tibshirani, R.J., 1994. An Introduction to the Bootstrap. CRC Press, Boca Raton, FL. doi:10.1201/9780429246593.

Guo, C., Pleiss, G., Sun, Y., Weinberger, K.Q., 2017. On calibration of modern neural networks, in: International Conference on Machine Learning (ICML), PMLR. pp. 1321–1330. doi:10.48550/arXiv.1706.04599.

Hendrycks, D., Gimpel, K., 2016. Gaussian error linear units (GELUs). arXiv preprint arXiv:1606.08415 doi:10.48550/arXiv.1606.08415.

Huang, G., Sun, Y., Liu, Z., Sedra, D., Weinberger, K.Q., 2016. Deep networks with stochastic depth, in: European Conference on Computer Vision (ECCV), pp. 646–661. doi:10.1007/978-3-319-46493-0_39.

Koo, T.K., Li, M.Y., 2016. A guideline of selecting and reporting intraclass correlation coefficients for reliability research. J. Chiropr. Med. 15, 155–163. doi:10.1016/j.jcm.2016.02.012.

Lin, T.Y., Goyal, P., Girshick, R., He, K., Dollár, P., 2017. Focal loss for dense object detection, in: IEEE International Conference on Computer Vision (ICCV), pp. 2980–2988. doi:10.1109/ICCV.2017.324.

Loshchilov, I., Hutter, F., 2019. Decoupled weight decay regularization, in: International Conference on Learning Representations (ICLR). doi:10.48550/arXiv.1711.05101.

Louis, D.N., Perry, A., Wesseling, P., Brat, D.J., Cree, I.A., Figarella-Branger, D., Hawkins, C., Ng, H.K., Pfister, S.M., Reifenberger, G., et al., 2021. The 2021 WHO classification of tumors of the central nervous system: a summary. Neuro-Oncol. 23, 1231–1251. doi:10.1093/neuonc/noab106.

Oktay, O., Schlemper, J., Folgoc, L.L., Lee, M., Heinrich, M., Misawa, K., Mori, K., McDonagh, S., Hammerla, N.Y., Kainz, B., Glocker, B., Rueckert, D., 2018. Attention U-Net: Learning where to look for the pancreas. arXiv preprint arXiv:1804.03999 doi:10.48550/arXiv.1804.03999.

Pedano, N., Flanders, A.E., Scarpace, L., Mikkelsen, T., Eschbacher, J.M., Hermes, B., Ostrom, Q., 2016. [dataset] the cancer genome atlas low grade glioma collection (TCGA-LGG) (version 3). The Cancer Imaging Archive. doi:10.7937/K9/TCIA.2016.L4LTD3TK.

Ronneberger, O., Fischer, P., Brox, T., 2015. U-Net: Convolutional networks for biomedical image segmentation, in: Medical Image Computing and Computer-Assisted Intervention (MICCAI), Springer. pp. 234–241. doi:10.1007/978-3-319-24574-4_28.

Salehi, S.S.M., Erdogmus, D., Gholipour, A., 2017. Tversky loss function for image segmentation using 3D fully convolutional deep networks, in: Machine Learning in Medical Imaging (MLMI), Springer. pp. 379–387. doi:10.1007/978-3-319-67389-9_44.

Shrout, P.E., Fleiss, J.L., 1979. Intraclass correlations: uses in assessing rater reliability. Psychol. Bull. 86, 420–428. doi:10.1037/0033-2909.86.2.420.

Wu, Y., He, K., 2018. Group normalization, in: European Conference on Computer Vision (ECCV), pp. 3–19. doi:10.1007/978-3-030-01261-8_1.

Yuan, W., Kang, H., 2025. Enhanced local feature extraction of lite network with scale-invariant CNN for precise segmentation of small brain tumors in MRI. PLoS ONE 20, e0334447. doi:10.1371/journal.pone.0334447.

Zhou, Z., Siddiquee, M.M.R., Tajbakhsh, N., Liang, J., 2018. UNet++: A nested U-Net architecture for medical image segmentation, in: Deep Learning in Medical Image Analysis and Multimodal Learning for Clinical Decision Support (DLMIA), Springer. pp. 3–11. doi:10.1007/978-3-030-00889-5_1.