

Spectral leakage spreads energy into neighboring

frequency bins when the observation window does not

contain an integer number of cycles

A tapered window reduces leakage but also widens the main

lobe.

CALIBRATION PROCEDURE

First remove sensor bias, then collect three reference

readings, and finally store the median correction.

Frequency resolution equals sample rate divided by sample

count