

# Label Proofing Report

Report ID: 202603170001

| CR NUMBER<br>(OPTIONAL) | SKU       | LABEL REVISION | CURRENT VERSION LABEL     | NEW VERSION LABEL         |
|-------------------------|-----------|----------------|---------------------------|---------------------------|
| -                       | 187301111 | Rev-D → Rev-E  | LCN-187301111_1_Rev-D.pdf | LCN-187301111_1_Rev-E.pdf |

## REQUIREMENTS SUMMARY

12 Requirements

| #  | ELEMENT | CHANGE TYPE | REQUIREMENT                                            | EXPECTED VALUE                    | ACTUAL VALUE                      | STATUS  |
|----|---------|-------------|--------------------------------------------------------|-----------------------------------|-----------------------------------|---------|
| 1  | Text    | MODIFIED    | Manufacturing date has to be changed                   | 2025-06-12                        | 2025-06-12                        | Matched |
| 2  | Symbol  | DELETED     | Remove CE mark                                         | CE mark removed                   | CE mark removed                   | Matched |
| 3  | Image   | MODIFIED    | Add background in the size of the implant (11mm, 7)    | Background added (11mm, 7)        | Background added (11mm, 7)        | Matched |
| 4  | Text    | MODIFIED    | Trademark ® change to ™                                | ™                                 | ™                                 | Matched |
| 5  | Symbol  | DELETED     | Remove EC REP symbol                                   | EC REP symbol removed             | EC REP symbol removed             | Matched |
| 6  | Symbol  | DELETED     | EC REP address from labels has to be removed.          | Address Removed                   | Address Removed                   | Matched |
| 7  | Symbol  | ADDED       | Add MR Conditional symbol                              | MR Conditional symbol present     | MR Conditional symbol present     | Matched |
| 8  | Text    | MODIFIED    | Change e-IFU symbol to e-IFU for US/Canada only        | e-IFU for US/Canada only          | e-IFU for US/Canada only          | Matched |
| 9  | Image   | MODIFIED    | Formatting change in the Data matrix                   | Size of data matrix to be changed | Size of data matrix to be changed | Matched |
| 10 | Text    | MODIFIED    | All Revisions change to the next consecutive character | Next consecutive character        | Next consecutive character        | Matched |
| 11 | Text    | MODIFIED    | Formatting change in POLYMER/CARBON FIBER COMPOSITE    | Updated formatting                | Updated formatting                | Matched |
| 12 | Text    | MODIFIED    | Formatting change in the eIFU symbol glossary URL      | Updated formatting                | Updated formatting                | Matched |

CURRENT VERSION LABEL

LCN-187301111\_1\_Rev-D

REF 187301111 LOT SAMPLE QTY 1  
BENGAL SYSTEM  
STANDARD IMPLANT

Modified  
11mm, 7°

IMPLANT STANDARD  
STANDARDIMPLANTAT  
STANDAARDIMPLANTAAT  
IMPIANTO STANDARD  
IMPLANTE ESTÁNDAR  
IMPLANTE PADRÃO  
STANDARDIMPLANTAT  
STANDARDIMPLANTAT  
VAKIOIMPLANTTI  
ΤΥΠΙΚΟ ΕΜΦΥΤΕΥΜΑ  
STANDARD IMPLANTÁTUM  
STANDARDNÍ IMPLANTÁT  
STANDARDOWY IMPLANT

DePuy Spine  
325 Paramount Drive  
Raynham, MA 02767, USA

Modified  
pat. www.depuysynthes.com/patentmarking  
https://www.e-ifu.com/symbols-glossary



(01)10705034137025(20)99



(10)SAMPLE

Deleted Deleted  
EC REP DePuy Ireland UC  
Loughbeg, Ringaskiddy  
Co. Cork, Ireland



Modified  
www.e-ifu.com  
US / Canada 800-255-2500  
EU: +800 8888 2020  
EU: +32 2 4037222



Modified  
REV. D

NEW VERSION LABEL

LCN-187301111\_1\_Rev-E

REF 187301111 LOT SAMPLE QTY 1  
BENGAL SYSTEM  
STANDARD IMPLANT

Modified  
11mm, 7°

IMPLANT STANDARD  
STANDARDIMPLANTAT  
STANDAARDIMPLANTAAT  
IMPIANTO STANDARD  
IMPLANTE ESTÁNDAR  
IMPLANTE PADRÃO  
STANDARDIMPLANTAT  
STANDARDIMPLANTAT  
VAKIOIMPLANTTI  
ΤΥΠΙΚΟ ΕΜΦΥΤΕΥΜΑ  
STANDARD IMPLANTÁTUM  
STANDARDNÍ IMPLANTÁT  
STANDARDOWY IMPLANT

DePuy Spine  
325 Paramount Drive  
Raynham, MA 02767, USA

Modified  
pat. www.depuysynthes.com/patentmarking  
https://www.e-ifu.com/symbols-glossary




(01)10705034137025(20)99



(10)SAMPLE



Modified  
Consult Instructions for Use  
www.e-ifu.com  
US / CANADA 800-255-2500



Modified  
REV. E

## INSPECTION SUMMARY

Total Differences: 12

Deleted: 3

Added: 1

Modified: 8

 Text: 0 Symbol: 3 Barcode: 0 Image: 0 Text: 0 Symbol: 1 Barcode: 0 Image: 0 Text: 6 Symbol: 0 Barcode: 0 Image: 2

## CHANGES MADE

## TEXT 5



| CHANGE TYPE | DETAILS                                                        |
|-------------|----------------------------------------------------------------|
| MODIFIED    | Manufacturing date has been changed                            |
| MODIFIED    | Trademark has been changed to ™                                |
| MODIFIED    | The e-IFU symbol has been changed to e-IFU for US/Canada only. |
| MODIFIED    | The Revisions was changed to the next consecutive character.   |
| MODIFIED    | Formatting change in the eIFU symbol glossary URL              |

## SYMBOL 4



| CHANGE TYPE | DETAILS                                      |
|-------------|----------------------------------------------|
| ADDED       | MR Conditional symbol has been added         |
| DELETED     | CE mark has been removed                     |
| DELETED     | EC REP symbol from labels has been removed.  |
| DELETED     | EC REP address from labels has been removed. |

## IMAGE 2



| CHANGE TYPE | DETAILS                                                        |
|-------------|----------------------------------------------------------------|
| MODIFIED    | Background has been added in the size of the implant (11mm, 7) |
| MODIFIED    | Formatting change has been done in the Data matrix             |