

From Eye to Insight



Leica Application Suite

Calibration Data



Configuration:	Cal - 05.03.2019
Microscope:	DM6 - 488215
Camera:	DMC4500 - 0168350219
Computer:	DESKTOP-58I50HT
Print date:	05-03-2019 14:40:29

Mag. Changer	Objective	Calibration State	Calibration Value	Adjustment
1	5	Measured	0.9765625 μm	1.0046
1	10	Measured	0.48685492 μm	1.0017
1	20	Measured	0.24342746 μm	1.0017

Master Calibration Certificate No.: **P/43/180053**

Acceptance Criteria: **$\pm 5\%$**

Calibration Done On: **05.03.2019**

Calibration Due On: **05.03.2020**

Calibration Done By: **ANAND DEO**

Signature: 

Leica Microsystems - A Division of DHR Holding India Pvt. Ltd., 3rd Floor, B-wing, Art Guild House, Phoenix Market City, LBS Road, Kurla (West), Mumbai - 400 070. Tel.: +91-22-3026 7000 Fax: 022 - 30267199 Email: contact.india@leica-microsystems.com

REGIONAL OFFICES AT:

Chennai - 044 24347008, **Delhi** - 011 66544900, **Lucknow** - 0522 2346535, **Kolkata** - 033 66376150, **Hyderabad** - 040 49145671, **Bangalore** - 080 23230912, **Thiruvananthapuram** - 0471 2324064. www.leica-microsystems.com