

I R TECHNOLOGY SERVICES PVT. LTD.

EL-91, TTC Industrial Area, Electronic Zone

MIDC, Mahape, Navi Mumbai- 400710

Tel. +91-022- 67896600, 27688817, 27611228.

Fax: 022-27681253.GST NO:-27AAACI1256K1ZL

Email: service_bom@irtech.in



CERTIFICATE OF CALIBRATION

NAME & ADDRESS OF CUSTOMER:	VE COMMERCIAL VEHICLES LTD 78-86, Industrial Area-3, A.B.Road, Dewas - 455001.
CALIBRATION CERTIFICATE ISSUED BY:	IR Technology Services Pvt Ltd. EL-91,TTC Industrial Area, Electronic Zone, Mahape, Navi Mumbai - 400710 Tel: 022-67896600/27688817/27681252 FAX: (022) 27681253 E-mail: service_bom@irtech.in Website: www.irtech.in
CERTIFICATE NO :	IRM152PA BX41-31052018
CALIBRATION STANDARD	AS PER ASTM E1951-14 (GUIDELINES)
DETAIL OF INSTRUMENT UNDER CALIBRATION:	OLYMPUS MICROSCOPE : BX41RF (S/N :2L14079) DIGITAL CAMERA : SC-30 (S/N: 57003778) IMAGE ANALYSER : STREAM ESS (S/N: 4ZDYMRTQAR) VER: 1.9.4; BUILD: 12868
DATE OF INSPECTION :	31-05-2018
EQUIPMENT & MASTERS USED FOR CALIBRATION:	STANDARD OBMM STAGE MICROMETER SCALE. REPORT NO : 17/53/01/352 -L/3/204 (AS PER ATTACHED)
DATE OF RECEIPT CALIBRATION PROCEDURE:	31-05-2018
NEXT CALIBRATION DUE:	30-05-2019
CALIBRATION RESULT :	☺ <u>Calibration of IMAGE ANALYSER Software has been calibrated w.r.t. 5X, 10X, 20X,50X ,100X Objectives using standard OBMM stage micrometer scale .</u>

AUTHORISED BY:



Mr. Jagdish Pagar

(I.R. TECHNOLOGY SERVICE PVT. LTD.)

Other Offices:

MUMBAI

BANGLORE

CHENNAI

NOIDA

HYDERABAD

JAMSHEDPUR

KOLKATTA