

Analytical Worksheet: Anion Exchange Chromatography, Percent Full Capsids

Master Batch Record worksheet. Percent full is reported by peak area. Empty capsids elute before full capsids on anion exchange.

1 SAMPLE AND METHOD

Ref	Field	Entry
1.1	Worksheet ID	WS-AEX-26052
1.2	Product / lot	OBX-201 / DS-26052
1.3	Method reference	AEX-HPLC, USP <1047>, SOP-3210
1.4	Test date	15JUL2026
1.5	Instrument (EIN)	HPLC system, EIN HPLC-014
1.6	Analyst	R. Okafor

2 SYSTEM SUITABILITY

Ref	Check	Criterion	Result	Meets criterion
2.1	Resolution, empty to full peak	Not less than 1.5	2.9	Yes
2.2	Percent RSD of replicate percent full (3 injections)	Not more than 5.0 percent	2.1	Yes
2.3	Percent RSD of retention time	Not more than 1.0 percent	0.4	Yes
2.4	Reference standard percent full recovery	90 to 110 percent of assigned	99.1	Yes

3 PEAK TABLE

Ref	Peak	Retention time (min)	Peak area	Percent area
3.1	Empty	4.1	1,890,000	18.9
3.2	Partial	7.0	870,000	8.7
3.3	Full	10.1	7,240,000	72.4
3.4	Total	N/A	10,000,000	100.0

4 RESULT

Ref	Acceptance criterion	Result	Meets criterion
4.1	Percent full capsids not less than 80.0 percent	72.4	No

5 ATTACHMENTS

Ref	Attachment
-----	------------

5.1	AEX chromatogram attached

6 WORKSHEET APPROVALS

Ref	Performed By / Date	Verified By / Date
6.1	R. Okafor 15JUL2026	M. Delgado 16JUL2026