

Analytical Worksheet: Anion Exchange Chromatography, Percent Full Capsids

Master Batch Record worksheet. Percent full is reported by peak area. Empty capsids elute before full capsids on anion exchange.

1 SAMPLE AND METHOD

Ref	Field	Entry
1.1	Worksheet ID	
1.2	Product / lot	
1.3	Method reference	AEX-HPLC, USP <1047>, SOP-3210
1.4	Test date	
1.5	Instrument (EIN)	
1.6	Analyst	

2 SYSTEM SUITABILITY

Ref	Check	Criterion	Result	Meets criterion
2.1	Resolution, empty to full peak	Not less than 1.5		
2.2	Percent RSD of replicate percent full (3 injections)	Not more than 5.0 percent		
2.3	Percent RSD of retention time	Not more than 1.0 percent		
2.4	Reference standard percent full recovery	90 to 110 percent of assigned		

3 PEAK TABLE

Ref	Peak	Retention time (min)	Peak area	Percent area
3.1	Empty			
3.2	Partial			
3.3	Full			
3.4	Total	N/A		

4 RESULT

Ref	Acceptance criterion	Result	Meets criterion
4.1	Percent full capsids not less than 80.0 percent		

5 ATTACHMENTS

Ref	Attachment
5.1	
6 WORKSHEET APPROVALS	

Ref	Performed By / Date	Verified By / Date
6.1		