



Date: 06/Aug/2024

## CERTIFICATE OF CALIBRATION

### VISCOMETER

MODEL NUMBER : Model 800 (OFITE)

SPRING : F1

SERIAL NUMBER: 12-4779-001

ROTOR: R1

CUSTOMER: Sava Energy Parseh

BOB : B1

This instrument has been calibrated using the procedure of API 13A , Appendix J ,and meets or exceeds the standard therein.

Calibration Fluid: 20 cP

Batch No : 904402

Exp. Date: 03/22/2025

Calibration Fluid: 50 cP

Batch No : 807918

Exp. Date: 12/18/2025

| RPM      | cP | TEMP,C | cP at Temperature | MEASURED cP<br>After Calibration(as left) | DEVIATION<br>Cp* |
|----------|----|--------|-------------------|---|------------------|
| 300      | 20 | 24.3   | 19.2              | 19.6                                      | 0.4              |
| 600/1.98 | 20 | 24.3   | 19.2              | 19.5                                      | 0.3              |
| 300      | 50 | 26.6   | 48                | 48.7                                      | 0.7              |
| 600/1.98 | 50 | 26.6   | 48                | 48.9                                      | 0.9              |

\*Record the deviation as +or - , The measured cP must not exceed+/- 1.5 cP of the standard cP .

  
**Hossein Seifi Nejad**  
Calibration Technician

Address: Amirkabir Technology Tower , NO.7 , Balavar Alley , Valiasr intersection , Tehran , Iran

Phone : (+98) 21 6640 3501    nfo@DiacoEnergy.com    www.DiacoEnergy.com