Magnetic Particle Imaging: A Survey

www.surveyx.cn

Abstract

Magnetic Particle Imaging (MPI) is a cutting-edge imaging modality that leverages the nonlinear magnetic response of superparamagnetic iron oxide nanoparticles (SPIONs) to produce high-resolution images without ionizing radiation. This survey paper provides a comprehensive overview of MPI, highlighting its core principles, such as magnetic field manipulation for spatial encoding and the use of specialized tracer materials. The paper delves into the technical advancements in hardware development and sophisticated image reconstruction algorithms, which are pivotal for enhancing image quality and accuracy. MPI's unique capabilities, including its high sensitivity and real-time imaging potential, make it particularly suitable for diverse biomedical applications, such as cardiovascular imaging and cancer detection. The integration of machine learning and advanced signal processing techniques further augments MPI's diagnostic and therapeutic potential. Despite its promising attributes, MPI faces challenges in clinical adoption, necessitating ongoing research to optimize tracer materials and improve spatial resolution and sensitivity. The paper concludes by exploring the future prospects of MPI in medical diagnostics, emphasizing its transformative potential in non-invasive imaging. As research progresses, MPI is poised to become a crucial tool in medical diagnostics and therapeutic interventions, offering novel possibilities for precise imaging solutions. This structured survey ensures a thorough understanding of MPI's capabilities and its potential impact on the future of biomedical imaging.

1 Introduction

1.1 Overview of Magnetic Particle Imaging (MPI)

Magnetic Particle Imaging (MPI) is an innovative tracer-based imaging modality that quantifies the spatial distribution of magnetic nanoparticles, particularly superparamagnetic iron oxide (SPIO) nanoparticles, making it a promising tool for various biomedical applications [1]. This technique has gained attention for its capacity to generate high-resolution images through the nonlinear magnetic response of SPIO nanoparticles, enabling precise visualization of their distribution within biological systems [2]. Unlike traditional imaging techniques, MPI operates without ionizing radiation, thus offering a safer alternative for patients and facilitating repeated imaging sessions without associated risks.

The fundamental principle of MPI involves applying dynamic magnetic fields to elicit a response from SPIO nanoparticles, which is measured to create high-resolution images. This method allows for positive contrast visualization, effectively distinguishing nanoparticles from surrounding biological tissues. The high sensitivity and rapid imaging capabilities of MPI render it particularly suitable for real-time imaging and dynamic studies, such as monitoring changes in cerebral blood volume during functional imaging and tracking the distribution of therapeutic agents, thanks to its ability to deliver high-contrast images devoid of background noise and signal attenuation in deep tissue [3, 4].

Despite its promising attributes, MPI remains in the early stages of clinical adoption, with ongoing research addressing technical challenges and expanding its application range [5]. The advancement of engineered nanoscale magnetic materials further enhances MPI's potential, contributing to its rapid



Figure 1: chapter structure

evolution in healthcare [6]. As research progresses, MPI is poised to become an essential tool in medical diagnostics and therapeutic interventions, presenting novel possibilities for non-invasive and precise imaging.

1.2 Significance and Impact on Biomedical Imaging

Magnetic Particle Imaging (MPI) represents a transformative modality in biomedical imaging, offering distinct advantages over established techniques such as computed tomography (CT), magnetic resonance imaging (MRI), positron emission tomography (PET), and single-photon emission computed tomography (SPECT) [7]. Its remarkable sensitivity and resolution, coupled with real-time imaging capabilities, make MPI invaluable for applications including vascular imaging and targeted drug delivery [8]. Operating without ionizing radiation, MPI minimizes patient exposure and reduces the need for nephrotoxic contrast agents, making it a safer option for frequent imaging [8].

In cancer diagnostics, MPI has shown significant promise for sensitive detection, staging, and monitoring, facilitating the tracking of cell populations and the assessment of functional brain physiology, thereby enhancing comprehensive disease management [9]. The integration of MPI with magnetic fluid hyperthermia (MFH) exemplifies its theranostic potential, enabling simultaneous imaging and localized treatment of cancer, which improves both diagnostic and therapeutic outcomes [10].

Recent technological advancements, such as single-sided field-free line (FFL) generators, have enabled multidimensional imaging with an expanded field of view, addressing previous limitations regarding object size and data acquisition [6]. The emergence of calibration-free multi-color MPI techniques, which differentiate nanoparticles based on relaxation responses, further emphasizes MPI's capabilities in functional imaging without requiring prior calibration [5].

Optimizing magnetic nanoparticles specifically for MPI, as opposed to those designed for MRI, is crucial for clinical translation, given that many MRI-approved nanoparticles may lack the necessary magnetic properties for effective MPI signal generation [11]. Advanced MPI systems have demonstrated the ability to achieve a 1D field of view of 97.6 mm at a magnetic gradient of 2.5 T/m/±0, underscoring its scalability and potential for future human applications [12].

MPI's unique capabilities enhance tumor targeting and imaging resolution, yielding superior outcomes compared to traditional imaging methods. Continuous advancements in signal processing methods, such as TranSMS, significantly improve the recovery of the system matrix and image reconstruction quality, highlighting MPI's evolving role in biomedical imaging [13]. The technique's capacity to directly measure cerebral blood volume (CBV) changes provides greater sensitivity and accuracy than conventional fMRI methods [14]. Moreover, addressing the challenge of accurately obtaining spatial distribution information of magnetic nanoparticles (MNPs) from their time-varying magnetization response enhances MPI's precision [2]. The proposed method for perfusion imaging without increasing the iron dose further enhances patient safety and allows for prolonged monitoring periods [1]. As research progresses, MPI is set to significantly improve medical diagnostics and treatment strategies.

1.3 Structure of the Survey

This survey is systematically organized to provide an in-depth exploration of Magnetic Particle Imaging (MPI), a pioneering imaging modality with substantial implications for biomedical applications. It begins with an introduction to MPI, elucidating its principles, significance, and potential impact on biomedical imaging. Following this foundation, the survey presents a comprehensive analysis of the background and essential principles of MPI, emphasizing key concepts such as magnetic field manipulation and the role of superparamagnetic nanoparticles as imaging tracers, which enable high-resolution, real-time 3D imaging without radiation exposure. This section also highlights MPI's diverse applications, including cell tracking, blood pool imaging, tumor detection, and visualized magnetic hyperthermia, underscoring its potential to enhance medical diagnostics and treatment [15, 16, 17, 18].

Subsequent sections focus on the imaging principles of MPI, detailing tracer material visualization and techniques for signal detection and noise reduction. The survey transitions to hardware development, discussing advancements and challenges in generating and detecting magnetic signals, complemented by an exploration of specialized tracer materials that optimize magnetic properties and enhance imaging quality.

The discussion then shifts to advanced image reconstruction algorithms, particularly the integration of sophisticated techniques such as deep learning and variational Bayesian inference to address the complexities of converting signal data into high-resolution images. Recent developments, including the use of a sparsity-driven observer for optimizing imaging hardware and the implementation of a deep plug-and-play prior in MPI, are highlighted for their contributions to reconstruction accuracy and efficiency. The survey also addresses calibration challenges and system imperfections in MPI, proposing innovative solutions that significantly enhance image quality and processing speed [19, 20, 13, 21].

Finally, the survey analyzes diverse biomedical applications, showcasing MPI's advantages in areas such as cardiovascular imaging and cancer detection, while examining the role of advanced signal processing techniques in improving image quality and accuracy. The conclusion reflects on the current

state and future prospects of MPI in biomedical imaging, discussing evaluation metrics, performance assessment, and potential avenues for future research and development. This structured approach facilitates a comprehensive exploration of MPI, emphasizing its advanced capabilities in high-contrast, real-time imaging of magnetic nanoparticles and its significant potential to revolutionize medical diagnostics and therapeutic interventions, particularly in cancer imaging, stem cell tracking, and targeted drug delivery [21, 15, 22, 23].The following sections are organized as shown in Figure 1.

2 Background and Core Concepts

2.1 Fundamental Principles of MPI

Magnetic Particle Imaging (MPI) is a state-of-the-art imaging modality that leverages the magnetic properties of superparamagnetic iron oxide nanoparticles (SPIONs) to achieve high-resolution, quantitative imaging without ionizing radiation [15]. The technique exploits the nonlinear magnetic response of SPIONs under oscillating external magnetic fields, enabling precise spatial visualization [24]. This characteristic is particularly advantageous for high-sensitivity applications and rapid imaging, such as real-time monitoring of cerebral blood volume changes and cellular dynamics [14].

Spatial encoding in MPI is accomplished through magnetic field manipulation, employing a selection field (SF) to establish a field-free region (FFR) where the magnetic response is most sensitive to field variations [25]. This setup enhances MPI's ability to detect low iron concentrations, facilitating physiological imaging without radiation risks [2]. Moreover, MPI's dynamic forward model incorporates time-dependent tracer concentrations, accurately representing dynamic behaviors during imaging sessions.

A significant advantage of MPI is its quantitative assessment of SPIO concentrations, which correlate with physiological parameters like cerebral blood volume changes [14]. This quantitative capability is crucial for enhancing disease diagnosis and treatment, especially in oncology, where precise imaging is essential. Advanced mathematical models, such as the Radon transform, enable accurate reconstruction of nanoparticle distributions from induced signals [25].

However, MPI faces challenges, including inconsistent data representation and evaluation metrics, complicating inter-study comparisons. Reduced sensitivity for human imaging limits the detection of small anatomical structures and tracer samples in vivo. Addressing these challenges through ongoing research is vital for advancing MPI technology in clinical contexts, particularly in overcoming limitations of existing imaging modalities in providing accurate, real-time, and non-invasive visualization of biological processes relevant to disease diagnosis and treatment monitoring [2].

2.2 Magnetic Field Manipulation for Spatial Encoding

Magnetic field manipulation is essential for spatial encoding in MPI, enabling precise localization and quantification of magnetic nanoparticle (MNP) distributions. This process uses varying magnetic fields to induce distinct MNP responses, facilitating differentiation and quantification [26]. Spatial encoding techniques have advanced to meet the demand for enhanced resolution and sensitivity, incorporating innovations like shifting fields and drive-field rotations [27].

A notable challenge in spatial encoding is the sensitivity drop with larger coil bore diameters, affecting low tracer concentration detection [28]. Techniques like Moving Table MPI (MT-MPI) allow data acquisition from multiple spatial positions without compromising temporal resolution, enabling accurate reconstructions of dynamic processes [3]. Lissajous scanning optimizes thermal effects on targeted tumor cells while minimizing damage to adjacent tissues [10].

Theoretical advancements, including Hilbert space analysis, provide a structured framework for understanding the MPI operator and addressing its ill-posedness, guiding the development of more robust spatial encoding techniques [29]. However, existing cylindrical MPI devices face challenges related to power consumption and geometrical constraints, hindering clinical application [6]. Current methods have yet to effectively analyze the influence of varying angles between drive and selection magnetic fields on signal harmonics, adding complexity to spatial encoding optimization [24].

Despite these challenges, magnetic field manipulation for spatial encoding in MPI remains an active research area, propelled by advancements such as deep learning methodologies and novel reconstruction algorithms that enhance image quality and processing speed [21, 15, 30]. Continued

innovation in magnetic field manipulation strategies is vital for overcoming existing limitations and improving MPI's clinical applicability.

2.3 Superparamagnetic Tracer Materials

Superparamagnetic tracer materials are crucial for MPI's efficacy, offering unique magnetic properties that enable high-contrast and high-sensitivity imaging. SPIONs generate a distinct, background-free signal that enhances image contrast and sensitivity, setting MPI apart from conventional imaging modalities like MRI and CT. This capability allows effective visualization of SPION distribution and concentration within biological tissues, facilitating high-resolution imaging without ionizing radiation [9, 4]. The magnetic behavior of SPIONs under external magnetic fields allows precise localization and quantification of tracer distributions, critical for accurate imaging. The selection and design of tracer materials significantly affect MPI systems' sensitivity and spatial resolution, particularly in measuring concentrations in blood vessels, thus enhancing diagnostic capabilities by quantifying stenosis.

The structural configuration of tracer materials, whether single-core or multi-core, plays a pivotal role in optimizing MPI systems for various biomedical applications. Functionalization for specific targeting, such as tumor localization, enhances tomographic image quality and enables targeted imaging approaches. Challenges persist, such as the inadequate sensitivity of current MPI systems to detect low SPION concentrations, crucial for numerous biomedical applications. The spatial distribution of MNPs is vital for applications like magnetic hyperthermia treatment (MHT), where localization is essential for effective treatment planning. Balancing tracer concentration while minimizing iron dosage poses significant challenges, necessitating careful optimization to ensure efficacy and safety [1]. Ongoing research and development in superparamagnetic tracer materials are crucial for advancing MPI technology and expanding its clinical applications.

2.4 Oscillating Magnetic Fields

Oscillating magnetic fields are fundamental to MPI, providing the dynamic environment necessary for the excitation and detection of superparamagnetic nanoparticles. These fields generate the nonlinear magnetic response of MNPs, crucial for achieving high-resolution imaging [2]. The field-free point (FFP) created by oscillating magnetic fields is particularly sensitive to high-frequency changes, facilitating accurate image reconstruction by leveraging the nonlinear magnetization response of MNPs, essential for differentiating and localizing tracer distributions within biological tissues.

The strategic manipulation of oscillating magnetic fields enhances MPI's spatial encoding capabilities. Techniques such as the Magnetic Particle Imaging/Navigation (MPIN) method utilize alternating imaging and navigation modes to control nanoparticle movement in fluid flows, improving tracer localization precision [31]. Additionally, the field-free line encoding scheme used in iMPI scanners enables real-time visualization of vascular structures, showcasing oscillating fields' ability to provide continuous and detailed imaging [8].

Advanced methodologies, including the integration of oscillating fields with stimulated emission depletion (STED) techniques, further enhance spatial resolution. The STED-MPI approach uses a donut-shaped focal spot to deplete signal intensity at the center, improving detection sensitivity and resolution [32]. Moreover, the interaction of oscillating magnetic fields with MNP relaxation behavior is critical for optimizing imaging processes, ensuring accurate representation of nanoparticle distributions [33].

Challenges such as direct feedthrough interference and isotropic resolution limitations persist, necessitating ongoing research to enhance hardware resolution and detection sensitivity [9]. The application of deep learning methods, such as the 3D-System Matrix Recovery Network (3D-SMRnet), offers promising solutions by leveraging complex data patterns to improve system matrix recovery and overall imaging quality [34]. Combining zero-shot denoising with 1-regularization effectively addresses noise and ill-conditioning in MPI data, promoting sparsity in reconstruction and enhancing image clarity [35].

Continuous advancement of oscillating magnetic fields, alongside cutting-edge computational techniques, is vital for addressing current challenges in MPI and broadening its clinical applications. Initiatives like the open-source project OS-MPI aim to democratize access to MPI technology by

providing detailed designs and resources for pre-clinical imaging systems, fostering innovation and collaboration within the research community. Novel image reconstruction methods for projection-based MPI have shown promise in accelerating imaging processes while maintaining high image quality, further supporting the expansion of MPI's clinical utility [15, 22]. These fields lay the groundwork for precise and high-resolution imaging capabilities, underscoring their significance in advancing MPI technology.

3 Imaging Principles

3.1 Visualization of Tracer Materials

Magnetic Particle Imaging (MPI) excels in visualizing superparamagnetic iron oxide nanoparticles (SPIONs) by leveraging their magnetic properties and nonlinear responses to external fields, crucial for real-time tumor imaging and vascular structure monitoring [8, 9]. MPI's high contrast and non-invasive nature enhance clinical diagnostics by capturing dynamic biological processes. The relaxation behavior of magnetic nanoparticles (MNPs) is pivotal in MPI visualization. Calibration-free relaxation-based multicolor MPI estimates relaxation time constants using signal symmetry during bidirectional scans, enhancing tracer visualization accuracy [5]. Accurate modeling of Neel rotation is essential for effective image reconstruction, impacting MNP magnetization behavior [36]. Lock-in amplifier models analyze signal harmonics under varying magnetic field orientations, providing insights into SPION dynamics [24]. Reconstructing three-dimensional particle distributions from 3D field-free line (FFL) scans addresses complex tracer distribution challenges [25]. Techniques like using physiological saline boluses create negative contrast, enhancing perfusion parameter measurement [1]. Integrating MPI with intravascular optical coherence tomography (IVOCT) improves catheter trajectory estimation, expanding vascular imaging applications [37]. MPI's high-resolution, high-sensitivity imaging, especially in real-time applications like vascular imaging and cell tracking, underscores its transformative potential in biomedical imaging [7]. Advancing visualization techniques is crucial for broadening MPI's clinical applicability, ensuring precise, non-invasive imaging solutions.



(a) A 3D rendering of a mechanical device with a cylindrical component and a rectangular box on a table[22]

(b) Comparison of Phantom and Adapted Window Results[38]

(c) The image depicts a flowchart illustrating the process of particle excitation and saturation in a fluid flow system.[39]

Figure 2: Examples of Visualization of Tracer Materials

As depicted in Figure 2, advanced visualization techniques enhance understanding of complex systems. The first example illustrates mechanical device visualization complexity. The second example compares "Phantom" and "Adapted Window" imaging techniques, aiding anatomical feature distinction. The third example presents a flowchart of particle excitation and saturation in fluid dynamics, showcasing graphical elements for dynamic processes. These examples highlight diverse tracer material visualization methodologies, enriching comprehension across scientific and engineering domains.

3.2 Signal Detection and Background Noise Elimination

Enhancing signal detection and background noise elimination is vital for optimizing Magnetic Particle Imaging (MPI) performance and accuracy. Lock-in amplifiers significantly improve MPI system

sensitivity by enhancing signal sensitivity and reducing noise interference, analyzing induced MNP signals in receiving coils [40, 24]. MPI's enhanced permeability and retention (EPR) effect aids early-stage cancer diagnosis, providing high sensitivity and resolution for detecting subtle tracer distribution changes [16]. The Signal Detection Operator (SDO) optimizes signal detection by computing a test statistic that ranks data-acquisition designs, offering insights applicable to MPI [20]. Addressing image blurriness involves blind deconvolution techniques to enhance resolution [41]. Magnetic particle spectroscopy (MPS) at various high frequencies is crucial for monitoring background noise, ensuring accurate signal detection [42]. MPI system performance is further enhanced by analyzing learned posterior distributions and uncertainty quantification, improving signal detection robustness [43]. However, existing noise reduction methods often compromise image quality or require high computational resources, limiting real-time applicability [44]. Integrating advanced signal processing techniques with innovative computational models is essential for overcoming challenges in signal detection and noise reduction, especially with diverse particle concentrations. Current MPI systems face dynamic range limitations, where highly concentrated signals can obscure nearby regions with lower concentrations. Recent advancements include a two-step algorithm enhancing dynamic range by a factor of four and a plug-and-play approach utilizing zero-shot denoising, improving reconstruction stability and spatial resolution [38, 21, 22, 35]. These efforts refine MPI's imaging capabilities, ensuring precise and reliable diagnostic outcomes.

3.3 Challenges in Multi-Sample Imaging

Multi-sample imaging in Magnetic Particle Imaging (MPI) presents significant challenges, particularly with samples exhibiting varying iron concentrations. The shadowing effect, where high iron concentrations obscure lower concentrations, complicates accurate tracer distribution visualization and quantification, essential for precise measurements in complex biological environments [38]. As illustrated in Figure 3, these primary challenges are visually represented, emphasizing the shadowing effect, design considerations for human use, and the handling of dynamic tracer concentrations, along with references to key studies. Advancing MPI scanners for human use requires design improvements to accommodate diverse clinical imaging requirements, including handling a wide range of tracer concentrations without sacrificing image quality [7]. Optimizing contrast agents for MPI enhances sensitivity and resolution, enabling effective differentiation between samples with varying iron content. Addressing imaging dynamic tracer concentrations necessitates integrating MPI with sophisticated computational models capable of managing wide dynamic range tracer concentrations. Recent research focuses on developing new MPI forward models accommodating dynamic tracer distributions, enhancing image reconstruction by overcoming traditional static model limitations. Innovative algorithms improve dynamic range and spatial resolution, allowing clearer differentiation between high and low tracer concentration areas, particularly in clinical applications like cancer imaging [45, 38, 46]. Innovations in hardware development and signal processing are crucial for mitigating the shadowing effect and improving multi-sample imaging accuracy, pivotal for expanding MPI's clinical applicability and ensuring reliable imaging outcomes across diverse biomedical scenarios.



Figure 3: This figure illustrates the primary challenges in multi-sample imaging within Magnetic Particle Imaging (MPI), highlighting the shadowing effect, design considerations for human use, and the handling of dynamic tracer concentrations, with references to key studies.

3.4 Dynamic Tracer Distribution and Quantification

Dynamic tracer distribution and quantification in Magnetic Particle Imaging (MPI) are crucial for accurately representing spatial and temporal tracer concentration changes, vital for various biomedical applications. A significant challenge is simultaneous imaging of samples with differing iron concentrations, leading to inaccuracies due to the shadowing effect [38]. A proposed two-step reconstruction algorithm effectively separates high and low concentration signals, enhancing tracer distribution visualization and quantification across samples [38]. Efficient MPI system calibration is addressed by a novel clustering approach balancing calibration time with image artifacts, enabling fewer calibration measurements without compromising image quality, facilitating accurate dynamic tracer distribution quantification [19]. This optimization reduces time and resource requirements while minimizing potential artifacts affecting dynamic tracer behavior interpretation. Advancements in reconstruction algorithms and calibration techniques are crucial for enhancing dynamic tracer distribution quantification accuracy and reliability in MPI. Innovations include a two-step algorithm increasing MPI's dynamic range by a factor of four, improving imaging of samples with varying concentrations. A deep plug-and-play prior (PP-MPI) within a model-based iterative optimization framework outperforms traditional methods, achieving higher peak signal-to-noise ratios and enabling real-time 3D imaging. Generalized multi-patch reconstruction techniques are optimized for non-ideal field conditions, reducing calibration time while minimizing artifacts. These advancements address inherent MPI challenges, like signal clipping and the ill-posed nature of the imaging operator, facilitating precise tracer distribution quantification in complex dynamic environments [19, 38, 21, 30]. Continued research and development are essential for expanding MPI's clinical utility, ensuring precise, real-time insights into complex biological processes.

In recent years, the field of Magnetic Particle Imaging (MPI) has experienced significant advancements, particularly in the development of its hardware. A comprehensive understanding of this progress is essential for identifying both the challenges and solutions that have emerged. Figure 4 illustrates the hierarchical structure of hardware development in MPI, highlighting key challenges, innovative solutions, open-source and modular approaches, as well as advancements in signal detection and sensitivity enhancements. This visual representation not only contextualizes the advancements discussed but also serves as a roadmap for future research directions in the field. By analyzing this structure, researchers can better appreciate the interconnectedness of various components and the innovative strategies that have been employed to overcome existing limitations.

4 Hardware Development

4.1 Challenges and Innovations in MPI

Developing Magnetic Particle Imaging (MPI) hardware involves overcoming significant challenges while fostering innovations that enhance clinical applicability and imaging performance. Transitioning from preclinical to clinical systems requires addressing technical issues like bore size and scalability for human use [8]. Current systems often lack the necessary size and capabilities for real-time visualization, crucial for human interventions [9]. A key obstacle is the inefficient and time-consuming calibration of the system matrix (SM), vital for precise image reconstruction [13]. Innovations such as using solid spherical harmonics for magnetic field representations improve reconstruction accuracy and reduce calibration time [47, 25].

As illustrated in Figure 5, the primary challenges and innovative solutions in the development of MPI are categorized into technical challenges, innovative solutions, and clinical challenges. Advancements in signal detection and noise reduction, including analyzing signal harmonics under various angular conditions, significantly enhance MPI image quality [24]. Methods that accelerate projection-based imaging improve image quality and noise reduction, especially at low projection numbers [15]. Interactive web-based MPI simulation platforms aid in optimizing MPI systems by enhancing user understanding and supporting real-time experimentation [2]. Using saline for negative contrast in MPI allows longer monitoring times without increasing the iron dose, enhancing patient safety [1].

Despite advancements, challenges remain, particularly in improving MPI sensitivity for detecting subtle cerebral blood volume changes compared to fMRI, crucial for identifying hemodynamic fluctuations [14]. Continued research is essential for overcoming these challenges, advancing MPI



Figure 4: This figure illustrates the hierarchical structure of hardware development in Magnetic Particle Imaging (MPI), highlighting key challenges, innovative solutions, open-source and modular approaches, as well as advancements in signal detection and sensitivity enhancements.

technology, and expanding its clinical applications, especially in high-sensitivity areas like tumor imaging.



Figure 5: This figure illustrates the primary challenges and innovative solutions in the development of Magnetic Particle Imaging (MPI). It categorizes the technical challenges, innovative solutions, and clinical challenges faced in advancing MPI technology.

4.2 Open-Source and Modular Approaches

Open-source and modular approaches in MPI hardware development represent a transformative shift towards accessible and customizable imaging solutions. These methods enable rapid prototyping and

iterative refinement, allowing researchers to tailor hardware configurations for various applications, such as cancer imaging and stem cell tracking. Incorporating advanced techniques like deep learning and open-source design resources enhances MPI systems' efficiency and adaptability, improving real-time imaging and image quality while reducing complexity and costs [15, 2, 21, 22, 23]. The open-source paradigm fosters collaboration across the scientific community, accelerating innovation and broadening MPI technology accessibility.

Modular design principles address limitations of traditional MPI systems, such as fixed configurations and high costs associated with bespoke hardware development. Implementing a modular framework allows researchers to interchange or upgrade individual components, enhancing flexibility and scalability. This approach facilitates the integration of specialized software packages tailored to specific research needs and the incorporation of third-party tools through dedicated interfaces, crucial for addressing MPI's complex requirements [22, 48, 17]. This adaptability is vital for tailoring systems to diverse biomedical applications, from small-scale preclinical studies to larger clinical implementations.

Open-source platforms play a critical role in advancing simulation tools and software frameworks essential for enhancing MPI performance. These platforms enable the development of flexible and modular simulation frameworks that accurately emulate the entire MPI experimental process, from magnetic field and particle dynamics to signal generation and data processing. By providing a comprehensive suite of interconnected software packages, open-source initiatives streamline workflows and lower barriers for new researchers, fostering innovation and collaboration within the MPI community [2, 21, 48, 22, 23]. These tools allow virtual testing of hardware configurations and imaging protocols, reducing the need for costly physical prototypes. Furthermore, open-source initiatives can standardize MPI components and protocols, facilitating interoperability and promoting broader technology adoption.

The increasing adoption of open-source and modular strategies in MPI hardware development is expected to significantly enhance technology adaptability, affordability, and collaborative opportunities. By providing detailed design specifications and fostering community-driven development through platforms like GitHub, initiatives such as OS-MPI lower barriers for researchers, enabling the construction of custom imaging systems tailored to specific needs. The introduction of flexible simulation frameworks and tools allows seamless integration of various software components, enhancing MPI research and applications across diverse fields [2, 21, 48, 22]. These strategies are instrumental in addressing current challenges and advancing the clinical translation of MPI, ensuring its position at the forefront of non-invasive imaging innovations.

4.3 Signal Detection and Sensitivity Enhancements

Advancements in MPI signal detection and sensitivity are crucial for improving imaging accuracy and reliability. One significant development involves using vector modulators to create compensation signals that actively eliminate direct feedthrough from the received MPI signal in the analog domain, enhancing signal clarity by isolating the magnetic nanoparticle response from background noise [49]. Specialized hardware design, such as a head surface coil for mouse brain imaging, improves the detection of small vessels and anatomical structures, essential for detailed preclinical imaging [28]. A gradiometric receive coil significantly reduces background noise, enhancing sensitivity to detect iron concentrations as low as 5 ng, crucial for accurate tracer quantification [50].

Innovative methods, including amplitude modulation with superconductor coils, achieve high magnetic gradients and a feasible field of view for human applications, addressing the challenge of maintaining high sensitivity in scaled-up MPI systems [12]. Furthermore, using a configuration of selection coils to create static and dynamic field-free lines (FFL) enhances imaging capabilities by providing precise spatial encoding [6].

The integration of deep learning techniques, such as the TranSMS method, enhances signal detection through super-resolution of low-resolution system matrix measurements. This hybrid model, combining convolutional and transformer modules, improves image reconstruction quality and signal clarity [13]. Applying solid harmonic expansions for magnetic field representation increases accuracy and computational efficiency, facilitating more precise and sensitive imaging [47].

Recent advancements in signal detection and sensitivity within MPI technology are crucial for enhancing imaging capabilities. These improvements expand the dynamic range of MPI, enabling

simultaneous imaging of samples with widely varying concentrations of superparamagnetic nanoparticles and allowing for high-resolution tumor imaging with minimal noise interference. By employing innovative algorithms and adaptive regularization techniques, MPI effectively differentiates between high and low nanoparticle concentration areas, maintaining its status as a premier non-invasive imaging modality for high-resolution biomedical applications, particularly in cancer diagnostics and treatment monitoring [38, 15, 45, 21].

5 Tracer Materials

5.1 Characteristics of Tracer Materials

Magnetic Particle Imaging (MPI) relies critically on the properties of tracer materials, predominantly superparamagnetic iron oxide nanoparticles (SPIONs). These nanoparticles exhibit superparamagnetic behavior, essential for producing strong magnetic responses under external fields, thus enabling high-resolution imaging [42]. Their nonlinear magnetization is leveraged by techniques like PoCT-MPI for enhanced imaging outcomes [51]. Figure 6 illustrates the key characteristics and considerations of tracer materials in MPI, focusing on the properties of SPIONs, their biocompatibility and safety for human use, and their integration with therapeutic applications.

Biocompatibility and safety are paramount in tracer development, particularly for human use. Current clinical tracer limitations necessitate the creation of specialized SPIONs that balance safety with optimal imaging performance [11]. Advances in synthesis have improved SPION biocompatibility and imaging capabilities, expanding diagnostic potential and MPI's clinical applications [52].

The structural configuration of SPIONs—single-core versus multi-core—affects MPI performance significantly. Single-core SPIONs are advantageous in volumetric assays facilitating wash-free detection, while multi-core variants excel in surface-based assays, each offering specific benefits depending on the imaging context [53]. Critical parameters such as size, zeta potential, saturation magnetization, and magnetic moment dictate their effectiveness as MPI tracers [54].

Integrating therapeutic functions with imaging, such as in magnetic fluid hyperthermia (MFH), requires specialized tracer design. SPIONs serve dual roles in imaging and therapy, enhancing clinical utility. Experiments with varying tracer concentrations in vessel phantoms highlight the impact of tracer characteristics on imaging quality, underscoring the need for precise tracer property control to optimize outcomes [37].



Figure 6: This figure illustrates the key characteristics and considerations of tracer materials in Magnetic Particle Imaging (MPI), focusing on the properties of superparamagnetic iron oxide nanoparticles (SPIONs), their biocompatibility and safety for human use, and their integration with therapeutic applications.

5.2 Optimization of Magnetic Properties

Enhancing MPI performance necessitates optimizing the magnetic properties of SPIONs, focusing on size, surface coating, and magnetic moment, which directly influence MPI sensitivity and resolution [54]. Smaller SPIONs, with higher surface-to-volume ratios, improve magnetic responses under external fields, boosting imaging quality [52].

Surface coatings, such as dextran or silica, improve SPION stability and biocompatibility, enhancing dispersibility in biological environments and preventing aggregation, crucial for maintaining consistent magnetic properties during imaging [52]. Tailored coatings enable SPIONs to target specific biological structures, increasing MPI's diagnostic specificity and effectiveness.

Optimizing the magnetic moment is vital for MPI performance. Enhanced synthesis processes can achieve higher saturation magnetization, improving signal-to-noise ratios and imaging resolution [54]. Techniques like thermal decomposition or co-precipitation significantly impact SPION magnetic properties, enabling production of nanoparticles with tailored characteristics for specific imaging needs.

5.3 Development of New Tracer Materials

Advancements in MPI capabilities are closely tied to the development of new tracer materials, which enhance sensitivity, resolution, and system performance. Research focuses on engineering SPIONs with improved magnetic properties and biocompatibility to meet MPI requirements [54], addressing limitations of existing tracers in terms of magnetic response and stability in biological environments.

Optimizing SPION core size and composition is crucial for achieving higher saturation magnetization and improved signal-to-noise ratios [52]. Fine-tuning synthesis processes, such as thermal decomposition and co-precipitation, allows control over particle size distribution and magnetic anisotropy, essential for enhancing imaging resolution and contrast.

Surface modification of SPIONs is key in new tracer development. Coatings like dextran, silica, and biocompatible polymers enhance stability and dispersibility in physiological conditions, enabling functionalization for targeted imaging [52]. These coatings can bind specific biomolecules, facilitating targeted diagnostics and therapeutic monitoring.

Recent developments have led to multifunctional tracers integrating diagnostic and therapeutic capabilities, particularly through SPIONs. These theranostic platforms enable simultaneous imaging and treatment modalities, such as magnetic fluid hyperthermia and drug delivery, enhancing cancer therapy effectiveness while minimizing invasiveness. MPI provides high sensitivity, excellent contrast without background interference, and real-time therapeutic process visualization, improving outcomes in neoplasmic disease management [16, 45, 55]. This approach leverages SPION magnetic properties for visualization and treatment, creating a comprehensive theranostic platform that enhances MPI's clinical utility.

Continuous innovation in tracer materials is vital for advancing MPI technology, ensuring its role as a leading non-invasive, high-resolution imaging modality, particularly in cancer detection and management. MPI's ability to deliver high contrast, sensitivity, and safety—without ionizing radiation—emphasizes the importance of developing tailored SPION tracers. These advancements promise improved diagnostic accuracy and integration of imaging with therapeutic strategies, contributing to precision medicine [56, 45]. Ongoing research is poised to expand MPI applications across diverse clinical scenarios, offering precise and effective diagnostic and therapeutic solutions.

5.4 Challenges in Tracer Material Development

Developing effective tracer materials for MPI involves overcoming significant challenges crucial for advancing clinical and research applications. A primary challenge is potential inaccuracies in MPI-guided methods due to external factors like noise and catheter movement complexity, which can compromise imaging precision [57]. This highlights the need for robust noise reduction strategies, especially when using vector modulators that may introduce additional noise in high-sensitivity applications [49].

The limited diversity of commercially available magnetic nanoparticle (MNP) products is another challenge, with research often focusing on specific suppliers, potentially overlooking broader options [54]. Comprehensive evaluations of various MNPs are necessary to identify those with optimal MPI properties. Additionally, the theoretical understanding of magnetic responses in multi-core superparamagnetic nanoparticles remains incomplete, complicating tracer development for diverse MPI conditions [53].

Existing benchmarks, typically designed for MRI, often fail to address MPI-specific magnetic characteristics, resulting in a lack of suitable tracers for clinical MPI use [11]. The reliance on long-circulating tracers to maintain necessary baseline concentrations for effective imaging is another challenge, as these tracers are not always readily available [1].

Similar to the limitations faced in training data quality for the DeepFold method, MPI tracer development is constrained by the need for high-quality data to inform design and optimization processes [58]. Addressing these challenges requires enhancing theoretical understanding, expanding MNP diversity, and developing benchmarks that accurately reflect MPI demands. This will ensure the creation of effective, versatile tracer materials, paving the way for broader MPI technology adoption and application.

6 Image Reconstruction

6.1 Challenges in MPI Image Reconstruction

Magnetic Particle Imaging (MPI) image reconstruction faces significant challenges impacting image quality and computational efficiency. A primary issue is the ill-posed nature of the MPI inverse problem, where exponential decay of singular values leads to unstable solutions, further exacerbated by signal noise [25]. Calibration, which defines the system function, is time-intensive, requiring multiple scans and frequent recalibrations due to variations in magnetic nanoparticle (MNP) types or scanning trajectories [24]. The behavior of signal harmonics, particularly with arbitrary drive magnetic field angles to the selection field, complicates MPI image modeling and reconstruction [24]. Additionally, dipole-dipole interactions among superparamagnetic iron oxide nanoparticles (SPIONs) pose challenges in accurately modeling magnetic responses for precise image reconstruction [14]. Real-time imaging of complex vascular structures demands further optimization of spatial resolution and sensitivity.

Motion-induced artifacts, especially in projection-based MPI, complicate reconstruction due to variations in projection data from the scanning direction of the field-free line (FFL), causing image blurring. Techniques like blind deconvolution show potential in improving image quality without prior knowledge of the point spread function (PSF), even amidst noise and artifacts [25].

Addressing these challenges requires advancements in computational methods, hardware capabilities, and calibration techniques. Innovative strategies, including deep learning-based TranSMS for super-resolution calibration and the PP-MPI method integrating deep plug-and-play priors, have demonstrated improvements in image reconstruction [15, 21, 13, 22]. Future research should focus on 3D MPI imaging tests of SPIONs to validate system performance and explore design enhancements, ensuring accuracy, speed, and applicability in clinical and research settings.

6.2 Advanced Reconstruction Algorithms

Advancements in image reconstruction algorithms for MPI are crucial for enhancing image quality and computational efficiency. MPI reconstruction methods can be categorized into system matrix-based and x-space approaches. System matrix-based methods focus on accurately modeling the relationship between magnetic particle distribution and the measured signal, often employing sophisticated computational techniques to manage the complexity and size of the system matrix [36]. These methods offer high resolution and sensitivity but demand extensive calibration and computational resources [5].

Conversely, x-space methods simplify reconstruction by directly mapping the measured signal into the spatial domain, resulting in faster computation times and reduced calibration needs, albeit potentially compromising some resolution and sensitivity [29]. Recent hybrid approaches have emerged, combining both methods' strengths and leveraging deep learning techniques to enhance reconstruction performance and address traditional algorithm limitations [13].

Innovative methods like 3d-SMRnet utilize deep learning frameworks to significantly improve matrix quality, reconstructed image quality, and processing time [34]. Integrating wavelet-based sparse Kaczmarz algorithms (SKA) with undecimated wavelet transforms has shown promise in enhancing image reconstruction speed and quality by efficiently managing large datasets and improving convergence rates [5].

The TranSMS method proposes a hybrid architecture combining convolutional neural networks (CNNs) and vision transformers to recover high-resolution system matrices from low-resolution inputs, demonstrating notable improvements in image quality and computational efficiency [13]. The iterative image reconstruction method introduced by [15] incorporates system function and total variation minimization to enhance image quality while reducing the number of projections, presenting a practical solution for improving MPI image reconstruction.

Exploring and developing advanced reconstruction algorithms remain critical for overcoming existing limitations in MPI, ensuring high-quality image outputs, and expanding the clinical applicability of this imaging modality. Future developments will prioritize enhancing documentation practices and delving into advanced algorithmic strategies, including integrating deep learning techniques like the proposed deep plug-and-play prior (PP-MPI) and optimizing reconstruction processes through modern inverse theory approaches.

6.3 System Matrix and x-Space Methods

In MPI, image reconstruction heavily relies on the System Matrix (SM) approach and the x-space method, each presenting unique advantages and challenges. The PP-MPI method enhances reconstruction quality by using a deep plug-and-play prior, achieving a notable increase in peak signal-to-noise ratio compared to traditional techniques and enabling real-time 3D imaging. The proposed data space for off-site image reconstruction addresses logistical challenges in acquiring and managing calibration data, improving the efficiency and accessibility of MPI systems in clinical settings [21, 23].

The System Matrix method is renowned for its high accuracy, derived from comprehensive modeling of the relationship between magnetic nanoparticle distribution and the resultant signal [59]. It involves constructing a detailed system matrix that captures the MPI scanner's response to various spatial positions of nanoparticles. However, the SM method is computationally intensive, requiring substantial resources for calibration and data processing, limiting its applicability in real-time imaging scenarios [59]. Recent advancements, such as the N´eel Rotation Model, enhance SM accuracy by simulating the magnetization behavior of magnetic nanoparticles (MNPs) using the Fokker-Planck equation, providing a more precise representation of the MPI system matrix [36].

To address SM approach computational challenges, methods leveraging the low-rank structure of the MPI imaging problem have been developed, yielding significant computational savings without sacrificing reconstruction quality [30]. Iterative methods like Maximum Likelihood Expectation Maximization (ML-EM) enhance image quality under low signal-to-noise ratio conditions by incorporating statistical properties of the data [60].

In contrast, the x-space method offers a more direct approach by mapping the acquired signal into the spatial domain, advantageous for real-time applications by simplifying reconstruction and reducing extensive calibration needs [59]. While faster, the x-space approach may sacrifice some resolution and accuracy compared to the SM method, necessitating further refinement.

Recent innovations in MPI reconstruction emphasize integrating deep learning techniques to enhance both SM and x-space methods, showing potential in improving reconstruction quality and reducing computational demands [61]. Additionally, developing new 3D MPI reconstruction models incorporating realistic magnetic field topologies allows for algebraic reconstruction of particle concentrations, further advancing MPI capabilities [62].

The ongoing development of system matrix and x-space methods, alongside cutting-edge computational techniques, is crucial for addressing current challenges in MPI image reconstruction. This includes improving calibration data acquisition efficiency, minimizing image artifacts under non-ideal magnetic field conditions, and enhancing dynamic image reconstruction capabilities amid motion. Innovative approaches such as off-site calibration data spaces, generalized multi-patch reconstruction, and deep learning-based regularization methods are being explored to optimize reconstruction processes, achieving higher image quality and faster processing times [21, 63, 23, 19, 30]. These advancements will ensure that MPI remains a leading modality for high-resolution, non-invasive imaging in biomedical applications.

6.4 Deep Learning and Probabilistic Approaches

Deep learning and probabilistic methods have become integral to advancing MPI, significantly enhancing image reconstruction quality and computational efficiency. Frameworks such as the 3d System Matrix Recovery Network (3d-SMRnet) have been employed to recover high-resolution system matrices from low-resolution samples, demonstrating marked improvements in image quality and processing time [34]. These frameworks utilize complex neural network architectures to refine reconstruction algorithms, yielding high-quality images with reduced artifacts and improved resolution.

Probabilistic approaches, including Deep Probabilistic Imaging (DPI), employ flow-based generative models to learn the posterior distribution of reconstructed images, facilitating efficient sampling and uncertainty characterization [43]. This capability is critical for capturing inherent uncertainties in MPI data, leading to more reliable imaging outcomes. Dynamic reconstruction methods like RESESOP-Kaczmarz integrate uncertainties from both measurement data and the model, effectively compensating for motion artifacts during reconstruction [63].

Integrating zero-shot denoisers, which utilize extensive datasets of natural images, has enhanced the robustness of image reconstruction in MPI, allowing for effective generalization across diverse imaging scenarios [35]. Moreover, applying the Maximum Likelihood Expectation Maximization (ML-EM) algorithm has proven effective in generating high-quality images with minimal artifacts, outperforming traditional techniques such as filtered back projection (FBP) [60].

Recent advancements in this area are visually summarized in Figure 7, which illustrates the integration of deep learning and probabilistic methods in Magnetic Particle Imaging (MPI). This figure highlights key methodologies, including deep learning methods such as 3d-SMRnet, ZeroShot 1-PnP, and Deep Image Prior, alongside probabilistic approaches like Deep Probabilistic Imaging, RESESOP-Kaczmarz, and the ML-EM algorithm. Additionally, it outlines future research directions that emphasize refining models, optimizing the N´eel rotation, and improving algorithm initialization.

Future research directions include refining models to incorporate complex interactions, exploring different excitation patterns for improved imaging, and developing joint reconstruction approaches for transfer functions and weight vectors, enhancing applicability in real-time scenarios [64]. Additionally, optimizing the N´eel rotation model by considering particle size distributions and exploring coupled Brownian and N´eel rotation models could further enhance reconstruction accuracy [36].

The Deep Image Prior (DIP) method presents another promising avenue, offering high-quality reconstructions without extensive training data, particularly beneficial in medical applications where such data may be limited [65]. Future research could also focus on refining algorithm initialization processes and incorporating additional regularization techniques to enhance performance [41].

The integration of deep learning and probabilistic methods in MPI signifies a substantial advancement, enhancing both the quality and reliability of imaging outcomes. Continued exploration and development in MPI are essential for improving the robustness and performance of the technology, particularly in addressing the challenges posed by diverse imaging scenarios, such as varying particle concentrations and complex anatomical structures. Recent advancements, including deep learning techniques like the plug-and-play prior (PP-MPI) and adaptive regularization algorithms, have demonstrated significant improvements in image quality and dynamic range, underscoring the importance of ongoing research in this rapidly evolving field [21, 22, 38].

7 Biomedical Applications

Magnetic Particle Imaging (MPI) represents a major advancement in biomedical imaging, offering high-resolution, real-time capabilities that significantly enhance the assessment of complex biological conditions. This section explores MPI's applications in cardiovascular imaging, highlighting its potential to improve diagnostic accuracy and patient care.

7.1 Cardiovascular Imaging

MPI is emerging as a powerful tool for cardiovascular imaging, offering exceptional sensitivity and resolution essential for precise vascular condition assessments. It accurately quantifies vascular stenosis, which is critical for monitoring atherosclerosis by measuring vessel narrowing [66]. The



Figure 7: This figure illustrates the integration of deep learning and probabilistic methods in Magnetic Particle Imaging (MPI), highlighting key methodologies and future research directions. The categorization includes deep learning methods such as 3d-SMRnet, ZeroShot 1-PnP, and Deep Image Prior, and probabilistic approaches like Deep Probabilistic Imaging, RESESOP-Kaczmarz, and the ML-EM algorithm. Future directions emphasize refining models, optimizing N´eel rotation, and improving algorithm initialization.

iMPI system, with its innovative field generator designs, supports real-time imaging in human-sized formats, enhancing MPI's clinical applicability [8]. This technology provides dynamic imaging of cardiovascular events, such as blood flow and perfusion, offering clinicians valuable insights into vascular health.

Pulsed excitation sequences enhance MPI's resolution and sensitivity, allowing detailed visualization of blood vessels and detection of subtle pathological changes [27]. Solid harmonic expansions offer a compact magnetic field representation, improving image reconstruction quality and cardiovascular imaging accuracy [47]. Modeling magnetization dynamics by considering magnetic easy axes orientation reduces calibration times and enhances system matrix estimation accuracy [64]. These advancements solidify MPI's role as a non-invasive imaging modality for cardiovascular applications, promising high-resolution insights into cardiovascular health and disease [38, 3, 15, 21, 59].

7.2 Cancer Imaging and Therapy

MPI is transforming cancer imaging and therapy with its exceptional sensitivity, specificity, and real-time monitoring capabilities. It achieves tumor-to-background ratios of up to 50, crucial for tracking superparamagnetic iron oxide (SPIO) nanoparticle dynamics within tumors [45]. Functionalized nanoparticles, particularly those conjugated with targeting molecules like lactoferrin, significantly enhance MPI's imaging capabilities, improving tumor specificity and resolution [67]. In therapy, MPI induces hyperthermia in targeted tissues, with Lissajous scanning enabling precise thermal ablation and real-time monitoring of cancer cells [10].

Recent technological advancements, including functionalized nanoparticles and innovative scanning techniques, position MPI as a powerful tool for cancer imaging and therapy. Using SPIO nanoparticles as tracers, MPI achieves exceptional image contrast and sensitivity, enabling precise tumor visualization and real-time monitoring, ultimately contributing to effective treatment planning and improved patient outcomes [21, 67, 45].

7.3 Cell Tracking and Functional Imaging

MPI leverages superparamagnetic iron oxide nanoparticles as tracers, providing a non-invasive, radiation-free approach to visualize cellular dynamics and physiological processes in real time. Its high spatial and temporal resolution, positive contrast, and quantitative assessment capabilities make it ideal for cell tracking, vascular imaging, and monitoring treatment responses, enhancing precision medicine integration [3, 15, 56, 4, 16].

SPIONs facilitate precise tracking of labeled cells, offering insights into cell migration, distribution, and viability crucial for regenerative medicine and cancer metastasis. Recent advancements focus on enhancing specificity and sensitivity through novel tracer materials, such as bacterial magneto-

somes, which exhibit strong magnetic properties and biocompatibility [52]. In functional imaging, MPI visualizes dynamic physiological processes with high temporal resolution, advantageous for cardiovascular and neurological disease studies [39, 3, 45, 8].

The integration of advanced tracer materials and innovative imaging techniques enhances MPI's utility in cell tracking and functional imaging, promising significant improvements in clinical and research applications. By optimizing imaging tracers and leveraging innovative reconstruction techniques, these developments aim to enhance diagnostic accuracy and treatment efficacy in precision medicine [56, 21, 16, 38].

7.4 Inflammation and Disease Diagnosis

MPI is a promising modality for diagnosing inflammation and various diseases due to its exceptional sensitivity and specificity in detecting SPIONs within tissues. Its advantages include linear quantitativity, positive contrast, unlimited tissue penetration, and the absence of radiation and background signals, making it effective for cell tracking, blood pool imaging, tumor imaging, and magnetic hyperthermia visualization. MPI achieves sub-millimeter spatial resolution and detects low iron concentrations, facilitating early and accurate disease diagnosis while supporting novel nanomedicine strategies in cancer therapy [67, 16]. It precisely visualizes inflammatory processes, aiding in differentiating inflamed tissues from healthy ones, crucial for conditions like arthritis.

MPI's real-time imaging capability, free from ionizing radiation, offers a safer alternative for frequent monitoring, beneficial for chronic inflammatory disease management. This feature allows repeated imaging sessions to monitor disease progression and evaluate treatment responses, facilitating timely therapeutic adjustments [3, 26, 20, 45].

Beyond inflammation, MPI's high-resolution imaging extends to diagnosing cardiovascular disorders and cancer. Its ability to monitor SPION distribution in targeted tissues enables early detection of pathological changes, crucial for timely interventions. This approach offers high image contrast and sensitivity, allowing clinicians to visualize tumor dynamics and assess treatment responses effectively [4, 45].

Advancements in MPI technology, such as long-circulating SPIO nanoparticles and improved imaging techniques, enhance its diagnostic capabilities. A novel reconstruction method integrates system functions with simultaneous algebraic reconstruction and total variation minimization, reducing noise and artifacts. These developments facilitate high-contrast, high-sensitivity imaging without ionizing radiation, positioning MPI as a promising tool for in vivo cancer detection, capable of visualizing tumor dynamics with tumor-to-background ratios of up to 50 [38, 15, 45]. Targeted SPIONs, functionalized to bind specific biomarkers associated with inflammation and disease, enhance MPI's specificity, allowing for more accurate pathological assessments.

MPI's integration in diagnosing inflammation and diseases represents a significant leap in medical imaging technology. By employing SPIONs as tracers, MPI enables non-invasive imaging with high spatial resolution and sensitivity while avoiding radiation exposure. Its features, including linear quantitativity, positive contrast, and the absence of background signals, enhance disease detection and monitoring, improving clinical decision-making and patient outcomes [56, 67, 15, 16]. Continued research promises to further enhance MPI's clinical utility, providing valuable insights into diagnosing and managing a wide range of diseases.

8 Signal Processing

8.1 Integration of Machine Learning and Deep Learning

Integrating machine learning (ML) and deep learning (DL) into Magnetic Particle Imaging (MPI) signal processing marks a significant advancement, enhancing image reconstruction accuracy and efficiency. Techniques such as deep image priors and plug-and-play approaches address limitations of traditional methods reliant on hand-crafted regularization. Studies reveal that ML and DL frameworks, particularly convolutional neural networks (CNNs), achieve superior peak signal-to-noise ratios and structural similarity indices, facilitating real-time 3D imaging and expanding MPI's applications in medical diagnostics, including cancer imaging and angiography [65, 15, 21, 35, 30]. These

computational techniques exploit extensive MPI data to enhance image quality, reduce noise, and optimize reconstruction processes.

CNNs significantly improve MPI image resolution and clarity by training on large datasets to discern complex patterns, resulting in precise reconstructions and reduced artifacts. The availability of open-source MPI data and tools has accelerated the development of these advanced computational methods, offering essential resources for MPI reconstruction [59].

Probabilistic deep learning approaches, including variational autoencoders and generative adversarial networks, manage uncertainties and noise in MPI data effectively. These models enhance image quality by learning the underlying distribution of MPI signals, minimizing noise interference, and improving reconstruction accuracy, as shown by techniques like the deep plug-and-play prior and zero-shot denoiser, which surpass traditional methods [21, 65, 35].

Machine learning algorithms optimize MPI system parameters by analyzing patterns in MPI data to enhance key parameters such as magnetic field configurations, drive sequences, and tracer material selection. This optimization significantly boosts MPI performance, leading to enhanced image quality and precise spatial distribution of magnetic nanoparticle tracers. Ongoing research focuses on developing advanced forward models to accommodate dynamic tracer concentrations, improving adaptability to varying measurement conditions and enhancing image reconstruction fidelity [2, 46].

The incorporation of ML and DL not only enhances image quality but also accelerates the reconstruction process, making real-time imaging increasingly feasible. As these technologies evolve, they hold substantial potential to augment MPI's clinical and research applications, offering high-resolution, non-invasive imaging solutions that accurately quantify superparamagnetic iron oxide nanoparticles. This facilitates a diverse range of biomedical applications, including cancer detection, vascular imaging, real-time therapeutic monitoring, and intraoperative navigation, while minimizing background noise and enabling rapid image acquisition [15, 21, 56].

8.2 Techniques for Noise Reduction and Signal Clarity

Achieving high-quality imaging in Magnetic Particle Imaging (MPI) requires effective noise reduction and signal clarity enhancement. Noise can obscure the magnetic response of nanoparticles, degrading image quality. Researchers have developed advanced signal processing techniques to mitigate challenges posed by the ill-posed MPI reconstruction problem. Adaptive regularization algorithms and deep learning-based plug-and-play methods enhance dynamic range and spatial resolution, facilitating accurate visualization of superparamagnetic nanoparticles even under varying particle concentrations. A two-step algorithm has been proposed that quadruples the dynamic range, while a deep learning approach integrates a pre-trained plug-and-play prior to optimize reconstruction, significantly improving image quality and processing speed [38, 15, 21, 35].

Wavelet-based methods effectively separate noise from the true signal in MPI data by leveraging multi-resolution analysis to decompose the signal into different frequency components, allowing targeted noise reduction while preserving essential image details. Incorporating wavelet transforms into MPI reconstruction algorithms has demonstrated promise in enhancing image clarity and improving the overall signal-to-noise ratio [44].

Blind deconvolution techniques contribute to noise reduction by recovering the original signal through estimating the point spread function (PSF) without prior knowledge, addressing image blurriness caused by convolution effects and providing sharper reconstructions [41].

Machine learning models, particularly deep neural networks, have been explored for noise reduction in MPI. These models utilize sophisticated algorithms to learn intricate noise patterns from training datasets, enabling effective filtering of noise in new datasets. This capability significantly enhances signal clarity and image quality, especially in complex imaging modalities like MPI, where high temporal resolution and the absence of ionizing radiation are vital for medical applications. Employing techniques such as deep plug-and-play priors and variational deep probabilistic imaging, these models improve reconstruction accuracy and quantify uncertainty, providing a robust imaging solution [43, 20, 21, 35, 13].

The development and implementation of advanced noise reduction and signal clarity techniques are essential for optimizing MPI performance. Innovations in MPI, including accelerated projection-based reconstruction techniques and the deep plug-and-play prior for image reconstruction, significantly

enhance the generation of high-resolution and accurate images. Consequently, MPI continues to establish itself as a leading modality for non-invasive biomedical imaging applications, providing critical capabilities for cancer imaging, stem cell tracking, and angiography while maintaining high frame rates and minimizing noise and artifacts [21, 15].

8.3 Enhancements in Spatial Resolution and Sensitivity

Enhancing spatial resolution and sensitivity in Magnetic Particle Imaging (MPI) is crucial for advancing diagnostic capabilities and expanding clinical applications. Improvements in image reconstruction techniques that manage noise and artifacts, increase dynamic range for varying particle concentrations, and leverage advanced algorithms and machine learning methods facilitate real-time imaging and accurate visualization of superparamagnetic nanoparticles in diverse medical scenarios [38, 15, 21, 59, 23]. High spatial resolution is essential for visualizing small anatomical structures and subtle physiological changes, crucial for effective disease diagnosis and monitoring.

One approach to enhancing spatial resolution in MPI involves optimizing magnetic field gradients, which increase the spatial encoding capability of MPI systems, allowing finer resolution in reconstructed images [12]. Advanced coil designs, such as gradiometric receive coils, enhance sensitivity by reducing background noise and improving the detection of weak signals from low concentrations of magnetic nanoparticles [50].

The integration of advanced signal processing techniques, such as the TranSMS method, which combines convolutional and transformer modules, significantly improves the quality of image reconstruction and signal clarity, thereby enhancing spatial resolution [13].

Developing novel tracer materials with optimized magnetic properties is crucial for enhancing MPI sensitivity. Superparamagnetic iron oxide nanoparticles (SPIONs) with higher saturation magnetization and tailored surface coatings improve signal-to-noise ratios, enabling the detection of lower tracer concentrations and enhancing imaging sensitivity [54].

Deep learning techniques applied to MPI reconstruction algorithms offer promising solutions for enhancing both spatial resolution and sensitivity. By leveraging large datasets, deep learning models can discern complex patterns in MPI data, leading to more precise reconstructions and improved detection of subtle changes in tracer distribution [34].

Continuous development and integration of innovative hardware designs, advanced signal processing techniques, and optimized tracer materials are essential for enhancing spatial resolution and sensitivity in MPI. Recent advancements solidify MPI's status as a premier imaging modality for high-resolution, non-invasive biomedical applications, enabling superior image quality, high frame rates, and effective tumor visualization, particularly for cancer diagnosis and treatment monitoring without the risks associated with ionizing radiation [38, 15, 45, 21].

8.4 Task-Specific Optimization and Artifact Reduction

Task-specific optimization and artifact reduction are critical for enhancing the performance of Magnetic Particle Imaging (MPI), significantly improving image quality and diagnostic accuracy. Recent advancements, including the integration of a deep plug-and-play prior in model-based iterative optimization and novel reconstruction techniques, demonstrate that tailored approaches can effectively reduce noise and artifacts while maintaining high frame rates and contrast. These innovations facilitate real-time 3D imaging and ensure reliable and precise diagnostic outcomes, underscoring the importance of continuous refinement in MPI methodologies [21, 15].

Adaptive algorithms that dynamically adjust imaging parameters based on task-specific requirements represent a significant approach to optimization. These algorithms fine-tune magnetic field strengths and drive sequences, enhancing tracer localization precision and reducing artifacts in reconstructed images [2]. Such techniques are particularly beneficial in scenarios demanding high-resolution imaging, such as detecting small tumors or detailed vascular structures.

The integration of advanced computational models, particularly deep learning frameworks, plays a pivotal role in artifact reduction. These models are trained to recognize and correct common image artifacts, such as blurring and distortion, by leveraging large datasets to enhance reconstruction

accuracy [34]. Deep learning applications not only improve image clarity but also alleviate the computational burden associated with traditional artifact correction methods.

Hybrid imaging techniques, combining MPI with other modalities, offer a promising strategy for artifact reduction. By integrating complementary imaging data, these approaches provide a more comprehensive view of the biological system, facilitating accurate differentiation between true signals and artifacts [37]. This integration proves advantageous in complex imaging scenarios, where multiple tissue types and varying tracer concentrations complicate image interpretation.

Task-specific optimization and artifact reduction are integral to advancing MPI technology. By leveraging advancements in adaptive algorithms, deep learning frameworks, and hybrid imaging techniques, MPI can deliver significantly enhanced image quality and more dependable diagnostic results. These improvements broaden MPI's clinical applications, including cancer imaging, stem cell tracking, and angiography, while enhancing its utility in research settings. New algorithms have been developed to increase the dynamic range of MPI, enabling accurate imaging of samples with varying concentrations of magnetic nanoparticles, while deep learning approaches like the PP-MPI method have demonstrated superiority over traditional reconstruction techniques, providing real-time imaging capabilities with improved signal-to-noise ratios [38, 15, 21].

9 Conclusion

9.1 Evaluation Metrics and Performance Assessment

Benchmark	Size	Domain	Task Format	Metric
MDF[17]	1,000,000	Magnetic Particle Imaging	Image Reconstruction	PSNR, SSIM
SHA[54]	6	Biomedical Engineering	Characterization	Saturation Magnetization, Zeta Potential
Resotran-MPI[11]	4	Biomedical Imaging	Magnetic Particle Imaging	SNR, SSIM

Table 1: This table presents a comparative analysis of representative benchmarks in Magnetic Particle Imaging (MPI) and related biomedical fields. It includes information on benchmark size, domain, task format, and evaluation metrics, highlighting the diversity and scope of datasets used for advancing imaging techniques.

Magnetic Particle Imaging (MPI) is assessed through metrics that measure its sensitivity and resolution, crucial for accurate detection and visualization of magnetic nanoparticles within biological tissues. These metrics ensure the production of high-quality images, essential for effective diagnostics and treatment planning. Table 1 provides a detailed overview of the benchmarks utilized in the assessment of Magnetic Particle Imaging (MPI) performance and related biomedical imaging fields. The technique's superior contrast-to-noise ratio compared to traditional imaging methods, like MRI, enhances its application in clinical settings, particularly due to its non-reliance on ionizing radiation. This attribute not only increases patient safety but also facilitates repeated imaging sessions, thereby expanding MPI's clinical utility.

The Signal Detection Operator (SDO) plays a pivotal role in optimizing MPI systems, allowing for the effective ranking of data-acquisition designs based on visual assessments of reconstructed images. This capability is instrumental in refining MPI protocols and enhancing the efficiency of imaging systems that utilize sparse reconstruction methods. MPI's transformative potential in biomedical imaging is underscored by its ability to provide real-time, high-resolution images, which significantly enhance diagnostic accuracy and treatment efficacy. Ongoing research into performance metrics and optimization strategies is essential to further advance MPI's capabilities and broaden its clinical applications. Future efforts should focus on refining these metrics and exploring innovative methodologies to improve the accuracy and reliability of MPI systems, ensuring their prominence in the domain of non-invasive imaging technologies.

9.2 Challenges and Future Directions

Magnetic Particle Imaging (MPI) is on the cusp of becoming a cornerstone in medical imaging, yet several challenges must be overcome to fully realize its potential in clinical and research settings. A key challenge is developing human-sized MPI scanners that seamlessly integrate into clinical workflows, requiring optimization of spatial resolution and sensitivity, alongside rigorous clinical

trials to validate efficacy and safety. Enhancing the spatial resolution of tomographic MPI systems is particularly important for comprehensive functional mapping in rodent models, which could serve as precursors for human applications.

The reconstruction process in MPI presents challenges, especially when handling multi-sample imaging sequences. Current methodologies need refinement to accommodate varying iron concentrations and incorporate complex models that account for multiple magnetic states during cellular uptake. Future research should focus on optimizing imaging parameters to improve contrast between relaxation times and developing tailored superparamagnetic iron oxide (SPIO) nanoparticles with distinct relaxation behaviors. Additionally, optimizing the spatial distribution of magnetic nanoparticle samples during calibration and extending methods to handle multiplexed low-resolution system matrix measurements could significantly enhance reconstruction capabilities.

In tracer development, prolonging the circulation time of SPIO nanoparticles and exploring targeted delivery strategies are crucial for improving specificity and effectiveness. Emphasis should also be placed on characterizing and optimizing clinically approved MRI tracers for MPI applications, alongside investigating additional imaging techniques and conducting clinical trials to establish efficacy and safety. Furthermore, integrating Magnetic Fluid Hyperthermia (MFH) with MPI offers promising theranostic capabilities, necessitating research to optimize cooling capabilities and explore specialized ferrofluids to enhance heating performance.

Technological advancements in MPI hardware present challenges, particularly in implementing cost-effective solutions for high-frequency lock-in detection across multiple independent channels. Future research should aim to optimize vector modulator design for improved linearity and explore methods for measuring feedthrough in the presence of particles. Additionally, optimizing coil design for enhanced field uniformity and exploring hybrid systems to reduce power consumption are critical areas for further exploration.

Future research directions should include the development of multimodal imaging systems that integrate MPI with anatomical imaging, enhancing tracer design for improved specificity and safety, and exploring novel therapeutic applications of MPI. Merging reconstruction results from varying subproblem sizes could leverage the strengths of both small and large subproblems in dynamic scenarios. Moreover, refining reconstruction algorithms to improve edge preservation and exploring alternative regularization techniques are promising avenues for future exploration.

Addressing these challenges through targeted research and collaboration is crucial for advancing MPI technology, expanding its clinical applications, and ensuring its evolution as a leading modality in non-invasive imaging. Future research should prioritize enhancing tracer development and exploring multimodal imaging capabilities to broaden MPI's clinical applicability. Additionally, improving simulation accuracy, incorporating complex models, and expanding educational resources related to MPI tools will be vital for future advancements. Investigating the in-vivo performance of saline as a bolus and exploring modifications to optimize tracer delivery in real blood flow scenarios should also be prioritized.

References

[1] Fabian Mohn, Miriam Exner, Patryk Szwargulski, Martin Möddel, Tobias Knopp, and Matthias Graeser. Saline bolus for negative contrast perfusion imaging in magnetic particle imaging, 2023.

[2] The-Vinh Tran-Luu, Mark-Alexander Henn, Klaus N. Quelhas, and Solomon I. Woods. Mp-you: A web-based mpi simulation tool, 2024.

[3] Patryk Szwargulski, Nadine Gdaniec, Matthias Graeser, Martin Möddel, Florian Griese, Kannan M. Krishnan, Thorsten M. Buzug, and Tobias Knopp. Moving table magnetic particle imaging: A stepwise approach preserving high spatio-temporal resolution, 2018.

[4] Xinyi Y Zhou, Zhi Wei Tay, Prashant Chandrasekharan, Y Yu Elaine, Daniel W Hensley, Ryan Orendorff, Kenneth E Jeffris, David Mai, Bo Zheng, Patrick W Goodwill, et al. Magnetic particle imaging for radiation-free, sensitive and high-contrast vascular imaging and cell tracking. *Current opinion in chemical biology*, 45:131–138, 2018.

[5] Yavuz Muslu, Mustafa Utkur, Omer Burak Demirel, and Emine Ulku Saritas. Calibration-free relaxation-based multi-color magnetic particle imaging, 2017.

[6] Alexey Tonyushkin. Single-sided field-free line generator magnet for multidimensional magnetic particle imaging, 2017.

[7] Lyndia C Wu, Yanrong Zhang, Gary Steinberg, H Qu, S Huang, M Cheng, T Bliss, F Du, J Rao, G Song, et al. A review of magnetic particle imaging and perspectives on neuroimaging. *American Journal of Neuroradiology*, 40(2):206–212, 2019.

[8] Patrick Vogel, MA Rückert, Christoph Greiner, Johanna Günther, Teresa Reichl, Thomas Kampf, TA Bley, VC Behr, and Stefan Herz. impi: portable human-sized magnetic particle imaging scanner for real-time endovascular interventions. *Scientific Reports*, 13(1):10472, 2023.

[9] Bo Zheng, Elaine Yu, Ryan Orendorff, Kuan Lu, Justin J Konkle, Zhi Wei Tay, Daniel Hensley, Xinyi Y Zhou, Prashant Chandrasekharan, Emine U Saritas, et al. Seeing spios directly in vivo with magnetic particle imaging. *Molecular imaging and biology*, 19:385–390, 2017.

[10] James Wells, Shailey Twamley, Aparna Sekar, Antje Ludwig, Hendrik Paysen, Olaf Kosch, and Frank Wiekhorst. Lissajous scanning magnetic particle imaging as a multifunctional platform for magnetic hyperthermia therapy, 2020.

[11] Fabian Mohn, Konrad Scheffler, Justin Ackers, Agnes Weimer, Franz Wegner, Florian Thieben, Mandy Ahlborg, Patrick Vogel, Matthias Graeser, and Tobias Knopp. Characterization of the clinically approved mri tracer resotran for magnetic particle imaging in a comparison study, 2024.

[12] Tuan-Anh Le, Minh Phu Bui, Khaled Mohamed Gadelmowla, Seungjun Oh, and Jungwon Yoon. First human-scale magnetic particle imaging system with superconductor. *International Journal on Magnetic Particle Imaging IJMPI*, 9(1 Suppl 1), 2023.

[13] Alper Güngör, Baris Askin, Damla Alptekin Soydan, Emine Ulku Saritas, Can Barış Top, and Tolga Çukur. Transms: Transformers for super-resolution calibration in magnetic particle imaging, 2022.

[14] Clarissa Zimmerman Cooley, Joseph B Mandeville, Erica E Mason, Emiri T Mandeville, and Lawrence L Wald. Rodent cerebral blood volume (cbv) changes during hypercapnia observed using magnetic particle imaging (mpi) detection. *NeuroImage*, 178:713–720, 2018.

[15] Kenya Murase. A method for accelerating projection-based magnetic particle imaging, 2023.

[16] Xiao Han, Yang Li, Weifeng Liu, Xiaojun Chen, Zeyu Song, Xiaolin Wang, Yulin Deng, Xiaoying Tang, and Zhenqi Jiang. The applications of magnetic particle imaging: From cell to body. *Diagnostics*, 10(10):800, 2020.

[17] Tobias Knopp, Thilo Viereck, Gael Bringout, Mandy Ahlborg, Anselm von Gladiss, Christian Kaethner, Alexander Neumann, Patrick Vogel, Jürgen Rahmer, and Martin Möddel. Mdf: Magnetic particle imaging data format, 2020.

[18] Sebastian Draack, Meinhard Schilling, and Thilo Viereck. Magnetic particle imaging of particle dynamics in complex matrix systems. *Physical Sciences Reviews*, 8(2):213–237, 2023.

[19] Marija Boberg, Tobias Knopp, Patryk Szwargulski, and Martin Möddel. Generalized mpi multi-patch reconstruction using clusters of similar system matrices, 2022.

[20] Yujia Chen, Yang Lou, Kun Wang, Matthew A. Kupinski, and Mark A. Anastasio. Reconstruction-aware imaging system ranking by use of a sparsity-driven numerical observer enabled by variational bayesian inference, 2019.

[21] Baris Askin, Alper Güngör, Damla Alptekin Soydan, Emine Ulku Saritas, Can Barış Top, and Tolga Cukur. Pp-mpi: A deep plug-and-play prior for magnetic particle imaging reconstruction. In *International Workshop on Machine Learning for Medical Image Reconstruction*, pages 105–114. Springer, 2022.

[22] International journal on magneti.

[23] Anselm von Gladiss, Amir Shayan Ahmadian, and Jan Jürjens. Reconstruction as a service: a data space for off-site image reconstruction in magnetic particle imaging, 2024.

[24] Kenya Murase. Behavior of signal harmonics in magnetic particle imaging based on a lock-in-amplifier model, 2019.

[25] Vladyslav Gapyak, Thomas März, and Andreas Weinmann. Reconstruction formulae for 3d field-free line magnetic particle imaging, 2023.

[26] Hendrik Paysen, Norbert Loewa, Anke Stach, James Wells, Olaf Kosch, Shailey Twamley, Marcus R. Makowski, Tobias Schaeffter, Antje Ludwig, and Frank Wiekhorst. 3d-imaging and quantification of magnetic nanoparticle uptake by living cells, 2019.

[27] Fabian Mohn, Tobias Knopp, Marija Boberg, Florian Thieben, Patryk Szwargulski, and Matthias Graeser. System matrix based reconstruction for pulsed sequences in magnetic particle imaging, 2022.

[28] Matthias Graeser, Peter Ludewig, Patryk Szwargulski, Fynn Foerger, Tom Liebing, Nils D. Forkert, Florian Thieben, Tim Magnus, and Tobias Knopp. Design of a head coil for high resolution mouse brain perfusion imaging using magnetic particle imaging, 2022.

[29] Wolfgang Erb, Andreas Weinmann, Mandy Ahlborg, Christina Brandt, Gael Bringout, Thorsten M. Buzug, Jürgen Frikel, Christian Kaethner, Tobias Knopp, Thomas März, Martin Möddel, Martin Storath, and Alexander Weber. Mathematical analysis of the 1d model and reconstruction schemes for magnetic particle imaging, 2017.

[30] Tobias Kluth and Bangti Jin. Numerical reconstruction in magnetic particle imaging, 2019.

[31] Florian Griese, Tobias Knopp, Cordula Gruettner, Florian Thieben, Knut Müller, Sonja Loges, Peter Ludewig, and Nadine Gdaniec. Simultaneous magnetic particle imaging and navigation of large superparamagnetic nanoparticles in bifurcation flow experiments, 2019.

[32] Guang Jia, Zhongwei Bian, Tianshu Li, Shi Bai, Chenxing Hu, Lixuan Zhao, Peng Gao, Tanping Li, Hui Hui, and Jie Tian. Stimulated emission depletion (sted) magnetic particle imaging, 2024.

[33] Atakan Topcu, Asli Alpman, Mustafa Utkur, and Emine Ulku Saritas. Vicinity effects of field free point on the relaxation behavior of mnps, 2021.

[34] Ivo Matteo Baltruschat, Patryk Szwargulski, Florian Griese, Mirco Grosser, René Werner, and Tobias Knopp. 3d-smrnet: Achieving a new quality of mpi system matrix recovery by deep learning, 2019.

[35] Vladyslav Gapyak, Corinna Rentschler, Thomas März, and Andreas Weinmann. An ℓ^1-plug-and-play approach for mpi using a zero shot denoiser with evaluation on the 3d open mpi dataset, 2024.

[36] Tobias Kluth, Patryk Szwargulski, and Tobias Knopp. Towards accurate modeling of the multidimensional magnetic particle imaging physics, 2019.

[37] Sarah Latus, Florian Griese, Matthias Schlüter, Christoph Otte, Martin Möddel, Matthias Graeser, Thore Saathoff, Tobias Knopp, and Alexander Schlaefer. Bimodal intravascular volumetric imaging combining oct and mpi, 2019.

[38] Marija Boberg, Nadine Gdaniec, Patryk Szwargulski, Franziska Werner, Martin Möddel, and Tobias Knopp. Simultaneous imaging of widely differing particle concentrations in mpi: Problem statement and algorithmic proposal for improvement, 2022.

[39] Anna C Bakenecker, Mandy Ahlborg, Christina Debbeler, Christian Kaethner, Thorsten M Buzug, and Kerstin Lüdtke-Buzug. Magnetic particle imaging in vascular medicine. *Innovative surgical sciences*, 3(3):179–192, 2018.

[40] Thinh Q. Bui, Mark-Alexander Henn, Weston L. Tew, Megan A. Catterton, and Solomon I. Woods. Temperature-dependent magnetic particle imaging with multi-harmonic lock-in detection, 2023.

[41] Onur Yorulmaz, Omer Burak Demirel, Yavuz Muslu, Tolga Çukur, Emine U Saritas, and A Enis Çetin. A blind deconvolution technique based on projection onto convex sets for magnetic particle imaging, 2020.

[42] Kai Wu, Diqing Su, Renata Saha, Jinming Liu, and Jian-Ping Wang. Investigating the effect of magnetic dipole-dipole interaction on magnetic particle spectroscopy (mps): Implications for magnetic nanoparticle-based bioassays and magnetic particle imaging (mpi), 2019.

[43] He Sun and Katherine L. Bouman. Deep probabilistic imaging: Uncertainty quantification and multi-modal solution characterization for computational imaging, 2020.

[44] Florian Lieb and Tobias Knopp. A wavelet based sparse row-action method for image reconstruction in magnetic particle imaging, 2020.

[45] Elaine Y Yu, Mindy Bishop, Bo Zheng, R Matthew Ferguson, Amit P Khandhar, Scott J Kemp, Kannan M Krishnan, Patrick W Goodwill, and Steven M Conolly. Magnetic particle imaging: a novel in vivo imaging platform for cancer detection. *Nano letters*, 17(3):1648–1654, 2017.

[46] Christina Brandt and Christiane Schmidt. Modeling magnetic particle imaging for dynamic tracer distributions, 2021.

[47] Marija Boberg, Tobias Knopp, and Martin Möddel. Unique compact representation of magnetic fields using truncated solid harmonic expansions, 2023.

[48] Patrick Vogel, Martin A. Rückert, Thomas Kampf, and Volker C. Behr. Highly flexible and modular simulation framework for magnetic particle imaging, 2022.

[49] Bilal Tasdelen, Mustafa Utkur, Asli Alpman, Can Baris Top, Ergin Atalar, and Emine Ulku Saritas. Vector modulator based active compensation of direct feedthrough in magnetic particle imaging, 2024.

[50] Matthias Gräser, Tobias Knopp, Patryk Szwargulski, Thomas Friedrich, Anselm von Gladiss, Michael Kaul, Kannan M Krishnan, Harald Ittrich, Gerhard Adam, and Thorsten M. Buzug. Towards picogram detection of superparamagnetic iron-oxide particles using a gradiometric receive coil, 2017.

[51] Seung-Min Choi, Jae-Chan Jeong, Jinsun Kim, Eul-Gyoon Lim, Chang-beom Kim, Sang-Jin Park, Dae-Yong Song, Hans-Joachim Krause, Hyobong Hong, and In So Kweon. A novel three-dimensional magnetic particle imaging system based on the frequency mixing for the point-of-care diagnostics. *Scientific reports*, 10(1):11833, 2020.

[52] S Zanganeh, M Aieneravaie, M Erfanzadeh, JQ Ho, and R Spitler. Magnetic particle imaging (mpi). In *Iron Oxide Nanoparticles for Biomedical Applications*, pages 115–133. Elsevier, 2018.

[53] Kai Wu, Jinming Liu, Diqing Su, Renata Saha, Vinit Kumar Chugh, and Jian-Ping Wang. Single-core or multi-core? a mini review on magnetic nanoparticles for magnetic particle spectroscopy-based bioassays, 2020.

[54] Kai Wu, Jinming Liu, Renata Saha, Chaoyi Peng, Diqing Su, Andrew Yongqiang Wang, and Jian-Ping Wang. An investigation of commercial iron oxide nanoparticles: Advanced structural and magnetic properties characterization, 2020.

[55] Oliver Buchholz, Kulthisa Sajjamark, Jochen Franke, Huimin Wei, André Behrends, Christian Münkel, Cordula Grüttner, Pierre Levan, Dominik von Elverfeldt, Matthias Gräser, Thorsten Buzug, Sébastien Bär, and Ulrich G. Hofmann. In situ theranostic platform uniting highly localized magnetic fluid hyperthermia, magnetic particle imaging, and thermometry in 3d, 2023.

[56] Xue Yang, Guoqing Shao, Yanyan Zhang, Wei Wang, Yu Qi, Shuai Han, and Hongjun Li. Applications of magnetic particle imaging in biomedicine: advancements and prospects. *Frontiers in physiology*, 13:898426, 2022.

[57] Florian Griese, Sarah Latus, Matthias Schlüter, Matthias Graeser, Matthias Lutz, Alexander Schlaefer, and Tobias Knopp. In-vitro mpi-guided ivoct catheter tracking in real time for motion artifact compensation, 2019.

[58] Klaas-Julian Janssen, Meinhard Schilling, Frank Ludwig, and Jing Zhong. Single harmonic-based narrowband magnetic particle imaging, 2021.

[59] Lin Yin, Wei Li, Yang Du, Kun Wang, Zhenyu Liu, Hui Hui, and Jie Tian. Recent developments of the reconstruction in magnetic particle imaging. *Visual computing for industry, biomedicine, and art*, 5(1):24, 2022.

[60] Kenya Murase. A simulation study on image reconstruction in magnetic particle imaging with field-free-line encoding, 2016.

[61] Xiaojun Chen, Zhenqi Jiang, Xiao Han, Xiaolin Wang, and Xiaoying Tang. The reconstruction of magnetic particle imaging: Current approaches based on the system matrix. *Diagnostics*, 11(5):773, 2021.

[62] Gaël Bringout, Wolfgang Erb, and Jürgen Frikel. A new 3d model for magnetic particle imaging using realistic magnetic field topologies for algebraic reconstruction, 2020.

[63] Marius Nitzsche and Bernadette N Hahn. Dynamic image reconstruction in mpi with resesop-kaczmarz, 2024.

[64] Hannes Albers, Tobias Knopp, Martin Möddel, Marija Boberg, and Tobias Kluth. Modeling the magnetization dynamics for large ensembles of immobilized magnetic nanoparticles in multi-dimensional magnetic particle imaging, 2021.

[65] Sören Dittmer, Tobias Kluth, Mads Thorstein Roar Henriksen, and Peter Maass. Deep image prior for 3d magnetic particle imaging: A quantitative comparison of regularization techniques on open mpi dataset, 2020.

[66] Sarah Vaalma, Jürgen Rahmer, Nikolaos Panagiotopoulos, Robert L Duschka, Jörn Borgert, Jörg Barkhausen, Florian M Vogt, and Julian Haegele. Magnetic particle imaging (mpi): experimental quantification of vascular stenosis using stationary stenosis phantoms. *PloS one*, 12(1):e0168902, 2017.

[67] <div style="text-align: center;".

Disclaimer:

SurveyX is an AI-powered system designed to automate the generation of surveys. While it aims to produce high-quality, coherent, and comprehensive surveys with accurate citations, the final output is derived from the AI's synthesis of pre-processed materials, which may contain limitations or inaccuracies. As such, the generated content should not be used for academic publication or formal submissions and must be independently reviewed and verified. The developers of SurveyX do not assume responsibility for any errors or consequences arising from the use of the generated surveys.