CO2xNxPhrag: Porewater SO4/Cl

July 2024 Samples

2025-10-14

Contents

0.1 Run Information 2
0.2 Assess Standard Curves 3
0.3 Assess Check Standards 4
0.4 Assess Blanks 5
0.5 Assess Duplicates 7
0.6 Calculate mmol/L concentrations & salinity, add dilutions 8
0.7 Assess Analytical Spikes 9
0.8 Check if samples within the range of the standard curve 9
0.9 Check to see if samples run match metadata & merge info 10
0.10 Visualize Data by Plot 11
0.11 Export Processed Data 11

##Add Required Packages

0.1 Run Information

```
###### Run information - PLEASE CHANGE
  Date_Run = "2025-08-20"  #Date that instrument was run
  Run_by = "Zoe Read"  #Instrument user
  Script_run_by = "Zoe Read" #Code user
  sample_year <- "2024"
  sample_month <- "July"
  Site = "GCReW"
  Project_name = "CO2xNxPhrag"
  run_notes = "
  " #any notes from the run
  samples <- c("GCW", "GWI", "MSM", "SWH") #whatever identifies your samples within the same names
  samples_pattern <- paste(samples, collapse = "|")
    #samples_pattern <- "GCW" #use this instead of the line above if you have only one site code

###### File Names - PLEASE CHANGE
#file path and name for raw summary data file
  raw_file_name_cl = "Raw Data/CO2xNxPhrag_202407_Cl.txt"
  raw_file_name_so4 = "Raw Data/CO2xNxPhrag_202407_SO4.txt"

#file path and name of processed data file
  processed_file_name = "Processed Data/GCReW_CO2xNxPhrag_PW_Processed_Cl_SO4_202407.csv"

###### Log Files - PLEASE CHECK
#downloaded metadata csv - downloaded from Google drive as csv for this year
  Raw_Metadata = here("..", "GCReW", "GCReW_Project_Treatment_Metadata.csv")

  #qaqc log file path for this year
  Log_path = "Raw Data/COMPASS_Synoptic_Cl_SO4_QAQClog_2024.csv"
```

##Set Up Code - constants and QAQC cutoffs

##Read in metadata and create similar sample IDs for matching to samples

##Import Sample Data

0.2 Assess Standard Curves






```
## [1] "QAQC log file exists and has been read into the code."
```




```
## [1] "Cl Curve r2 GOOD"

## [1] "SO4 Curve r2 GOOD"
```

0.3 Assess Check Standards

```
## # A tibble: 5 x 5
##   sample_ID  mean_Cl  sd_Cl    cv_Cl flag_Cl
##   <chr>        <dbl>  <dbl>    <dbl> <chr>
## 1 Standard 1    4.69 0.0646 0.0138   Chloride Check Standard RSD within Range -~
## 2 Standard 2    9.61 0.102  0.0106   Chloride Check Standard RSD within Range -~
## 3 Standard 3   19.3  0.110  0.00567  Chloride Check Standard RSD within Range -~
## 4 Standard 4   99.3  0.0530 0.000534 Chloride Check Standard RSD within Range -~
## 5 Standard 5  200.   0.0946 0.000472 Chloride Check Standard RSD within Range -~

## # A tibble: 5 x 5
##   sample_ID  mean_SO4    sd_SO4    cv_SO4 flag_SO4
##   <chr>         <dbl>     <dbl>     <dbl> <chr>
## 1 Standard 1    0.436 0.00155   0.00354   Sulfate Check Standard RSD within Ran~
## 2 Standard 2    0.972 0.00502   0.00517   Sulfate Check Standard RSD within Ran~
## 3 Standard 3    2.02  0.0000707 0.0000349 Sulfate Check Standard RSD within Ran~
## 4 Standard 4   10.7   0.00997   0.000934  Sulfate Check Standard RSD within Ran~
## 5 Standard 5   19.7   0.000919  0.0000467 Sulfate Check Standard RSD within Ran~

## [1] ">80% of Chloride Check Standards have RSD within range - PROCEED"

## [1] ">80% of Sulfate Check Standards have RSD within range - PROCEED"
```




```
## [1] ">80% of Chloride Check Standards are within range of expected concentration - PROCEED"
```

```
## [1] ">80% of Sulfate Check Standards are within range of expected concentration - PROCEED"
```

0.4 Assess Blanks

```
## [1] ">80% of Chloride Blank concentrations are lower 25% quartile of samples"
```

```
## [1] ">80% of Sulfate Blank concentrations are lower 25% quartile of samples"
```






```
## Chloride blanks mean ppm:

## [1] 0

## Sulfate blanks mean ppm:

## [1] 0
```

0.5 Assess Duplicates




```
## [1] ">80% of Chloride Duplicates have a CV <10% - PROCEED"

## [1] ">80% of Sulfate Duplicates have a CV <10% - PROCEED"
```




```
## [1] ">80% of Chloride Duplicates have a percent difference <10% - PROCEED"
```

```
## [1] ">80% of Sulfate Duplicates have a percent difference <10% - PROCEED"
```

0.6 Calculate mmol/L concentrations & salinity, add dilutions

```
# Convert ppm to mmol/L
all_dat$SO4_Conc_mM <- (all_dat$SO4_ppm / s_mw)
all_dat$Cl_Conc_mM <- (all_dat$Cl_ppm / cl_mw)

# Calculate Salinity
# calculated using the Knudsen equation
# Salinity = 0.03 + 1.8050 * Chlorinity
# Ref: A Practical Handbook of Seawater Analysis by Strickland & Parsons (P. 11)
# =((1.807*Cl_ppm)+0.026)/1000
all_dat$salinity <- ((1.8070 * all_dat$Cl_ppm) + 0.026) / 1000

#Need to determine dilution factors for your samples
#for Steph / COMPASS this depends on the site so...
all_dat$Dilution <- Dilution
# head(all_dat)
```

0.7 Assess Analytical Spikes




```
## [1] ">80% of SO4 spikes have a recovery between the high and low cutoff - PROCEED"
```

0.8 Check if samples within the range of the standard curve

```
## Sample Flagging
```



Table 1: SO4 samples

SO4_Conc_flag	Percent_samples
Within_Range	100

Table 2: Cl samples

Cl_Conc_flag	Percent_samples
Within_Range	100

0.9 Check to see if samples run match metadata & merge info

```
## All sample IDs are present in metadata.
```

0.10 Visualize Data by Plot





0.11 Export Processed Data

#end