


COMMISSIONING LICENCE FOR RADIOTHERAPY
COMMISSIONING LICENCE FOR DIAGNOSTIC RADIOLOGY

NIGERIAN NUCLEAR REGULATORY AUTHORITY
(NUCLEAR SAFETY AND RADIATION PROTECTION ACT 1995)



The type of machine
should appear here.

COMMISSIONING LICENCE FOR LINEAR ACCELERATOR
No. NNRA/RT/AUT/COM/464/2019

The Nigerian Nuclear Regulatory Authority (hereinafter referred to as the **Authority**) pursuant to Sections 6 (a) and 22 of the Nuclear Safety and Radiation Protection Act (the **Act**) hereby grants Commissioning Licence No. NNRA/RT/AUT/COM/464/2019

To

NIGERIAN TEACHING HOSPITAL, represented by the
Chief Medical Director, **Prof. Doctor** (hereinafter referred to as the **Licensee**),
whose registered address is Idi – Araba, Lagos, Nigeria

To

Commission The Machine with the following specifications:

Equipment Type	Serial No.		Photon Energies	Electron Energies	Manufacturer/Country/Year
	Machine	Tube			
LINAC VITABEAM	3628	Vitalbeam	6, 10, 15MV	6, 8, 10, 12, 15MeV	Varian Medical Systems, USA/2018

Machine Type	Machine Model No.	Serial No.	Isotope	Activity	Manufacturer/Country/Year
HDR Brachytherapy	GammaMed HDR	1175	Ir-192	10.0Ci	Varian Medical Systems/Germany/2018

Machine Type	Model No.	Serial No.	Max. kVp	Max. mAs	Manufacturer/Country/Year
⁶⁰ Co Simulator	CT590RT	237468113	-	-	General Electric Inc., /2018

This Licence is not transferable and is subject to the Terms and Conditions herein

TERMS AND CONDITIONS

- 1 The Licensee is authorized to commission the specified Linear Accelerators, HDR Brachytherapy and Computed Tomography Simulator Machines at the **Lagos University Teaching Hospital, Idi Araba, Lagos State**.
- 2 The commissioning of the Linear Accelerators, HDR Brachytherapy and Computed Tomography Simulator Machines at an unauthorized location is illegal;
- 3 The Licensee shall perform the commissioning using phantoms only;
- 4 The Licensee shall comply with the Nigeria Basic Ionizing Radiation Regulations, 2003 and the Act;

All fields are compulsory