TITLE: Certificate of Conformity of TSSC Elegance Silver Edition – 5413424-R

Descri	ption	Supplier P/N	Test document
TSSC Elegance Silver Edit	tion CDRH - 5413424-R		PSP 5413422
Date of Manufacturing:	Serial Number:	Supplier ID:	Country of origin: Made in Poland

Description	Reference	CTQ N°	PASS/FAIL
Functional test	Product tested with fixture N°8F1356CT0001	NA	
References of joystick bushing	Bushing WSM 1618-20 supplier: IGUS P/N: 8FM000098	CTQ 122	
Reference of buzzer	SMA24-P10 supplier: Sonitron	CTQ 123	
Reference of the emergency stop	PVA2 OA H5 1.7N P/N: 283-2005-025	CTQ 124	
CDRH mandatory records	Buzzer test	CTQ 50	
CDRH mandatory records	Collimator II1 left test	CTQ 1	
CDRH mandatory records	Collimator II1 right test	CTQ 2	
CDRH mandatory records	Collimator II1 down test	CTQ 3	
CDRH mandatory records	Collimator II1 up test	CTQ 4	
CDRH mandatory records	Collimator II2c	CTQ 5	
CDRH mandatory records	Collimator II3a left test	CTQ 6	
CDRH mandatory records	Collimator II3a right test	CTQ 7	
CDRH mandatory records	Collimator II3a down test	CTQ 8	
CDRH mandatory records	Collimator II3a up test	CTQ 9	
CDRH mandatory records	Collimator II3b 1 test	CTQ 10	
CDRH mandatory records	Collimator II3b 2 test	CTQ 11	
CDRH mandatory records	Collimator II3b 3 test	CTQ 12	
CDRH mandatory records	Collimator II3c test	CTQ 13	
CDRH mandatory records	Timer (Fluoro) II16	CTQ 45	
Harness wire (DTC P/N: 8F1356M0804)	I housing III 2464) Supplier CIFF III registration		
Board (DTC P/N 8F1356B0103)	PCB; UL certified (UL: 94; Flammability rating V0) Supplier SANMINA-SCI UL 94V0	NA	

	Source	Model	Rating	Annual Test Nur	nber	PASS/FAIL
Motion		PVA2 OA H5	32V DC Max	Report	Date of test	
Emergency	C&K	1,7N	100mA DC	number		
			Max			

Dérogation : □ Oui □ Non		n	N° de dérogation :				
5	П О:	D.1-T-				.	Γ
Product	□ Oui	DATE		Name		Signature	
tested	☐ Non						
Product	□ Oui	DATE		Name		Signature	
released	□ Non						

Service réparation SRO BUC	DOCXYZ	
----------------------------	--------	--