



Customer Repair Report

Model OTHER	Serial Number YEZ15859TC	Manufacture Date	RMA Number 02793	IRIDEX Service Contact Mark W	Completion Date
Customer Name Topcon Europe Medical B.V.		Customer Number 10868	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

still On while IQ577 console is shutting down.

Technician Product Evaluate:

CONTROL BOX STILL ON WHILE IQ577 CONSOLE IS SHUTTING DOWN. INETRLOCK BOARD AND SC500 CABLE CAUSING THE PROBLEM.

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
70014		1	PCBA SCANNER LASER INTERFACE & INTERLOCK	\$ 1,364.56
70077		1	ASSY CABLE SC500 TO INTERFACE	\$ 155.68
TECHNICIAN HQ		5	Technician Labor	\$ 1,425.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ -

TOTAL CHARGES **\$ 2,945.24**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: