

## Customer Repair Report

Model	Serial Number		Manufacture Date		RMA Number	IRIDEX Service Contact	Completion Date
OTHER	2721-VE	Γ			02043	Mark W	
Customer Name		Custo	mer Number	Repair Type			
Garden State Veterinary Specialists			24144	☐ Warranty ☐ Upgrade ☐ Other			

Creatarnan	Description	of Duoblem
Customer	<b>Description</b>	of Problem

NO ILLUMINATION TRIED 2 LIO

## **Technician Product Evaluate:**

Need to replace power supply. Will calibrate and test to meet MFG standards.

## **Service Performed on Unit:** Waiting for customer approval

## **ITEMIZED REPAIR CHARGES**

Part Number	Price/Unit	Quantity	Description	Total Price	
11081		1	POWER SUPPLY BOARD IR SL SLx DioVet	\$ 1,	,970.00
TECHNICIAN HQ		6	Technician Labor	<b>\$</b> 1,	710.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$	206.50

TOTAL CHARGES \$ 3,886.50

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Name:
Name:

Signature: Date: