

Customer Repair Report

Model	Serial Number		Manufacture Date		RMA Number	IRIDEX Service Contact	Completion Date	
OTHER	YEZ145237	ГС			03035	MARK W		
Customer Name			mer Number	Repair Type				
Retina Consultants of PR			25369	☐ Warranty ☑ Upgrade ☐ Other				

Customer Description of Problem

E05111

Technician Product Evaluate:

UNABLE TO TEST CONTROL BOX NON-STOP BLINKING. ASSY, PIVOT BASE IS MISSINGN.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number Price/Unit		Quantity	Description	Total Price	
30943-02		1	ASSY PIVOT BASE IRIDEX: 30943-XX	\$	651.52
88356		1	COMPUTER HDD	\$	1,950.00
88351		2	COMPUTER MOUNT	\$	138.00
88352		1	CABLE ADAPTER, TxCell CPU	\$	46.16
88261		1	Cable Assy. USB Two Ports.	\$	83.00
TECHNICIAN HQ		6	Technician Labor	\$	1,710.00
BILLABLE FREIGHT	1	1	BILLABLE FREIGHT	\$	_

TOTAL CHARGES

4,578.68

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Signature: Date: