

Customer Repair Report

Model	Serial Number		Manufacture Date	e	RMA Number	IRIDEX Service Contact	Completion Date
OTHER	K1227				02871	Mark W	
Customer Name		Custo	mer Number	Repair Type			
Banner Desert Medical Center		13148	Warranty Non-warranty Upgrade Other				

Customor	Description	of Problem
Customer	Describtion	oi Problem

Broken Stem

Technician Product Evaluate:

Cannot test due to Twisted and burnt assembly fiber optic, burnt halogen bulb and will need alignment.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number Price/Unit		Quantity	Description	Total Price	
10784		1	LIO Halogen Bulb-10W/6V Heine: X-004.88.068	\$	135.00
88080		1	Jumper Assy Mulitlam LIO+	\$	984.00
11481		1	DUST CAP W/LANYARD DART: IRIS DWG:	\$	3.00
TECHNICIAN HQ		4.5	Technician Labor	\$	1,282.50
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$	114.55

TOTAL CHARGES

2,519.05

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature: Date: