⊛ IRIDEX

Customer Repair Report

Model	Serial Number	Manufacture Date	RMA Number	IRIDEX Service Contact	Completion Date
OTHER	YEZ13552TC		03102	Mark w	

Customer Name		Customer Number	Repair Type			
Baylor Scott & White Memorial Hospital		14915	☐ Warranty ☒ non-warranty ☐ Upgrade ☐ Other			

Customer Description of Problem

LED KEEP BLINKING & aux MEESAGE

Technician Product Evaluate:

UNABLE TO TESTED CONTROL BOX NON-STOP BLINKING, COMPUTER BOARD NOT WORKING.

Service Performed on Unit:
Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
12757		1	THUMBSCREW M3.5 ZEISS SLAMM	$ 62.50
70014		1	PCBA SCANNER LASER INTERFACE & INTERLOCK	$ 1,364.56
88356		1	COMPUTER HDD	$ 1,950.00
88351		2	COMPUTER MOUNT	$ 138.00
88352		1	CABLE ADAPTER, TxCell CPU	$ 46.16
88261		1	Cable Assy. USB Two Ports.	$ 83.00
70208		1	ASSY JUMPER RFID SLA 9FT EXTENDED JACKET	$ 984.00
TECHNICIAN HQ		6	Technician Labor	$ 1,710.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	$ 201.47

TOTAL CHARGES | $ 6,539.69

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges.
Freight and sales tax will be added where applicable.

Print Name:

Signature: Date: