



Customer Repair Report

Model OTHER	Serial Number IQG81269C	Manufacture Date	RMA Number 02004	IRIDEX Service Contact Mark W	Completion Date
Customer Name Lawton Indian Hospital		Customer Number 24407	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

No communicatin between the IQ& TxCell E05

Technician Product Evaluate:

The laser module was degrading powers, needed to replace OPS diode.

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
65513		1	LASER OPS 532 5W GENSIS 532-5000 M PN:1103776	\$ 9,102.72
31720		1	ASSEMBLY CABLE TE COOLER	\$ 303.88
65989		1	INSULATION OPSL MFG :ARMACELL # F-05031	\$ 34.28
TECHNICIAN HQ		10	Technician Labor	\$ 2,850.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 192.26

TOTAL CHARGES	\$ 12,483.14
----------------------	---------------------

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges.
Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: