



Customer Repair Report

Model OTHER	Serial Number TX1464REVB	Manufacture Date	RMA Number 02847	IRIDEX Service Contact Mark w	Completion Date
Customer Name St. Marks Hospital		Customer Number 23374	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

getting an Error 2

Technician Product Evaluate:

Error2 and error 5 .ckicking noise on CTRL board . CTRL board #31586-1 is bad . Laser head require Eval

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
31586-1		1	PCBA LASER CONTROL JENOPTIK	\$ 2,631.48
30264		1	ASSY VANADATE GL2	\$ 1,563.56
10153		1	Fiber Adapter for Laser Indirect Ophthalmoscope	\$ 29.80
TECHNICIAN HQ		10	Technician Labor	\$ 2,850.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 201.47

TOTAL CHARGES **\$ 7,276.31**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: