



## Customer Repair Report

Model OTHER	Serial Number K1924	Manufacture Date	RMA Number 02785	IRIDEX Service Contact Mark W	Completion Date
Customer Name Memorial Hermann Hospital System		Customer Number 10808	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

### Customer Description of Problem

NO ILLUMINATION NEED ALIGNMENT

### Technician Product Evaluate:

Cannot test due to Burnt assembly fiber optic, missing eyepiece rubber, loose viewing assembly from T-slot adapter and will need alignment.

### Service Performed on Unit:

Waiting for customer approval

### ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
30628		2	EYEPiece RUBBER P/N: 11.65 29	\$ 50.00
11481		1	DUST CAP W/LANYARD DART: IRIS DWG:	\$ 3.00
TECHNICIAN HQ		4	Technician Labor	\$ 1,140.00
10784		1	LIO Halogen Bulb-10W/6V Heine: X-004.88.068	\$ 135.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 191.60

---

**TOTAL CHARGES** **\$ 1,519.60**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: