

Customer Repair Report

Model	Serial Number	Manufacture Date	9	RMA Number	IRIDEX Service Contact	Completion Date	
OTHER	YH13623TC-D)		02802	MArk W		
Customer Name	Cu	stomer Number	Repa	ir Type			
Ridley Creek Retina		25748		Warranty Marranty Upgrade Other			

Customer Description of Problem

IQ wasn't recognizing fiber, check fiber for any issue

Technician Product Evaluate:

UNABLE TO TESTED CONTROX NON-STOP BLINKING, BAD COMPUTER BOARD INTERFACE BOARD NEED TO BE UPGRADED.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description		Total Price	
70014		1	PCBA SCANNER LASER INTERFACE & INTERLOCK	\$	1,364.56	
88356		1	COMPUTER HDD	\$	1,950.00	
88351		2	COMPUTER MOUNT	\$	138.00	
88352		1	CABLE ADAPTER, TxCell CPU	\$	43.16	
88261		1	Cable Assy. USB Two Ports.	\$	83.00	
TECHNICIAN HQ		5	Technician Labor	\$	1,425.00	
BILLABLE FREIGHT	•	1	BILLABLE FREIGHT	\$	206.50	

TOTAL CHARGES

5,210.22

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Signature: Date: