



Customer Repair Report

Model OTHER	Serial Number GH16090LM	Manufacture Date	RMA Number 02048	IRIDEX Service Contact Mark W	Completion Date
Customer Name Southwestern Eye Sun City Surgery		Customer Number 11217	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

LOW AIMBEAM AND IMAGE NOT CENTERED

Technician Product Evaluate:

Fiber Jumper was not attached to SLA, ESF Filter is Scratched, Align and Calibrate, Clean all Mirrors, Lenses and Test.

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
30604		2	FILTER 532MM EYE SFTY #2	\$ 1,480.00
11510		1	PLATE ESF TOPCON HS	\$ 57.50
TECHNICIAN HQ		4	Technician Labor	\$ 1,140.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 112.36

TOTAL CHARGES	\$ 2,789.86
----------------------	--------------------

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges.
Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: