

Customer Repair Report

Model	Serial Number		Manufacture Date)	RMA Number	IRIDEX Service Contact	Completion Date
OTHER	1361977	,			02603	Mark W	
Customer Name		Custo	mer Number	Repair Type			
Medical Park Tower Surgery Center			25113	☐ Warranty ☐ Upgrade ☐ Other			

Customer Description of Problem

BRKEN NEED ALIGNMENT

Technician Product Evaluate:

Cannot test due to Twisted and burnt assembly fiber optic, loose housing illumination, delivery mirror coating delaminated, bent control stem, loose viewing assembly from T-slot adapter and will need aligenment.

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity Description		Total Price	
30792		1	MIRROR BROADBAND LIO+IRIDEX: 30792	\$	95.60
10784		1	LIO Halogen Bulb-10W/6V Heine: X-004.88.068	\$	135.00
11481		1	DUST CAP W/LANYARD DART: IRIS DWG:	\$	3.00
12687		1	BEAM CONTROL STEMS HEINE SUPPORT KNOBS P/N: 11.65 15	\$	125.00
12685		1	CONTROL STEM KNOB L & R VIEWING P/N: 11.65 38	\$	110.00
88080		1	Jumper Assy Mulitlam LIO+	\$	984.00
TECHNICIAN HQ		5	Technician Labor	\$	1,425.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$	191.60

TOTAL CHARGES \$

3,069.20

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:	
Signature:	Date: