

## Customer Repair Report

| Model                             | Serial Number | ]      | Manufacture Date | 2                            | RMA Number | IRIDEX Service Contact | Completion Date |
|-----------------------------------|---------------|--------|------------------|------------------------------|------------|------------------------|-----------------|
| OTHER                             | GH13669R      | M      |                  |                              | 02128      | Mark W                 |                 |
| Customer Name                     |               | Custor | mer Number       | Repair Type                  |            |                        |                 |
| The Eye Care Center of New Jersey |               |        | 21949            | ☐ Warranty ☐ Upgrade ☐ Other |            |                        |                 |

| Customor | <b>Description</b> | of Droblem |
|----------|--------------------|------------|
| Customer | Describuon         | of Problem |

Preventative Maintenance

## **Technician Product Evaluate:**

Re-tightening all scew. Need to replace fiber. Test ,clean and inspect.

## **Service Performed on Unit:** Waiting for customer approval

## **ITEMIZED REPAIR CHARGES**

| Part Number      | Part Number Price/Unit ( |     | Description              | Total Price |          |
|------------------|--------------------------|-----|--------------------------|-------------|----------|
| 30604            |                          | 2   | FILTER 532MM EYE SFTY #2 | \$          | 1,480.00 |
| 11510            |                          | 1   | PLATE ESF TOPCON HS      | \$          | 57.50    |
| TECHNICIAN HQ    |                          | 2.5 | Technician Labor         | \$          | 712.50   |
| BILLABLE FREIGHT |                          | 1   | BILLABLE FREIGHT         | \$          | 132.24   |

TOTAL CHARGES

2,382.24

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Signature: Date: