



## Customer Repair Report

Model OTHER	Serial Number YEZ15476TC	Manufacture Date	RMA Number 02926	IRIDEX Service Contact Mark W	Completion Date
Customer Name Salient Medical Solutions		Customer Number 24360	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

### Customer Description of Problem

PSI with low power

### Technician Product Evaluate:

UNABLE TO TESTED BURNED FIBER, BROKEN FLEX PLATE,INTERFACE BOARD NEED TO BE UPGR

### Service Performed on Unit:

Waiting for customer approval

### ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
70208		1	ASSY JUMPER RFID SLA 9FT EXTENDED JACKET	\$ 984.00
70014		1	PCBA SCANNER LASER INTERFACE & INTERLOCK	\$ 1,364.56
70117		2	AIMING BEAM FLEX PLATE	\$ 216.64
TECHNICIAN HQ		5	Technician Labor	\$ 1,425.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ -

---

**TOTAL CHARGES** **\$ 3,990.20**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: