

Customer Repair Report

		Custome	Repair	Report	
Model	Serial Number	Manufacture Dat	e RMA Numb	er IRIDEX Service Contact	Completion Date
OTHER	IQI91957	С	0168	34 Mark W	
Customer Name	•	Customer Number	Repair Type	•	-
Topcon Europe Medical B.V.		14251			
Customer Description UIM application failed to		m			
Technician Product DAMAGED ENCODER		E REPLACE			
Service Performed of Waiting for customer a					
Part Number	Price/Unit		REPAIR CH	HARGES Description	Total Price
65297	111cc/Cint	1		PCBA ENCODER	\$ 329.64
TECHNICIAN PASCAI		5		TECHNICIAN PASCAL	\$ 1,425.00
			-	TOTAL C	HARGES \$ 1,754.64
and are warranted as such. T	he labor and part ranty Statement.	s are guaranteed for 90 Your signature below a	days after the ship o	ecification and quality control state of the product. Items not co sponsibility for payment of the l	
Print Name:					
Signature			Date:		
Signature:			Date:		