



## Customer Repair Report

Model OTHER	Serial Number YH13001TC	Manufacture Date	RMA Number 02238	IRIDEX Service Contact Mark W	Completion Date
Customer Name New York Ophthalmology		Customer Number 25406	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

### Customer Description of Problem

LED KEPT BLINKING

### Technician Product Evaluate:

CONTROL BOX NON-STOP BLINKING UP POSITION IS HIGH X PO 6.9 X PI 7.9 GALVO'S BAD AND COMPUTER BOARD NOT WORKING TO SOME OF THE PARTS NEED TO BE UPGRADED,BEZEL AND ASSY, CABLE ARE DEMMAGE,TO

### Service Performed on Unit:

Waiting for customer approval

### ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
70358		1	INTERFACE BOX BEZEL LEDPLASTIC	\$ 373.00
70431		1	Bracket Cable Guide	\$ -
70036		1	ASSY CABLE AC INPUT SCANNER	\$ -
70171		1	GALVO DRIVER SET (6SIRIDEX2)	\$ 2,751.60
88356		1	COMPUTER HDD	\$ 1,950.00
88351		2	COMPUTER MOUNT	\$ -
88352		1	CABLE ADAPTER, TxCell CPU	\$ -
88261		1	Cable Assy. USB Two Ports.	\$ -
70406		1	EXTENSION COLLAR STRAIN RELIEF TxCELL	\$ -
70117		2	AIMING BEAM FLEX PLATE	\$ 433.28
70407		1	SEPARATOR CABLE	\$ 120.56
70318		1	STRAIN RELIEF CABLE CLAMP	\$ -
66085		2	SCREW SET 2-56 BALL END SSS	\$ -
70413		8	ASSY SCANNING INPUT LENS MOUNT HS	\$ 580.28
TECHNICIAN HQ		6	Technician Labor	\$ 1,710.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 206.50
TOTAL CHARGES				\$ 8,125.22

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: