

Customer Repair Report

Customer Report											
Model	Serial Number M		Manufacture Date		RMA Number	IRIDEX Service Contact	Completion Date				
OTHER	IQI92080	QI92080C			02843	Mark W					
Customer Name		Custo	omer Number	Repai	ir Type						
Madison Surgery Center			14652		Warranty Non-warranty Upgrade Other						
Customer Description of Problem											
Preventative Maintenanc	e										
Technician Product	Evaluate:										
UIM application failed to	load. Need to	repl	ace part.								

Service Performed on Unit: Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price		
65297		1	PCBA ENCODER	\$	329.64	
TECHNICIAN HQ		1	Technician Labor	\$	2,500.00	
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$	201.47	

TOTAL CHARGES \$ 3,031.11

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Signature: Date: