



Customer Repair Report

Model OTHER	Serial Number YH14126TC	Manufacture Date	RMA Number 02246	IRIDEX Service Contact Mark W	Completion Date
Customer Name Medical Faculty Associates Inc.		Customer Number 16488	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

CONNECTIVITY ISSUE NOT RECOGNIZED,

Technician Product Evaluate:

UNABLE TO TESTED CONTROL BOX NON-STOP BLINKING COMPUTER BOARD NOT WORKING CAUSING THE PROBLEM.

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
88356		1	COMPUTER HDD	\$ 1,950.00
88351		2	COMPUTER MOUNT	\$ 138.00
88352		1	CABLE ADAPTER, TxCell CPU	\$ 43.16
88261		1	Cable Assy. USB Two Ports.	\$ 83.00
TECHNICIAN HQ		5	Technician Labor	\$ 1,425.00

TOTAL CHARGES **\$ 3,639.16**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: