



Customer Repair Report

Model OTHER	Serial Number IQI91130C	Manufacture Date	RMA Number 02895	IRIDEX Service Contact Mark W	Completion Date
Customer Name Medical University Hospital Authority		Customer Number 13716	Repair Type <input type="checkbox"/> Warranty <input checked="" type="checkbox"/> non-warranty <input type="checkbox"/> Upgrade <input type="checkbox"/> Other		

Customer Description of Problem

NOT RECOGNIZING FIBERS TRY 2 DIFFRENT LOT

Technician Product Evaluate:

Replace PCBA fiber detect and retested all the delivery devices

Service Performed on Unit:

Waiting for customer approval

ITEMIZED REPAIR CHARGES

Part Number	Price/Unit	Quantity	Description	Total Price
66231		1	PCBA FIBER DETECT G6	\$ 362.16
66298		1	KNOB SMALL UIM	\$ 63.40
TECHNICIAN HQ		6	Technician Labor	\$ 1,710.00
BILLABLE FREIGHT		1	BILLABLE FREIGHT	\$ 206.50

TOTAL CHARGES **\$ 2,342.06**

Repaired products may contain reconditioned parts, but those are subject to the same specification and quality control standards applied to new materials and are warranted as such. The labor and parts are guaranteed for 90 days after the ship date of the product. Items not covered in the warranty are stated in the IRIDEX Standard Warranty Statement. Your signature below acknowledges the responsibility for payment of the listed charges. Freight and sales tax will be added where applicable.

Print Name:

Signature:

Date: